
**Solar energy — Specification and
classification of instruments for
measuring hemispherical solar and
direct solar radiation**

*Énergie solaire — Spécification et classification des instruments de
mesurage du rayonnement solaire hémisphérique et direct*

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

ISO 9060:2018

<https://standards.iteh.ai/catalog/standards/iso/1ec73ab8-4e64-4e2e-91e4-7f9a0a2f3b43/iso-9060-2018>



iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

ISO 9060:2018

<https://standards.iteh.ai/catalog/standards/iso/1ec73ab8-4e64-4e2e-91e4-7f9a0a2f3b43/iso-9060-2018>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2018

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Instruments to measure hemispherical solar radiation — Pyranometers	3
4.1 General physical design	3
4.2 Types	4
4.3 Classification	4
4.3.1 General	4
4.3.2 Pyranometer specifications	5
4.3.3 Classification criteria	7
4.3.4 Identification of classification	8
5 Instruments to measure direct solar radiation—Pyrheliometers	8
5.1 General physical design	8
5.2 Types	9
5.2.1 Absolute pyrheliometer	9
5.2.2 Compensation pyrheliometer	9
5.2.3 Pyrheliometers without self-calibration capability	9
5.3 Classification	9
5.3.1 General	9
5.3.2 Pyrheliometer specifications	10
5.3.3 Classification criteria	10
5.3.4 Identification of classification	11
6 Final remarks	12
Annex A (informative) Comments on the specifications given in Tables 1 to 2	14
Bibliography	18

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 180, *Solar energy*, Subcommittee SC 1, *Climate — Measurement and data*.

This second edition cancels and replaces the first edition (ISO 9060:1990), which has been technically revised. The main changes compared to the previous edition are as follows:

- in addition to thermopile radiometers, other technology options have been included such as photoelectric sensors as long as they fulfil the requirements specified in this document;
- the spectral error is used to characterize the spectral responsivity;
- to further characterize the radiometers, the additional properties “spectrally flat” and “fast response” can be added to the classification if the radiometers fulfil specific criteria;
- more intuitive names have been introduced for the classes: “A”, “B”, “C”.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

This document is one of a series of standards that specify methods and instruments for the measurement of solar radiation in support to solar energy utilization.

Accurate solar radiation data are used in meteorology and are needed for developing solar energy appliances, in particular for performance testing, solar radiation simulation and resource assessment.

The measurement of radiation is needed for determination of the conversion efficiencies of solar appliances. The specification and classification of these instruments are needed in order to enable the comparison of solar radiation data on a worldwide basis. In addition, this classification is intended to assist end users/consumers and entities requiring and tendering radiometers with the choice or comparison of instruments, to protect end users/consumers and to offer a level playing field for manufacturers.

The specification and classification of solar radiometers specified in this document provides an accuracy ranking and focuses on application specific requirements and qualities. However, solar radiometers are used in a wide range of applications with often conflicting requirements. The best radiometer for one application may be inadequate for a different application. In order to address this issue at least partly, a sensor of a given class can be assigned the additional properties “fast response” and/or “spectrally flat” to further characterize the radiometers.

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO 9060:2018](https://standards.iteh.ai/catalog/standards/iso/1ec73ab8-4e64-4e2e-91e4-7f9a0a2f3b43/iso-9060-2018)

<https://standards.iteh.ai/catalog/standards/iso/1ec73ab8-4e64-4e2e-91e4-7f9a0a2f3b43/iso-9060-2018>

Solar energy — Specification and classification of instruments for measuring hemispherical solar and direct solar radiation

1 Scope

This document establishes a classification and specification of instruments for the measurement of hemispherical solar and direct solar radiation integrated over the spectral range from approximately 0,3 μm to about 3 μm to 4 μm .

Instruments for the measurement of hemispherical solar radiation and direct solar radiation are classified according to the results obtained from indoor or outdoor performance tests. This document does not specify the test procedures.

2 Normative references

There are no normative references in this document.

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

3.1

hemispherical solar radiation

solar radiation received by a plane surface from a solid angle of 2π sr

Note 1 to entry: Approximately 97 % to 99 % of the hemispherical solar radiation incident at the Earth's surface is contained within the wavelength range from 0,3 μm to 3 μm ^[1]. Generally, hemispherical solar radiation is composed of direct solar radiation and diffuse solar radiation (solar radiation scattered in the atmosphere) as well as solar radiation reflected by the ground.

3.2

global horizontal irradiance

hemispherical solar radiation received by a horizontal plane surface

Note 1 to entry: The tilt angle and the azimuth of the receiver surface should be specified, e.g. horizontal.

3.3

direct solar radiation

radiation received from a small solid angle centred on the sun's disc, on a given plane

Note 1 to entry: In general, direct solar radiation is measured by instruments with field-of-view angles of up to 6°. Therefore a part of the scattered radiation around the sun's disc (circumsolar radiation or aureole) is also included (see 5.1). Historic pyrheliometers of the Angström type (compensation pyrheliometer) have a larger field of view of up to 15°. A more detailed definition of circumsolar radiation and related parameters can be found in Reference [2].

Note 2 to entry: Approximately 97 % to 99 % of the direct solar radiation received at the ground is contained within the wavelength range from 0 μm to 3 μm ^[1].

Note 3 to entry: The tilt angle of the receiver surface should be specified, e.g. horizontal or normal to the direct solar radiation.

3.4 diffuse solar radiation diffuse radiation

hemispherical solar radiation minus coplanar direct solar radiation

Note 1 to entry: For the purposes of solar energy technology, diffuse radiation includes solar radiation scattered in the atmosphere as well as solar radiation reflected by the ground, depending on the inclination of the receiver surface.

Note 2 to entry: The tilt angle and the azimuth of the receiver surface should be specified, e.g. horizontal.

3.5 pyranometer

radiometer designed for measuring the irradiance on a plane receiver surface which results from the radiant fluxes incident from the hemisphere above within the wavelength range from approximately 0,3 μm to about 3 μm to 4 μm

Note 1 to entry: The spectral range (50 % transmittance points) given is only nominal. Depending on the radiometer design, the spectral limits of its responsivity can be different from the limits mentioned above.

3.6 pyrheliometer

radiometer designed for measuring the irradiance which results from the solar radiant flux from a well-defined solid angle the axis of which is perpendicular to the plane receiver surface

Note 1 to entry: It follows from this definition that pyrheliometers are used to measure direct solar radiation at normal incidence. Typical opening half angles of common and historical pyrheliometers range from 2,5° to 7,5°. Reference [3] recommends that the opening half-angle is 2,5° ($6 \cdot 10^{-3}$ sr) and the slope angle 1° for all new designs of direct solar radiation instruments. The opening half-angle is measured from the centre of the (circular) receiver aperture to the edge of the view-limiting aperture. The slope angle is the opening half-angle of the cone defined by both apertures. For mathematical definitions of the angles, see 5.1 b). A more detailed description of the influence of circumsolar radiation on the pyrheliometers can be found in Reference [2].

Note 2 to entry: The spectral responsivity of field pyrheliometers is often limited to the range of approximately 0,3 μm to 3 μm , depending on the radiometer properties. The spectral range (50 % points) given is only nominal. Depending on the radiometer design, the spectral limits of its responsivity can be different from the limits mentioned above.

3.7 diffusometer

radiometer designed for measuring the diffuse solar radiation, consisting of a pyranometer and a shading structure which can be a shading ball, a shading disk, a shading ring, a rotating shadowband or a shading mask

Note 1 to entry: Shading balls and disks shall be tracked to the sun, so that the pyranometer is shaded. Shading disks and their tracking are defined in ISO 9846[4]. The centre of a shading ball is tracked to the same point as the centre of a shading disk. The diameter of the ball corresponds to the diameter of the disk. The shaded opening angle and slope angle of shading balls and -disks for the sun in the zenith shall be 2,5° and 1°.

Note 2 to entry: Shading rings are positioned such that the pyranometer is shaded for all solar positions occurring throughout approximately two days. Shading rings shall be adjusted approximately every two days. Shading rings therefore prevent not only the direct radiation, but also a part of the diffuse radiation from reaching the pyranometer and only an approximation of the diffuse radiation can be measured.

Note 3 to entry: A rotating shadowband is rotated around the pyranometer so that this pyranometer is shaded for some time during the rotation. The pyranometer measures an approximation of the diffuse radiation when the shadowband shades the sensor. The pyranometer measures the hemispherical radiation when the shadowband is below the pyranometer's field-of-view. When the shadowband's shadow is close to the sensor, but not on the sensor the hemispherical radiation except of the blocked diffuse radiation is measured. With these three measurements so-called rotating shadowband irradiometers determine the diffuse radiation.

Note 4 to entry: Shading masks throw a shadow on one or various pyranometers depending on the solar position.

3.8

offset correction

value added algebraically to the uncorrected result of a measurement to compensate for systematic error

Note 1 to entry: The offset correction is equal to the negative of the estimated systematic error.

Note 2 to entry: Since the systematic error cannot be known perfectly, the compensation cannot be complete.

3.9

correction factor

numerical factor by which the uncorrected result of a measurement is multiplied to compensate for systematic error

Note 1 to entry: Since the systematic error cannot be known perfectly, the compensation cannot be complete.

[SOURCE: ISO/IEC Guide 98-3:2008, B.2.24]

3.10

acceptance interval

interval of permissible measured quantity values

[SOURCE: BIPM, 2012^[6], 3.3.9]

3.11

tolerance interval

interval of permissible values of a property

[SOURCE: BIPM, 2012^[6], 3.3.5]

3.12

guard band

interval between a tolerance limit and a corresponding acceptance limit

[SOURCE: BIPM, 2012^[6], 3.3.11]

3.13

accuracy class

class of measuring instruments or measuring systems that meet stated metrological requirements that are intended to keep measurement errors or instrumental uncertainties within specified limits under specified operating conditions

[SOURCE: ISO/IEC Guide 99:2007, 4.25, modified — Notes have been deleted.]

4 Instruments to measure hemispherical solar radiation — Pyranometers

4.1 General physical design

Pyranometers are radiometers used to measure hemispherical solar radiation (see 3.1, 3.2, 3.4, 3.5 and 3.7).

Thermal sensors transform radiant energy into thermal energy with a consequent rise in the temperature of the receiving surface. This rise in temperature is balanced by various kinds of heat losses to thermal sinks (e.g. the body of the pyranometer and ambient air).

The thermal sensor of a pyranometer is protected from wind, rain and dust as well as the exchange of thermal radiation by one or two transparent domes and/or a diffusor whose spectral transmittance confines the spectral range of responsivity to the interval between approximately 0,3 μm and 3 μm (50 % transmittance points).

Photodiode pyranometers use photodiodes as sensors that convert the incoming radiation in electrical energy. The photodiodes are often placed below a diffuser.

Because of the spectral limits of the measurement, thermal sensors have an advantage compared to photodiode sensors as they can achieve a nearly uniform spectral responsivity required for low spectral errors. The spectral irradiance error is the error introduced by the change in the spectral distribution of the incident solar radiation and the difference between the spectral responsivity of the radiometer with respect to a radiometer with completely homogeneous spectral responsivity in the wavelength range of interest.

Other technologies exist that are not mentioned in this document.

The main parts of common pyranometers are:

- a) the sensor;
- b) the transparent dome(s) or diffuser, which cover(s) concentrically the receiving surface; and
- c) the body, which is often shielded by a sun-screen, and used as a thermal reference.

4.2 Types

One type of pyranometer is the “thermoelectric” pyranometer which is equipped with a thermopile (sometimes called a thermobattery) measuring the difference in temperature between the receiving surface (active junctions) and the body (passive junctions). The position and number of the active and passive junctions vary depending on the different pyranometer models. Generally, these sensors are covered by one or two concentric glass dome(s) or a diffuser.

Another type of pyranometer is the “photoelectric” pyranometer which is equipped with a photoelectric receiver (using e.g. silicon photodiode or photovoltaic cell) measuring photovoltaic power. This type of pyranometer is usually called “Si-pyranometer”. Often these sensors are placed below a diffuser. The diffuser can have the shape of a cylinder or other shapes.

4.3 Classification

ISO 9060:2018

<https://standards.iteh.ai/catalog/standards/iso/1ec73ab8-4e64-4e2e-91e4-7f9a0a2f3b43/iso-9060-2018>

4.3.1 General

The classification of pyranometers is based exclusively on the measuring specifications of the instruments. The classification is not based upon manufacturing technologies but rather on criteria deduced from the various applications of pyranometers. Following this principle, any technical device which produces a signal when irradiated (e.g. a photovoltaic cell) could be classified as a pyranometer according to this document.

Most of the classification criteria (see [Table 1](#)) are of general relevance, whereas others may be important only for specific applications.

Therefore statements about the overall measurement uncertainty can only be made on an individual basis, taking all relevant factors into account.

The classification scheme is based on various specifications, as given in [4.3.2](#) and various classification criteria, as given in [4.3.3](#).

The classification can be understood as an accuracy ranking. The letters indicate the typically reached accuracy for well-maintained measurements when compared under the same measurement conditions. The accuracy decreases in alphabetic order (A reaches a better accuracy than B or C). However, the accuracy ranking does not mean that a radiometer of higher accuracy class is more accurate than another radiometer of lower class under all conditions. First of all, as different radiometers can have different maintenance requirements and e.g. susceptibility to soiling, the term “well-maintained” is important in the previous statement. Furthermore, depending on the application and the measurement conditions, a sensor of a lower class can be more appropriate in some cases. For example, radiometers