



SLOVENSKI STANDARD
oSIST prEN ISO 16645:2019
01-januar-2019

Radiološka zaščita - Medicinski elektronski pospeševalniki - Zahteve in priporočila za snovanje in ocenjevanje zaščitnih zaslonov (ISO 16645:2016)

Radiological protection - Medical electron accelerators - Requirements and recommendations for shielding design and evaluation (ISO 16645:2016)

Strahlenschutz - Medizinische Elektronenbeschleuniger-Anlagen - Anforderungen und Empfehlungen an die Ausführung der Abschirmung und deren Bewertung (ISO 16645:2016)

Radioprotection - Accélérateurs médicaux d'électrons - Exigences et recommandations pour la conception et l'évaluation du blindage (ISO 16645:2016)

Ta slovenski standard je istoveten z: prEN ISO 16645

ICS:

11.040.99	Druga medicinska oprema	Other medical equipment
13.280	Varstvo pred sevanjem	Radiation protection

oSIST prEN ISO 16645:2019

en,fr,de

INTERNATIONAL
STANDARD

ISO
16645

First edition
2016-10-01

Corrected version
2016-11-15

**Radiological protection — Medical
electron accelerators — Requirements
and recommendations for shielding
design and evaluation**

*Radioprotection — Accélérateurs médicaux d'électrons — Exigences
et recommandations pour la conception et l'évaluation du blindage*

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 16645:2019

<https://standards.iteh.ai/catalog/standards/sist/b6c1d492-5a62-493c-b144-e79f8e93d99f/sist-en-iso-16645-2019>



Reference number
ISO 16645:2016(E)

© ISO 2016

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO 16645:2019

<https://standards.iteh.ai/catalog/standards/sist/b6c1d492-5a62-493c-b144-e79f8e93d99f/sist-en-iso-16645-2019>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2016, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

	Page
Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
3.1 Quantities.....	1
3.2 Definitions.....	4
4 Shielding design goals and other design criteria	6
4.1 Shielding design goals.....	6
4.2 Shielding design assumptions.....	7
5 Role of the manufacturers, of the radiation protection officer or qualified expert and interactions between stakeholders	8
5.1 General.....	8
5.2 Linear accelerator manufacturer.....	8
5.3 Shielding material vendor.....	9
5.4 Architectural firm/general contractor.....	10
5.5 Radiation protection officer or qualified expert.....	10
5.6 The licensee.....	11
6 Radiation fields around a linear electron accelerator	11
6.1 General.....	11
6.2 X-ray radiation.....	11
6.2.1 Primary X-ray beam.....	11
6.2.2 Primary electron beam bremsstrahlung.....	12
6.2.3 Secondary X-ray radiation.....	12
6.2.4 Tertiary X-ray radiation.....	13
6.3 Neutron radiation.....	13
6.3.1 General.....	13
6.3.2 Direct neutron radiation.....	14
6.3.3 Scattered and thermal neutron radiation.....	14
6.3.4 Primary barrier neutron radiation.....	15
6.4 γ radiation.....	15
6.4.1 General.....	15
6.4.2 Maze γ radiation.....	15
6.4.3 Door γ radiation.....	15
6.4.4 Primary barrier γ radiation.....	15
6.4.5 Air γ radiation.....	16
7 Shielding materials and transmission values	16
8 General formalism for shielding calculation	18
9 Shielding calculation for conventional devices	20
9.1 General.....	20
9.2 Primary barriers.....	20
9.2.1 Radiation components.....	20
9.2.2 Barrier with a unique material.....	21
9.2.3 Barrier with multiple layers.....	21
9.3 Secondary barriers.....	22
9.3.1 Radiation components.....	22
9.3.2 Barrier with a unique material.....	23
9.3.3 Barriers with multiple layers.....	24
10 Doors and mazes	24
10.1 General.....	24

ISO 16645:2016(E)

10.2	Radiation components	25
10.3	Standard maze.....	25
10.3.1	Maze X-ray scatter calculations.....	25
10.3.2	X-ray direct Leakage.....	30
10.3.3	Maze neutron and capture gamma calculations.....	31
10.4	Two legged maze.....	33
10.5	No maze - Direct-shielded doors.....	34
10.5.1	General.....	34
10.5.2	Shielding at the far side of a direct-shielded door entrance.....	35
10.5.3	Shielding at the near side of a direct-shielded door entrance.....	37
10.6	No door at maze entrance.....	39
10.7	Door Calculations.....	40
10.7.1	General.....	40
10.7.2	Maze door calculations.....	40
10.7.3	Direct Shielded Door Calculations.....	41
11	Shielding calculations for special devices.....	41
11.1	General.....	41
11.2	Robotic arm accelerator.....	41
11.3	Helical intensity modulated radiotherapy.....	42
11.4	Dedicated device for intra operative radiotherapy with electrons.....	42
12	Ducts.....	43
12.1	Duct impact on radiation protection.....	43
12.2	Recommended location and geometry.....	43
12.3	Additional shielding.....	44
12.3.1	General.....	44
12.3.2	Neutron and capture gamma radiation passing through the interior of the shielded duct.....	44
12.3.3	X scattered radiation passing through the interior of the shielded duct.....	45
12.3.4	Scattered radiation passing through the walls of the duct shielding.....	46
12.3.5	Dose equivalent at HVAC duct exterior opening.....	46
13	Special considerations.....	46
13.1	Skyshine.....	46
13.1.1	General.....	46
13.1.2	X-ray skyshine radiation.....	46
13.1.3	Neutron skyshine radiation.....	48
13.2	Groundshine radiations.....	49
13.3	Joints and junctions.....	49
14	Shielding evaluation (experimental verification).....	49
14.1	General.....	49
14.2	Measuring equipment and methodology.....	50
14.3	Evaluation.....	50
15	Indication, warning signs, interlocks.....	52
Annex A (informative) Tenth value layers for the most common shielding materials.....		53
Annex B (informative) Supporting data for shielding calculations.....		66
Annex C (informative) Example of calculation for conventional device and standard maze.....		68
Bibliography.....		75

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2. www.iso.org/directives

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received. www.iso.org/patents

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT), see the following URL: [Foreword - Supplementary information](#)

The committee responsible for this document is ISO/TC 85, *Nuclear energy, nuclear technologies, and radiological protection*, Subcommittee SC 2, *Radiological protection*.

This corrected version of ISO 16645:2016 incorporates the correction of [Tables A.9](#) and [C.6](#).

<https://standards.iteh.ai/catalog/standards/sist/b6c1d492-5a62-493c-b144-e79f8e93d99f/sist-en-iso-16645-2019>

ISO 16645:2016(E)

Introduction

Radiotherapy uses external beam radiation to kill cancer cells and shrink tumours. The use of electron linear accelerators to administer external beam radiation has spread during recent decades and is now common throughout the world. These accelerators deliver high energy electron and photon beams with increasingly high dose rates. Although the use of radiotherapy is well established, irradiation techniques have continued to evolve and are becoming increasingly complex. Examples include modulation of beam intensity, availability of high dose rate modes, arctherapy, helical intensity modulated radiotherapy, robotic arm accelerators, and dedicated devices for intra-operative radiotherapy. The shielding design of treatment rooms has been evolving with these changes. The higher radiation workload associated with most of these techniques can impact the shielding materials used. The irradiation technique can also impact the geometry to be considered in the shielding calculations.

IEC 60601-2-1 relates to the design and the construction of the accelerators in order to ensure the safety of their operation^[1]. In addition, several national^{[2][3]} or international (IAEA Safety Reports Series Report No. 47, 2006) reports propose recommendations concerning the installation and the exploitation of these accelerators, the safety devices, the design and the calculation of protections, the radiological control and monitoring. National standards have been established in certain countries^[4] ^[5]. Moreover national regulations impose particular rules of protection against radiation, in particular relating to the definition of the controlled areas and the calculation of shielding.

Taking into account the developments of new irradiation techniques and of new designs of treatment room facilities on the one hand, and the variety of guides or normative documents on the other hand, it appeared judicious to establish an international standard to be used as a general framework. This standard is intended to be complementary to the other international standards (IEC and IAEA).

The following items are discussed in the standard:

- types of accelerators: conventional accelerators with and without flattening filter (FF and FFF operating modes), devices for helical intensity modulated radiotherapy and robotic arm accelerator, dedicated machines for intra-operative radiotherapy;
- radiation fields: electrons, X photons and neutrons (direct, scattered, leakage), neutron capture gamma rays;
- Treatment room geometry: maze without and with door, no maze with direct door;
- materials of protection: concrete (ordinary or high density), metals, laminated barriers (concrete and metal), hydrogenated materials, earth and others;
- design of the radiotherapy facility:
- calculation methods of the shielding, including neutrons, various types of installations and shielding geometries;
- evaluation of the impact of the maze and calculation of the protection of the entrance door;
- evaluation of the impact of the ducts (ventilation and air-conditioning, high voltage and fluids) and additional protections;
- shielding design assumption and goals;
- Radiation survey of the completed installation to ensure national requirements have been met and the shielding and design is fit for purpose after installation of the accelerator.

Radiological protection — Medical electron accelerators — Requirements and recommendations for shielding design and evaluation

1 Scope

This International Standard is applicable to medical electron linear accelerators i.e. linear accelerators with nominal energies of the beam ranging from 4 MV to 30 MV, including particular installations such as robotic arm, helical intensity modulated radiotherapy devices and dedicated devices for intra operative radiotherapy (IORT) with electrons.

The cyclotrons and the synchrotrons used for hadrontherapy are not considered.

The radiation protection requirements and recommendations given in this International Standard cover the aspects relating to regulations, shielding design goals and other design criteria, role of the manufacturers, of the radiation protection officer or qualified expert and interactions between stakeholders, radiations around a linear accelerator, shielding for conventional and special devices (including shielding materials and transmission values, calculations for various treatment room configurations, duct impact on radiation protection) and the radiological monitoring (measurements).

NOTE 1 [Annex A](#) provides transmission values for the most common shielding materials.

NOTE 2 [Annex B](#) provides supporting data for shielding calculation.

NOTE 3 [Annex C](#) provides an example of calculation for conventional device and standard maze.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60976, *Medical electrical equipment — Medical electron accelerators — Functional performance characteristics*

IAEA Safety Reports Series Report No. 47, *Radiation protection in the Design of Radiotherapy Facilities (2006)*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 60976 and the following apply.

3.1 Quantities

3.1.1

absorbed dose

D

quotient of $d\bar{\varepsilon}$ by dm , where $d\bar{\varepsilon}$ is the mean energy imparted to matter of mass dm thus

$$D = \frac{d\bar{\varepsilon}}{dm}$$

ISO 16645:2016(E)

Note 1 to entry: In this document, the absorbed dose is defined for radiation produced by a linear accelerator at a specific location: the absorbed dose to water at the isocentre (at 1 m from the source for conventional devices) at a reference depth in water in electron equilibrium conditions (for example at the depth of maximum absorbed dose).

Note 2 to entry: The unit of absorbed dose is joule per kilogram ($\text{J}\cdot\text{kg}^{-1}$), and its special name is gray (Gy).

[SOURCE: ISO 12749-2:2013, 4.1.6.7]^[6]

3.1.2 absorbed dose rate output rate

DR_0

dose absorbed per unit of time

Note 1 to entry: In this International Standard, in the absence of specific indication, the absorbed dose rate is defined for radiation produced by a linear accelerator at a specific location: the absorbed dose rate to water at the isocentre (at 1 m from the source for conventional devices) at a reference depth in water in electron equilibrium conditions (for example at the depth of maximum absorbed dose).

Note 2 to entry: The unit of absorbed dose rate is gray per second ($\text{Gy}\cdot\text{s}^{-1}$). The usual unit for medical accelerators is gray per hour ($\text{Gy}\cdot\text{h}^{-1}$).

3.1.3 dose equivalent

H

product of D and Q at a point in tissue, where D is the absorbed dose (3.1.1) and Q is the quality factor for the specific radiation at this point, thus: $H = D \times Q$

Note 1 to entry: The unit of dose equivalent is joule per kilogram ($\text{J}\cdot\text{kg}^{-1}$), and its special name is sievert (Sv).

[SOURCE: ISO 12749-2:2013, 4.1.6.8]^[6]

3.1.4 IMRT ratio

C_1

ratio of the average monitor unit per unit prescribed absorbed dose needed for IMRT (MU_{IMRT}) and the monitor unit per unit absorbed dose for conventional treatment (MU_{CONV})

$$C_1 = \frac{MU_{\text{IMRT}}}{MU_{\text{CONV}}}$$

3.1.5 instantaneous dose-equivalent rate

IDR

“ambient/personal” dose-equivalent rate ($\text{Sv}\cdot\text{h}^{-1}$) as measured with the linear accelerator operating at the absorbed dose rate DR_0 ($\text{Gy}\cdot\text{h}^{-1}$)

Note 1 to entry: This is the direct reading of the ratemeter that gives a stable reading in dose-equivalent per hour. IDR is specified at a reference point (30 cm) beyond the penetrated barrier.

3.1.6 effective dose

E

summation of all the tissue equivalent doses, each multiplied by the appropriate tissue weighting factor

3.1.7 occupancy factor

T

fraction of time the areas adjacent to the treatment room are occupied by an individual or group during linear accelerator operation

3.1.8**orientation or use factor***U*

fraction of the time during which the radiation under consideration is directed at a particular barrier

3.1.9**reflection coefficient** α

fraction of radiation (e.g., fluence, energy fluence) expressed by the ratio of the amount backscattered to that incident

3.1.10**shielding design goal***P*

practical values of dose equivalent, for a single radiotherapy source or set of sources, evaluated at a reference point beyond a protective barrier

Note 1 to entry: The shielding design goals ensure that the respective annual values for effective dose limit defined by national regulation or IAEA/ICRP for controlled and uncontrolled areas are not exceeded.

3.1.11**(patient) scatter fraction** $a(\theta)$

ratio of absorbed dose at 1 m from a tissue-equivalent scattering object to the absorbed dose measured at the isocentre with the object removed

Note 1 to entry: This quantity is a function of the scatter angle (θ), incident beam quality, and beam area. A scattering phantom is typically a water-equivalent volume representing a standard human being.

3.1.12**tenth-value distance***TVD*

distance that a specified radiation travels under broad beam condition in order to reduce the radiation field intensity to one-tenth of its original value

3.1.13**tenth-value layer***TVL*

thickness of a specific material that reduces a specified radiation field intensity by a factor of 10 of its original value, under broad beam condition

Note 1 to entry: *TVL* is expressed in m or cm of a defined material or in kg/m^2 (thickness \times density).

Note 2 to entry: TVL_1 and TVL_2 are the first and the second tenth-value layer thicknesses, respectively.

Note 3 to entry: TVL_e is the equilibrium tenth-value layer, thickness for each subsequent tenth-value layer in the region in which the directional and spectral distributions of the radiation field are practically independent of thickness.

Note 4 to entry: TVL_c is the cumulative tenth-value layer, approximate value based on large attenuation measurements: for a given thickness t , $TVL_c = -t/\log(B)$.

3.1.14**time averaged dose-equivalent rate***TADR*

barrier attenuated dose-equivalent rate averaged over a specified period of accelerator operation

Note 1 to entry: *TADR* is proportional to instantaneous dose-equivalent rate (*IDR*), and depends on the values of workload (*W*) and orientation or use factor (*U*).

ISO 16645:2016(E)**3.1.15****transmission factor (or barrier transmission)*****B***

ratio of radiation field intensity at a location behind the barrier on which radiation is incident to the field intensity at the same location without the presence of the shield, for a given radiation type and quality

Note 1 to entry: *B* is a measure of the shielding effectiveness of the barrier.

3.1.16**workload*****W***

average absorbed dose to water of radiation produced by a linear accelerator, at the isocentre at a reference depth in water in electron equilibrium conditions, over a specified period averaged over a year

Note 1 to entry: The workload is specified in Gray (Gy).

Note 2 to entry: The time period should be consistent between shielding design goals and workload.

Note 3 to entry: The isocentre is at 1 m from the source for conventional devices.

Note 4 to entry: The reference depth in water is for example the depth of maximum absorbed dose.

3.2 Definitions**3.2.1****barrier (or protective barrier)**

protective wall of radiation attenuating material(s) used to reduce the dose equivalent on the side beyond the radiation source to an acceptable level compatible with national legislation or international guidance

3.2.2**primary barrier**

wall, ceiling, floor or other structure designed to attenuate the direct radiation emitted from the target or source that passes through the collimator opening (useful beam) to an acceptable level compatible with national legislation or international guidance

3.2.3**secondary barrier**

wall, ceiling, floor or other structure not struck by the primary beam and designed to attenuate the leakage and scattered radiations to an acceptable level compatible with national legislation or international guidance

3.2.4**controlled area**

defined area in which specific protection measures and safety provisions are or could be required for controlling exposures or preventing the spread of contamination in normal working conditions, and preventing or limiting the likelihood and magnitude of potential exposures

Note 1 to entry: This implies that access, occupancy, and working conditions are controlled for radiation protection purposes.

[SOURCE: IAEA BSS]^[7]

3.2.5**geometrical field size**

geometrical projection as seen from the centre of the front surface of the radiation source on a plane perpendicular to the axis of the beam of the distal end of the beam limiting device or collimator

Note 1 to entry: The field is thus of the same shape as the aperture of the beam limiting device.

Note 2 to entry: The projected field size is specified at a particular distance from the target, e.g. at the isocentre 1 m from the target or at the reference distance of the device.

[SOURCE: IEC 60976:2007]

3.2.6

helical intensity modulated radiotherapy

radiotherapy using a linear accelerator that delivers treatment with a slit beam adjustable by a multileaf collimator (MLC) and that rotates continuously around patient with geometry resembling diagnostic computed tomography (CT), with concomitant motion of the couch

Note 1 to entry: Helical intensity modulated radiotherapy is often called tomotherapy.

3.2.7

intensity-modulated radiation therapy

IMRT

treatment procedure requiring, in general, the coordinated control of photon or electron fluence, beam orientation relative to the patient, and beam size, either in a continuous or a discrete manner, and as pre-determined by a treatment plan

Note 1 to entry: The primary purpose of IMRT is to improve the conformity of the dose distribution to the planned target volume, while minimizing dose to surrounding healthy tissue.

[SOURCE: IEC 60976:2007]

3.2.8

isocentre

point defined by intersection of the gantry axis of rotation and the beam centerline of a linear accelerator

Note 1 to entry: For conventional linear accelerator, the isocentre is located at 1 m from the radiation source.

3.2.9

leakage radiation

radiation, except the useful beam, coming from the accelerator head and other beam-line components

Note 1 to entry: It is attenuated by shielding in the treatment head as specified by IEC 60601-2-1^[1].

3.2.10

members of the public

persons who are not occupationally exposed by a source or practice under consideration

Note 1 to entry: When being irradiated as a result of medical care, patients are not considered as members of the public.

3.2.11

nominal energy

energy stated by the manufacturer to characterize the radiation beam

Note 1 to entry: "MV" is used when referring to accelerating voltages and the end point energy of a bremsstrahlung spectrum, while "MeV" is used when referring to monoenergetic photons or electrons

[SOURCE: IEC 60976:2007]

3.2.12

occupied area

room or other space, indoors or outdoors, that is likely to be occupied by any person, either regularly or periodically during the course of the person's work, habitation or recreation, and in which an ionizing radiation field exists because of radiation sources in the vicinity

3.2.13

radiation protection officer

person technically competent in radiation protection matters relevant for a given type of practice who is designated by the registrant, licensee or employer to oversee the application of relevant requirements

[SOURCE: IAEA BSS]^[2]

ISO 16645:2016(E)**3.2.14****qualified expert**

individual who, by virtue of certification by appropriate boards or societies, professional licence or academic qualifications and experience, is duly recognized as having expertise in a relevant field of specialization, e.g. medical physics, radiation protection, occupational health, fire safety, quality management or any relevant engineering or safety specialty

[SOURCE: IAEA BSS]^[Z]

3.2.15**robotic arm accelerator**

device composed by a linear accelerator mounted on a 6D robotic arm allowing a multi-directional delivery of the dose

Note 1 to entry: The robotic arm is referred to as 6 degrees of freedom because movements are made for 3 translational motions (X, Y and Z) and 3 rotational motions.

Note 2 to entry: The “geometric isocentre” is a reference point in the room that serves as the origin for several coordinates systems related to robot and imaging calibration.

3.2.16**scattered radiation**

radiation that, during passage through matter, is changed in direction, and the change is usually accompanied by a decrease in energy and intensity

3.2.17**secondary radiation**

radiation produced by scattering from the areas struck by the primary X-ray beam or leakage radiation through the treatment head of the linear accelerator

3.2.18**supervised area**

defined area in which specific protection measures and safety provisions are or could be required for controlling normal exposures during normal working conditions, and preventing or limiting the extent of potential exposures

3.2.19**tertiary radiation**

radiation produced by scattering from areas struck by leakage radiation, secondary radiation and primary electron beam bremsstrahlung

4 Shielding design goals and other design criteria**4.1 Shielding design goals**

Shielding design goals (P) are levels of dose equivalent (H) used in the design calculations and evaluation of barriers constructed for the protection of workers or members of the public. Different shielding design goals shall be defined for supervised, controlled and public areas. They have to be in accordance with existing national regulation or if not available according to IAEA basic safety standards on radiation protection related to effective dose limits for workers and members of the public.

The P value (Sv) set by national authorities should be a fraction of dose limits for workers or for members of the public. They can be expressed as weekly values (mSv/week) since the workload (W) for a radiotherapy equipment has traditionally a weekly format. But according to national regulation, other time periods can be used.

Shielding design goals (P) are practical values that are evaluated at a reference point beyond a protective barrier for the workload W proposed by the radiotherapy department.

4.2 Shielding design assumptions

A radiotherapy facility that uses the assumption proposed above would produce effective dose values lower than the regulation statements for supervised, controlled and public areas. This is the result of the conservatively safe nature of the shielding design methodology recommended.

Some design assumption should be made:


- the minimum distance to the occupied area from a shielded wall (or barrier) is assumed to be 0,3 m (considered as representative for whole body exposure);
- attenuation of the primary beam by the patient should be neglected;
- calculations of recommended barrier thickness for primary barrier assume perpendicular incidence of the radiation except for robotic arm accelerator for which oblique incidence should be considered;
- leakage radiation from radiotherapy equipment is assumed to be at the maximum value recommended by IEC 60601-2-1[4] for the radiotherapy device, although in practice the leakage radiation is often less than this value;
- the primary barrier width (w) shall be sufficient to extend at least 0,3 m beyond the edge of a geometrical field of maximum size rotated 45°. For a square f m x f m field, the minimum value of w is given by:

$$w = (f / 1 \text{ m}) \sqrt{2} d_N + 0,6 \text{ m} \quad (1)$$

where d_N is the distance from the target to the far side of the most narrow part of the barrier, in metres.

The distance from the target to the narrow point of the primary barrier should be measured at the top of the barrier, not at the same height as isocentre.

[Figure 1](#) illustrates the typical locations for the narrowest point of the barrier.



Key

- 1 target
- 2 isocentre
- 3 metal

Figure 1 — Primary barrier width