



**SLOVENSKI STANDARD**  
**oSIST prEN 14625:2022**

**01-december-2022**

---

**Zunanji zrak - Standardna metoda za določanje koncentracije ozona z ultravijolično fotometrijo**

Ambient air - Standard method for the measurement of the concentration of ozone by ultraviolet photometry

Außenluft - Messverfahren zur Bestimmung der Konzentration von Ozon mit Ultraviolett-Photometrie

Air ambiant - Méthode normalisée de mesurage de la concentration en ozone par photométrie U.V.

**Ta slovenski standard je istoveten z: prEN 14625**

---

**ICS:**

13.040.20      Kakovost okoljskega zraka      Ambient atmospheres

**oSIST prEN 14625:2022**

**en,fr,de**



EUROPEAN STANDARD  
NORME EUROPÉENNE  
EUROPÄISCHE NORM

**DRAFT**  
**prEN 14625**

October 2022

ICS 13.040.20

Will supersede EN 14625:2012

English Version

## Ambient air - Standard method for the measurement of the concentration of ozone by ultraviolet photometry

Air ambiant - Méthode normalisée de mesurage de la concentration en ozone par photométrie U.V.

Außenluft - Messverfahren zur Bestimmung der Konzentration von Ozon mit Ultraviolett-Photometrie

This draft European Standard is submitted to CEN members for enquiry. It has been drawn up by the Technical Committee CEN/TC 264.

If this draft becomes a European Standard, CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

This draft European Standard was established by CEN in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.

Recipients of this draft are invited to submit, with their comments, notification of any relevant patent rights of which they are aware and to provide supporting documentation.

**Warning** : This document is not a European Standard. It is distributed for review and comments. It is subject to change without notice and shall not be referred to as a European Standard.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

## Contents

	Page
European foreword.....	7
<b>1 Scope .....</b>	<b>8</b>
<b>2 Normative references .....</b>	<b>9</b>
<b>3 Terms and definitions .....</b>	<b>9</b>
<b>4 Abbreviated terms .....</b>	<b>15</b>
<b>5 Principle .....</b>	<b>15</b>
5.1 General.....	15
5.2 Measuring principle .....	15
5.3 Type testing.....	17
5.4 Field operation and quality control.....	18
<b>6 Sampling.....</b>	<b>18</b>
6.1 General.....	18
6.2 Sampling location.....	18
6.3 Sampling system.....	19
6.3.1 Construction.....	19
6.3.2 Particle filter.....	19
6.3.3 Loss of ozone.....	20
6.3.4 Conditioning.....	20
6.4 Control and regulation of sample flow rate.....	20
6.5 Sampling pump for the manifold.....	20
<b>7 Analyser equipment.....</b>	<b>20</b>
7.1 General.....	20
7.2 Ultraviolet absorption cell.....	21
7.3 Ultraviolet source lamp .....	21
7.4 UV detector .....	21
7.5 Ozone-specific scrubber .....	21
7.6 Switching valve .....	21
7.7 Temperature indicator .....	21
7.8 Pressure indicator .....	22
7.9 Flow rate indicator .....	22
7.10 Sampling pump for the analyser.....	22
7.11 Internal ozone span source .....	22
7.12 Particle filter.....	22
<b>8 Type testing of ultraviolet photometric ozone analysers .....</b>	<b>22</b>
8.1 General.....	22
8.2 Relevant performance characteristics and performance criteria.....	23
8.3 Design change.....	26
8.4 Procedures for determination of the performance characteristics during the laboratory test.....	26
8.4.1 General.....	26
8.4.2 Test conditions.....	26
8.4.3 Response time .....	28
8.4.4 Short-term drift .....	30
8.4.5 Repeatability standard deviation.....	32

8.4.6	Lack of fit of linearity of the calibration function .....	32
8.4.7	Sensitivity coefficient to ambient gas pressure .....	34
8.4.8	Sensitivity coefficient to sample gas pressure .....	35
8.4.9	Sensitivity coefficient to the surrounding temperature .....	35
8.4.10	Sensitivity coefficient to electrical voltage .....	36
8.4.11	Interferents .....	37
8.4.12	Averaging test .....	37
8.4.13	Difference sample/calibration port .....	38
8.4.14	Residence time in the analyser .....	39
8.5	Determination of the performance characteristics during the field test .....	39
8.5.1	General .....	39
8.5.2	Selection of a monitoring station for the field test .....	39
8.5.3	Operational requirements .....	40
8.5.4	Long-term drift .....	41
8.5.5	Reproducibility standard deviation under field conditions .....	41
8.5.6	Period of unattended operation .....	42
8.5.7	Period of availability of the analyser .....	42
8.6	Type testing and uncertainty calculation .....	42
9	Field operation and ongoing quality control .....	43
9.1	General .....	43
9.2	Suitability evaluation .....	43
9.2.1	General .....	43
9.2.2	Analyser for a monitoring station or task .....	44
9.3	Initial installation .....	45
9.4	Ongoing quality assurance/quality control .....	46
9.4.1	General .....	46
9.4.2	Frequency of calibrations, checks and maintenance .....	47
9.5	Calibration of the analyser .....	50
9.5.1	General .....	50
9.5.2	Calibration gases .....	51
9.5.3	Data adjustment function .....	51
9.6	Checks .....	51
9.6.1	Zero and span checks .....	51
9.6.2	Lack of fit .....	53
9.6.3	Testing the sample manifold .....	53
9.6.4	Treatment of data after exceedance of performance criteria .....	54
9.7	Maintenance .....	55
9.7.1	Change of particle filters .....	55
9.7.2	Change of sampling lines .....	56
9.7.3	Change of consumables as applicable .....	56
9.7.4	Preventive/routine maintenance of components of the analyser .....	56
9.8	Data handling and data reports .....	56
9.9	Measurement uncertainty .....	56
10	Expression of results .....	57
11	Test reports and documentation .....	57
11.1	Type testing .....	57
11.2	Field operation .....	59
11.2.1	Suitability evaluation .....	59
11.2.2	Documentation .....	59
11.2.3	Ambient air quality data reports .....	59
Annex A (normative)	Test of lack of fit .....	60

## prEN 14625:2022 (E)

A.1	Establishment of the regression line .....	60
A.2	Calculation of the residuals of the averages.....	61
Annex B (informative) Sampling equipment.....		62
Annex C (informative) Ultraviolet photometric analyser.....		64
Annex D (informative) Manifold testing equipment .....		66
Annex E (normative) Type testing.....		68
E.1	Type testing and uncertainty calculation.....	68
E.1.1	Type testing.....	68
E.1.2	Uncertainty calculation.....	68
E.2	Type testing Requirement a) .....	68
E.3	Type testing Requirement b) .....	70
E.3.1	General.....	70
E.3.2	Calculation of standard uncertainties .....	72
E.3.2.1	Repeatability at zero .....	72
E.3.2.2	Repeatability at the hourly alert threshold value of ozone .....	73
E.3.2.3	Lack of fit.....	74
E.3.2.4	Influence quantities .....	74
E.3.3	Example calculation .....	79
E.4	Type testing Requirement c).....	80
E.5	Type testing Requirement d) .....	80
E.5.1	General.....	80
E.5.2	Combined standard uncertainty.....	82
E.5.3	Absolute expanded uncertainty.....	82
E.5.4	Relative expanded uncertainty .....	82
E.5.5	Calculation of standard uncertainties .....	83
E.5.6	Example calculation .....	85
Annex F (informative) Calculation of uncertainty in field operation at the hourly alert threshold value .....		87
F.1	General.....	87
F.2	Combined standard uncertainty.....	87
F.3	Standard uncertainties .....	88
F.3.1	General.....	88
F.3.1.1	Repeatability at zero.....	88
F.3.1.2	Repeatability at the hourly alert threshold value .....	88
F.3.1.3	Lack of fit.....	88
F.3.2	Influence quantities .....	89
F.3.2.1	General.....	89

F.3.2.2	Sample gas pressure .....	89
F.3.2.3	Sample gas temperature .....	89
F.3.2.4	Surrounding temperature .....	90
F.3.2.5	Electrical voltage.....	91
F.3.3	Interferents.....	91
F.3.3.1	Water vapour .....	91
F.3.3.2	Other interferents .....	91
F.3.4	Averaging effect.....	92
F.3.5	Reproducibility under field conditions .....	92
F.3.6	Long-term drift at zero .....	92
F.3.7	Long-term drift at level of the hourly alert threshold value .....	93
F.3.8	Zero gas .....	93
F.3.9	Calibration gas.....	93
F.3.10	Difference sample/calibration port.....	93
F.4	Example calculation.....	94
Annex G (informative)	Calculation of uncertainty in field operation at the 8-hour target value .....	96
G.1	General .....	96
G.2	Combined standard uncertainty .....	96
G.3	Standard uncertainties.....	97
G.3.1	General .....	97
G.3.1.1	Repeatability at zero .....	97
G.3.1.2	Repeatability at the 8-hour target value .....	98
G.3.1.3	Lack of fit.....	98
G.3.2	Influence quantities.....	98
G.3.2.1	General .....	98
G.3.2.2	Sample gas pressure.....	99
G.3.2.3	Sample gas temperature .....	99
G.3.2.4	Surrounding temperature .....	100
G.3.2.5	Electrical voltage.....	100
G.3.3	Interferents.....	101
G.3.3.1	Water vapour .....	101
G.3.3.2	Other interferents .....	101
G.3.4	Averaging effect.....	102
G.3.5	Zero gas .....	103
G.3.6	Calibration gas.....	103
G.3.7	Difference sample/calibration port.....	103

**prEN 14625:2022 (E)**

<b>G.3.8</b>	<b>Reproducibility under field conditions.....</b>	<b>103</b>
<b>G.3.9</b>	<b>Long-term drift at zero.....</b>	<b>104</b>
<b>G.3.10</b>	<b>Long-term drift at level of the 8-hour target value.....</b>	<b>104</b>
<b>G.4</b>	<b>Example calculation .....</b>	<b>105</b>
<b>Annex H (informative)</b>	<b>Test stand for the test point "sensitivity coefficient of sample gas pressure".....</b>	<b>107</b>
<b>Annex I (informative)</b>	<b>Significant changes.....</b>	<b>109</b>
<b>Bibliography.....</b>		<b>110</b>

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[oSIST prEN 14625:2022](https://standards.iteh.ai/catalog/standards/sist/cae07bb4-3481-4737-9068-13265082d5bc/osist-pren-14625-2022)

<https://standards.iteh.ai/catalog/standards/sist/cae07bb4-3481-4737-9068-13265082d5bc/osist-pren-14625-2022>



## European foreword

This document (prEN 14625:2022) has been prepared by Technical Committee CEN/TC 264 “Air quality”, the secretariat of which is held by DIN.

This document is currently submitted to the CEN Enquiry.

This document will supersede EN 14625:2012.

In comparison with the previous edition, the technical modifications listed in Annex I of this document have been made.

iTeh STANDARD PREVIEW  
(standards.iteh.ai)

<https://standards.iteh.ai/catalog/standards/sist/cae07bb4-3481-4737-9068-13265082d5bc/osist-pren-14625-2022>

**prEN 14625:2022 (E)****1 Scope**

This document specifies a continuous measurement method for the determination of the concentrations of ozone present in ambient air based on the ultraviolet photometric measuring principle. This document describes the performance characteristics and sets the relevant minimum criteria required to select an appropriate ultraviolet photometric analyser by means of type testing. It also includes the evaluation of the suitability of an analyser for use in a specific fixed site in order to meet the data quality requirements specified in Annex I of Directive 2008/50/EC [1] and requirements during sampling, calibration and quality assurance for use.

The method is applicable to the determination of the concentration of ozone present in ambient air up to 500 µg/m<sup>3</sup>. This concentration range represents the certification range for ozone for type testing.

NOTE 1 Other ranges may be used for measurement systems applied at rural locations monitoring ecosystems.

NOTE 2 When this document is used for other purposes than Directive 2008/50/EC, the ranges and uncertainty requirements may not apply.

The method covers the determination of ambient air concentrations of ozone in zones classified as rural areas, urban and urban-background areas.

The results are expressed in µg/m<sup>3</sup> (at 20 °C and 101,3 kPa).

NOTE 3 500 µg/m<sup>3</sup> of O<sub>3</sub> corresponds to 250 nmol/mol of O<sub>3</sub> at 20 °C and 101,3 kPa.

This document contains information for different groups of users.

Clause 5 to Clause 7 and Annex B and Annex C contain general information about the principles of ozone measurement by ultraviolet photometric analyser and sampling equipment.

Clause 8 and Annex E are specifically directed towards test houses and laboratories that perform type testing of ozone analysers. These sections contain information about:

- type testing conditions, test procedures and test requirements;
- analyser performance requirements;
- evaluation of the type testing results;
- evaluation of the uncertainty of the measurement results of the ozone analyser based on the type testing results.

Clause 9 to Clause 11 and Annex F and Annex G are directed towards monitoring networks performing the practical measurements of ozone in ambient air. These sections contain information about:

- initial installation of the analyser in the monitoring network and acceptance testing;
- ongoing quality assurance/quality control;
- calculation and reporting of measurement results;
- evaluation of the uncertainty of measurement results under practical monitoring conditions.

The present document represents an evolution of earlier editions (EN 14625:2005 and EN 14625:2012). It is recommended that when equipment is procured it complies fully with the present document.

NOTE 4 Type testing performed prior to the publication of this document for the purpose of demonstrating equivalence are still valid.

NOTE 5 Analysers type tested prior to the publication of this document remain valid for use for regulated monitoring purposes.

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 15267-1, *Air quality - Certification of automated measuring systems - Part 1: General principles*

EN 15267-2, *Air quality - Certification of automated measuring systems - Part 2: Initial assessment of the AMS manufacturer's quality management system and post certification surveillance for the manufacturing process*

EN ISO 6145-6, *Gas analysis - Preparation of calibration gas mixtures using dynamic methods - Part 6: Critical flow orifices (ISO 6145-6)*

EN ISO 6145-7, *Gas analysis - Preparation of calibration gas mixtures using dynamic methods - Part 7: Thermal mass-flow controllers (ISO 6145-7)*

EN ISO 14956, *Air quality - Evaluation of the suitability of a measurement procedure by comparison with a required measurement uncertainty (ISO 14956)*

## 3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

### 3.1

#### **adjustment**

#### **adjustment of a measuring system**

set of operations carried out on a measuring system so that it provides prescribed indications corresponding to given values of a quantity to be measured

Note 1 to entry: Types of adjustment of a measuring system include zero adjustment of a measuring system, offset adjustment, and span adjustment (sometimes called gain adjustment).

Note 2 to entry: Adjustment of a measuring system should not be confused with calibration, which is a prerequisite for adjustment.

Note 3 to entry: In the context of this document, adjustment is performed on measurement data rather than on the analyser.

[SOURCE: JCGM 200:2012 (VIM) [2]]

### 3.2

#### **alert threshold**

level beyond which there is a risk to human health from brief exposure for the population as a whole and at which immediate steps are to be taken by the Member States

[SOURCE: Council Directive 2008/50/EC [1]]

**prEN 14625:2022 (E)****3.3****ambient air**

outdoor air in the troposphere, where provisions concerning health and safety at work apply and to which members of the public do not have regular access

Note 1 to entry: This excludes workplaces as defined by Directive 89/654/EEC

[SOURCE: Council Directive 2008/50/EC [1]]

**3.4****analyser**

measuring system that provides an output signal which is a function of the concentration, partial pressure, flow or temperature of one or more components of a gas mixture

**3.5****availability of the analyser**

fraction of the total time period for which valid measuring data of the ambient air concentration is available from an analyser

**3.6****calibration**

operation that, under specified conditions, in a first step, establishes a relation between the quantity values with measurement uncertainties provided by measurement standards and corresponding indications with associated measurement uncertainties and, in a second step, uses this information to establish a relation for obtaining a measurement result from an indication

Note 1 to entry: A calibration may be expressed by a statement, calibration function, calibration diagram, calibration curve, or calibration table. In some cases, it may consist of an additive or multiplicative correction of the indication with associated measurement uncertainty.

Note 2 to entry: Calibration should not be confused with adjustment of a measuring system, often mistakenly called “self-calibration”, nor with verification of a calibration.

Note 3 to entry: Often, the first step alone in the above definition is perceived as being calibration.

Note 4 to entry: In the context of this document, calibration is a comparison of the analyser response to a known gas concentration with a known uncertainty when the information obtained from the comparison is used for the successive adjustment (if needed) of the analyser.

[SOURCE: JCGM 200, 2.39 [2]]

**3.7****certification range**

concentration range for which the analyser is type tested

**3.8****check**

verification that the analyser is still operating within specified performance limits

**3.9****combined standard uncertainty**

standard uncertainty of the result of a measurement when that result is obtained from the values of a number of other quantities, equal to the positive square root of a sum of terms, the terms being the variances or co-variances of these other quantities weighted according to how the measurement result varies with changes in these quantities

[SOURCE: JCGM 100: 2.3.4 [3]]

**3.10****coverage factor**

numerical factor used as a multiplier of the combined standard uncertainty in order to obtain an expanded uncertainty

[SOURCE: JCGM 100: 2.3.6 [3]]

**3.11****designated body**

body which has been designated for a specific task (type testing and/or QA/QC activities in the field) by the competent authority in the Member States

**3.12****detection limit**

smallest concentration of a measurand that can be reliably detected by a specific measurement process

Note 1 to entry: The detection limit is calculated as  $3,3 \times (s_z/B)$  where  $s_z$  is the standard deviation of analyser response at zero measurand concentration (see 8.4.5) and  $B$  is the slope of the calibration function [4].

**3.13****expanded uncertainty**

quantity defining an interval about the result of a measurement that may be expected to encompass a large fraction of the distribution of values that could reasonably be attributed to the measurand

Note 1 to entry: The fraction may be viewed as the coverage probability or level of confidence of the interval.

Note 2 to entry: To associate a specific level of confidence with the interval defined by the expanded uncertainty requires explicit or implicit assumptions regarding the probability distribution characterised by the measurement result and its combined standard uncertainty. The level of confidence that may be attributed to this interval can be known only to the extent to which such assumptions may be justified.

Note 3 to entry: For the purpose of this document, the expanded uncertainty is the combined standard uncertainty multiplied by a coverage factor  $k=2$  resulting in an interval with a level of confidence of 95 %.

[SOURCE: JCGM 100: 2.3.5 [3]]

**3.14****fall time**

difference between the response time (fall) and the lag time (fall)

**3.15****independent measurement**

individual measurement that is not influenced by a previous individual measurement by separating two individual measurements by at least four response times

Note 1 to entry: The largest value of response time (rise) and response time (fall) are intended.

**prEN 14625:2022 (E)****3.16****individual measurement**

measurement averaged over a time period equal to the response time of the analyser

Note 1 to entry: The largest value of response time (rise) and response time (fall) are intended.

Note 2 to entry: This definition differs from the meaning of the concept “individual measurement” in Directive 2008/50/EC [1].

**3.17****influence quantity**

quantity that is not the measurand but that affects the result of the measurement

[SOURCE: JCGM 100, B.2.10] [3]]

**3.18****interferent**

component of the air sample, excluding the measured constituent, that affects the output signal

**3.19****lack of fit**

maximum deviation from the linear regression line of the average of a series of measurement results at the same concentration

**3.20****lag time**

time interval from the moment at which a step change of sample concentration occurs at the inlet of the analyser to the moment at which the output reading reaches a level corresponding to a predefined change of the stable output reading

**3.21****long term drift**

difference between zero or span readings over a determined period of time (e.g. period of unattended operation)

**3.22****monitoring station**

enclosure located in the field in which an analyser has been installed to monitor concentrations of one or more ambient air pollutants

**3.23****parallel measurements**

measurements from different analysers, sampling from one and the same sampling manifold, starting at the same time and ending at the same time

**3.24****performance characteristic**

one of the parameters assigned to equipment in order to define its performance

**3.25****performance criterion**

limiting quantitative numerical value assigned to a performance characteristic, to which conformance is tested

**3.26****period of unattended operation**

time period over which the drift is within the performance criterion for long term drift

**3.27****repeatability (of results of measurement)**

closeness of the agreement between the results of successive individual measurements of ozone carried out under the same conditions of measurement

Note 1 to entry: These conditions are called laboratory repeatability conditions and include:

- a) the same measurement procedure;
- b) the same observer;
- c) the same analyser, used under the same conditions;
- d) at the same location;
- e) repetition over a short period of time.

**3.28****reproducibility under field conditions**

closeness of the agreement between the results of simultaneous measurements with two analysers in ambient air carried out under the same conditions of measurement

Note 1 to entry: These conditions are called field reproducibility conditions and include:

- a) the same measurement procedure;
- b) two identical analysers, used under the same conditions;
- c) at the same monitoring station;
- d) the period of unattended operation.

**3.29****residence time inside the analyser**

time period for the sampled air to be transported from the inlet of the analyser to the UV absorption cell

**3.30****residence time in the sampling system**

time period for the sampled air to be transferred to the inlet of the analyser

**3.31****response time**

time interval from the instant at which a step change of sample concentration occurs at the inlet of the analyser to the instant at which the output reading reaches a level corresponding to a predefined change of the output reading

**3.32****sampled air**

part of ambient air that is transferred through the sampling inlet and sampling system for subsequent measurement