



**SLOVENSKI STANDARD**  
**SIST EN 14625:2025**

**01-april-2025**

**Nadomešča:**  
**SIST EN 14625:2012**

---

**Zunanji zrak - Standardna metoda za določanje koncentracije ozona z ultravijolično fotometrijo**

Ambient air - Standard method for the measurement of the concentration of ozone by ultraviolet photometry

Außenluft - Messverfahren zur Bestimmung von Ozon in Luft mit dem UV-Verfahren

Air ambient - Méthode normalisée de mesure de la concentration en ozone par photométrie U.V.

**Ta slovenski standard je istoveten z: EN 14625:2024**

<https://standards.iteh.ai/catalog/standards/sist/cae07bb4-3481-4737-9068-13265082d5bc/sist-en-14625-2025>

**ICS:**

13.040.20      Kakovost okoljskega zraka      Ambient atmospheres

**SIST EN 14625:2025**

**en,fr,de**



EUROPEAN STANDARD

EN 14625

NORME EUROPÉENNE

EUROPÄISCHE NORM

December 2024

ICS 13.040.20

Supersedes EN 14625:2012

English Version

## Ambient air - Standard method for the measurement of the concentration of ozone by ultraviolet photometry

Air ambient - Méthode normalisée de mesurage de la concentration en ozone par photométrie U.V.

Außenluft - Messverfahren zur Bestimmung von Ozon in Luft mit dem UV-Verfahren

This European Standard was approved by CEN on 11 November 2024.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.

[SIST EN 14625:2025](https://standards.iteh.ai/catalog/standards/sist/cae07bb4-3481-4737-9068-13265082d5bc/sist-en-14625-2025)

<https://standards.iteh.ai/catalog/standards/sist/cae07bb4-3481-4737-9068-13265082d5bc/sist-en-14625-2025>



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

## Contents

	Page
European foreword.....	8
<b>1</b> <b>Scope</b> .....	<b>9</b>
<b>2</b> <b>Normative references</b> .....	<b>10</b>
<b>3</b> <b>Terms and definitions</b> .....	<b>10</b>
<b>4</b> <b>Abbreviated terms</b> .....	<b>16</b>
<b>5</b> <b>Principle</b> .....	<b>16</b>
5.1 <b>General</b> .....	<b>16</b>
5.2 <b>Measuring principle</b> .....	<b>16</b>
5.3 <b>Type testing</b> .....	<b>17</b>
5.4 <b>Field operation and quality control</b> .....	<b>18</b>
<b>6</b> <b>Sampling</b> .....	<b>18</b>
6.1 <b>General</b> .....	<b>18</b>
6.2 <b>Sampling location</b> .....	<b>19</b>
6.3 <b>Sampling system</b> .....	<b>19</b>
6.3.1 <b>Construction</b> .....	<b>19</b>
6.3.2 <b>Particle filter</b> .....	<b>19</b>
6.4 <b>Loss of ozone</b> .....	<b>20</b>
6.5 <b>Control and regulation of sample flow rate</b> .....	<b>20</b>
6.6 <b>Sampling pump for the manifold</b> .....	<b>20</b>
<b>7</b> <b>Analyser equipment</b> .....	<b>20</b>
7.1 <b>General</b> .....	<b>20</b>
7.2 <b>Ultraviolet absorption cell</b> .....	<b>21</b>
7.3 <b>Ultraviolet source lamp</b> .....	<b>21</b>
7.4 <b>UV detector</b> .....	<b>21</b>
7.5 <b>Ozone-specific scrubber</b> .....	<b>21</b>
7.6 <b>Switching valve</b> .....	<b>21</b>
7.7 <b>Temperature indicator</b> .....	<b>22</b>
7.8 <b>Pressure indicator</b> .....	<b>22</b>
7.9 <b>Flow rate indicator</b> .....	<b>22</b>
7.10 <b>Sampling pump for the analyser</b> .....	<b>22</b>
7.11 <b>Internal ozone span source</b> .....	<b>22</b>
7.12 <b>Particle filter</b> .....	<b>22</b>
<b>8</b> <b>Type testing of ultraviolet photometric ozone analysers</b> .....	<b>22</b>
8.1 <b>General</b> .....	<b>22</b>
8.2 <b>Relevant performance characteristics and performance criteria</b> .....	<b>23</b>
8.3 <b>Design change</b> .....	<b>27</b>
8.4 <b>Procedures for determination of the performance characteristics during the laboratory test</b> .....	<b>27</b>
8.4.1 <b>General</b> .....	<b>27</b>
8.4.2 <b>Test conditions</b> .....	<b>27</b>
8.4.3 <b>Response time</b> .....	<b>29</b>
8.4.4 <b>Short-term drift</b> .....	<b>31</b>
8.4.5 <b>Repeatability standard deviation</b> .....	<b>32</b>
8.4.6 <b>Lack of fit of linearity of the calibration function</b> .....	<b>32</b>
8.4.7 <b>Sensitivity coefficient to sample gas pressure</b> .....	<b>33</b>

8.4.8	Sensitivity coefficient to sample gas temperature .....	34
8.4.9	Sensitivity coefficient to the surrounding temperature.....	34
8.4.10	Sensitivity coefficient to electrical voltage.....	35
8.4.11	Interferents.....	36
8.4.12	Averaging test .....	36
8.4.13	Difference sample/calibration port.....	37
8.4.14	Residence time in the analyser.....	38
8.5	Determination of the performance characteristics during the field test.....	38
8.5.1	General .....	38
8.5.2	Selection of a monitoring station for the field test.....	38
8.5.3	Operational requirements .....	39
8.5.4	Long-term drift.....	40
8.5.5	Reproducibility standard deviation under field conditions .....	40
8.5.6	Period of unattended operation.....	41
8.5.7	Period of availability of the analyser.....	41
8.6	Type testing and uncertainty calculation .....	42
9	Field operation and ongoing quality control.....	42
9.1	General .....	42
9.2	Suitability evaluation .....	43
9.2.1	General .....	43
9.2.2	Analyser for a monitoring station or task.....	43
9.3	Initial installation .....	44
9.4	Ongoing quality assurance/quality control .....	46
9.4.1	General .....	46
9.4.2	Frequency of calibrations, checks and maintenance.....	47
9.5	Calibration of the analyser .....	50
9.5.1	General .....	50
9.5.2	Calibration gases.....	51
9.5.3	Data adjustment function .....	51
9.5.4	Testing the sampling system .....	51
9.5.5	Treatment of data after exceedance of performance criteria .....	53
9.6	Checks .....	55
9.6.1	Zero and span checks .....	55
9.6.2	Lack of fit.....	56
9.6.3	Testing the sample manifold .....	57
9.7	Maintenance .....	58
9.7.1	Change of particle filters .....	58
9.7.2	Maintenance of sampling system.....	58
9.7.3	Change of consumables as applicable .....	58
9.7.4	Preventive/routine maintenance of components of the analyser.....	58
9.8	Data handling and data reports.....	59
9.9	Measurement uncertainty .....	59
10	Expression of results .....	59
11	Test reports and documentation .....	60
11.1	Type testing .....	60
11.2	Field operation .....	61
11.2.1	Suitability evaluation .....	61
11.2.2	Documentation .....	61
11.2.3	Ambient air quality data reports .....	61
Annex A (normative)	Test of lack of fit .....	63
A.1	Establishment of the regression line.....	63

## EN 14625:2024 (E)

A.2	Calculation of the residuals of the averages.....	64
	Annex B (informative) Sampling equipment.....	65
	Annex C (informative) Ultraviolet photometric analyser.....	67
	Annex D (informative) Manifold testing .....	69
D.1	Procedure for applying test gas .....	69
D.2	Procedure for the cross test .....	70
D.2.1	General.....	70
D.2.2	Initial stage .....	70
D.2.3	Stage 1 .....	71
D.2.4	Stage 2 .....	71
D.2.5	Data Processing .....	71
D.2.6	Evaluation .....	71
	Annex E (normative) Type testing.....	73
E.1	Type testing and uncertainty calculation.....	73
E.1.1	Type testing.....	73
E.1.2	Uncertainty calculation.....	73
E.2	Type testing Requirement a) .....	73
E.3	Type testing Requirement b) .....	75
E.3.1	General.....	75
E.3.2	Calculation of standard uncertainties .....	78
E.3.2.1	General.....	78
E.3.2.2	Repeatability at zero .....	78
E.3.2.3	Repeatability at the hourly alert threshold value of ozone .....	78
E.3.2.4	Lack of fit.....	79
E.3.2.5	Influence quantities .....	79
E.3.2.5.1	..... General	79
E.3.2.5.2	..... Sample gas pressure.....	80
E.3.2.5.3	..... Sample gas temperature .....	80
E.3.2.5.4	..... Surrounding temperature .....	80
E.3.2.5.5	..... Electrical voltage.....	81
E.3.2.5.6	..... Water vapour .....	81

E.3.2.5.7.....	Other interferents.....	82
E.3.2.5.8.....	Averaging effect.....	83
E.3.2.5.9.....	Calibration gas.....	83
E.3.2.5.10.....	Difference sample/calibration port.....	83
E.3.3	Example calculation.....	84
E.4	Type testing Requirement c).....	85
E.5	Type testing Requirement d).....	85
E.5.1	General.....	85
E.5.2	Combined standard uncertainty.....	87
E.5.3	Absolute expanded uncertainty.....	87
E.5.4	Relative expanded uncertainty.....	87
E.5.5	Calculation of standard uncertainties.....	88
E.5.6	Example calculation.....	90
Annex F (informative)	Calculation of uncertainty in field operation at the hourly alert threshold value.....	92
F.1	General.....	92
F.2	Combined standard uncertainty.....	92
F.3	Standard uncertainties.....	93
F.3.1	General.....	93
F.3.1.1	Introduction.....	93
F.3.1.2	Repeatability at zero.....	93
F.3.1.3	Repeatability at the hourly alert threshold value.....	93
F.3.1.4	Lack of fit.....	93
F.3.2	Influence quantities.....	94
F.3.2.1	General.....	94
F.3.2.2	Sample gas pressure.....	94
F.3.2.3	Sample gas temperature.....	94
F.3.2.4	Surrounding temperature.....	95
F.3.2.5	Electrical voltage.....	96
F.3.3	Interferents.....	96
F.3.3.1	General.....	96
F.3.3.2	Water vapour.....	96
F.3.3.3	Other interferents.....	97
F.3.4	Averaging effect.....	97

## EN 14625:2024 (E)

F.3.5	Reproducibility under field conditions.....	97
F.3.6	Long-term drift at zero.....	97
F.3.7	Long-term drift at level of the hourly alert threshold value .....	98
F.3.8	Zero gas.....	98
F.3.9	Calibration gas .....	98
F.3.10	Difference sample/calibration port .....	98
F.4	Example calculation .....	99
<b>Annex G (informative) Calculation of uncertainty in field operation at the 8-hour target value.....</b>		<b>101</b>
G.1	General.....	101
G.2	Combined standard uncertainty.....	101
G.3	Standard uncertainties .....	102
G.3.1	General.....	102
G.3.1.1	Introduction .....	102
G.3.1.2	Repeatability at zero.....	102
G.3.1.3	Repeatability at the 8-hour target value.....	103
G.3.1.4	Lack of fit.....	103
G.3.2	Influence quantities .....	103
G.3.2.1	General.....	103
G.3.2.2	Sample gas pressure .....	104
G.3.2.3	Sample gas temperature.....	104
G.3.2.4	Surrounding temperature.....	105
G.3.2.5	Electrical voltage .....	105
G.3.3	Interferents .....	106
G.3.3.1	General.....	106
G.3.3.2	Water vapour.....	106
G.3.3.3	Other interferents.....	107
G.3.4	Averaging effect .....	108
G.3.5	Zero gas.....	108
G.3.6	Calibration gas .....	108
G.3.7	Difference sample/calibration port .....	108
G.3.8	Reproducibility under field conditions.....	108
G.3.9	Long-term drift at zero.....	109
G.3.10	Long-term drift at level of the 8-hour target value.....	109
G.4	Example calculation .....	110
<b>Annex H (informative) Test stand for the test point "sensitivity coefficient of sample gas pressure" .....</b>		<b>112</b>

**Annex I (informative) Significant changes..... 113**  
**Bibliography ..... 114**

**iTeh Standards**  
**(<https://standards.iteh.ai>)**  
**Document Preview**

[SIST EN 14625:2025](https://standards.iteh.ai/catalog/standards/sist/cae07bb4-3481-4737-9068-13265082d5bc/sist-en-14625-2025)

<https://standards.iteh.ai/catalog/standards/sist/cae07bb4-3481-4737-9068-13265082d5bc/sist-en-14625-2025>

## EN 14625:2024 (E)

### European foreword

This document (EN 14625:2024) has been prepared by Technical Committee CEN/TC 264 “Air quality”, the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by June 2025, and conflicting national standards shall be withdrawn at the latest by June 2025.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 14625:2012.

In comparison with the previous edition, the technical modifications listed in Annex I of this document have been made.

Any feedback and questions on this document should be directed to the users’ national standards body. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

(<https://standards.iteh.ai>)  
Document Preview

[SIST EN 14625:2025](https://standards.iteh.ai)

<https://standards.iteh.ai/catalog/standards/sist/cae07bb4-3481-4737-9068-13265082d5bc/sist-en-14625-2025>

## 1 Scope

This document specifies a continuous measurement method for the determination of the concentrations of ozone present in ambient air based on the ultraviolet photometric measuring principle. This document describes the performance characteristics and sets the relevant minimum criteria required to select an appropriate ultraviolet photometric analyser by means of type testing. It also includes the evaluation of the suitability of an analyser for use in a specific fixed site in order to meet the data quality requirements (see Annex I of Directive 2008/50/EC [1] for additional information) and requirements during sampling, calibration and quality assurance for use.

The method is applicable to the determination of the concentration of ozone present in ambient air up to 500 µg/m<sup>3</sup>. This concentration range represents the certification range for ozone for type testing.

NOTE 1 Other ranges may be used for measurement systems applied at rural locations monitoring ecosystems.

NOTE 2 When this document is used for other purposes than Directive 2008/50/EC, the ranges and uncertainty requirements may not apply.

The method covers the determination of ambient air concentrations of ozone in zones classified as rural areas, urban and urban-background areas.

The results are expressed in µg/m<sup>3</sup> (at 20 °C and 101,3 kPa).

NOTE 3 500 µg/m<sup>3</sup> of O<sub>3</sub> corresponds to 250 nmol/mol of O<sub>3</sub> at 20 °C and 101,3 kPa.

This document contains information for different groups of users.

Clause 5 to Clause 7 and Annex B and Annex C contain general information about the principles of ozone measurement by ultraviolet photometric analyser and sampling equipment.

Clause 8 and Annex E are specifically directed towards test houses and laboratories that perform type testing of ozone analysers. These sections contain information about:

- type testing conditions, test procedures and test requirements;
- analyser performance requirements;
- evaluation of the type testing results;
- evaluation of the associated uncertainty of the measurement performed by the ozone analyser based on the type testing results.

Clause 9 to Clause 11 and Annex F and Annex G are directed towards monitoring networks performing the practical measurements of ozone in ambient air. These sections contain information about:

- initial installation of the analyser in the monitoring network and acceptance testing;
- ongoing quality assurance/quality control;
- calculation and reporting of measurement results;
- evaluation of the uncertainty of measurement results under practical monitoring conditions.

This document represents an evolution of earlier editions (EN 14625:2005 and EN 14625:2012).

NOTE 4 Analysers type tested prior to the publication of this document can still be used for regulated monitoring purposes. As newer versions of analysers tested under this document become available, discontinue the use of older reference analysers.

## EN 14625:2024 (E)

### 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 15267-1, *Air quality - Assessment of air quality monitoring equipment - Part 1: General principles of certification*

EN 15267-2, *Air quality - Assessment of air quality monitoring equipment - Part 2: Initial assessment of the manufacturer's quality management system and post certification surveillance for the manufacturing process*

EN ISO 6145-6, *Gas analysis - Preparation of calibration gas mixtures using dynamic methods - Part 6: Critical flow orifices (ISO 6145-6)*

EN ISO 6145-7, *Gas analysis - Preparation of calibration gas mixtures using dynamic methods - Part 7: Thermal mass-flow controllers (ISO 6145-7)*

EN ISO 14956, *Air quality - Evaluation of the suitability of a measurement procedure by comparison with a required measurement uncertainty (ISO 14956)*

ISO 19229, *Gas analysis - Purity analysis and the treatment of purity data*

### 3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

#### 3.1

##### **adjustment of a measuring system**

set of operations carried out on a measuring system so that it provides prescribed indications corresponding to given values of a quantity to be measured

Note 1 to entry: Types of adjustment of a measuring system include zero adjustment of a measuring system, offset adjustment, and span adjustment (sometimes called gain adjustment).

Note 2 to entry: Adjustment of a measuring system should not be confused with calibration, which is a prerequisite for adjustment.

Note 3 to entry: In the context of this document, adjustment is performed on measurement data rather than on the analyser.

[SOURCE: JCGM 200:2012 (VIM) [2]]

#### 3.2

##### **alert threshold**

level beyond which there is a risk to human health from brief exposure for the population as a whole and at which immediate steps are to be taken by the Member States

[SOURCE: Council Directive 2008/50/EC [1]]

### 3.3

#### **ambient air**

outdoor air in the troposphere, where provisions concerning health and safety at work apply and to which members of the public do not have regular access

Note 1 to entry: This excludes workplaces as defined by Directive 89/654/EEC

[SOURCE: Council Directive 2008/50/EC [1]]

### 3.4

#### **analyser**

measuring system that provides an output signal which is a function of the concentration or partial pressure of one or more components in a gas mixture and flow or temperature of this gas mixture

### 3.5

#### **availability of the analyser**

fraction of the total time period for which valid measuring data of the ambient air concentration is available from an analyser

### 3.6

#### **calibration**

operation that, under specified conditions, in a first step, establishes a relation between the quantity values with measurement uncertainties provided by measurement standards and corresponding indications with associated measurement uncertainties and, in a second step, uses this information to establish a relation for obtaining a measurement result from an indication

Note 1 to entry: A calibration may be expressed by a statement, calibration function, calibration diagram, calibration curve, or calibration table. In some cases, it may consist of an additive or multiplicative correction of the indication with associated measurement uncertainty.

Note 2 to entry: Calibration should not be confused with adjustment of a measuring system, often mistakenly called “self-calibration”, nor with verification of a calibration.

Note 3 to entry: Often, the first step alone in the above definition is perceived as being calibration.

Note 4 to entry: In the context of this document, calibration is a comparison of the analyser response to a known gas concentration with a known uncertainty when the information obtained from the comparison is used for the successive adjustment (if needed) of the analyser.

[SOURCE: JCGM 200, 2.39 [2]]

### 3.7

#### **certification range**

concentration range for which the analyser is type tested

### 3.8

#### **check**

verification that the analyser is still operating within specified performance limits

**EN 14625:2024 (E)****3.9****combined standard uncertainty**

standard uncertainty of the result of a measurement when that result is obtained from the values of a number of other quantities, equal to the positive square root of a sum of terms, the terms being the variances or co-variances of these other quantities weighted according to how the measurement result varies with changes in these quantities

[SOURCE: JCGM 100: 2.3.4 [3]]

**3.10****coverage factor**

numerical factor used as a multiplier of the combined standard uncertainty in order to obtain an expanded uncertainty

[SOURCE: JCGM 100: 2.3.6 [3]]

**3.11****competent body**

organization which can demonstrate its competence for a specific task to the national competent authority of the Member State

**3.12****detection limit**

smallest concentration of a measurand that can be reliably detected by a specific measurement process

Note 1 to entry: The detection limit is calculated as  $3,3 \times (s_{r,z}/B)$  where  $s_{r,z}$  is the standard deviation of analyser response at zero measurand concentration (see 8.4.5) and  $B$  is the slope of the calibration function [4].

**3.13****expanded uncertainty**

quantity defining an interval about the result of a measurement that may be expected to encompass a large fraction of the distribution of values that could reasonably be attributed to the measurand

Note 1 to entry: The fraction may be viewed as the coverage probability or level of confidence of the interval.

Note 2 to entry: To associate a specific level of confidence with the interval specified by the expanded uncertainty requires explicit or implicit assumptions regarding the probability distribution characterised by the measurement result and its combined standard uncertainty. The level of confidence that may be attributed to this interval can be known only to the extent to which such assumptions may be justified.

Note 3 to entry: For the purpose of this document, the expanded uncertainty is the combined standard uncertainty multiplied by a coverage factor  $k=2$  resulting in an interval with a level of confidence of 95 %.

[SOURCE: JCGM 100: 2.3.5 [3]]

**3.14****fall time**

difference between the response time (fall) and the lag time (fall)

**3.15****independent measurement**

individual measurement that is not influenced by a previous individual measurement by separating two individual measurements by at least four response times

Note 1 to entry: The largest value of response time (rise) and response time (fall) are intended.