



SLOVENSKI STANDARD

SIST EN 14212:2025

01-april-2025

Nadomešča:

SIST EN 14212:2012

SIST EN 14212:2012/AC:2014

Zunanji zrak - Standardna metoda za določanje koncentracije žveplovega dioksida z ultravijolično fluorescenco

Ambient air - Standard method for the measurement of the concentration of sulphur dioxide by ultraviolet fluorescence

Außenluft - Messverfahren zur Bestimmung der Konzentration von Schwefeldioxid mit Ultraviolett-Fluoreszenz

Air ambiant - Méthode normalisée pour le mesurage de la concentration en dioxyde de soufre par fluorescence U.V.

[SIST EN 14212:2025](#)

<https://standards.iteh.ai/catalog/standards/sist/e35904cf-db5e-4cd0-b69d-e01b2d4acb77/sist-en-14212-2025>
Ta slovenski standard je istoveten z: [EN 14212:2024](#)

ICS:

13.040.20 Kakovost okoljskega zraka Ambient atmospheres

SIST EN 14212:2025

en,fr,de

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN 14212

December 2024

ICS 13.040.20

Supersedes EN 14212:2012

English Version

Ambient air - Standard method for the measurement of the concentration of sulphur dioxide by ultraviolet fluorescence

Air ambiant - Méthode normalisée pour le mesurage de la concentration en dioxyde de soufre par fluorescence U.V.

Außenluft - Messverfahren zur Bestimmung der Konzentration von Schwefeldioxid mit Ultraviolett-Fluoreszenz

This European Standard was approved by CEN on 11 November 2024.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.

SIST EN 14212:2025

<https://standards.iteh.ai/catalog/standards/sist/e35904af-db5c-4cd0-b69d-e01b2d4acb77/sist-en-14212-2025>



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents

	Page
European foreword	8
1 Scope	9
2 Normative references	10
3 Terms and definitions	11
4 Abbreviated terms	16
5 Principle	17
5.1 General	17
5.2 Measuring principle	17
5.3 Type testing	18
5.4 Field operation and quality control	18
6 Sampling	18
6.1 General	18
6.2 Sampling location	19
6.3 Sampling system	19
6.4 Control and regulation of sample flow rate	20
6.5 Sampling pump for the manifold	20
7 Analyser equipment	21
7.1 General	21
7.2 Selective traps for interfering agents	21
7.3 Optical assembly	21
7.4 Pressure measurement	21
7.5 Flow rate indicator	22
7.6 Sampling pump for the analyser	22
7.7 Internal sulfur dioxide span source	22
7.8 Particle filter	22
8 Type testing of ultraviolet fluorescence sulfur dioxide analysers	22
8.1 General	22
8.2 Relevant performance characteristics and criteria	23
Table 1 — Relevant performance characteristics and criteria	24
8.3 Design change	26
8.4 Procedures for determination of the performance characteristics during the laboratory test	26
Table 2 — Set points and stability of test parameters	27
Table 3 — Methods for preparation of test gases	27
Table 4a — Specification for purity of test gas	28
Table 4b — Specification for purity of zero gas for interferences testing [9]	28
Table 4c — Specification for purity of zero gas for other tests [9]	28
Figure 1 — Diagram illustrating the response time	30
Figure 2 — Concentration variations for the averaging effect test	37
8.5 Determination of the performance characteristics during the field test	38
8.6 Type testing and uncertainty calculation	42
9 Field operation and ongoing quality control	42

9.1 General	42
9.2 Suitability evaluation	43
Table 5 — Site-specific conditions to be evaluated	43
9.3 Initial installation	44
9.4 Ongoing quality assurance/quality control	46
Table 6 — Required frequency of calibration, checks and maintenance	47
9.5 Calibration of the analyser	49
Figure 3 — Flow scheme for performance of evaluation of effects of violation of performance requirements and possibilities for data correction	52
9.6 Checks	53
9.7 Maintenance	56
9.8 Data handling and data reports	57
9.9 Measurement uncertainty	57
10 Expression of results	58
11 Test reports and documentation	58
11.1 Type testing	58
11.2 Field operation	59
Annex A (normative) Test of lack of fit	61
A.1 Establishment of the regression line	61
A.2 Calculation of the residuals of the averages	61
Annex B (informative) Sampling equipment	63
Figure B.1 — Sampling layout with a main sampling manifold	63
Figure B.2 — Sampling layout with individual lines	64
Annex C (informative) Ultraviolet fluorescence analyser	65
Figure C.1 — Schematic diagram of a UV fluorescence sulfur dioxide analyser	65
Annex D (informative) Manifold testing	66
D.1 Procedure for applying test gas	66
Figure D.1 — Schematic diagram of manifold testing equipment	67
D.2 Procedure for the cross test	67
D.2.1 General	67
D.2.2 Initial stage	67
D.2.3 Stage 1	68
D.2.4 Stage 2	68
D.2.5 Data Processing	68
D.2.6 Evaluation	68
Annex E (normative) Type testing	70
E.1 Type testing and uncertainty calculation	70
E.1.1 Type testing	70
E.1.2 Uncertainty calculation	70
E.2 Type testing Requirement a)	70

EN 14212:2024 (E)

Table E.1 — Relevant performance characteristics and criteria	71
E.3 Type testing Requirement b)	72
E.3.1 General.....	72
Table E.2 — Standard uncertainties to be incorporated in the calculation of the expanded uncertainty after the laboratory tests	73
Table E.3 — Standard uncertainty of the calibration gas to be incorporated in the calculation of the expanded uncertainty after the laboratory tests.....	73
E.3.2 Calculation of standard uncertainties	75
E.3.2.1 General.....	75
E.3.2.2 Repeatability at zero.....	75
E.3.2.3 Repeatability at the hourly limit value of sulfur dioxide.....	75
E.3.2.4 Lack of fit.....	76
E.3.2.5 Influence quantities	76
E.3.2.5.1 General	76
E.3.2.5.2 Sample gas pressure	77
E.3.2.5.3 Sample gas temperature	77
E.3.2.5.4 Surrounding temperature	78
E.3.2.5.5 Electrical voltage.....	78
E.3.2.5.6 Water vapour	79
E.3.2.5.7 Other interferents.....	79
E.3.2.5.8 Averaging effect.....	80
E.3.2.5.9 Calibration gas	81
E.3.2.5.10 Difference sample/calibration port.....	81
E.3.3 Example calculation	82
Table E.4 —Example calculation.....	82
E.4 Type testing Requirement c).....	84
Table E.5 — Relevant performance characteristics and criteria	84
E.5 Type testing Requirement d)	84
E.5.1 General.....	84
Table E.6 — Standard uncertainties to be incorporated in the calculation of the expanded uncertainty after the laboratory and field tests.....	85
Table E.7 — Standard uncertainty of the calibration gas to be incorporated in the calculation of the expanded uncertainty after the laboratory and field tests	85
E.5.2 Combined standard uncertainty.....	86
E.5.3 Absolute expanded uncertainty.....	86
E.5.4 Relative expanded uncertainty	87
E.5.5 Calculation of standard uncertainties	87
E.5.6 Example calculation	89

Table E.8 —Example calculation	89
Annex F (informative) Calculation of uncertainty in field operation at the hourly limit value	91
F.1 General	91
F.2 Combined standard uncertainty	91
F.3 Standard uncertainties	92
F.3.1 General	92
F.3.1.1 General	92
F.3.1.2 Repeatability at zero	92
F.3.1.3 Repeatability at the hourly limit value	92
F.3.1.4 Lack of fit.....	92
F.3.2 Influence quantities.....	93
F.3.2.1 General	93
F.3.2.2 Sample gas pressure.....	93
F.3.2.3 Sample gas temperature	93
F.3.2.4 Surrounding temperature	94
F.3.2.5 Electrical voltage.....	95
F.3.3 Interferents.....	95
F.3.3.1 Water vapour	95
F.3.3.2 Other interferents	96
F.3.4 Averaging effect.....	96
F.3.5 Reproducibility under field conditions	96
F.3.6 Long-term drift at zero	97
F.3.7 Long-term drift at level of the hourly limit value	97
F.3.8 Zero gas	97
F.3.9 Calibration gas.....	97
F.3.10 Difference sample/calibration port.....	97
F.4 Example calculation.....	98
Table F.1 —Example calculation	98
Annex G (informative) Calculation of uncertainty in field operation at the daily limit value	100
G.1 General	100
G.2 Combined standard uncertainty	100
G.3 Standard uncertainties	101
G.3.1 General	101
G.3.1.1 General	101
G.3.1.2 Repeatability at zero	101
G.3.1.3 Repeatability at the daily limit value	102

EN 14212:2024 (E)

G.3.1.4 Lack of fit	102
G.3.2 Influence quantities	102
G.3.2.1 General.....	102
G.3.2.2 Sample gas pressure	103
G.3.2.3 Sample gas temperature.....	103
G.3.2.4 Surrounding temperature.....	104
G.3.2.5 Electrical voltage.....	104
G.3.3 Interferents	105
G.3.3.1 General.....	105
G.3.3.2 Water vapour.....	105
G.3.3.3 Other interferents.....	106
G.3.4 Averaging effect.....	107
G.3.5 Zero gas.....	107
G.3.6 Calibration gas	107
G.3.7 Difference sample/calibration port	107
G.3.8 Reproducibility under field conditions.....	108
G.3.9 Long-term drift at zero.....	108
G.3.10 Long-term drift at level of the daily limit value	108
G.4 Example calculation	109
Table G.1 —Example calculation	109
Annex H (informative) Calculation of uncertainty in field operation at the annual critical level.....	111
https://standards.iteh.ai/catalog/standards/sist/035904-ai-d03c-4cd0-009d-e0102d4ac077/sist-en-14212-2025	
H.1 General.....	111
H.2 Combined standard uncertainty.....	111
H.3 H.3 Standard uncertainties.....	113
H.3.1 General.....	113
H.3.2 Repeatability at zero.....	113
H.3.3 Repeatability at the annual limit value	113
H.3.4 Lack of fit	114
H.3.5 Influence quantities	114
H.3.5.1 General.....	114
H.3.5.2 Sample gas pressure	114
H.3.5.3 Sample gas temperature.....	115
H.3.5.4 Surrounding temperature.....	115
H.3.5.5 Electrical voltage	116
H.3.5.6 Interferents	117
H.3.5.7 Water vapour	117

H.3.5.8 Other interferents	117
H.3.6 Averaging effect.....	118
H.3.7 Zero gas	119
H.3.8 Calibration gas.....	119
H.3.9 Difference sample/calibration port.....	119
H.3.10 Reproducibility under field conditions	120
H.3.11 Long term drift at zero.....	120
H.3.12 Long term drift at level of the annual critical level	121
Annex I (informative) Test stand for the test point “sensitivity coefficient of sample gas pressure”.....	122
Figure I.1 — Test stand for the test point “sensitivity coefficient of sample gas pressure”.....	122
Annex J (informative) Significant technical changes	123
Table J.1 —Details of significant technical changes between this document and the previous edition.....	123
Bibliography	124

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[SIST EN 14212:2025](#)

<https://standards.iteh.ai/catalog/standards/sist/e35904af-db5c-4cd0-b69d-e01b2d4acb77/sist-en-14212-2025>

EN 14212:2024 (E)**European foreword**

This document (EN 14212:2024) has been prepared by Technical Committee CEN/TC 264 "Air quality", the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by June 2025, and conflicting national standards shall be withdrawn at the latest by June 2025.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 14212:2012.

The technical modifications in comparison with the previous edition are listed in Annex J of this document.

Any feedback and questions on this document should be directed to the users' national standards body. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Document Preview

[SIST EN 14212:2025](#)

<https://standards.iteh.ai/catalog/standards/sist/e35904af-db5c-4cd0-b69d-e01b2d4acb77/sist-en-14212-2025>

1 Scope

This document specifies a continuous measurement method for the determination of the concentration of sulfur dioxide present in ambient air based on the ultraviolet fluorescence measuring principle. This document describes the performance characteristics and sets the relevant minimum criteria required to select an appropriate ultraviolet fluorescence analyser by means of type testing. It also includes the evaluation of the suitability of an analyser for use in a specific fixed site so as to meet the data quality requirements (see Annex I of Directive 2008/50/EC [1] for additional information) and requirements during sampling, calibration and quality assurance for use.

The method is applicable to the determination of the mass concentration of sulfur dioxide present in ambient air up to 1000 µg/m³. This concentration range represents the certification range for sulfur dioxide for type testing.

NOTE 1 It is possible to use other ranges depending on the levels present in ambient air.

NOTE 2 Exemplar uncertainty budget calculations are given in Annexes E to H referring to Directive 2008/50/EC [1]. In the event that the Limit Values are updated in future iterations of Directive 2008/50/EC [1], the user can use these new values to calculate measurement uncertainties.

The method covers the determination of ambient air concentrations of sulfur dioxide in zones classified as rural areas, urban-background areas, traffic-oriented locations and locations influenced by industrial sources.

The results are expressed in µg/m³ (at 20 °C and 101,3 kPa).

NOTE 3 1 000 µg/m³ of SO₂ corresponds to 376 nmol/mol of SO₂.

This document contains information for different groups of users.

Clause 5 to Clause 7 and Annex C and Annex D contain general information about the principles of sulfur dioxide measurement by ultraviolet fluorescence analyser and sampling equipment.

Clause 8 and Annex E are specifically directed towards test houses and laboratories that perform type testing of sulfur dioxide analysers. These sections contain information about:

- type testing conditions, test procedures and test requirements;
- analyser performance requirements;
- evaluation of the type testing results;
- evaluation of the associated uncertainty of the measurement performed by the sulfur dioxide analyser based on the type testing results.

Clause 9 to Clause 11 and Annex F, Annex G and Annex H are directed towards monitoring networks performing the practical measurements of sulfur dioxide in ambient air. These sections contain information about:

- initial installation of the analyser in the monitoring network and acceptance testing;
- ongoing quality assurance/quality control;
- calculation and reporting of measurement results;
- evaluation of the uncertainty of the measurement results under practical monitoring conditions.

This document represents an evolution of earlier editions (EN 14212:2005 and EN 14212:2012).

EN 14212:2024 (E)

NOTE 4 Analysers type tested prior to the publication of this document can still be used for regulated monitoring purposes. As newer versions of analysers tested under this document become available, discontinue the use of older reference analysers.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 15267-1, *Air quality - Assessment of air quality monitoring equipment - Part 1: General principles of certification*

EN 15267-2, *Air quality - Assessment of air quality monitoring equipment - Part 2: Initial assessment of the manufacturer's quality management system and post certification surveillance for the manufacturing process*

EN ISO 6142, *Gas analysis — Preparation of calibration gas mixtures — Gravimetric method (ISO 6142)*

EN ISO 6143, *Gas analysis - Comparison methods for determining and checking the composition of calibration gas mixtures (ISO 6143)*

EN ISO 6144, *Gas analysis - Preparation of calibration gas mixtures - Static volumetric method (ISO 6144)*

EN ISO 6145-6, *Gas analysis - Preparation of calibration gas mixtures using dynamic methods - Part 6: Critical flow orifices (ISO 6145-6)*

EN ISO 6145-7, *Gas analysis - Preparation of calibration gas mixtures using dynamic methods - Part 7: Thermal mass-flow controllers (ISO 6145-7)*

EN ISO 6145-10, *Gas analysis - Preparation of calibration gas mixtures using dynamic volumetric methods - Part 10: Permeation method (ISO 6145-10)*

EN ISO 14956, *Air quality - Evaluation of the suitability of a measurement procedure by comparison with a required measurement uncertainty (ISO 14956)*

EN ISO 19229, *Gas analysis - Purity analysis and the treatment of purity data*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

3.1

adjustment of a measuring system

set of operations carried out on a measuring system so that it provides prescribed indications corresponding to given values of a quantity to be measured

Note 1 to entry: Types of adjustment of a measuring system include zero adjustment of a measuring system, offset adjustment, and span adjustment (sometimes called gain adjustment).

Note 2 to entry: Adjustment of a measuring system should not be confused with calibration, which is a prerequisite for adjustment.

Note 3 to entry: In the context of this document, adjustment is generally performed on measurement data rather than on the analyser.

[SOURCE: JCGM 200:2012, 3.11 [2]]

ITEH Standards
(<https://standards.iteh.ai>)
Document Preview

Note 1 to entry: This excludes workplaces as defined by Directive 89/654/EEC.

<https://standards.iteh.ai/04af-db5c-4cd0-b69d-e01b2d4acb77/sist-en-14212-2025>

3.2

ambient air

outdoor air in the troposphere where provisions concerning health and safety at work apply and to which members of the public do not have regular access

3.3

analyser

measuring system that provides an output signal which is a function of the concentration or partial pressure of one or more components in a gas mixture and flow or temperature of this gas mixture

3.4

availability of the analyser

fraction of the time period for which valid measuring data of the ambient air concentration is available from an analyser

EN 14212:2024 (E)**3.5****calibration**

operation that, under specified conditions, in a first step, establishes a relation between the quantity values with measurement uncertainties provided by measurement standards and corresponding indications with associated measurement uncertainties and, in a second step, uses this information to establish a relation for obtaining a measurement result from an indication

Note 1 to entry: A calibration may be expressed by a statement, calibration function, calibration diagram, calibration curve, or calibration table. In some cases, it may consist of an additive or multiplicative correction of the indication with associated measurement uncertainty.

Note 2 to entry: Calibration should not be confused with adjustment of a measuring system, often mistakenly called "self-calibration", nor with verification of a calibration.

Note 3 to entry: Often, the first step alone in the above definition is perceived as being calibration.

Note 4 to entry: In the context of this document, calibration is a comparison of the analyser response to a known gas concentration with a known uncertainty when the information obtained from the comparison is used for the successive adjustment (if needed) of the analyser.

[SOURCE: JCGM 200:2012, 2.39 [2]]

3.6**certification range**

concentration range for which the analyser is type tested

3.7**check**

verification that the analyser is still operating within specified performance limits

3.8**combined standard uncertainty**

standard uncertainty of the result of a measurement when that result is obtained from the values of a number of other quantities, equal to the positive square root of a sum of terms, the terms being the variances or co-variances of these other quantities weighted according to how the measurement result varies with changes in these quantities

[SOURCE: JCGM 100:2008, 2.3.4 [3]]

3.9**coverage factor**

numerical factor used as a multiplier of the combined standard uncertainty in order to obtain an expanded uncertainty

[SOURCE: JCGM 100:2008, 2.3.6 [3]]

3.10**competent body**

organization which can demonstrate its competence for a specific task to the national competent authority of the Member State