ISO 20391-2:2019(E)

ISO,20391-2:2019(E)

Deleted: /FDIS

Deleted: /FDIS

i

<u>2019-08</u>

ISO/TC 276/WG 3

Secretariat: DIN

Biotechnology — Cell counting — Part 2: Experimental design and statistical analysis to quantify counting method performance

Biotechnologie — Dénombrement des cellules — Partie 2: Conception expérimentale et analyse statistique pour quantifier les performances de la méthode de dénombrement

iTeh STANDARD PREVIEW (standards.iteh.ai)

<u>ISO 20391-2:2019</u>

https://standards.iteh.ai/catalog/standards/sist/fe480b56-54bf-4816-99d2-e2d873c15bfc/iso-20391-2-2019

IS	GQ_20391-2:2019(E)	 Deleted: /FDIS
	© ISO 2019	
	All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.	
	ISO copyright office	
	CP 401 • Ch. de Blandonnet 8	
	CH-1214 Vernier, Geneva	
	Phone: +41 22 749 01 11	
	Fax: +41 22 749 09 47	
	Email: copyright@iso.org	
	Website: <u>www.iso.org</u>	 Deleted: <u>www.iso.org</u>
	Published in Switzerland Teh STANDARD PREVIE	

(standards.iteh.ai)

ISO 20391-2:2019

https://standards.iteh.ai/catalog/standards/sist/fe480b56-54bf-4816-99d2-e2d873c15bfc/iso-20391-2-2019

--- Deleted: /FDIS

Contents

<u>Forew</u>	<u>ord</u>	ł
Introd	uction	
1	<u>Scope</u> <u>1</u>	
2	Normative references	
3	Terms definitions symbols and abbreviated terms	
<u>3.1</u>	Terms and definitions	l
3.2	List of abbreviated terms and symbols7	l
4	Principle 9	l
4.1	General	l
4.2	Proportionality	l
4.3	Deviation from proportionality	l
5	Experimental design	l
5.1	General11	l
5.2	Considerations for the cell counting measurement process	ł
5.3	Preparation of samples for the experimental design	l
<u>5.3.1</u>	General	l
<u>5.3.2</u>	Stock cell solution	l
<u>5.3.3</u>	Dilution fraction experimental design	l
<u>5.3.4</u>	Considerations for generating dilution fractions	l
<u>5.4</u>	Test sample labelling	l
5.5	Measurement of the test sample	l
6	Statistical methods	ľ
<u>6.1</u>	<u>General</u> <u>16</u>	l
<u>6.2</u>	Mean cell count 16	l
<u>6.3</u>	Measurement precision	l
<u>6.4</u>	Proportional model fit	l
0.5 6.6	10 10	l
<u>0.0</u> 6 6 1	Proportionality muck (Pr)	l
<u>0.0.1</u> 662	Calculation of the smoothed residual (emotion) 19	l
6.6.3	Calculation of proportionality index (P)	l
6.7	Additional statistical analysis and quality metrics	l
6.8	Data Interpretation	l
6.8.1	General 20	l
6.8.2	Interpretation of %CV	l
<u>6.8.3</u>	Interpretation of R ²	l
<u>6.8.4</u>	Interpretation of PI values	l
<u>6.8.5</u>	Comparison of PI values	1
7	Reporting	I
7.1	Reporting of quality indicators	1
7.2	Documentation of experimental design parameters and statistical analysis method22	1
7.3	Additional reporting elements on the cell counting measurement process	
Annex	A (informative) Method to assess pipetting error contributions to dilution integrity24	I
A.1	General 24	l
		L

Deleted: Error! Hyperlink reference not valid.¶ Introduction 6¶ Error! Hyperlink reference not valid.¶ 2 Normative references 1¶ Error! Hyperlink reference not valid.¶ 3.1 Terms and definitions 19 Error! Hyperlink reference not valid.¶ 4 Principle 9¶ Error! Hyperlink reference not valid.¶ 4.2 Proportionality 99 Error! Hyperlink reference not valid.¶ 5 Experimental design 119 Error! Hyperlink reference not valid.¶ 5.2 Considerations for the cell counting measurement process 11¶ Error! Hyperlink reference not valid.¶ 5.3.1 General 12¶ Error! Hyperlink reference not valid.¶ 5.3.3 Dilution fraction experimental design 13¶ Error! Hyperlink reference not valid.¶ 5.4 Test sample labelling 15¶ Error! Hyperlink reference not valid.¶ 6 Statistical methods 16¶ Error! Hyperlink reference not valid.¶ 6.2 Mean cell count 16¶ Error! Hyperlink reference not valid.¶ 6.4 Proportional model fit . 17¶ Error! Hyperlink reference not valid.¶ 6.6 Proportionality Index (PI) 18¶ Error! Hyperlink reference not valid.¶ 6.6.2 Calculation of the smoothed residual (esmoothed) 19¶ Error! Hyperlink reference not valid.¶ 6.7 Additional statistical analysis and quality metrics 20¶ Error! Hyperlink reference not valid.¶ 6.8.1 General 20¶ Error! Hyperlink reference not valid.¶ 6.8.3 Interpretation of R² 20¶ ... [1]

ISO 20)391-2:2019(E)	Deleted: /FDIS
<u>A.2</u>	Example procedure for pre-evaluating pipetting error contributions to dilution integrity using a calibrated scale to obtain accurate volume estimates upon pipetting24	
<u>A.3</u>	Example procedure for obtaining a measured DF during the evaluation of PI using a calibrated scale to obtain accurate volume estimates	
<u>Anne</u>	<u>x B (normative) Method to calculate smoothed residual (e_{smoothed}) when a set of measured dilution fractions (DF_{measured}) is obtained</u>	
Annex	<u>x C (informative) Example formulae for calculating PI</u>	
<u>C.1</u>	Detailed example of the calculation of <i>PI</i> based on sum of the absolute value of scaled smoothed residuals	
<u>C.2</u>	Detailed example of the calculation of <i>PI</i> based on <i>R</i> ² of smoothed residuals	
<u>C.3</u>	Additional examples of the calculation of <i>PI</i> when measured dilution fraction is utilized	
<u>Anne</u>	<u>x D (informative) Use case 1 — Evaluating the quality of a single cell counting measurement process</u>	
<u>D.1</u>	General	
<u>D.2</u>	Description of experimental design and statistical analysis	
D.3	Raw Data and Data Analysis for Use Case 1	
<u>D.4</u>	Example Report for Use Case 1	
<u>D.5</u>	Interpretation	
<u>Anne</u>	<u>x E (informative) Use case 2 — Comparing the quality of several cell counting</u> <u>measurement processes</u>	
<u>E.1</u>	<u>General</u>	
<u>E.2</u>	Description of experimental design and statistical analysis	
<u>E.3</u>	Raw Data and Data Analysis for Use Case 2	
<u>E.4</u>	Example Report for Use Case 2	
<u>E.5</u>	Comparison of cell counting methods evaluated in Use Case 2	
<u>E.5.1</u>	Comparing quality indicators between cell counting methods	
<u>E.6</u>	Interpretation	
<u>E.7</u>	Consistency of results with underlying simulated model	
<u>Biblio</u>	graphy	

© ISO 2019 - All rights reserved

ISO,20391-2:2019(E)	Deleted: /FDIS
Foreword	
ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.	
The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see <u>www.iso.org/directives</u>).	Deleted: <u>www.iso.org/directives</u>
Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).	Deleted: <u>www.iso.org/patents</u>
Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.	I
For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT),	
This document was prepared by Technical Committee ISO/TC 276, <i>Biotechnology</i> 14816, 9942, 2248	d.html
A list of all parts in the ISO 20391 series can be found on the ISO website.	
Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.	Deleted: <u>www.iso.org/members.ht</u>

v

Deleted: /FDIS

Introduction

Cell counting impacts many aspects of biotechnology, from biomanufacturing to medical diagnosis and advanced therapy. The cell count can serve as an in-process quality control or be used in decision-making. Cell count is also an important parameter in many cell-based assays, including activity and potency, which are often normalized to the cell count to allow data comparison.

Cell count is generally expressed as a concentration and can reflect the total cell count of a cell population (total cell count) or subpopulation (differential cell count). Advances in instrumentation have resulted in a wide range of cell counting techniques/instruments for total and/or differential cell counts. In the absence of a readily available reference material or ground truth, the accuracy of a measurement method has been difficult to ascertain. This has been confounded by the complexity of the biological preparation (e.g. cell type, sources, preparation, etc.). Several standards that address sector/application-specific cell counting or the use of a specific measurement system exist (See ISO 20391-1 and Reference [16] for further information). Some of these methods use a comparability approach whereby the result from a newer cell counting test method is traced to the results obtained from a more established cell counting method. While the comparability approach allows the data from the second instrument to be benchmarked against those obtained from a primary (more established) instrument, it does not address the quality of either measurement process^[17]. There remains a need to develop strategies to provide assurance for the quality of a cell counting measurement process in the absence of a reference material or reference method^[17].

This document provides a method for evaluating aspects of the quality of a cell counting measurement process through the use of a dilution series experimental design. From this experimental design, a set of quality indicators are derived to assess the performance of a cell counting measurement process. Specifically, the quality indicators assess precision and proportionality of cell counting measurement processes. This approach is particularly useful when accuracy cannot be determined (i.e. in the absence of a traceable reference method or traceable reference material) and is also relevant in aspects of validating and monitoring the quality of cell counting measurement processes in general^[17].

Information in this document is intended to provide confidence in the data produced by a chosen cell counting measurement process. This approach can be useful for selecting or optimizing a measurement process for a given cell preparation. This approach can also provide supporting performance parameters that can be utilized during performance qualification of a particular cell counting measurement process.

873c15bfc/iso

ISO 20391-2:2019(E)

Deleted: FINAL DRAFT
Deleted: /FDIS

Biotechnology — Cell counting — Part 2: Experimental design and statistical analysis to quantify counting method performance

1 Scope

This document provides a method for evaluating aspects of the quality of a cell counting measurement process for a specific cell preparation through a set of quality indicators derived from a dilution series experimental design and statistical analysis. The quality indicators are based on repeatability of the measurement and the degree to which the results conform to an ideal proportional response to dilution. This method is applicable to total, differential, direct and indirect cell counting measurement processes, provided that the measurement process meets the criteria of the experimental design (e.g. cells are suspended in a solution).

This method is most suitable during cell counting method development, optimization, validation, evaluation and/or verification of cell counting measurement processes.

This method is especially applicable in cases where an appropriate reference material to assess accuracy is not readily available. This method does not directly provide the accuracy of the cell count.

This method is primarily applicable to eukaryotic cells.

NOTE Several sector/application specific international and national standards for cell counting exist. Where applicable, consulting existing standards when operating within their scope can be helpful.

2 Normative references en ai/catalog/standards/sist/fe480b56-54bf-4816-99d2-e2d873c15bfc/iso-

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 20391-1, Biotechnology — Cell counting — Part 1: General guidance on cell counting methods

3 Terms, definitions, symbols and abbreviated terms

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <u>https://www.iso.org/obp</u>
- IEC Electropedia: available at.<u>http://www.electropedia.org/</u>____

3.1 Terms and definitions

3.1.1

accuracy

<measurement> closeness of agreement between a measured quantity value and a true quantity value of a measurand

© ISO 2019 - All rights reserved

p Deleted: <u>http://www.electropedi</u> a.org/

Deleted: https://www.iso.org/ob

Deleted: /FDIS

ISO_20391-2:2019(E)

[SOURCE: ISO/IEC Guide 99:2007, 2.13, modified — Notes deleted]

3.1.2

bias

<measurement> estimate of a systematic measurement error

[SOURCE: ISO/IEC Guide 99:2007, 2.18]

Note 1 to entry: Systematic measurement error is a component of measurement error that in replicate measurements remains constant or varies in a predictable manner. A reference quantity value for a systematic measurement error is a true quantity value, or a measured quantity value of a measurement standard of negligible measurement uncertainty, or a conventional quantity value.

Note 2 to entry: Also defined as the difference between the expectation of the test results and an accepted reference value (ISO 3534-1).

3.1.3 cell concentration cell count per volume

Note 1 to entry: Typically used for cells in suspension (e.g. cell number per ml).

Note 2 to entry: Cell concentration can refer to the total cell count or the count of a specific subset of cells within the volume (e.g. viable cell number per ml).

3.1.4 cell count discrete number of measured cells

Note 1 to entry: Cell count for cells in suspension is typically expressed as cell concentration.

3.1.5 https://standards.iteh.ai/catalog/standards/sist/fe480b56-54bf-4816-99d2-e2d873c15bfc/iso-

cell counting

20

measurement process to determine the cell count

3.1.6

cell suspension

single cells or aggregates of cells dispersed in a liquid matrix

3.1.7

debris

<cell suspensions> fragments of cells and/or particles of biological or non-biological origin

3.1.8

differential cell count

cell count of a subset of cells, which have been distinguished from other cell subpopulations by at least one distinct cell attribute identified in the measurement

Note 1 to entry: The concentrations derived from a differential cell count can be expressed in absolute concentration or as a relative measure (i.e. percentage) with respect to the total cell number or another predefined population.

3.1.9 dilution fraction

ISO/FDIS 20391-2:2019(E)

ratio by which the concentration of solute in a solution has been reduced from an original concentration

Note 1 to entry: Dilution fraction can range from 0 to 1.

Note 2 to entry: Dilution fraction is also sometimes referred to as "dilution ratio" or "dilution factor".

EXAMPLE The ratio by which the concentration of cells (solute) in a cell suspension (solution) has been reduced from a starting concentration of cells in suspension.

3.1.10

dilution series

group of solutions that have increasing or decreasing concentrations of the same substance

Note 1 to entry: A dilution series can be generated by serial dilution or by independent dilution.

Note 2 to entry: For a cell suspension, a dilution series is a group of suspensions that have increasing or decreasing concentrations of cells.

3.1.11

experimental design

process of planning a study to meet specified objectives

Note 1 to entry: Plan for assigning experimental conditions to participants and the statistical analysis associated with the plan. Typically, this includes a specification of the independent variables, dependent variables, number of participants and sampling strategy, procedure for assigning participants to experimental conditions, and order in which test tasks are given.

3.1.12

independent dilution

dilution series where each dilution is conducted independently of other dilutions

Note 1 to entry: Generally independent dilution series are generated directly from a common stock solution at a pre-specified (or target) dilution fraction.

3.1.13

intermediate precision

condition of measurement, out of a set of conditions that includes the same measurement procedure, same location, and replicate measurements on the same or similar objects over an extended period of time, but may include other conditions involving changes

Note 1 to entry: The changes can include new calibrations, calibrators, operators, and measuring.

Note 2 to entry: Operator bias refers specifically to error introduced by human operator experience.

[SOURCE: ISO/IEC Guide 99:2007, 2.22, modified — Note 3 deleted]

3.1.14 limit of quantitation

LOQ

<cell counting> lowest cell count in a sample that can be quantitatively determined with a suitable precision and accuracy using a specific analytical method

Note 1 to entry: The limit of quantification describes quantitative assay for low levels of cells in sample matrices.

3.1.15 linearity

Deleted: /FDIS

ISO 20391-2:2019(E)

within a given range, ability of an analytical procedure to obtain test results which are directly proportional to the concentration (amount) of analyte in the sample

[SOURCE: Reference (14), modified.]

Note 1 to entry: In cell counting the concentration of analyte refers to the concentration of cells (total or differential) in the sample.

Note 2 to entry: When a set of measurements exhibits linearity over a range of a given input (while all other inputs and measurement conditions are held constant), the expected value of the measurand can be expressed as the sum of a constant bias term and the input parameter multiplied by a fixed constant.

3.1.16 measurand quantity intended to be measured

[SOURCE: ISO/IEC Guide 99:2007, 2.3, modified — Notes and examples deleted.]

3.17

measured dilution fraction

dilution fraction verified by a traceable measurement

Note 1 to entry: For example, the volume of liquid can be verified by measuring the mass of the liquid (taking into density) using a calibrated and traceable scale with appropriate sensitivity.

3.1.18

measurement process

<cell counting> entire process for obtaining a cell count

Note 1 to entry: A measurement process can include sample preparation procedures, the measuring system, its settings (e.g. aperture choice, cell size gating, magnification, light exposure time etc.), and data analysis.

3.1.19

measurement precision

20391-2-2019

closeness of agreement between indications or measured quantity values obtained by replicate measurements on the same or similar objects under specified conditions

Note 1 to entry: Measurement precision is usually expressed numerically by measures of imprecision, such as standard deviation, variance, or coefficient of variation (CV) under the specified conditions of measurement.

Note 2 to entry: The 'specified conditions' can be, for example, repeatability conditions of measurement, intermediate precision conditions of measurement, or reproducibility conditions of measurement (see ISO 5725-1).

[SOURCE: ISO/IEC Guide 99:2007, 2.15, modified — Notes 3 and 4 deleted.]

3.1.20

proportionality

ability of an analytical procedure, irrespective of range, to obtain test results which are directly proportional to the concentration (amount) of analyte in the sample

Note 1 to entry: In cell counting the concentration of analyte refers to the concentration of cells (total or differential) in the sample.

Note 2 to entry: A collection of measurements exhibit proportionality with respect to a given input parameter when the ratio of the expected value of the measurement to the value of the input parameter at which the measurements were taken remains constant as the value of the input parameter changes (while all other inputs and measurement conditions are held constant).

ISO/FDIS 20391-2:2019(E)

Note 3 to entry: When a set of measurements exhibits proportionality over a range of a given input, then, Y = cX where Y, the expected value of the measurements is expressed as the input parameter (X) multiplied by a fixed constant (c), with no bias term.

3.1.21

proportionality constant

constant multiplier that directly relates the measurand to an input parameter

3.1.22

proportionality index

<cell counting> measure of deviation from proportionality for a dilution series experimental design

Note 1 to entry: The proportionality index (*PI*) is specific to the cell preparation and cell counting measurement process being evaluated.

3.1.23

p-value

output of a statistical hypothesis test

Note 1 to entry: The *p*-value is obtained in the following manner: The distribution of the test statistic under the assumption that the null hypothesis is true, called the null distribution, is determined. The *p*-value is computed from the null distribution as the probability of observing a test statistic that is as or more extreme than the test statistic obtained from the actual data.

3.1.24

quantity property of a phenomenon, body, or substance, where the property has a magnitude that can be expressed as a number and a reference

[SOURCE: ISO/IEC Guide 99:2007, 1.1, modified — Notes and example deleted.]

https://standards.iteh.ai/catalog/standards/sist/fe480b56-54bf-4816-99d2-e2d873c15bfc

3.1.25 range

20391-2-2019

quantity interval bounded by rounded or approximate extreme indications

3.1.26

reference material

reference standard

material, sufficiently homogeneous and stable with reference to specified properties, which has been established to be fit for its intended use in measurement or in examination of nominal properties

[SOURCE: ISO/IEC Guide 99:2007, 5.13, modified — Notes and examples deleted.]

3.1.27

reference method reference measurement procedure

measurement procedure accepted as providing measurement results fit for their intended use in assessing measurement trueness of measured quantity values obtained from other measurement procedures for quantities of the same kind, in calibration, or in characterizing reference materials

[SOURCE: ISO/IEC Guide 99:2007, 2.7]

3.1.28

repeatability

precision of the results of measurement under defined conditions of measurement

ISO 20391-2:2019(E)

Deleted: /FDIS

Note 1 to entry: Repeatability can also be considered as the closeness of the agreement between results of successive measurements of the same measurand carried out under the same conditions of the measurement^[17].

3.1.29 residual

<numerical analysis> numerical difference between the observed value of a dependent variable and the predicted value

3.1.30

sample

one or more parts taken from a system and intended to provide information on the system

Note 1 to entry: Often the sample serves as a basis for decision on the system or its production.

Note 2 to entry: For example, a smaller volume or aliquot of cell suspension taken from a larger volume of cell suspension^[17].

[SOURCE: ISO 15198:2004, 3.22, modified — "population" replaced by "system", Notes added.]

3.1.31

serial dilution

stepwise dilution of a substance in solution where the reduction of concentration is cumulative, lessening with each subsequent dilution

Note 1 to entry: In a serial dilution series, all dilutions except for the first are dependent on the preceding dilution.

3.1.32

stock cell solution

sufficiently stable (over time) cell suspension at sufficiently high concentration to allow dilution into working concentrations during experimentation

3.1.33

systematic error

20391-2-2019

component of measurement error that in replicate measurements remains constant or varies in a predictable manner

Note 1 to entry: A reference quantity value for a systematic measurement error is a true quantity value, or a measured quantity value of a measurement standard of negligible measurement uncertainty, or a conventional quantity value.

Note 2 to entry: Systematic measurement error, and its causes, can be known or unknown. A correction can be applied to compensate for a known systematic measurement error.

Note 3 to entry: Systematic measurement error equals measurement error minus random measurement error.

[SOURCE: ISO/IEC Guide 99:2007, 2.17]

3.1.34

target dilution fraction

dilution fraction that is trying to be achieved by diluting with a specified volume of solution

3.1.35

test sample

small aliquot of the sample that is prepared for measurement in the method of interest

ISO/FDIS 20391-2:2019(E)

Note 1 to entry: Generally, test samples are representative of the sample they are prepared from and are sometimes referred to as "representative test sample(s)".

3.1.36

total cell count cell count of all cells, independent of the attribute(s) of the cell

3.1.37 true count true quantity value quantity value consistent with the definition of a quantity

Note 1 to entry: In the error approach to describing measurement, a true quantity value is considered unique and, in practice, unknowable. The uncertainty approach is to recognize that, owing to the inherently incomplete amount of detail in the definition of a quantity, there is not a single true quantity value but rather a set of true quantity values consistent with the definition. However, this set of values is, in principle and in practice, unknowable. Other approaches dispense altogether with the concept of true quantity value and rely on the concept of metrological compatibility of measurement results for assessing their validity.

Note 2 to entry: In the special case of a fundamental constant, the quantity is considered to have a single true quantity value.

Note 3 to entry: When the definitional uncertainty associated with the measurand is considered to be negligible compared to the other components of the measurement uncertainty, the measurand can be considered to have an "essentially unique" true quantity value. This is the approach taken by the ISO/IEC Guide 98-3 and associated documents, where the word "true" is considered to be redundant.

[SOURCE: ISO/IEC Guide 99:2007, 2.11, modified — "GUM" replaced by "ISO/IEC Guide 98-3".]

3.1.38 validation

SO 20391-2:2019

confirmation, through the provision of objective evidence, that the requirements for a specific intended 873c15bfc/isouse or application have been fulfilled 20391-2-2019

[SOURCE: ISO 9000:2015, 3.8.13, modified — Notes deleted.]

3.1.39

variability quantification of probability distribution function for variable, parameter, or condition

[SOURCE: ISO 16732-1:2012, 3.29]

3.2 List of abbreviated terms and symbols

List of abbreviations in order of citation.

Abbreviated term or symbol	Description
β	proportionality constant that can differ from $k_{ m ideal}$
β_1	scalar coefficient estimated from the proportional model fitting
CV	coefficient of variation
CV _{ij}	coefficient of variation for a set of K_{ij} repeated observations of representative test sample j , at target dilution fraction df_i
$\%\overline{CV}_{df_i}$	mean percent CV for a set of n_i representative test samples, with target dilution

ISO₂20391-2:2019(E)

Deleted: /FDIS

	fractions <i>df</i> _i	
Cideal	ideal proportionality constant	
tc _j	theoretical/true count of sample <i>j</i>	
DF	dilution fraction	
DFj	dilution fraction of sample <i>j</i> controlled by the measurement process or determined experimentally	
tDF	theoretical/true dilution fraction	
DF	set of unique target dilution fractions	
DF measured	set of measured dilution fractions	
df _i	targeted dilution fraction	
df _{ij}	measured dilution fraction	
e	residual between data and modelled fit	
e ^{smoothed}	smoothed residual between processed cell count and proportional model fit	
e_i^{smoothed}	smoothed residual when target dilution fraction is used in the analysis of proportionality (smoothed residual at each target DF)	
$e_{ij}^{ m smoothed}$	smoothed residual when measured dilution fraction is used in the analysis of proportionality (smoothed residual for each representative test sample)	
$E(oc_j)$	expected value of observed counts	
i	index for target dilution fraction	
i	index for replicate representative test sample	
k	index for replicate measurement made on a representative test sample	
K _{ij}	number of repeated measurements of the representative test sample	
nups	number of target dilution fractions	
n _i	number of replicate representative test samples at the target dilution fraction	
PI	proportionality index	
R ²	coefficient of determination	
Y _{ijk}	observed value from measurement <i>k</i> of representative test sample <i>j</i> at target dilution fraction <i>i</i>	
\overline{Y}_{df_i}	mean cell count for a set of n_i representative test samples, with target dilution fractions (df_i)	
\overline{Y}_{ij} .	mean over the set of K_{ij} repeated observations for the j^{th} representative test sample of df_i	
-	mean of \overline{Y}_{ii} , over independent representative test samples <i>j</i> for a set of n_i replicate	
Y	representative test samples	
$\lambda_{DF_k}^{\mathrm{proportional}}$	estimated cell count at DF_k using β_1 obtained from proportional model fit $\lambda_i^{\text{proportional}}$	
$\lambda_{ij}^{\text{proportional}}$	proportional model fit to \overline{Y}_{ij} versus df_{ij}	

4 Principle

4.1 General

Achieving high confidence in cell counting implies that the measurement is both accurate and precise^[15]. For a well-controlled dilution fraction series, the concept of proportionality may be used as an internal reference and deviation from proportionality can serve as an alternative to the direct evaluation of accuracy^[16]. Specifically, using experimental design and statistical analysis, quality indicators that describe deviation from proportionality and coefficient of variation (*CV*) can be evaluated to assess aspects of the quality of a cell counting measurement process.

The quality indicators evaluate the overall quality of a cell counting measurement process, where the measurement process includes sample preparation and handling, data acquisition, and data processing/correction.

Accuracy is ideally evaluated using a reference method and/or reference material with a known "true" value (see ISO 5725-1 and ISO 5725-2 for further information). In the absence of an appropriate reference material or reference method, the quality of a cell counting measurement can be indirectly assessed through its adherence to or deviation from the fundamental principle of proportionality, which implies that the measured cell count shall be proportional to the dilution fraction (DF) under ideal experimental conditions. Deviation from proportionality would indicate that a systematic measurement error has occurred to reduce the overall measurement confidence. This approach however does not directly provide the accuracy of the cell count.

The precision of a cell counting measurement indicates the closeness of agreement between cell counts obtained by replicate measurements on the same or similar cell preparation under specified conditions. Experimental data with low precision but with average cell counts fitting well to proportionality would reduce the quality of the measurement process. Importantly, low measurement precision (i.e. large random measurement error) can mask deviations from proportionality.

4.2 Proportionality

20391-2-2019

The theoretical true counts of samples extracted from a common, ideally homogenized, stock solution are related by their respective dilution fractions in accordance with the expression shown in Formula (1):

$$tc_{i} = c_{ideal} \times tDF_{i} \tag{1}$$

where

tc_j is the theoretical true count for the sample *j*;

*c*_{ideal} is an unknown proportionality constant equal to the theoretical true count for an undiluted sample;

 tDF_j is the true dilution fraction for sample *j*.

By rigorously controlling the dilution fraction, the theoretical tDF_j may be approximated by DF_j . See Formula (2):

$$tDF_i \cong DF_i \tag{2}$$

where DF_j is the dilution fraction of sample *j* controlled by the measurement process or determined experimentally.