



SLOVENSKI STANDARD SIST EN ISO 81060-3:2023

01-april-2023

Neinvazivni sfigmomanometri - 3. del: Klinična validacija kontinuirnih avtomatiziranih merilnikov krvnega tlaka (ISO 81060-3:2022)

Non-invasive sphygmomanometers - Part 3: Clinical investigation of continuous automated measurement type (ISO 81060-3:2022)

Nicht-invasive Blutdruckmessgeräte - Teil 3: Klinische Prüfung der kontinuierlichen automatisierten Bauart (ISO 81060-3:2022)

Sphygmomanomètres non invasifs - Partie 3: Investigation clinique pour type à mesure automatique continu (ISO 81060-3:2022)

Ta slovenski standard je istoveten z: EN ISO 81060-3:2023

ICS:

11.040.55	Diagnostična oprema	Diagnostic equipment
-----------	---------------------	----------------------

SIST EN ISO 81060-3:2023

en

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN ISO 81060-3

January 2023

ICS 11.040.10

English Version

Non-invasive sphygmomanometers - Part 3: Clinical
investigation of continuous automated measurement type
(ISO 81060-3:2022)

Sphygmomanomètres non invasifs - Partie 3:
Investigation clinique pour type à mesurage
automatique continu (ISO 81060-3:2022)

Nicht-invasive Blutdruckmessgeräte - Teil 3: Klinische
Prüfung der kontinuierlichen automatisierten Bauart
(ISO 81060-3:2022)

This European Standard was approved by CEN on 20 October 2022.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents

Page

European foreword.....	3
------------------------	---

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 81060-3:2023

<https://standards.iteh.ai/catalog/standards/sist/3a78279b-1dda-4b64-964d-8838f18f411a/sist-en-iso-81060-3-2023>

European foreword

This document (EN ISO 81060-3:2023) has been prepared by Technical Committee ISO/TC 121 "Anaesthetic and respiratory equipment" in collaboration with Technical Committee CEN/TC 205 "Non-active medical devices" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by July 2023, and conflicting national standards shall be withdrawn at the latest by July 2023.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Endorsement notice

The text of ISO 81060-3:2022 has been approved by CEN as EN ISO 81060-3:2023 without any modification.

<https://standards.iteh.ai/catalog/standards/sist/3a78279b-1dda-4b64-964d-8838f18f411a/sist-en-iso-81060-3-2023>

INTERNATIONAL STANDARD

ISO
81060-3

First edition
2022-12

Non-invasive sphygmomanometers — Part 3: Clinical investigation of continuous automated measurement type

Sphygmomanomètres non invasifs —

*Partie 3: Investigation clinique pour type à mesurage automatique
continu*

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 81060-3:2023

<https://standards.iteh.ai/catalog/standards/sist/3a78279b-1dda-4b64-964d-8838f18f411a/sist-en-iso-81060-3-2023>



Reference number
ISO 81060-3:2022(E)

© ISO 2022

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO 81060-3:2023

<https://standards.iteh.ai/catalog/standards/sist/3a78279b-1dda-4b64-964d-8838f18f411a/sist-en-iso-81060-3-2023>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2022

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword.....	iv
Introduction.....	v
1 Scope.....	1
2 Normative references.....	1
3 Terms and definitions.....	2
4 General requirements for the <i>clinical investigation</i>.....	4
4.1 Good clinical practice.....	4
4.2 General.....	4
4.3 <i>Reference method</i>	5
4.3.1 <i>Reference invasive blood pressure monitoring equipment</i>	5
4.3.2 Subject requirements.....	5
4.3.3 <i>Blood pressure distribution</i>	7
4.3.4 <i>Arterial reference site</i>	8
4.4 Lateral difference.....	9
4.4.1 General.....	9
4.4.2 <i>Sequential procedure</i>	10
4.4.3 <i>Simultaneous procedure</i>	10
4.4.4 Application of lateral difference.....	11
4.5 Statistical considerations.....	11
4.5.1 General.....	11
4.5.2 Intra-class correlation coefficient and number of independent measurements.....	12
4.5.3 <i>Procedure to derive the number of subjects and number of repeated paired measurements per subject</i>	12
5 Methods for <i>clinical investigation</i>.....	13
5.1 Method for the accuracy of <i>blood pressure determination</i>	13
5.1.1 General.....	13
5.1.2 <i>Procedure</i>	13
5.1.3 Data analysis.....	13
5.1.4 Acceptance criteria.....	14
5.2 Method for stability.....	16
5.2.1 General.....	16
5.2.2 <i>Procedure</i>	17
5.2.3 Data analysis.....	17
5.2.4 Acceptance criteria.....	19
5.3 Method for <i>blood pressure changes</i>	19
5.3.1 <i>Change evaluation interval</i>	19
5.3.2 General.....	20
5.3.3 <i>Procedure</i>	20
5.3.4 Data analysis.....	20
5.3.5 Acceptance criteria.....	22
Annex A (informative) Rationale and guidance.....	23
Bibliography.....	35
Terminology — alphabetized index of defined terms.....	36

ISO 81060-3:2022(E)

Foreword

ISO (the International Organization for Standardization) and IEC (the International Electrotechnical Commission) form the specialized system for worldwide standardization. National bodies that are members of ISO or IEC participate in the development of International Standards through technical committees established by the respective organization to deal with particular fields of technical activity. ISO and IEC technical committees collaborate in fields of mutual interest. Other international organizations, governmental and non-governmental, in liaison with ISO and IEC, also take part in the work.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives or www.iec.ch/members_experts/refdocs).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO and IEC shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents) or the IEC list of patent declarations received (see <https://patents.iec.ch>).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html. In the IEC, see www.iec.ch/understanding-standards.

This document was prepared jointly by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 3, *Respiratory devices and related equipment used for patient care*, and Technical Committee IEC/TC 62, *Electrical equipment in medical practice*, Subcommittee SC 62D, *Electromedical equipment*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 205, *Non-active medical devices*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

A list of all parts in the ISO 81060 series can be found on the ISO and IEC websites.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html and www.iec.ch/national-committees.

Introduction

The number of continuously measuring non-invasive *automated sphygmomanometers* has increased significantly in the last 10 years. This document is intended to provide the necessary requirements for *clinical investigation* to ensure that the *essential performance* of these *sphygmomanometers* is at an adequate level, similar to those standards on *intermittent automated non-invasive sphygmomanometer*.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 81060-3:2023

<https://standards.iteh.ai/catalog/standards/sist/3a78279b-1dda-4b64-964d-8838f18f411a/sist-en-iso-81060-3-2023>

Non-invasive sphygmomanometers —

Part 3: Clinical investigation of continuous automated measurement type

1 Scope

This document specifies the requirements and methods for the *clinical investigation of continuous automated non-invasive sphygmomanometers* used for the measurement of the *blood pressure* of a patient.

This document does not cover usability aspects such as the form and manner of the data display or output. This document does not specify a numerical threshold on the *minimum output period*. A *continuous automated non-invasive sphygmomanometer* providing *blood pressure* parameters (e.g., *systolic blood pressure*, *diastolic blood pressure* or *mean arterial pressure*) with an *output period* considerably larger than 30 s is not typically considered a *continuous automated non-invasive sphygmomanometer*.

This document covers both trending *continuous automated non-invasive sphygmomanometers* and absolute accuracy *continuous automated non-invasive sphygmomanometers* and focuses solely on requirements for the *clinical investigation*. Representation of output is not covered by this document.

NOTE 1 IEC 62366-1 provides requirements on the application of usability engineering to medical devices. The usability engineering *process* can be used to clarify for the intended user whether the displayed data concerns absolute accurate values or trending values.

The requirements and methods for the *clinical investigation of continuous automated non-invasive sphygmomanometers* provided in this document are applicable to any subject population, and any condition of use of the *continuous automated non-invasive sphygmomanometers*.

NOTE 2 Subject populations can, for example, be represented by age or weight ranges.

NOTE 3 This document does not provide a method to assess the effect of artefacts during the *clinical investigation* (e.g. motion artefacts induced by the movement of the subject or the movement of the platform supporting the subject).

This document specifies additional disclosure requirements for the *accompanying documents of continuous automated non-invasive sphygmomanometers* that have undergone *clinical investigation* according to this document.

This document is not applicable to:

- the *clinical investigation* of a *non-automated sphygmomanometer* as given in ISO 81060-1,
- the *clinical investigation* of an *intermittent automated non-invasive sphygmomanometer* as given in ISO 81060-2,
- an *automated non-invasive sphygmomanometer* as given in IEC 80601-2-30, or
- *invasive blood pressure monitoring equipment* as given in IEC 60601-2-34.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.