



SLOVENSKI STANDARD

SIST ISO 20456:2019

01-januar-2019

Meritve pretoka tekočin v zaprtih cevovodih - Navodilo za uporabo elektromagnetnih merilnikov pretoka za prevodne tekočine

Measurement of fluid flow in closed conduits -- Guidance for the use of electromagnetic flowmeters for conductive liquids

iTeh STANDARD PREVIEW

Mesurage du débit des fluides dans les conduites fermées -- Lignes directrices pour l'utilisation des débitmètres électromagnétiques dans les liquides conducteurs

[SIST ISO 20456:2019](https://standards.iteh.ai/catalog/standards/sist/76ad5513-d42a-4dcf-aed0-b01887522af/sist-iso-20456-2019)

Ta slovenski standard je istoveten z:

ISO 20456:2017

ICS:

17.120.10

Pretok v zaprtih vodih

Flow in closed conduits

SIST ISO 20456:2019

en,fr

iTeh STANDARD PREVIEW **(standards.iteh.ai)**

SIST ISO 20456:2019

<https://standards.iteh.ai/catalog/standards/sist/76ad5513-d42a-4dcf-aed0-bbf8889522af/sist-iso-20456-2019>

INTERNATIONAL STANDARD

ISO
20456

First edition
2017-09

Measurement of fluid flow in closed conduits — Guidance for the use of electromagnetic flowmeters for conductive liquids

*Mesurage du débit des fluides dans les conduites fermées — Lignes
directrices pour l'utilisation des débitmètres électromagnétiques dans
les liquides conducteurs*

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST ISO 20456:2019

<https://standards.iteh.ai/catalog/standards/sist/76ad5513-d42a-4dcf-aed0-bbf8889522af/sist-iso-20456-2019>



Reference number
ISO 20456:2017(E)

© ISO 2017

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST ISO 20456:2019

<https://standards.iteh.ai/catalog/standards/sist/76ad5513-d42a-4dcf-aed0-bbf8889522af/sist-iso-20456-2019>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2017, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Symbols	3
5 Theory and basic formulae	4
6 Construction and principle of operation	4
6.1 General	4
6.2 Sensor	5
6.3 Transmitter	7
6.3.1 General	7
6.3.2 Alternating magnetic field in the measuring system	7
6.3.3 Measuring system with applied pulsed DC excitation (simplified model)	7
6.3.4 Measuring system with applied AC excitation (simplified model)	8
6.4 Flowmeter/Transmitter output	9
7 Equipment marking	9
7.1 Recommended data	9
7.1.1 Sensor	9
7.1.2 Transmitter	10
8 Installation design and practice	10
8.1 Sensor	10
8.1.1 Sizing	10
8.1.2 Mounting conditions	11
8.1.3 Potential equalization — General requirements	12
8.1.4 Electrical connections	13
8.1.5 Sensor mounting	13
8.1.6 Installation dimensions for flanged connections	14
8.2 Transmitter location	15
8.3 Operational considerations	16
8.3.1 General	16
8.3.2 Effect of the liquid conductivity	16
8.3.3 Reynolds number effect	16
8.3.4 Velocity profile effect	16
9 Flowmeter calibration, validation, and verification	16
9.1 Flowmeter calibration	16
9.2 Flowmeter verification (<i>in-situ</i> electronic verification)	16
10 Evaluation of flowmeter performance	17
10.1 General	17
10.2 Applications within the scope of other standards	17
11 Uncertainty analysis	17
Annex A (informative) Materials for construction of sensors	19
Annex B (informative) Practical considerations for measuring system with AC and DC excitation	22
Annex C (informative) Cathodic protection	23
Annex D (informative) Conversion of nominal diameters from metric to US units	24
Annex E (informative) Manufacturers' accuracy specifications	25
Bibliography	29

ISO 20456:2017(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html. (standards.iteh.ai)

This document was prepared by Technical Committee ISO/TC 30, *Measurement of fluid flow in closed conduits*, Subcommittee SC 5, *Velocity and mass methods*.

This first edition of ISO 20456 cancels and replaces ISO 6817:1992, ISO 9104:1991 and ISO 13359:1998, which has been technically revised.

Introduction

[Clauses 3](#) to [7](#) cover the definitions, symbols and basic theory of electromagnetic flowmeters. This document does not cover insertion type meters, partially filled meters or meters for non-conductive and highly conductive fluids.

[Clause 8](#) covers installation types and practice, the different types of meter construction, transmitters, lay lengths and sizing, in order to achieve the best performance of the electromagnetic flowmeter in the field.

[Clauses 9](#) to [11](#) cover some methods of calibration, verification, evaluation, and uncertainty analysis, which can be useful for users or independent testing establishments to verify manufacturer's relative performance and to demonstrate suitability of application

The tests specified in this document are not necessarily sufficient for instruments specifically designed for unusually difficult duties. Conversely, a restricted series of tests may be suitable for instruments designed to perform within a limited range of conditions.

This document is for users and manufacturers.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST ISO 20456:2019](#)

<https://standards.iteh.ai/catalog/standards/sist/76ad5513-d42a-4dcf-aed0-bbf8889522af/sist-iso-20456-2019>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST ISO 20456:2019

<https://standards.iteh.ai/catalog/standards/sist/76ad5513-d42a-4dcf-aed0-bbf8889522af/sist-iso-20456-2019>

Measurement of fluid flow in closed conduits — Guidance for the use of electromagnetic flowmeters for conductive liquids

1 Scope

This document applies to industrial electromagnetic flowmeters used for the measurement of flowrate of a conductive liquid in a closed conduit running full. It covers flowmeter types utilizing both alternating current (AC) and pulsed direct current (DC) circuits to drive the field coils and meters running from a mains power supply and those operating from batteries or other sources of power.

This document is not applicable to insertion-type flowmeters or electromagnetic flowmeters designed to work in open channels or pipes running partially full, nor does it apply to the measurement of magnetically permeable slurries or liquid metal applications.

This document does not specify safety requirements in relation to hazardous environmental usage of the flowmeter.

2 Normative references

There are no normative references in this document.

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

3.1

electromagnetic flowmeter

flowmeter which creates a magnetic field perpendicular to the direction of flow, so enabling the flowrate to be deduced from the induced voltage, U_v , produced by the motion of a conducting fluid through the magnetic field

Note 1 to entry: The electromagnetic flowmeter consists of a *sensor* (3.2) and a *transmitter* (3.3).

3.2

sensor

device containing at least the following elements:

- an electrically insulating meter tube through which the conductive fluid to be measured flows;
- one pair of electrodes across which the signal generated in the fluid is measured;
- an electromagnet for producing a magnetic field in the *meter tube* (3.4)

Note 1 to entry: The sensor produces a signal proportional to the flowrate and, in some cases, a *reference signal* (3.9). See 6.2.

Note 2 to entry: For a sensor, the wording primary device or flowtube has previously been used.

ISO 20456:2017(E)

Note 3 to entry: In some cases, further electrodes are used such as grounding electrodes, full pipe detection electrodes (empty pipe detection) (see [3.5](#)).

3.3 transmitter

equipment which contains the circuitry which drives the field coils and extracts the flow signal

Note 1 to entry: This equipment may be mounted directly onto the *sensor* ([3.2](#)) or remotely, connected to the sensor by a cable.

Note 2 to entry: For a transmitter, the wording secondary device, converter or electronic unit has previously been used.

3.4 meter tube

pipe section of the *sensor* ([3.2](#)) through which the liquid flows, at least part of whose inner surface is electrically insulating

3.5 measuring electrodes

one or more pairs of electrical contacts or capacitor plates by means of which the induced voltage is detected

3.6 lower range value

lowest value of the measured variable that a device is set to measure

3.7 upper range value

highest value of the measured variable that a device is set to measure

3.8 span

difference between the upper and *lower range values* ([3.6](#))

3.9 reference signal

signal which is proportional to the magnetic flux created in the *sensor* ([3.2](#)) and which is compared in the *transmitter* ([3.3](#)) with the flow signal

3.10 output signal

signal from the *transmitter* ([3.3](#)) which is a function of the flowrate

3.11 Reynolds number

dimensionless parameter expressing the ratio between the inertial and the viscous forces

Note 1 to entry: For closed pipe flow through an *electromagnetic flowmeter* ([3.1](#)), Reynolds number should be based on the nominal diameter of the meter and corresponding mean velocity through a section of that size.

3.12 accuracy

closeness of the agreement between the result of a measurement and the (conventional) true value of the measurement

Note 1 to entry: The quantitative expression of accuracy should be in terms of uncertainty (see [Annex E](#)).

Note 2 to entry: The use of the term precision for accuracy should be avoided.

3.13**uncertainty**

<of measurement> range within which the true value of the measured quantity can be expected to lie with a specified value and confidence level

Note 1 to entry: See [Clause 11](#).

3.14**calibration factor**

number, determined by liquid calibration, that enables the *output signal* ([3.10](#)) to be related to the volumetric flowrate

3.15**calibration**

operation that, under specified conditions, in a first step, establishes a relation between the quantity values with measurement uncertainties provided by measurement standards and corresponding indications with associated measurement uncertainties and, in a second step, uses this information to establish a relation for obtaining a measurement result from an indication

3.16**verification**

<in-situ electronic verification> means of verifying that an *electromagnetic flowmeter* ([3.1](#)) is operating correctly, normally with a poorer uncertainty than under controlled laboratory conditions

3.17**calibration validation**

number of runs (one or more) at flowrates between zero and the *upper range value* ([3.7](#)) in order to verify that the flowmeter does perform in the expected way and within the manufacturer's specification

3.18**measuring window**

period of time during which the voltage representing the flow velocity is measured

3.19**ideal flow conditions**

conditions that exist when a pipe is infinitely long and straight with no internal disturbances

Note 1 to entry: For *electromagnetic flowmeters* ([3.1](#)), it may, in addition, also be assumed that the metering liquid has a viscosity and density similar to water. Under these conditions, the flow is axisymmetric and will be fully developed and turbulent at flowrates and pipe sizes most often found in industry.

4 Symbols

Symbol	Quantity	Units (SI)
$\vec{B}$	magnetic field strength	Tesla (T)
$\bar{B}$	mean magnetic field strength	Tesla (T)
d	inside diameter of meter tube	metres (m) ^a
$\vec{E}$	electric field strength	volt per metre (V/m)
U_c	electrochemical voltage	volt (V)
U_t	transformer voltage	volt (V)
U_v	velocity related voltage	volt (V)
F_{Lorentz}	Lorentz force	newton (N)
k_1	constant	dimensionless (—)
k_2	constant	dimensionless (—)

Symbol	Quantity	Units (SI)
L_e	distance between measuring electrodes	metres (m)
q_v	volumetric flowrate of the liquid	cubic meters per second (m ³ /s)
$\bar{v}$	mean axial liquid velocity	metres per second (m/s)
∇	Nabla or Del operator	dimensionless (—)

^a See [Annex D](#) for a conversion table of nominal diameters from metric to US units.

5 Theory and basic formulae

When a conductive liquid moves through a magnetic field, voltage(s), U_v , are generated in accordance with Faraday's law (see [Formula 2](#)). The strength of the induced voltages is given by the simplified expression shown in [Formula \(1\)](#):

$$F_{\text{Lorentz}} = q(\vec{E} + \vec{v} \times \vec{B}) = 0 \quad (1)$$

$$\vec{E} = -\vec{v} \times \vec{B} = \nabla(U_v);$$

$$\nabla(U_v) = -\vec{v} \times \vec{B}$$

Spatial integration of [Formula \(1\)](#) results in [Formula \(2\)](#):

$$U_v = k_1 \bar{B} L_e \bar{v} \quad (2)$$

The volume flowrate in the case of a circular pipe is given in [Formula \(3\)](#):

$$q = \frac{\pi d^2}{4} \bar{v} \quad (3)$$

Which, combined with [Formula \(2\)](#), gives [Formula \(4\)](#):

$$q = \frac{\pi d^2}{4 k_1 L_e} \left(\frac{U_v}{\bar{B}} \right) \quad (4)$$

Or [Formula \(5\)](#):

$$q = k_2 U_v \quad (5)$$

[Formula \(5\)](#) may be interpreted in various ways to produce a calibration factor which in practice is usually determined by wet calibration, as described in [9.1](#).

6 Construction and principle of operation

6.1 General

As indicated schematically in [Figure 1](#), the magnetic field is so placed with respect to a lined meter tube that the path of the conductive liquid, flowing in the meter tube, is normal to the magnetic field. In accordance with Faraday's law, motion of the liquid through the magnetic field induces a voltage, U_v , in the liquid in a path mutually normal both to the field and the direction of liquid motion. By placing electrodes which contact the liquid in insulated mountings or by using insulated electrodes with capacitance-type coupling in the meter tube in a diametrical plane normal to the magnetic field, a