

---

**Gnojila - Določevanje različnih oblik dušika v istem vzorcu, ki vsebuje dušik v nitratni, amonijevi, sečninski in cianoamidni obliki (ISO 15604:2016)**

Fertilizers - Determination of different forms of nitrogen in the same sample, containing nitrogen as nitric, ammoniacal, urea and cyanamide nitrogen (ISO 15604:2016)

Düngemittel - Bestimmung verschiedener, nebeneinander anwesender Stickstoff-Formen in derselben Probe mit Stickstoff in Form von Ammonium, Nitrat, Harnstoff und Cyanamid; Deutsche und Englische Fassung prEN ISO 15604:2018

(standards.iteh.ai)

Engrais - Détermination des différentes formes d'azote dans un même échantillon d'engrais contenant l'azote sous forme nitrique ammoniacale, uréique et cyanamidique (ISO 15604:2016)

**Ta slovenski standard je istoveten z: prEN ISO 15604**

---

**ICS:**

65.080                      Gnojila                                      Fertilizers

**oSIST prEN ISO 15604:2019**                                      **en,fr,de**

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[oSIST prEN ISO 15604:2019](https://standards.iteh.ai/catalog/standards/sist/94061e41-5019-4b6b-a94a-e4576b09a9d1/osist-pren-iso-15604-2019)

<https://standards.iteh.ai/catalog/standards/sist/94061e41-5019-4b6b-a94a-e4576b09a9d1/osist-pren-iso-15604-2019>

INTERNATIONAL  
STANDARD

ISO  
15604

First edition  
2016-05-01

---

---

**Fertilizers — Determination of  
different forms of nitrogen in the same  
sample, containing nitrogen as nitric,  
ammoniacal, urea and cyanamide  
nitrogen**

*Engrais — Détermination des différentes formes d'azote dans  
un même échantillon contenant l'azote sous forme nitrique,  
ammoniacale, uréique et cyanamidique*

iTeh STANDARD PREVIEW  
(standards.iteh.ai)

[oSIST prEN ISO 15604:2019](https://standards.iteh.ai/catalog/standards/sist/94061e41-5019-4b6b-a94a-e4576b09a9d1/osist-pren-iso-15604-2019)

[https://standards.iteh.ai/catalog/standards/sist/94061e41-5019-4b6b-a94a-  
e4576b09a9d1/osist-pren-iso-15604-2019](https://standards.iteh.ai/catalog/standards/sist/94061e41-5019-4b6b-a94a-e4576b09a9d1/osist-pren-iso-15604-2019)



Reference number  
ISO 15604:2016(E)

© ISO 2016

## iTeh STANDARD PREVIEW (standards.iteh.ai)

<https://standards.iteh.ai/catalog/standards/sist/94061e41-5019-4b6b-a94a-e4576b09a9d1/osist-pren-iso-15604-2019>



### **COPYRIGHT PROTECTED DOCUMENT**

© ISO 2016, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Ch. de Blandonnet 8 • CP 401  
CH-1214 Vernier, Geneva, Switzerland  
Tel. +41 22 749 01 11  
Fax +41 22 749 09 47  
copyright@iso.org  
www.iso.org

# Contents

	Page
<b>Foreword</b> .....	<b>iv</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms and definitions</b> .....	<b>1</b>
<b>4 Principle</b> .....	<b>1</b>
4.1 Total soluble and insoluble nitrogen.....	1
4.2 Forms of soluble nitrogen.....	2
4.2.1 General.....	2
4.2.2 Total soluble nitrogen.....	2
4.2.3 Total soluble nitrogen with the exception of nitrate nitrogen.....	2
4.2.4 Nitrate nitrogen by difference.....	2
4.2.5 Ammoniacal nitrogen.....	2
4.2.6 Urea nitrogen.....	2
4.2.7 Cyanamide nitrogen.....	3
<b>5 Reagents</b> .....	<b>3</b>
<b>6 Apparatus</b> .....	<b>5</b>
<b>7 Sampling and sample preparation</b> .....	<b>10</b>
<b>8 Procedure</b> .....	<b>10</b>
8.1 Total soluble and insoluble nitrogen.....	10
8.1.1 In the absence of nitrates.....	10
8.1.2 In the presence of nitrate.....	11
8.2 Forms of soluble nitrogen.....	12
8.2.1 Preparation of the solution to be analysed.....	12
8.2.2 Total soluble nitrogen.....	13
8.2.3 Total soluble nitrogen with the exception of nitrate nitrogen.....	14
8.2.4 Nitrate nitrogen.....	14
8.2.5 Ammoniacal nitrogen.....	15
8.2.6 Urea nitrogen.....	16
8.2.7 Cyanamide nitrogen.....	18
<b>9 Verification of the result</b> .....	<b>18</b>
<b>10 Test report</b> .....	<b>18</b>
<b>Bibliography</b> .....	<b>19</b>

## ISO 15604:2016(E)

### Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

ISO 15604 was prepared by CEN/TC 260 as EN 15604:2009 and was adopted (without modification other than those stipulated below) by Technical Committee ISO/TC 134, *Fertilizers and soil conditioners*.

Modifications were made as follows:

- a) [5.2](#): p.a. = pro analysis = analytical grade;
- b) [6.2](#): add "Refer to [Figure 1](#)".

# Fertilizers — Determination of different forms of nitrogen in the same sample, containing nitrogen as nitric, ammoniacal, urea and cyanamide nitrogen

## 1 Scope

This International Standard specifies a method for the determination of any one form of nitrogen in the presence of any other form.

The method is applicable to any fertilizer provided for in the Regulation (EC) No 2003/2003, Annex I<sup>[2]</sup> containing nitrogen in various forms.

## 2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

ISO 14820-2, *Fertilizers and liming materials — Sampling and sample preparation — Part 2: Sample preparation*

ISO 25475, *Fertilizers — Determination of ammoniacal nitrogen*

EN 12944-1, *Fertilizers and liming materials — Vocabulary — Part 1: General terms*

EN 12944-2, *Fertilizers and liming materials — Vocabulary — Part 2: Terms relating to fertilizers*

EN 15562, *Fertilizers — Determination of cyanamide nitrogen*

## 3 Terms and definitions

For the purposes of this document, the terms and definitions given in EN 12944-1 and EN 12944-2 apply.

## 4 Principle

### 4.1 Total soluble and insoluble nitrogen

According to the list of standard fertilizers given in Regulation (EC) No 2003/2003, Annex I,<sup>[2]</sup> this determination is applicable to products containing calcium cyanamide.

In the absence of nitrates, the test sample is mineralized by direct Kjeldahl digestion.

In the presence of nitrates, the test sample is mineralized by Kjeldahl digestion after reduction with the aid of metallic iron and stannous chloride.

In both cases, the ammonia is determined according to ISO 25475.

**NOTE** If analysis shows an insoluble nitrogen content of more than 0,5 %, one concludes that the fertilizer contains other forms of insoluble nitrogen not included in the list in Regulation (EC) No 2003/2003, Annex I.<sup>[2]</sup>

## ISO 15604:2016(E)

### 4.2 Forms of soluble nitrogen

#### 4.2.1 General

The forms of soluble nitrogen referred to in [4.2.2](#) to [4.2.7](#) are determined from different aliquots taken from the same solution of the test sample.

#### 4.2.2 Total soluble nitrogen

**4.2.2.1** In the absence of nitrates, by direct Kjeldahl digestion. The ammonia is then determined (by the same method as that described in ISO 25475).

**4.2.2.2** In the presence of nitrates, by Kjeldahl digestion on an aliquot part taken from the solution after reduction according to Ulsch. The ammonia is then determined (by the same method as that described in ISO 25475).

#### 4.2.3 Total soluble nitrogen with the exception of nitrate nitrogen

By Kjeldahl digestion after elimination in an acid medium of nitrate nitrogen with ferrous sulfate. The ammonia is then determined (by the same method as that described in ISO 25475).

#### 4.2.4 Nitrate nitrogen by difference

**4.2.4.1** In the absence of calcium cyanamide, by determining the difference between the nitrogen determined as summarized in [4.2.2.2](#) and that determined as summarized in [4.2.3](#) or between total soluble nitrogen (see [4.2.2](#)) and the sum of ammoniacal nitrogen and ureic organic nitrogen ([4.2.5](#) + [4.2.6](#)).

**4.2.4.2** In the presence of calcium cyanamide, by determining the difference between the nitrogen determined as summarized in [4.2.2.2](#) and that determined as summarized in [4.2.3](#) or between the nitrogen determined as summarized in [4.2.2.2](#) and the sum of that determined as summarized in [4.2.5](#), [4.2.6](#) and [4.2.7](#).

#### 4.2.5 Ammoniacal nitrogen

**4.2.5.1** Solely in the presence of ammoniacal nitrogen and ammoniacal plus nitrate nitrogen, according to ISO 25475.

**4.2.5.2** In the presence of urea nitrogen and/or cyanamide nitrogen by cold distillation after making slightly alkaline, the ammonia is absorbed in a standard solution of sulfuric acid and determined according to ISO 25475.

#### 4.2.6 Urea nitrogen

**4.2.6.1** By conversion using urease into ammonia which is titrated with a standard solution of hydrochloric acid.

or

**4.2.6.2** By gravimetry with xanthidrol: the co-precipitated biuret can be counted with urea nitrogen without great error, its content remaining generally low in absolute value in compound fertilizers.

or

**4.2.6.3** By difference according to [Table 1](#).



Table 1 — Determination of urea nitrogen by difference

Case	Nitrate nitrogen	Ammoniacal nitrogen	Cyanamidic nitrogen	Difference
1	Absent	Present	Present	(4.2.2.1) - (4.2.5.2 + 4.2.7)
2	Present	Present	Present	(4.2.3) - (4.2.5.2 + 4.2.7)
3	Absent	Present	Absent	(4.2.2.1) - (4.2.5.2)
4	Present	Present	Absent	(4.2.3) - (4.2.5.2)

#### 4.2.7 Cyanamide nitrogen

By precipitation as a silver compound, the nitrogen being determined in the precipitate by the Kjeldahl method.

## 5 Reagents

### 5.1 General.

Use only reagents of recognized analytical grade and distilled or de-mineralized water of grade 3 according to ISO 3696.

5.2 **Potassium sulfate**, p.a. (p.a. = pro analysis = analytical grade).

5.3 **Iron powder**, reduced with hydrogen.

The prescribed quantity of iron shall be able to reduce at least 50 mg of nitrate nitrogen.

5.4 **Potassium thiocyanate**, p.a.

5.5 **Potassium nitrate**, p.a.

5.6 **Ammonium sulfate**, p.a.

5.7 **Urea**, p.a.

5.8 **Sulfuric acid diluted**.

Dilute one volume of sulfuric acid ( $\rho_{20} = 1,84$  g/ml) in one volume of water.

5.9 **Sulfuric acid, standard solution**,  $c = 0,1$  mol/l.

5.10 **Sodium hydroxide solution**, aqueous solution of about 30 % (mass concentration), free from ammonia.

5.11 **Sodium or potassium hydroxide, standard solution**,  $c = 0,2$  mol/l, free from carbonates.

5.12 **Stannous chloride solution**.

Dissolve 120 g of  $\text{SnCl}_2 \cdot 2\text{H}_2\text{O}$  in 400 ml of concentrated hydrochloric acid ( $\rho_{20} = 1,18$  g/ml) and make up to 1 l with water. The solution shall be perfectly clear and prepared immediately before use.

It is essential to check the reducing power of stannous chloride: dissolve 0,5 g of  $\text{SnCl}_2 \cdot 2\text{H}_2\text{O}$  in 2 ml of concentrated hydrochloric acid ( $\rho_{20} = 1,18$  g/ml) and make up to 50 ml with water. Then, add 5 g

**ISO 15604:2016(E)**

of Rochelle salt (potassium sodium tartrate), then a sufficient quantity of sodium bicarbonate for the solution to be alkaline to litmus paper.

Titrate with an iodine solution ( $I_2$ ) of  $c = 0,05$  mol/l in the presence of a starch solution as an indicator.

1 ml of iodine solution ( $I_2$ ) of  $c = 0,05$  mol/l corresponds to 0,011 28 g of  $SnCl_2 \cdot 2H_2O$ .

At least 80 % of the total tin present in the solution thus prepared shall be in bivalent form. For the titration, at least 35 ml of the  $c = 0,05$  mol/l iodine solution ( $I_2$ ) should be used.

**5.13 Sulfuric acid**,  $\rho_{20} = 1,84$  g/ml.

**5.14 Hydrochloric acid diluted.**

Mix one volume of hydrochloric acid ( $\rho_{20} = 1,18$  g/ml) with one volume of water.

**5.15 Acetic acid**, 96 % to 100 %.

**5.16 Sulfuric acid solution**, containing about 30 % of  $H_2SO_4$  (mass concentration).

**5.17 Ferrous sulfate**, crystalline,  $FeSO_4 \cdot 7H_2O$ .

**5.18 Sulfuric acid standard solution**,  $c = 0,05$  mol/l.

**5.19 Octyl alcohol.**

**5.20 Potassium carbonate, saturated solution..**

**5.21 Sodium or potassium hydroxide standard solution**,  $c = 0,1$  mol/l (free from carbonates).

**5.22 Barium hydroxide saturated solution.**

**5.23 Sodium carbonate solution**, at 10 % (mass concentration).

**5.24 Hydrochloric acid**,  $c = 2$  mol/l.

**5.25 Hydrochloric acid standard solution**,  $c = 0,1$  mol/l.

**5.26 Urease solution.**

Suspend 0,5 g of active urease in 100 ml of water. Using hydrochloric acid 0,1 mol/l (5.25), adjust the pH to 5,4, measured by a pH-meter.

**5.27 Xanthidrol.**

Use a solution at 5 % in ethanol or methanol (5.32) (do not use products giving a high proportion of insoluble matter). The solution may be kept for three months in a well-stoppered bottle, away from the light.

**5.28 Catalyst.**

Use 0,3 g to 0,4 g of copper oxide per determination or an equivalent quantity of copper sulfate pentahydrate of 0,95 g to 1,25 g per determination.

**5.29 Anti-bump granules**, washed in hydrochloric acid and calcined.

iTeh STANDARD PREVIEW  
(standards.iteh.ai)

oSIST prEN ISO 15604:2019

<https://standards.iteh.ai/catalog/standards/sist/94061e41-5019-4b6b-a94a-e4576b09a9d1/osist-pr-en-iso-15604-2019>

### 5.30 Indicator solutions.

#### 5.30.1 Solution A.

Dissolve 1 g of methyl red in 37 ml of sodium hydroxide solution 0,1 mol/l and make up to 1 l with water.

#### 5.30.2 Solution B.

Dissolve 1 g of methylene blue in water and make up to 1 l.

#### 5.30.3 Combined indicator solution.

Mix one volume of solution A with two volumes of solution B.

This indicator is violet in acid solution, grey in neutral solution and green in alkaline solution. Use 0,5 ml (10 drops) of this indicator solution.

#### 5.30.4 Methyl red indicator solution.

Dissolve 0,1 g of methyl red in 50 ml of 95 % ethanol. Make up to 100 ml with water and filter if necessary. This indicator (four or five drops) may be used instead of that described in [5.30.3](#).

**5.31 Indicator papers**, litmus bromothymol blue (or other papers sensitive to pH = 6 to pH = 8).

**5.32 Ethanol or methanol**, solution 95 %.

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

## 6 Apparatus

### 6.1 Distillation apparatus.

oSIST prEN ISO 15604:2019

<https://standards.iteh.ai/catalog/standards/sist/94061e41-5019-4b6b-a94a-e4576b09a9d1/osist-pren-iso-15604-2019>

Consisting of a round-bottomed flask of suitable capacity connected to a condenser by means of a splash head. The equipment is made of borosilicate glass.

NOTE The different types of equipment recommended for this determination are reproduced, showing all the features of construction, in [Figures 1, 2, 3 and 4](#).

An automatic distillation apparatus may also be used, provided that the results are statistically equivalent.