
**Electrical apparatus for the detection and measurement of combustible gases -
Performance requirements for Group I apparatus indicating up to 100 % (v/v)
methane in air**

Electrical apparatus for the detection and measurement of combustible gases -
Performance requirements for Group I apparatus indicating up to 100 % (v/v) methane in
air

Elektrische Geräte für die Detektion und die Messung brennbarer Gase - Anforderungen
an das Betriebsverhalten von Geräten der Gruppe I mit einem Meßbereich bis zu 100 %
(V/V) Methan in Luft

[SIST EN 50056:2000](https://standards.iteh.ai/catalog/standards/sist/40734ed5-16f6-448c-8f6c-1a1b2c3d4e5f/sist-en-50056-2000)

Appareils électriques de détection et de mesure des gaz combustibles - Règles de
performances des appareils du Groupe I pouvant indiquer jusqu'à 100 % (v/v) de
méthane dans l'air

Ta slovenski standard je istoveten z: EN 50056:1998

ICS:

13.320	Alarmni in opozorilni sistemi	Alarm and warning systems
29.260.20	Električni aparati za eksplozivna ozračja	Electrical apparatus for explosive atmospheres

SIST EN 50056:2000**en**

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 50056:2000

<https://standards.iteh.ai/catalog/standards/sist/40734ed5-16f6-448c-8f6c-4ae6b25eae04/sist-en-50056-2000>

EUROPEAN STANDARD

EN 50056

NORME EUROPÉENNE

EUROPÄISCHE NORM

July 1998

ICS 13.220.60;19.080

Supersedes EN 50056:1991 + A1:1995

Descriptors: Electrical apparatus, explosive atmosphere, mine susceptible to firedamp, detector, measuring apparatus, flammable gas, combustible gas, characteristic

English version

**Electrical apparatus for the detection and measurement
of combustible gases
Performance requirements for Group I apparatus indicating
up to 100 % (v/v) methane in air**

Appareils électriques de détection et de
mesure des gaz combustibles

Règles de performances des appareils
du Groupe I pouvant indiquer jusqu'à
100 % (v/v) de méthane dans l'air

Elektrische Geräte für das Aufspüren
und die Messung brennbarer Gase

Anforderungen an das Betriebsverhalten
von Geräten der Gruppe I mit einem
Meßbereich bis zu 100 % (V/V) Methan
in Luft

[SIST EN 50056:2000](https://standards.iteh.ai/catalog/standards/sist/40734ed5-16f6-448c-8f6c-4ae6b25eae04/sist-en-50056-2000)

<https://standards.iteh.ai/catalog/standards/sist/40734ed5-16f6-448c-8f6c-4ae6b25eae04/sist-en-50056-2000>

This European Standard was approved by CENELEC on 1996-12-09. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the Central Secretariat has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.

CENELEC

European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

Central Secretariat: rue de Stassart 35, B - 1050 Brussels

Foreword

This second edition, of the European Standard was prepared by SC 31-9, Electrical apparatus for the detection and measurement of combustible gases to be used in industrial and commercial potentially explosive atmospheres, of Technical Committee CENELEC TC 31, Electrical apparatus for explosive atmosphere, on the basis of EN 50056:1991, its amendment A1:1995 and a second amendment.

This second amendment was approved by CENELEC on 1996-12-09 for incorporation into a new edition of EN 50056.

This European Standard replaces EN 50056:1991 + A1:1995.

The following dates were fixed:

- latest date by which the EN has to be implemented
at national level by publication of an identical
national standard or by endorsement (dop) 1999-01-01
- latest date by which the national standards conflicting
with the EN have to be withdrawn (dow) 1999-01-01

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and covers essential requirements of EC Directive 94/9/EC.

iteh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN 50056:2000](https://standards.iteh.ai/catalog/standards/sist/40734ed5-16f6-448c-8f6c-4ae6b25eae04/sist-en-50056-2000)

<https://standards.iteh.ai/catalog/standards/sist/40734ed5-16f6-448c-8f6c-4ae6b25eae04/sist-en-50056-2000>

Contents

1 Scope	4
2 Normative references.....	4
3 Definitions	4
4 General requirements	4
5 Performance requirements	4
5.1 General.....	4
5.2 Unpowered storage	4
5.3 Calibration curve (not applicable to alarm-only apparatus)	4
5.4 Drift (continuous duty apparatus)	4
5.5 Drift (spot reading apparatus).....	4
5.6 Alarm	5
5.7 Temperature	5
5.8 Pressure	5
5.9 Humidity.....	5
5.10 Air speed.....	5
5.11 Pumping rate	5
5.12 Orientation	5
5.12.1 Portable apparatus.....	5
5.12.2 Fixed and transportable apparatus.....	5
5.13 Vibration (applicable only to machine-mounted apparatus).....	5
5.14 Drop test (applicable to portable apparatus and remote sensors).....	5
5.15 Warm-up time (not applicable to spot-reading apparatus).....	6
5.15.1 Fixed and transportable apparatus.....	6
5.15.2 Continuous duty portable apparatus.....	6
5.16 Time of response (not applicable to spot-reading apparatus)	6
5.17 Minimum time of operation (spot reading apparatus)	6
5.18 High gas concentration above the measuring range	6
5.19 Battery capacity	6
5.19.1 Battery powered portable continuous duty apparatus.....	6
5.19.2 Battery-powered portable spot-reading apparatus.....	6
5.20 Power supply variations.....	6
5.20.1 A.C. powered apparatus.....	6
5.20.2 External d.c. powered apparatus.....	7
5.20.3 Battery powered apparatus	7
5.21 Power supply interruptions, voltage transients and step changes of voltage	7
5.22 Addition of sampling probe	7
5.23 Dust	7
5.24 Poisons and other gases	7
5.24.1 Poisons.....	7
5.24.2 Other gases.....	7
5.25 Electromagnetic immunity.....	7
6 Field calibration kit	7
7 Information for use	7

1 Scope

1.1 This European Standard specifies performance requirements for Group I (as defined in EN 50054) portable, transportable and fixed apparatus for the detection and measurement of methane concentrations in mine air. The apparatus, or parts thereof, are intended for use in mines susceptible to firedamp and shall meet the general requirements and test methods specified in EN 50054.

1.2 This European Standard is restricted to apparatus intended for the detection and measurement of volume ratios of methane in air from 0 % (v/v) up to 100 % (v/v).

NOTE: Apparatus covered by this European Standard will normally be intended to operate in volume ratios greater than 5 % (v/v).

2 Normative references

This European Standard incorporates by dated or undated reference, provisions from other publications. These normative references are cited at appropriate places in the text and the publications are listed hereafter. For dated references, subsequent amendments to or revisions of any of these publications apply to the European Standard only when incorporated in it by amendment or revision. For undated references the latest edition of the publication referred to applies.

EN 50054 1998 Electrical apparatus for the detection and measurement of combustible gases
General requirements and test methods.

3 Definitions

For the purposes of this European Standard, the definitions given in EN 50054 apply.

4 General requirements

The apparatus shall comply with the general requirements specified in clause 4 of EN 50054.

5 Performance requirements

5.1 General

The normal conditions for test are specified in subclause 5.3 of EN 50054. Compliance shall be determined in accordance with the test methods specified in subclause 5.4 of EN 50054.

5.2 Unpowered storage

After being submitted to the conditions specified in subclause 5.4.2 of EN 50054, the apparatus shall meet the requirements specified in 5.3 to 5.25 of this European Standard.

5.3 Calibration curve (not applicable to alarm-only apparatus)

Each of the three indications (after correction using the manufacturer's calibration curve, if necessary) obtained for each of the four gas volume ratios shall not differ from these volume ratios by more than ± 3 % (v/v) methane or ± 5 % of the indication, whichever is the greater.

5.4 Drift (continuous duty apparatus)

The medium term variation shall not exceed ± 3 % (v/v) methane or ± 5 % of the indication, whichever is the greater, in air and in the standard test gas.

5.5 Drift (spot reading apparatus)

The variation of the indication shall not exceed ± 3 % (v/v) methane or ± 5 % of the indication, whichever is the greater, in air and in the standard test gas.

5.6 Alarm

The alarm shall operate during every cycle of the test. If a latching alarm is provided, the manual reset action shall be checked during every cycle.

5.7 Temperature

The variation of the indication from that at 20 °C, over the temperature range:

- a) -10 °C to 0 °C, shall not exceed ± 7 % (v/v) methane or ± 15 % of the indication and,
- b) 0 °C to + 40 °C, shall not exceed ± 5 % (v/v) methane or ± 10 % of the indication.

whichever is the greater.

5.8 Pressure

The variation of the indication from that at 101,3 kPa, over a pressure range of 92 kPa to 115 kPa, shall not exceed $\pm 7,5$ % (v/v) methane or ± 15 % of the indication, whichever is the greater.

5.9 Humidity

At 40° C, the variation of the indication from that at 55 % relative humidity (r.h.) over a humidity range of 5 % r.h. to 90 % r.h., shall not exceed ± 5 % (v/v) methane or ± 10 % of the indication, whichever is the greater.

5.10 Air speed

The variation of the indication over a range of air speeds from 0 m/s to 6 m/s shall not exceed ± 3 % (v/v) methane or ± 5 % of the indication, whichever is the greater.

5.11 Pumping rate

The variation of the indication shall not exceed ± 3 % (v/v) methane or 5% of the variation, whichever is the greater.

5.12 Orientation

5.12.1 Portable apparatus

The variation in the indication shall not exceed ± 5 % (v/v) methane or ± 10 % of the indication, whichever is the greater.

5.12.2 Fixed and transportable apparatus

The variation in the indication shall not exceed ± 3 % (v/v) methane or ± 5 % of the indication, whichever is the greater.

5.13 Vibration (applicable only to machine-mounted apparatus)

During the vibration test, the apparatus shall not suffer any loss of function and shall not give a false alarm or fault signal. The apparatus shall not suffer damage resulting in hazard or loss of function.

At the conclusion of the vibration test and after the apparatus sensor has then been exposed to clean air followed by the standard test gas, the variation of the indication from that determined prior to the test shall not exceed ± 3 % (v/v) methane or ± 5 % of the indication, whichever is the greater.

5.14 Drop test (applicable to portable apparatus and remote sensors)

The apparatus shall not suffer damage resulting in hazard or loss of function.

The variation of the indication shall not exceed + 3 % (v/v) methane or ± 5 % of the indication, whichever is the greater.

5.15 Warm-up time (not applicable to spot-reading apparatus)**5.15.1 Fixed and transportable apparatus**

The apparatus shall warm-up in clean air to indicate zero to within $\pm 3\%$ (v/v) methane, in a time not exceeding 5 min, or longer if specified by the manufacturer. The apparatus shall warm-up in the standard test gas to give a final indication to within $\pm 3\%$ (v/v) methane in a time not exceeding 5 min, or longer if specified by the manufacturer.

5.15.2 Continuous duty portable apparatus

The apparatus shall warm-up in clean air to indicate zero to within $\pm 3\%$ (v/v) methane in a time not exceeding 60s. The apparatus shall warm-up in the standard test gas to give a final indication to within $\pm 3\%$ (v/v) methane in a time not exceeding 60 s.

5.16 Time of response (not applicable to spot-reading apparatus)

The time of response t_{50} shall not be greater than 10 s, and t_{90} shall not be greater than 30 s.

The specified times of response shall apply to apparatus in the as-supplied condition and without optional accessories attached to the sensors for special purposes, e.g. collecting cones, weather protection.

5.17 Minimum time of operation (spot reading apparatus)

For apparatus without probe or sample line, the indication shall reach 90 % of the final value in a time not exceeding 15 s.

For aspirated apparatus using a sample line or probe length of not more than 3 m, an additional 10 s is permitted.

5.18 High gas concentration above the measuring range

Not applicable.

5.19 Battery capacity**5.19.1 Battery powered portable continuous duty apparatus**

The variation shall not exceed $\pm 3\%$ (v/v) methane or $\pm 5\%$ of the indication, whichever is the greater, at the end of the 8 h or 10 h period as appropriate.

At the end of the further 15 min following the indication of low battery condition, the variation shall not exceed $\pm 6\%$ (v/v) methane or $\pm 10\%$ of the indication, whichever is the greater.

5.19.2 Battery-powered portable spot-reading apparatus

The variation shall not exceed $\pm 3\%$ (v/v) methane, or $\pm 5\%$ of the indication, whichever is the greater, at the end of 200 operations.

After a further 10 operations following the indication of low battery condition, the variation shall not exceed $\pm 6\%$ (v/v) methane or $\pm 10\%$ of the indication, whichever is the greater.

5.20 Power supply variations**5.20.1 A.C. powered apparatus**

The variation of the indication shall not exceed $\pm 3\%$ (v/v) methane or $\pm 5\%$ of the indication, whichever is the greater.