

SLOVENSKI STANDARD
SIST EN ISO 5356-2:2013/oprA1:2019
01-marec-2019

Anestezijska in respiratorna oprema - Konični priključki - 2. del: Nosilni priključki z navojem - Dopolnilo A1 (ISO 5356-2:2012/DAM1:2018)

Anaesthetic and respiratory equipment - Conical connectors - Part 2: Screw-threaded weight-bearing connectors - Amendment 1 (ISO 5356-2:2012/DAM1:2018)

Anästhesie- und Beatmungsgeräte - Konische Konnektoren - Teil 2: Gewichtstragende Konnektoren mit Schraubgewinde - Änderung 1 (ISO 5356-2:2012/DAM1:2018)

Matériel d'anesthésie et de réanimation respiratoire - Raccords coniques - Partie 2: Raccords à vis pouvant supporter un certain poids - Amendement 1 (ISO 5356-2:2012/DAM1:2018)

Ta slovenski standard je istoveten z: EN ISO 5356-2:2012/prA1

ICS:

11.040.10	Anestezijska, respiratorna in reanimacijska oprema	Anaesthetic, respiratory and reanimation equipment
-----------	--	--

SIST EN ISO 5356-2:2013/oprA1:2019 en,fr,de

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Full standard:
<https://standards.iteh.ai/catalog/standards/sist/f52fc32d-6b81-444b-abb2-40993a1a562c/sist-en-iso-5356-2-2013-oprA1-2019>

DRAFT AMENDMENT

ISO 5356-2:2012/DAM 1

ISO/TC 121/SC 1

Secretariat: DIN

Voting begins on:
2018-12-20Voting terminates on:
2019-03-14

Anaesthetic and respiratory equipment — Conical connectors —

Part 2: Screw-threaded weight-bearing connectors

AMENDMENT 1

*Matériel d'anesthésie et de réanimation respiratoire — Raccords coniques —**Partie 2: Raccords à vis pouvant supporter un certain poids**AMENDEMENT 1*

ICS: 11.040.10

ITEH STANDARD PREVIEW
(standards.iteh.ai)
Full standard:
<https://standards.iteh.ai/catalog/standards/sist/f52fc32d-6b81-444b-a6b2-40993a1a562c/sist-en-iso-5356-2-2013-oprA1-2019>

THIS DOCUMENT IS A DRAFT CIRCULATED FOR COMMENT AND APPROVAL. IT IS THEREFORE SUBJECT TO CHANGE AND MAY NOT BE REFERRED TO AS AN INTERNATIONAL STANDARD UNTIL PUBLISHED AS SUCH.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

This document is circulated as received from the committee secretariat.

ISO/CEN PARALLEL PROCESSING



Reference number
ISO 5356-2:2012/DAM 1:2018(E)

© ISO 2018

ISO 5356-2:2012/DAM 1:2018(E)

ITEH STANDARD PREVIEW
(standards.iteh.ai)
Full standard:
<https://standards.iteh.ai/catalog/standards/sist/f52fc32d-6b81-444b-abb2-40993a1a562c/sist-en-iso-5356-2-2013-oprA1-2019>

**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2018

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 121 *Anaesthetic and respiratory equipment*, Subcommittee SC 1, *Breathing attachments and anaesthetic machines*.

A list of all parts in the ISO 5356- series can be found on the ISO website.

ISO 5356-2:2012/DAM 1:2018(E)

Introduction

Amendment 1 to ISO 5356-2:2012 was prepared to align this document with the new edition of ISO 5356-1:2015, *Anaesthetic and respiratory equipment — Conical connectors — Part 1: Cones and sockets*.

ITEH STANDARD PREVIEW
(standards.iteh.ai)
Full standard:
<https://standards.iteh.ai/catalog/standards/sist/f52fc32d-6b81-444b-abb2-40993a1a562c/sist-en-iso-5356-2-2013-oprA1-2019>

Anaesthetic and respiratory equipment — Conical connectors —

Part 2: Screw-threaded weight-bearing connectors

AMENDMENT 1

Introduction

First paragraph, first sentence

Replace "conical connectors" by "cones and sockets" so that the sentence reads:

Whereas the cones and sockets specified in ISO 5356-1 are satisfactory for lightweight breathing attachments, there is also a need for more substantial connectors to support heavy or fragile components.

First paragraph, last sentence

Replace "connectors" by "cones and sockets" so that the sentence reads:

They should also be compatible with cones and sockets meeting the requirements specified in ISO 5356-1 and should, if possible, eliminate the likelihood of inadvertent misconnection, although it has to remain the responsibility of the user to check that any system in which they are used is correctly assembled before use.

Normative references

Add the year of publication for ISO 5356-1 and delete footnote 1 so that the reference reads.

ISO 5356-1:2015, *Anaesthetic and respiratory equipment — Conical connectors — Part 1: Cones and sockets*

4 Design

Add the following sentence after the last paragraph:

Check compliance by measurement and functional testing.

Figure 1 — Components for screw-threaded, weight-bearing connectors

In the key to Figure 1, replace item 1 and item 4 by the following:

- 1 22 mm cone as specified in ISO 5356-1:2015, Figure 2, with or without a 15 mm coaxial socket
- 4 22 mm socket as specified in ISO 5356-1:2015