



# SLOVENSKI STANDARD SIST EN ISO 15189:2023

01-marec-2023

Nadomešča:

SIST EN ISO 15189:2013

SIST EN ISO 22870:2017

---

## Medicinski laboratoriji - Zahteve za kakovost in kompetentnost (ISO 15189:2022)

Medical laboratories - Requirements for quality and competence (ISO 15189:2022)

Medizinische Laboratorien - Anforderungen an die Qualität und Kompetenz (ISO 15189:2022)

Laboratoires de biologie médicale - Exigences concernant la qualité et la compétence (ISO 15189:2022)

**Ta slovenski standard je istoveten z: EN ISO 15189:2022**

---

### ICS:

03.120.10	Vodenje in zagotavljanje kakovosti	Quality management and quality assurance
11.100.01	Laboratorijska medicina na splošno	Laboratory medicine in general

**SIST EN ISO 15189:2023**

**en,fr,de**



EUROPEAN STANDARD

EN ISO 15189

NORME EUROPÉENNE

EUROPÄISCHE NORM

December 2022

ICS 03.120.10; 11.100.01

Supersedes EN ISO 15189:2012, EN ISO 22870:2016

English Version

## Medical laboratories - Requirements for quality and competence (ISO 15189:2022)

Laboratoires de biologie médicale - Exigences  
concernant la qualité et la compétence (ISO  
15189:2022)

Medizinische Laboratorien - Anforderungen an die  
Qualität und Kompetenz (ISO 15189:2022)

This European Standard was approved by CEN on 15 November 2022.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.

<https://standards.iteh.ai/catalog/standards/sist/8d07f2e4-b830-4619-8a36-92fda9d3c7ea/sist-en-iso-15189-2023>



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents	Page
European foreword.....	3

iTeh STANDARD PREVIEW  
(standards.iteh.ai)

[SIST EN ISO 15189:2023](https://standards.iteh.ai/catalog/standards/sist/8d07f2e4-b830-4619-8a36-92fda9d3c7ea/sist-en-iso-15189-2023)

<https://standards.iteh.ai/catalog/standards/sist/8d07f2e4-b830-4619-8a36-92fda9d3c7ea/sist-en-iso-15189-2023>

## European foreword

This document (EN ISO 15189:2022) has been prepared by Technical Committee ISO/TC 212 "Clinical laboratory testing and in vitro diagnostic test systems" in collaboration with Technical Committee CEN/TC 140 "In vitro diagnostic medical devices" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by June 2023, and conflicting national standards shall be withdrawn at the latest by December 2025.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 15189:2012 and EN ISO 22870:2016.

This document has been prepared under a Standardization Request given to CEN by the European Commission and the European Free Trade Association.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

[SIST EN ISO 15189:2023](https://standards.iteh.ai/catalog/standards/sist/8d07f2e4-b830-4619-8a36-92fda9d3c7ea/sist-en-15189-2022)

[https://standards.iteh.ai/catalog/standards/sist/8d07f2e4-b830-4619-8a36-92fda9d3c7ea/sist-en-](https://standards.iteh.ai/catalog/standards/sist/8d07f2e4-b830-4619-8a36-92fda9d3c7ea/sist-en-15189-2022)

### Endorsement notice

The text of ISO 15189:2022 has been approved by CEN as EN ISO 15189:2022 without any modification.



INTERNATIONAL  
STANDARD

ISO  
15189

Fourth edition  
2022-12

---

---

**Medical laboratories — Requirements  
for quality and competence**

*Laboratoires de biologie médicale — Exigences concernant la qualité  
et la compétence*

iTeh STANDARD PREVIEW  
(standards.iteh.ai)

[SIST EN ISO 15189:2023](https://standards.iteh.ai/catalog/standards/sist/8d07f2e4-b830-4619-8a36-92fda9d3c7ea/sist-en-iso-15189-2023)

<https://standards.iteh.ai/catalog/standards/sist/8d07f2e4-b830-4619-8a36-92fda9d3c7ea/sist-en-iso-15189-2023>



Reference number  
ISO 15189:2022(E)

© ISO 2022

iTeh STANDARD PREVIEW  
(standards.iteh.ai)

SIST EN ISO 15189:2023

<https://standards.iteh.ai/catalog/standards/sist/8d07f2e4-b830-4619-8a36-92fda9d3c7ea/sist-en-iso-15189-2023>



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2022

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland



# Contents

Page

<b>Foreword</b> .....	<b>vi</b>
<b>Introduction</b> .....	<b>vii</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms and definitions</b> .....	<b>1</b>
<b>4 General requirements</b> .....	<b>8</b>
4.1 Impartiality.....	8
4.2 Confidentiality.....	8
4.2.1 Management of information.....	8
4.2.2 Release of information.....	9
4.2.3 Personnel responsibility.....	9
4.3 Requirements regarding patients.....	9
<b>5 Structural and governance requirements</b> .....	<b>9</b>
5.1 Legal entity.....	9
5.2 Laboratory director.....	10
5.2.1 Laboratory director competence.....	10
5.2.2 Laboratory director responsibilities.....	10
5.2.3 Delegation of duties.....	10
5.3 Laboratory activities.....	10
5.3.1 General.....	10
5.3.2 Conformance with requirements.....	10
5.3.3 Advisory activities.....	10
5.4 Structure and authority.....	11
5.4.1 General.....	11
5.4.2 Quality management.....	11
5.5 Objectives and policies.....	11
5.6 Risk management.....	12
<b>6 Resource requirements</b> .....	<b>12</b>
6.1 General.....	12
6.2 Personnel.....	12
6.2.1 General.....	12
6.2.2 Competence requirements.....	12
6.2.3 Authorization.....	13
6.2.4 Continuing education and professional development.....	13
6.2.5 Personnel records.....	13
6.3 Facilities and environmental conditions.....	13
6.3.1 General.....	13
6.3.2 Facility controls.....	14
6.3.3 Storage facilities.....	14
6.3.4 Personnel facilities.....	14
6.3.5 Sample collection facilities.....	14
6.4 Equipment.....	15
6.4.1 General.....	15
6.4.2 Equipment requirements.....	15
6.4.3 Equipment acceptance procedure.....	15
6.4.4 Equipment instructions for use.....	15
6.4.5 Equipment maintenance and repair.....	15
6.4.6 Equipment adverse incident reporting.....	16
6.4.7 Equipment records.....	16
6.5 Equipment calibration and metrological traceability.....	17
6.5.1 General.....	17
6.5.2 Equipment calibration.....	17

## ISO 15189:2022(E)

6.5.3	Metrological traceability of measurement results .....	17
6.6	Reagents and consumables .....	18
6.6.1	General .....	18
6.6.2	Reagents and consumables — Receipt and storage .....	18
6.6.3	Reagents and consumables — Acceptance testing .....	18
6.6.4	Reagents and consumables — Inventory management .....	18
6.6.5	Reagents and consumables — Instructions for use .....	19
6.6.6	Reagents and consumables — Adverse incident reporting .....	19
6.6.7	Reagents and consumables — Records .....	19
6.7	Service agreements .....	19
6.7.1	Agreements with laboratory users .....	19
6.7.2	Agreements with POCT operators .....	19
6.8	Externally provided products and services .....	20
6.8.1	General .....	20
6.8.2	Referral laboratories and consultants .....	20
6.8.3	Review and approval of externally provided products and services .....	20
<b>7</b>	<b>Process requirements .....</b>	<b>21</b>
7.1	General .....	21
7.2	Pre-examination processes .....	21
7.2.1	General .....	21
7.2.2	Laboratory information for patients and users .....	21
7.2.3	Requests for providing laboratory examinations .....	21
7.2.4	Primary sample collection and handling .....	22
7.2.5	Sample transportation .....	23
7.2.6	Sample receipt .....	24
7.2.7	Pre-examination handling, preparation, and storage .....	24
7.3	Examination processes .....	25
7.3.1	General .....	25
7.3.2	Verification of examination methods .....	25
7.3.3	Validation of examination methods .....	25
7.3.4	Evaluation of measurement uncertainty (MU) .....	26
7.3.5	Biological reference intervals and clinical decision limits .....	26
7.3.6	Documentation of examination procedures .....	27
7.3.7	Ensuring the validity of examination results .....	27
7.4	Post-examination processes .....	30
7.4.1	Reporting of results .....	30
7.4.2	Post-examination handling of samples .....	32
7.5	Nonconforming work .....	33
7.6	Control of data and information management .....	33
7.6.1	General .....	33
7.6.2	Authorities and responsibilities for information management .....	33
7.6.3	Information systems management .....	34
7.6.4	Downtime plans .....	34
7.6.5	Off site management .....	34
7.7	Complaints .....	34
7.7.1	Process .....	34
7.7.2	Receipt of complaint .....	35
7.7.3	Resolution of complaint .....	35
7.8	Continuity and emergency preparedness planning .....	35
<b>8</b>	<b>Management system requirements .....</b>	<b>35</b>
8.1	General requirements .....	35
8.1.1	General .....	35
8.1.2	Fulfilment of management system requirements .....	36
8.1.3	Management system awareness .....	36
8.2	Management system documentation .....	36
8.2.1	General .....	36
8.2.2	Competence and quality .....	36

8.2.3	Evidence of commitment.....	36
8.2.4	Documentation.....	36
8.2.5	Personnel access.....	36
8.3	Control of management system documents.....	37
8.3.1	General.....	37
8.3.2	Control of documents.....	37
8.4	Control of records.....	37
8.4.1	Creation of records.....	37
8.4.2	Amendment of records.....	37
8.4.3	Retention of records.....	38
8.5	Actions to address risks and opportunities for improvement.....	38
8.5.1	Identification of risks and opportunities for improvement.....	38
8.5.2	Acting on risks and opportunities for improvement.....	38
8.6	Improvement.....	39
8.6.1	Continual improvement.....	39
8.6.2	Laboratory patients, user, and personnel feedback.....	39
8.7	Nonconformities and corrective actions.....	39
8.7.1	Actions when nonconformity occurs.....	39
8.7.2	Corrective action effectiveness.....	40
8.7.3	Records of nonconformities and corrective actions.....	40
8.8	Evaluations.....	40
8.8.1	General.....	40
8.8.2	Quality indicators.....	40
8.8.3	Internal audits.....	40
8.9	Management reviews.....	41
8.9.1	General.....	41
8.9.2	Review input.....	41
8.9.3	Review output.....	41
<b>Annex A (normative) Additional requirements for Point-of-Care Testing (POCT).....</b>		<b>43</b>
<b>Annex B (informative) Comparison between ISO 9001:2015 and ISO 15189:2022 (this document).....</b>		<b>44</b>
<b>Annex C (informative) Comparison between ISO 15189:2012 and ISO 15189:2022 (this document).....</b>		<b>54</b>
<b>Bibliography.....</b>		<b>61</b>

## ISO 15189:2022(E)

### Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 212, *Clinical laboratory testing and in vitro diagnostic test systems*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 140, *In vitro diagnostic medical devices*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This fourth edition cancels and replaces the third edition (ISO 15189:2012), which has been technically revised. It also replaces ISO 22870:2016.

The main changes are as follows:

- Alignment with ISO/IEC 17025:2017 resulted in the management requirements now appearing at the end of the document;
- Requirements for point-of-care testing (POCT), previously in ISO 22870, have been incorporated;
- Increased emphasis on risk management.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

## Introduction

The objective of this document is to promote the welfare of patients and satisfaction of laboratory users through confidence in the quality and competence of medical laboratories.

This document contains requirements for the medical laboratory to plan and implement actions to address risks and opportunities for improvement. Benefits of this approach include: increasing the effectiveness of the management system, decreasing probability of invalid results, and reducing potential harm to patients, laboratory personnel, the public and the environment.

The requirements for risk management are aligned with the principles of ISO 22367.

The requirements for laboratory safety are aligned with the principles of ISO 15190.

The requirements for sample collection and transport are aligned with ISO 20658.<sup>1)</sup>

This document contains the requirements for point-of-care testing (POCT) and supersedes ISO 22870, which will be withdrawn upon publication of this document.

The format of this document is based on ISO/IEC 17025:2017.

The medical laboratory is essential to patient care; activities are provided within an ethical and governance framework, that recognizes the obligations of healthcare providers to the patient. These activities are undertaken in a timely manner to meet the needs of all patients and the personnel responsible for the care of those patients. Activities include arrangements for examination requests, patient preparation, patient identification, collection of samples, transportation, processing of patient samples, selection of examinations that are fit for intended use, examination of samples, sample storage, as well as subsequent interpretation, result reporting and advice to laboratory users. This may also include the provision of results to the patient, arrangements for urgent testing and the notification of critical results.

While this document is intended for use throughout the currently recognized medical laboratory disciplines, it can effectively be applied to other healthcare services, such as diagnostic imaging, respiratory therapy, physiological sciences, blood banks and transfusion services.

The use of this document facilitates cooperation between medical laboratories and other healthcare services, assists in the exchange of information, and in the harmonization of methods and procedures.

The comparability of patient examination results between medical laboratories, regardless of city or country, is facilitated when medical laboratories conform to this document.

When a laboratory seeks accreditation, it should select an accreditation body which operates in accordance with ISO/IEC 17011, and which takes into account the particular requirements of medical laboratories.

Comparisons between this document, ISO 9001:2015 and ISO/IEC 17025:2017 are in [Annex B](#). The comparison of ISO 15189:2012 to ISO 15189:2022 (this document) is in [Annex C](#).

---

1) First edition under preparation (previous edition was a Technical Specification). Stage at the time of publication: ISO/DIS 20658:2022.



# Medical laboratories — Requirements for quality and competence

## 1 Scope

This document specifies requirements for quality and competence in medical laboratories.

This document is applicable to medical laboratories in developing their management systems and assessing their competence. It is also applicable for confirming or recognizing the competence of medical laboratories by laboratory users, regulatory authorities and accreditation bodies.

This document is also applicable to point-of-care testing (POCT).

NOTE International, national, or regional regulations or requirements can also apply to specific topics covered in this document.

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO/IEC Guide 99:2007, *International vocabulary of metrology — Basic and general concepts and associated terms (VIM)*

NOTE ISO/IEC Guide 99 is also known as the Joint Committee for Guides in Metrology (JCGM) 200.

ISO/IEC 17000:2020, *Conformity assessment — Vocabulary and general principles*

ISO/IEC 17025:2017, *General requirements for the competence of testing and calibration laboratories*

## 3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO/IEC Guide 99 and ISO/IEC 17000 and the following apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

### 3.1

#### **bias measurement bias**

estimate of a systematic measurement error

Note 1 to entry: This definition only applies to quantitative measurements

[SOURCE: ISO/IEC Guide 99:2007, 2.18, modified — Note 1 to entry has been added.]