



SLOVENSKI STANDARD
oSIST prEN ISO 80601-2-85:2020
01-maj-2020

Medicinska električna oprema - 2-85. del: Posebne zahteve za osnovno varnost in bistvene lastnosti opreme za cerebralno oksimetrijo (ISO/DIS 80601-2-85:2020)

Medical electrical equipment - Part 2-85: Particular requirements for basic safety and essential performance of cerebral tissue oximeter equipment (ISO/DIS 80601-2-85:2020)

Medizinische elektrische Geräte - Teil 2-85: Besondere Anforderungen für die Sicherheit einschließlich der wesentlichen Leistungsmerkmale von Geräten für die zerebrale Oxymetrie (ISO/DIS 80601-2-85:2020)

(standards.iteh.ai)

Appareils électromédicaux - Partie 2-85: Exigences particulières pour la sécurité de base et les performances essentielles des oxymètres pour tissu cérébral (ISO/DIS 80601-2-85:2020)

oSIST prEN ISO 80601-2-85:2020
<https://standards.iteh.ai/catalog/standards/sist/d4b3e4fa-5ece-4cae-bedd-d44875ccca6f/ksist-fpren-iso-80601-2-85-2020>

Ta slovenski standard je istoveten z: prEN ISO 80601-2-85

ICS:

11.040.10	Anestezijska, respiratorna in reanimacijska oprema	Anaesthetic, respiratory and reanimation equipment
-----------	--	--

oSIST prEN ISO 80601-2-85:2020 en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[ksIST FprEN ISO 80601-2-85:2020](https://standards.iteh.ai/catalog/standards/sist/d4b3e4fa-5ece-4cae-bedd-d44875ccca6f/ksist-fpren-iso-80601-2-85-2020)

<https://standards.iteh.ai/catalog/standards/sist/d4b3e4fa-5ece-4cae-bedd-d44875ccca6f/ksist-fpren-iso-80601-2-85-2020>

DRAFT INTERNATIONAL STANDARD

ISO/DIS 80601-2-85

ISO/TC 121/SC 3

Secretariat: ANSI

Voting begins on:
2020-03-09Voting terminates on:
2020-06-01

Medical electrical equipment —

Part 2-85:

Particular requirements for basic safety and essential performance of cerebral tissue oximeter equipment

ICS: 11.040.10

iTeh STANDARD PREVIEW (standards.iteh.ai)

[ksIST FprEN ISO 80601-2-85:2020](https://standards.iteh.ai/catalog/standards/sist/d4b3e4fa-5ece-4cae-bedd-d44875ccca6f/ksist-fpren-iso-80601-2-85-2020)<https://standards.iteh.ai/catalog/standards/sist/d4b3e4fa-5ece-4cae-bedd-d44875ccca6f/ksist-fpren-iso-80601-2-85-2020>

This document is circulated as received from the committee secretariat.

This draft is submitted to a parallel vote in ISO and in IEC.

ISO/CEN PARALLEL PROCESSING

THIS DOCUMENT IS A DRAFT CIRCULATED FOR COMMENT AND APPROVAL. IT IS THEREFORE SUBJECT TO CHANGE AND MAY NOT BE REFERRED TO AS AN INTERNATIONAL STANDARD UNTIL PUBLISHED AS SUCH.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.



Reference number
ISO/DIS 80601-2-85:2020(E)

© ISO 2020

iTeh STANDARD PREVIEW (standards.iteh.ai)

[ksIST FprEN ISO 80601-2-85:2020](https://standards.iteh.ai/catalog/standards/sist/d4b3e4fa-5ece-4cae-bedd-d44875ccca6f/ksist-fpren-iso-80601-2-85-2020)
<https://standards.iteh.ai/catalog/standards/sist/d4b3e4fa-5ece-4cae-bedd-d44875ccca6f/ksist-fpren-iso-80601-2-85-2020>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

CONTENTS

36			
37	201.1	Scope, object and related standards.....	1
38	201.2	Normative references	3
39	201.3	Terms and definitions.....	5
40	201.4	General requirements	10
41	201.5	General requirements for testing of <i>ME equipment</i>.....	12
42	201.6	Classification of <i>ME equipment</i> and <i>ME systems</i>	12
43	201.7	<i>ME equipment</i> identification, marking and documents.....	12
44	201.8	Protection against electrical <i>hazards</i> from <i>ME equipment</i>.....	16
45	201.9	Protection against mechanical <i>hazards</i> of <i>ME equipment</i> and <i>ME systems</i>	17
46	201.10	Protection against unwanted and excessive radiation <i>hazards</i>.....	17
47	201.11	Protection against excessive temperatures and other <i>hazards</i>	17
48	201.12	<i>Accuracy</i> of controls and instruments and protection against hazardous outputs	
49		21
50	201.13	<i>Hazardous situations</i> and fault conditions for <i>ME equipment</i>.....	25
51	201.14	<i>Programmable electrical medical systems (PEMS)</i>	26
52	201.15	Construction of <i>ME equipment</i>.....	26
53	201.16	<i>ME systems</i>	28
54	201.17	Electromagnetic compatibility of <i>ME equipment</i> and <i>ME systems</i>	28
55	201.101	<i>Cerebral tissue oximeter probes</i> and <i>probe cable extenders</i>	28
56	201.102	<i>Functional connection</i>.....	29
57	202	Electromagnetic disturbances – Requirements and tests	29
58	206	Usability	31
59	208	General requirements, tests and guidance for alarm systems in medical	
60		electrical equipment and medical electrical systems	31
61	211	Requirements for medical electrical equipment and medical	
62		electrical systems used in the home healthcare environment	32
63	212	Requirements for medical electrical equipment and medical	
64		electrical systems used in the emergency medical services environment.....	32
65	Annex C (informative)	Guide to marking and labelling requirements for	
66		<i>ME equipment</i> and <i>ME systems</i>.....	33
67	Annex D (informative)	Symbols on marking	37
68	Annex AA (informative)	Particular guidance and rationale	39
69	Annex BB (informative)	Skin temperature at the <i>cerebral tissue oximeter probe</i>	50
70	Annex CC (informative)	Determination of <i>accuracy</i>	52
71	Annex DD (informative)	Characteristics of a <i>tissue haemoglobin phantom</i> for the	
72		<i>verification</i> of the <i>accuracy</i> of <i>cerebral tissue oximeter equipment</i>.....	58

ISO/DIS 80601-2-85:2020(E)

73	Annex EE (informative) Guideline for evaluating and documenting <i>StO₂ accuracy</i>	
74	in human subjects	68
75	Annex FF (informative) <i>Functional testers for cerebral tissue oximeter equipment</i>	74
76	Annex GG (informative) Concepts of <i>ME equipment</i> response time.....	78
77	Annex HH (normative) Data interface requirements.....	83
78	Annex II (informative) Comparison of methods of performance evaluation	87
79	Annex JJ (informative) Reference to the <i>essential principles</i>	92
80	Annex KK (informative) Terminology — alphabetized index of defined terms.....	96

81

82

iTeh STANDARD PREVIEW (standards.iteh.ai)

[ksIST FprEN ISO 80601-2-85:2020](https://standards.iteh.ai/catalog/standards/sist/d4b3e4fa-5ece-4cae-bedd-d44875ccca6f/ksist-fpren-iso-80601-2-85-2020)

<https://standards.iteh.ai/catalog/standards/sist/d4b3e4fa-5ece-4cae-bedd-d44875ccca6f/ksist-fpren-iso-80601-2-85-2020>

ISO/DIS 80601-2-85:2020(E)

83

Foreword

84 ISO (the International Organization for Standardization) is a worldwide federation of national standards
85 bodies (ISO member bodies). The work of preparing International Standards is normally carried out
86 through ISO technical committees. Each member body interested in a subject for which a technical
87 committee has been established has the right to be represented on that committee. International
88 organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO
89 collaborates closely with the International Electrotechnical Commission (IEC) on all matters of
90 electrotechnical standardization.

91 The procedures used to develop this document and those intended for its further maintenance are
92 described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the
93 different types of ISO documents should be noted. This document was drafted in accordance with the
94 editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

95 Attention is drawn to the possibility that some of the elements of this document may be the subject of
96 patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any
97 patent rights identified during the development of the document will be in the Introduction and/or on
98 the ISO list of patent declarations received (see www.iso.org/patents).

99 Any trade name used in this document is information given for the convenience of users and does not
100 constitute an endorsement.

iTeh STANDARD PREVIEW

101 For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and
102 expressions related to conformity assessment, as well as information about ISO's adherence to the World
103 Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL:
104 www.iso.org/iso/foreword.html. [https://standards.iteh.ai/catalog/standards/sist/d4b3e4fa-5ece-4cae-bedd-](https://standards.iteh.ai/catalog/standards/sist/d4b3e4fa-5ece-4cae-bedd-d44875ccca6f/ksist-fpren-iso-80601-2-85-2020)

105 This document was prepared jointly by Technical Committee ISO/TC 121, *Anaesthetic and respiratory*
106 *equipment*, Subcommittee SC 3, *Respiratory devices and related equipment used for patient care*, and
107 Technical Committee IEC/TC 62, *Electrical equipment in medical practice*, Subcommittee SC 62D,
108 *Electromedical equipment*. The draft was circulated for voting to the national bodies of both ISO and IEC.

109 This first edition of ISO 80601-2-85 includes an alignment with Amendment 1 of both the third edition of
110 IEC 60601-1 and the second edition of IEC 60601-1-8, as well as the fourth edition of IEC 60601-1-2, the
111 third edition of IEC 60601-1-6, the second edition of IEC 60601-1-11 and IEC 60601-1-12.

112 A list of all the parts of the ISO/IEC 80601 series is available on the ISO website.

ISO/DIS 80601-2-85:2020(E)

Introduction

The approximation of cerebral tissue oximetry is increasingly used in many areas of medicine. This document covers *basic safety* and *essential performance* requirements achievable within the limits of existing technology.

Annex AA contains a rationale for some of the requirements. It is included to provide additional insight into the reasoning of the committees that led to a requirement and identifying the *hazards* that the requirement addresses.

Annex BB is a literature survey relevant to the determination of the maximum safe temperature of the interface between a *cerebral tissue oximeter probe* and a *patient's* tissue.

Annex CC discusses both the formulae used to evaluate the *StO₂ accuracy* of *cerebral tissue oximeter equipment* measurements, and the names that are assigned to those formulae.

Annex DD presents guidance on using in-vitro methods for assessing the performance of *cerebral tissue oximeter equipment*.

Annex EE presents a guideline for a *controlled desaturation study* for the *verification* of *cerebral tissue oximeter equipment*.

Annex FF is a description of *functional testers* for use with *cerebral tissue oximeter equipment*.

Annex GG describes concepts of *cerebral tissue oximeter equipment* response time.

Annex HH describes data interface requirements.

Annex II is a comparison between human desaturations (in-vivo) and *tissue haemoglobin phantom* desaturations (in-vitro) for assessing *StO₂ accuracy*.

Annex JJ contains Reference to the *essential principles*.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this document, the following print types are used:

- requirements and definitions: roman type;
- *Instructions, test specifications and terms defined in Clause 3 of the general standard, in this document or as noted: italic type;*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type; normative text of tables is also in a smaller type;

In referring to the structure of this document, the term

- “clause” means one of the numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 201.7 includes subclauses 201.7.1, 201.7.2) and
- “subclause” means a numbered subdivision of a clause (e.g. 201.7.1, 7.2 and 201.7.2.1 are all subclauses of Clause 201.7).

ISO/DIS 80601-2-85:2020(E)

146 References to clauses within this document are preceded by the term “Clause” followed by the clause
147 number. References to subclauses within this document are by number only.

148 In this document, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination
149 of the conditions is true.

150 The verbal forms used in this document conform to usage described in Annex H of the ISO/IEC Directives,
151 Part 2. For the purposes of this document, the auxiliary verb:

152 – “shall” means that compliance with a requirement or a test is mandatory for compliance with this
153 document;

154 – “should” means that compliance with a requirement or a test is recommended but is not mandatory
155 for compliance with this document; and

156 – “may” is used to describe permission (e.g. a permissible way to achieve compliance with a
157 requirement or test);

158 – “can” is used to describe a possibility or capability; and

159 – “must” is used to express an external constraint.

160 An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates
161 that there is guidance or rationale related to that item in AA.

162

163

ITEH STANDARD PREVIEW
(standards.iteh.ai)
[kSIST FprEN ISO 80601-2-85:2020](https://standards.iteh.ai/catalog/standards/sist/d4b3e4fa-5ece-4cae-bedd-d44875ccca6f/ksist-fpren-iso-80601-2-85-2020)
<https://standards.iteh.ai/catalog/standards/sist/d4b3e4fa-5ece-4cae-bedd-d44875ccca6f/ksist-fpren-iso-80601-2-85-2020>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[ksIST FprEN ISO 80601-2-85:2020](https://standards.iteh.ai/catalog/standards/sist/d4b3e4fa-5ece-4cae-bedd-d44875ccca6f/ksist-fpren-iso-80601-2-85-2020)

<https://standards.iteh.ai/catalog/standards/sist/d4b3e4fa-5ece-4cae-bedd-d44875ccca6f/ksist-fpren-iso-80601-2-85-2020>

164 **Medical electrical equipment — Part 2-85: Particular**
 165 **requirements for basic safety and essential performance of**
 166 **cerebral tissue oximeter equipment**

167 **201.1 Scope, object and related standards**

168 Clause 1 of the general standard¹ applies, except as follows:

169 **201.1.1 * Scope**

170 *Replacement:*

171 This document applies to *basic safety and essential performance of cerebral tissue oximeter equipment*,
 172 that employs light at multiple wavelengths to derive a quantitative measure of oxygen saturation of
 173 haemoglobin within the volume of tissue sampled under the *probe* attached to the head. The *cerebral*
 174 *tissue oximeter equipment* can be based on continuous light, frequency domain or time domain
 175 technologies. This document applies to *ME equipment* used in a hospital environment as well as when
 176 used outside the hospital environment, such as in ambulances and air transport. Additional standards
 177 may apply to *ME equipment* for those environments of use.

178 NOTE 1 *Cerebral tissue oximeters* are sometimes referred to as near infrared spectroscopy equipment in medical
 179 literature.

180 Not included within the scope of this document are:

- 181 — *invasive tissue or vascular oximeters*;
<http://standards.iteh.ai/catalog/standards/sist/d4b3e4fa-5ece-4cae-bedd-d44875ccca6f/ksist-fpren-iso-80601-2-85-2020>
- 182 — *oximeters that require a blood sample from the patient*;
- 183 — *equipment measuring dissolved oxygen*;
- 184 — *ME equipment*, or part thereof, that measures path-length-dependent haemoglobin change. The
 185 requirements for functional near-infrared spectroscopy equipment are found in ISO 80601-2-71 [2];
- 186 — *ME equipment*, or part thereof, that measures arterial saturation based on pulsatile changes in tissue
 187 optical properties (SpO_2). The requirements for pulse oximeter equipment are found in
 188 ISO 80601-2-61 [2];
- 189 — *ME equipment*, or part thereof, that claims to monitor tissue in other parts of the body other than the
 190 head.

191 These requirements also apply to *cerebral tissue oximeter equipment*, including *cerebral tissue oximeter*
 192 *monitors*, *cerebral tissue oximeter probes* and *probe cable extenders*, which have been *remanufactured*.

193 If a clause or subclause is specifically intended to be applicable to *ME equipment* only, or to *ME systems*
 194 only, the title and content of that clause or subclause will say so. If that is not the case, the clause or
 195 subclause applies both to *ME equipment* and to *ME systems*, as relevant.

¹ The general standard is IEC 60601-1:2005+AMD1:2012.

ISO/DIS 80601-2-85:2020(E)

196 *Hazards* inherent in the intended physiological function of *ME equipment* or *ME systems* within the scope
 197 of this document are not covered by specific requirements in this document except in 201.11 and in
 198 201.7.2.13 and 201.8.4.1 of the general standard.

199 NOTE 2 See also 4.2 of the general standard. “The general standard” is IEC 60601-1:2005+AMD1:2012, Medical
 200 electrical equipment – Part 1: General requirements for basic safety and essential performance.

201 This document can also be applied to *ME equipment* and their *accessories* used for compensation or
 202 alleviation of disease, injury or disability.

203 This document is not applicable to remote or slave (secondary) equipment that displays StO_2 values that
 204 are located outside of the *patient environment*.

205 NOTE 3 *ME equipment* that provides selection between diagnostic and monitoring functions is expected to meet
 206 the requirements of the appropriate document when configured for that function.

207 This document is a particular standard in the IEC 60601-1 and ISO/IEC 80601 series of standards.

201.1.2 Object

208 *Replacement:*

210 The object of this document is to establish particular *basic safety* and *essential performance* requirements
 211 for *cerebral tissue oximeter equipment* [as defined in 201.3.202] and its *accessories*.

212 NOTE *Accessories* are included because the combination of the *cerebral tissue oximeter monitor* and the
 213 *accessories* needs to be adequately safe. *Accessories* can have a significant impact on the *basic safety* or *essential*
 214 *performance of cerebral tissue oximeter equipment*.

201.1.3 Collateral standards

215 *Amendment (add after existing text):*

217 This document refers to those applicable collateral standards that are listed in Clause 2 of the general
 218 standard and Clause 201.2 of this document.

219 IEC 60601-1-2, IEC 60601-1-6, IEC 60601-1-8, IEC 60601-1-11 and IEC 60601-1-12 apply as modified in
 220 Clauses 202, 206, 208, 211 and 212 respectively. IEC 60601-1-3^[4] does not apply. All other published
 221 collateral standards in the IEC 60601-1 series apply as published.

201.1.4 Particular standards

222 *Replacement:*

224 In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in
 225 the general standard, including the collateral standards, as appropriate for the particular *ME equipment*
 226 under consideration, and may add other *basic safety* or *essential performance* requirements.

227 A requirement of a particular standard takes priority over the general standard or the collateral
 228 standards.

229 For brevity, IEC 60601-1:2005+AMD1:2012 is referred to in this document as the general standard.
 230 Collateral standards are referred to by their document number.

231 The numbering of clauses and subclauses of this document corresponds to those of the general standard
 232 with the prefix “201” (e.g. 201.1 in this document addresses the content of Clause 1 of the general

ISO/DIS 80601-2-85:2020(E)

233 standard) or applicable collateral standard with the prefix "2xx" where xx is the final digits of the
 234 collateral standard document number (e.g. 202.4 in this document addresses the content of Clause 4 of
 235 the IEC 60601-1-2 collateral standard, 208.4 in this document addresses the content of Clause 4 of the
 236 IEC 60601-1-8 collateral standard, etc.). The changes to the text of the general standard are specified by
 237 the use of the following words:

238 "Replacement" means that the clause or subclause of the general standard or applicable collateral
 239 standard is replaced completely by the text of this document.

240 "Addition" means that the text of this document is additional to the requirements of the general standard
 241 or applicable collateral standard.

242 "Amendment" means that the clause or subclause of the general standard or applicable collateral
 243 standard is amended as indicated by the text of this document.

244 Subclauses or figures that are additional to those of the general standard are numbered starting from
 245 201.101, additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

246 Subclauses or figures that are additional to those of a collateral standard are numbered starting from 2xx,
 247 where "x" is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

248 The term "this document" is used to make reference to the general standard, any applicable collateral
 249 standards and this particular document taken together.

250 Where there is no corresponding clause or subclause in this document, the section, clause or subclause
 251 of the general standard or applicable collateral standard, although possibly not relevant, applies without
 252 modification; where it is intended that any part of the general standard or applicable collateral standard,
 253 although possibly relevant, is not to be applied, a statement to that effect is given in this particular
 254 document.

255 **201.2 Normative references**

256 The following documents are referred to in the text in such a way that some or all of their content
 257 constitutes requirements of this document. For dated references, only the edition cited applies. For
 258 undated references, the latest edition of the referenced document (including any amendments) applies.

259 NOTE 1 The way in which these referenced documents are cited in normative requirements determines the extent
 260 (in whole or in part) to which they apply.

261 NOTE 2 Informative references are listed in the Bibliography.

262 Clause 2 of the general standard applies, except as follows:

263 *Replacement:*

264 ISO 7000:2019, *Graphical symbols for use on equipment — Registered symbols*

265 ISO 7010:2019, *Graphical symbols — Safety colours and safety signs — Registered safety signs*

ISO/DIS 80601-2-85:2020(E)

266 ISO 15223-1:—², *Medical devices — Symbols to be used with medical device labels, labelling and*
 267 *information to be supplied — Part 1: General requirements*

268 IEC 60529:2013, *Degrees of protection provided by enclosures (IP code)*

269 IEC 60601-1-2:2014, *Medical electrical equipment — Part 1-2: General requirements for basic safety and*
 270 *essential performance — Collateral Standard: Electromagnetic disturbances — Requirements and tests*

271 IEC 60601-1-6:2010+AMD1:2013, *Medical electrical equipment — Part 1-6: General requirements for*
 272 *basic safety and essential performance — Collateral standard: Usability*

273 IEC 60601-1-8:2006+AMD1:2012, *Medical electrical equipment — Part 1-8: General requirements for*
 274 *basic safety and essential performance — Collateral Standard: General requirements, tests and guidance for*
 275 *alarm systems in medical electrical equipment and medical electrical systems*

276 IEC 60825-1:2014, *Safety of laser products - Part 1: Equipment classification and requirements*

277 *Addition:*

278 ISO 14155:2011, *Clinical investigation of medical devices for human subjects — Good clinical practice*

279 ISO 14937:2009, *Sterilization of health care products — General requirements for characterization of a*
 280 *sterilizing agent and the development, validation and routine control of a sterilization process for medical*
 281 *devices*

282 ISO 16142-1:2016, *Medical devices — Recognized essential principles of safety and performance of medical*
 283 *devices — Part 1: General essential principles and additional specific essential principles for all non-IVD*
 284 *medical devices and guidance on the selection of standards*

285 ISO 17664:2017, *Processing of health care products — Information to be provided by the medical device*
 286 *manufacturer for the processing of medical devices*

287 IEC 60068-2-27:2008+AMD1:2013, *Environmental testing — Part 2-27: Tests — Test Ea and guidance:*
 288 *Shock*

289 IEC 60068-2-31:2008, *Environmental testing — Part 2-31: Tests — Test Ec: Rough handling shocks,*
 290 *primarily for equipment-type specimens*

291 IEC 60068-2-64:2008, *Environmental testing — Part 2-64: Tests — Test Fh: Vibration, broadband random*
 292 *and guidance*

293 IEC 60601-1-11:2015, *Medical electrical equipment — Part 1-11: General requirements for basic safety and*
 294 *essential performance — Collateral Standard: Requirements for medical electrical equipment and medical*
 295 *electrical systems used in the home healthcare environment*

296 IEC 60601-1-12:2014, *Medical electrical equipment — Part 1-12: General requirements for basic safety and*
 297 *essential performance — Collateral Standard: Requirements for medical electrical equipment and medical*
 298 *electrical systems used in the emergency medical services environment*

² Under preparation. Stage at the time of publication: ISO DIS 15223-1:2020.

ISO/DIS 80601-2-85:2020(E)

- 299 IEC 62366-1:2015, *Medical devices — Part 1: Application of usability engineering to medical devices*
- 300 IEC 60601-2-2:2009, *Medical electrical equipment - Part 2-2: Particular requirements for the basic safety*
301 *and essential performance of high frequency surgical equipment and high frequency surgical accessories*
- 302 IEC 62471:2006, *Photobiological safety of lamps and lamp systems*
- 303 AAMI 2700-1:2019 (formerly ASTM F2761-09), *Medical devices and medical systems — Essential safety*
304 *requirements for equipment comprising the patient-centric integrated clinical environment (ICE) — Part 1:*
305 *General requirements and conceptual model*

201.3 Terms and definitions

307 For the purposes of this document, the terms and definitions given in ISO 16142-1:2016, ISO 17664:2017,
308 IEC 60601-1:2005+AMD 1:2012, IEC 60601-1-2:2014, IEC 60601-1-6:2010+AMD 1:2013,
309 IEC 60601-1-8:2006+AMD 1:2012, IEC 60601-1-11:2015, IEC 60601-1-12:2014, IEC 62366-1:2015,
310 IEC 60601-2-2:2009, AAMI 2700-1:2019 and the following apply.

311 ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- 312 – IEC Electropedia: available at <http://www.electropedia.org/>
- 313 – ISO Online browsing platform: available at <http://www.iso.org/obp>

314 NOTE An alphabetized index of (defined terms is found in Annex KK.

201.3.201**accuracy** **A_{rms}**

318 closeness of agreement between a test result and the true value

319 Note 1 to entry: 201.12.1.101.2 contains methods for estimating the StO_2 accuracy of cerebral tissue oximeter
320 equipment.

321 Note 2 to entry: Additional information is found in Annexes CC, DD, EE and II.

322 Note 3 to entry: In this document, accuracy (A_{rms}) is stated in terms of the root mean square difference. See
323 201.12.1.101.3.

324 [SOURCE: ISO 3534-2:2006 ^[5] 3.3.1, modified, with Notes to entry replaced]

201.3.202**cerebral tissue oximeter****cerebral tissue oximeter equipment**

328 *ME equipment* for the noninvasive estimation of functional oxygen saturation of haemoglobin in cerebral
329 tissue below the probe (StO_2 or rSO_2), based on light interacting with tissue

330 Note 1 to entry: Cerebral tissue oximeter equipment comprises a cerebral tissue oximeter monitor, a probe cable
331 extender, if provided, and a cerebral tissue oximeter probe, which can be combined in a single assembly.

332 Note 2 to entry: Light is more technically referred to as electromagnetic radiation (optical radiation). This document
333 uses the common term.

334 Note 3 to entry: Measurements are based upon light interacting with all tissue under the probe to determine StO_2 .