



SLOVENSKI STANDARD
oSIST prEN 17701-1:2023
01-maj-2023

Rastlinski biostimulanti - Določevanje specifičnih elementov - 1. del: Razklop z zlatotopko za določevanje elementov

Plant biostimulants - Determination of specific elements - Part 1: Digestion by aqua regia for subsequent determination of elements

Pflanzen-Biostimulanzien - Bestimmung spezifischer Elemente - Teil 1: Aufschluss durch Königswasser zur anschließenden Bestimmung der Elemente

Biostimulants des végétaux - Détermination des éléments spécifiques - Partie 1 : Digestion par eau régale pour la détermination subséquente des éléments

Ta slovenski standard je istoveten z: prEN 17701-1

ICS:

65.080 Gnojila Fertilizers

oSIST prEN 17701-1:2023 **en,fr,de**

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

DRAFT
prEN 17701-1

March 2023

ICS 65.080

Will supersede CEN/TS 17701-1:2022

English Version

Plant biostimulants - Determination of specific elements - Part 1: Digestion by aqua regia for subsequent determination of elements

Biostimulants des végétaux - Détermination des
éléments spécifiques - Partie 1 : Digestion par eau
régale pour la détermination subséquente des
éléments

Pflanzen-Biostimulanzien - Bestimmung spezifischer
Elemente - Teil 1: Aufschluss durch Königswasser zur
anschließenden Bestimmung der Elemente

This draft European Standard is submitted to CEN members for enquiry. It has been drawn up by the Technical Committee CEN/TC 455.

If this draft becomes a European Standard, CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

This draft European Standard was established by CEN in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.

Recipients of this draft are invited to submit, with their comments, notification of any relevant patent rights of which they are aware and to provide supporting documentation.

Warning : This document is not a European Standard. It is distributed for review and comments. It is subject to change without notice and shall not be referred to as a European Standard.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents	Page
European foreword	3
Introduction	4
1 Scope	5
2 Normative references	5
3 Terms and definitions	5
4 Principle	5
5 Sampling and sample preparation	5
6 Reagents	5
7 Apparatus	6
8 Procedure	6
8.1 Sample digestion	6
8.2 Sample pre-concentration	7
8.3 Blank	7
9 Test report	7
Annex A (informative) Adjustment to different digestion procedures and measurement apparatuses	8
A.1 Adjustment of the procedure to different final volumes	8
A.2 Calculation of the sample weight	8
A.2.1 General	8
A.2.2 Calculation of the weight of the sample (dry matter)	9
A.2.3 Calculation of the weight of the sample for analysis	9
Annex ZA (informative) Relationship of this European Standard and the essential requirements of Regulation (EU) 2019/1009 making available on the market of EU fertilising products aimed to be covered	10
Bibliography	11

European foreword

This document (prEN 17701-1:2023) has been prepared by Technical Committee CEN/TC 455 “Plant biostimulants”, the secretariat of which is held by AFNOR.

This document is currently submitted to the CEN Enquiry.

This document will supersede CEN/TS 17701-1:2022.

This document has been prepared under a Standardization Request given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s) / Regulation(s).

For relationship with EU Directive(s) / Regulation(s), see informative Annex ZA, which is an integral part of this document.

EN 17701 series, *Plant biostimulants — Determination of specific elements*, consists of the following parts:

- *Part 1: Digestion by aqua regia for subsequent determination of elements;*
- *Part 2: Determination of total content of Cd, Pb, Ni, As, Cr, Cu and Zn;*
- *Part 3: Determination of mercury.*

ITEH STANDARD PREVIEW
(standards.iteh.ai)

[oSIST prEN 17701-1:2023](https://standards.iteh.ai/catalog/standards/sist/903c0f67-a05d-46ef-bdad-cc52787677c7/osist-pren-17701-1-2023)

<https://standards.iteh.ai/catalog/standards/sist/903c0f67-a05d-46ef-bdad-cc52787677c7/osist-pren-17701-1-2023>

Introduction

This document was prepared by the experts of CEN/TC 455 “Plant biostimulants”. The European Committee for Standardization (CEN) was requested by the European Commission (EC) to draft European standards or European standardization deliverables to support the implementation of Regulation (EU) 2019/1009 of the European Parliament and of the Council of 5 June 2019 laying down rules on the making available on the market of EU fertilising products (“FPR” or “Fertilising Products Regulation”).

This standardization request, presented as M/564 and M/564/Amd1, also contributes to the Communication on “Innovating for Sustainable Growth: A Bio economy for Europe”. Working Group 4 “Other safety parameters” was created to develop a work programme as part of this standardization request. Technical Committee CEN/TC 455 “Plant biostimulants” was established to carry out the work programme that will prepare a series of standards. The interest in biostimulants has increased significantly in Europe as a valuable tool to use in agriculture. Standardization was identified as having an important role in order to promote the use of biostimulants. The work of CEN/TC 455 seeks to improve the reliability of the supply chain, thereby boosting the confidence of farmers, industry, and consumers in biostimulants, and will promote and support commercialisation of the European biostimulant industry.

The preparation of this document is based on a standardization request to CEN by the European Commission and the European Free Trade Association (Mandate M/564 and M/564/Amd1) concerning the modernization of methods of analysis of fertilizers in the framework of Regulation (EU) 2019/1009 of the European Parliament and of the Council.

Aqua regia is applied to digest different matrices for the subsequent determination of many elements. For example, a multi-matrix standard for aqua regia extraction of soils, sludges and biowaste was prepared by CEN/TC 444. A similar procedure was applied for determination of aqua regia extractable contents of arsenic, mercury, cadmium, chromium, nickel and lead in fertilizers and liming materials (standard prepared by CEN/TC 260). CEN/TC 223 published a standard for a similar procedure for soil improvers and growing media. Wide use of the aqua regia digestion, availability of the instruments and the possibility to merge the standards for different matrices in future, were the main reasons for also applying this method of digestion for plant biostimulants. A simple pre-concentration procedure is a part of this document to also cover the digestion of liquid samples with low dry matter content in the case that the measurement method is not sufficiently sensitive. Annex A describes in detail the calculation of the initial sample weight to be taken for the pre-concentration step to achieve concentrations of the individual elements above the limit of quantification (LOQ).

WARNING — Persons using this document should be familiar with usual laboratory practice. This document does not purport to address all of the safety issues, if any, associated with its use. It is the responsibility of the user to establish appropriate health and safety practices and to ensure compliance with any national regulatory conditions.

IMPORTANT — It is absolutely essential that tests conducted according to this document are carried out by suitably trained staff.

1 Scope

This document specifies the method for the digestion of different plant biostimulants with aqua regia to enable a subsequent determination of arsenic (As), cadmium (Cd), copper (Cu), chromium (Cr), mercury (Hg), nickel (Ni), lead (Pb) and zinc (Zn). The method can be also applied for determination of other elements. The method is applicable for all solid and/or liquid plant biostimulants.

This document is also applicable to the fertilizing product blends belonging to PFC 7 where the EU fertilising product plant biostimulant contained in the blend represents the highest % by mass in the blend. In case of equal shares, the user can apply either this or the standard(s) applicable to the other component product(s).

The extracts are suitable for analysis using EN 17701-2:—¹ (ICP-AES) and EN 17701-3:—² (Hg analysis).

NOTE Alternatively, inductively coupled plasma mass spectrometry (ICP-MS) can be used for the determination of the elements in the aqua regia digests if the user proves that the method gives the same results.

2 Normative references

There are no normative references in this document.

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

digestion

mineralization of the organic matter of a sample and dissolution of its mineral part, more or less completely, when reacting with a reagent mixture

3.2

blank test solution

solution prepared in the same way as the test sample solution but omitting the test portion

4 Principle

The samples are digested by boiling in aqua regia for two hours under reflux conditions.

5 Sampling and sample preparation

Sampling and sample preparation are not part of this document. A recommended sampling method and sample preparation method are given in EN 17702-1:—³ and EN 17702-2:—⁴.

6 Reagents

6.1 The concentration of the analyte or interfering substances in the reagents and the water should be negligible compared to the lowest concentration to be determined. Water with a specific conductivity not higher than 0,2 mS/m at 25 °C, free from the elements to be determined.

6.2 Hydrochloric acid 37 %, $c(\text{HCl}) \approx 12 \text{ mol/l}$, $\rho \approx 1,18 \text{ g/ml}$.

¹ Under preparation

² Under preparation

³ Under preparation

⁴ Under preparation

prEN 17701-1:2023 (E)

6.3 Nitric acid 65 %, $c(\text{HNO}_3) \approx 14,3 \text{ mol/l}$, $\rho \approx 1,4 \text{ g/ml}$.

6.4 Antifoaming agent, e.g. *n*-dodecane ($\text{C}_{12}\text{H}_{26}$) or octanol ($\text{C}_8\text{H}_{18}\text{O}$) are suitable.

7 Apparatus

7.1 Common laboratory glassware, all glassware and plastic ware shall be adequately cleaned and stored to avoid any contamination.

7.2 Apparatus for thermal heating digestion, temperature controlled with reaction vessel and reflux condenser. The capacity of the reaction vessel should be at least 5 times of the volume of the aqua regia used.

7.3 Filter paper, ash free and of recognized and tested quality.

7.4 Analytical balance capable of weighing to an accuracy of 1 mg or better.

8 Procedure

CAUTION — Suitable precautions shall be taken to avoid any contact of laboratory staff with acid fumes. The digestion procedure shall be carried out in a well-ventilated fume cupboard.

8.1 Sample digestion

This procedure can be used for all samples of plant biostimulants if a sufficiently sensitive method for determination is used. For some liquid samples with a low dry matter content, a pre-concentration step (8.2) may be necessary.

Weigh 2,5 g to 5,0 g of the solid, or liquid sample representing approximately 1,5 g to 3,0 g of dry sample, to the nearest 0,001 g. Transfer the sample quantitatively to the reaction vessel (7.2). Liquid samples are preferably weighed directly into the reaction vessel (7.2). Moisten the solid sample with about 0,5 ml to 1,0 ml of water and add, while gently mixing, 21 ml of hydrochloric acid (6.2) followed by 7 ml of nitric acid (6.3) drop wise if necessary. Add one drop of the antifoaming agent (6.4) in the case of excessive foaming. Connect condenser to the reaction vessel and let the mixture stand at laboratory temperature until any effervescence almost ceases. Transfer to the heating device and slowly raise the temperature of the reaction mixture to reflux conditions and maintain for 2 h ensuring that the condensation zone is lower than $\frac{1}{3}$ of the height of the condenser. Then allow to cool and rinse the condenser with 10 ml of water (6.1).

Transfer the content of the reaction vessel quantitatively into a 150 ml volumetric flask, dilute to the mark with water (6.1) and mix well.

Test solutions may be filtered or centrifuged if necessary. Use ash-free filter paper of recognized and tested quality (7.3) and discard the first portion of the filtrate (approximately 20 ml).

Carry out the measurement immediately or store the extracts in tightly closed vessels for up to 15 days.

The procedure may be modified for the use of 100 ml or 50 ml volumetric flasks or graduated plastic tubes. In this case the weight of the sample and volumes of the acids should be changed accordingly. For more information see Annex A.

NOTE 1 Weigh the samples for dry matter determination at the same time and from the same sub-sample to ensure that the dry matter is determined on the samples identical to those used for determination of parameters that relates to dry matter.

NOTE 2 If excessive foaming occurs despite the addition of an anti-foaming reagent, it is possible to leave the samples overnight after the addition of the acids to allow a slow oxidation or to use a digestion block with two independent heating zones.

NOTE 3 The presence of small quantities of organic matter after digestion will not affect determination of the elements by ICP-AES and it is not necessary to apply additional organic matter removal.

8.2 Sample pre-concentration

This procedure is necessary for liquid samples with low dry matter content in the case that the measurement method is not sufficiently sensitive. See Annex A for calculation of the initial sample weight to be taken for the pre-concentration step to achieve concentrations of the individual elements above the limit of quantification (LOQ).

Weigh approximately 10 g to 50 g of the liquid sample representing 1 g to 3 g of dry sample to the nearest 0,001 g directly into the reaction vessel (7.2) and heat the vessel to approximately 75 °C to evaporate excessive water. After evaporation to approximately 5 ml and cooling, follow procedure 8.1. Adjustment of the pre-concentration step for different final volumes is given in Annex A.

Evaporation of a large volume of sample usually takes several hours. Higher temperature (up to 95 °C) for faster evaporation may be used if the final digest is not intended for the determination of mercury. The fastest evaporation can be achieved by using a 250 ml beaker placed on a hot plate for evaporation and consequently also for digestion under a watch glass in the same beaker.

8.3 Blank

Prepare a blank test solution following the same procedure as for samples. The measurement of a blank test solution is necessary to determine the contribution of the extracting solution, glassware and filter paper used.

9 Test report

The test report shall contain at least the following information:

- a) all information necessary for the complete identification of the sample;
- b) test method used with reference to this document;
- c) date of sampling and sampling procedure (if known);
- d) date when the extraction was finished;
- e) all operating details not specified in this document, or regarded as optional, together with details of any incidents that occurred when performing the method, which might have influenced the test result(s).

Annex A (informative)

Adjustment to different digestion procedures and measurement apparatuses

A.1 Adjustment of the procedure to different final volumes

Volumetric flasks or graduated plastic tubes for 150 ml, 100 ml, or 50 ml may be used for the final volume adjustment. Table A.1 gives recalculated parameters for the different final volumes.

Table A.1 — Recalculated parameters for different final volumes

V_F ml	V_{HNO_3} ml	V_{HCl} ml	m_{DM} g	V_{max} ml
150	7	21	1,5 to 3	6
100	4,7	14	1 to 2	4
50	2,3	7	0,5 to 1	2

V_F – final volume of a volumetric flask or a graduated plastic tube
 m_{DM} – weight of the sample recalculated to dry matter
 V_{max} – maximum volume after pre-concentration step (approximately)

A.2 Calculation of the sample weight

A.2.1 General

According to the given legislative limits and the limits of quantification for the individual elements and the measurement apparatus, the weight of the sample should be adjusted to ensure that the concentration of the element in the final digest is above the LOQ. The calculation of the weight of the sample is necessary only for the elements with a low legislation limit and/or very high LOQ. Usually only for cadmium, in some cases also for arsenic and nickel. The highest volume calculated for different elements will be applied.

Table A.2 — Legislative limits and typical limits of the quantitative determination

Element	Cd	Pb	Hg	As	Cr	Cu	Ni	Zn
Legislative limit [mg/kg]	1,5	120	1	40	200	600	50	1 500
Legislative limit [μ g/l]	30	2 400	20	800	4 000	12 000	1000	30 000
LOQ [μ g/l]	7,5	100	1,0	75	75	75	75	250

Legislative limits are from Regulation (EU) 2019/1009 of the European Parliament and of the Council
 Legislative limit in [μ g/l] was calculated for 1 g sample (dry matter) in 50 ml
 LOQ – example of typical values for an ICP-AES instrument, radial view
 LOQ – example of a limit for Hg is given for a direct amalgamation technique