

**SLOVENSKI STANDARD
SIST EN IEC 62127-1:2022****01-julij-2022****Nadomešča:****SIST EN 62127-1:2008****SIST EN 62127-1:2008/A1:2014**

Ultrazvok - Hidrofoni - 1. del: Meritve in karakterizacija medicinskih ultrazvočnih polj (IEC 62127-1:2022)

Ultrasonics - Hydrophones - Part 1: Measurement and characterization of medical ultrasonic fields (IEC 62127-1:2022)

Ultraschall - Hydrophone - Teil 1: Messung und Charakterisierung von medizinischen Ultraschallfeldern (IEC 62127-1:2022)

Ultrasons - Hydrophones - Partie 1: Mesurage et caractérisation des champs ultrasoniques médicaux jusqu'à (IEC 62127-1:2022)

<https://standards.iteh.ai/catalog/standards/sist/0d399001-f0ad-45a6-ac82-32650061c7d2/sist-en-iec-62127-1-2022>

Ta slovenski standard je istoveten z: EN IEC 62127-1:2022**ICS:**

11.040.01	Medicinska oprema na splošno	Medical equipment in general
17.140.50	Elektroakustika	Electroacoustics

SIST EN IEC 62127-1:2022**en**

**iTeh STANDARD
PREVIEW
(standards.iteh.ai)**

SIST EN IEC 62127-1:2022

<https://standards.iteh.ai/catalog/standards/sist/0d399001-f0ad-45a6-ac82-32650061c7d2/sist-en-iec-62127-1-2022>

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN IEC 62127-1

May 2022

ICS 17.140.50

Supersedes EN 62127-1:2007 + A1:2013

English Version

**Ultrasonics - Hydrophones - Part 1: Measurement and
characterization of medical ultrasonic fields
(IEC 62127-1:2022)**

Ultrasons - Hydrophones - Partie 1: Mesurage et
caractérisation des champs ultrasoniques médicaux
(IEC 62127-1:2022)

Ultraschall - Hydrophone - Teil 1: Messung und
Charakterisierung von medizinischen Ultraschallfeldern bis
zu 40 MHz
(IEC 62127-1:2022)

This European Standard was approved by CENELEC on 2022-04-29. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

road-45a6-ac82-32650061c7d2/sist-en-iec-62127-1-2022



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

EN IEC 62127-1:2022 (E)**European foreword**

The text of document 87/783/FDIS, future edition 2 of IEC 62127-1, prepared by IEC/TC 87 "Ultrasonics" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN IEC 62127-1:2022.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2023-01-29
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2025-04-29

This document supersedes EN 62127-1:2007 and all of its amendments and corrigenda (if any).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC shall not be held responsible for identifying any or all such patent rights.

Any feedback and questions on this document should be directed to the users' national committee. A complete listing of these bodies can be found on the CENELEC website.

iteh STANDARD
PREVIEW
Endorsement notice
(standards.iteh.ai)

The text of the International Standard IEC 62127-1:2022 was approved by CENELEC as a European Standard without any modification. [SIST EN IEC 62127-1:2022](https://standards.iteh.ai/catalog/standards/sist/0d399001-f0ad-43ac-ab32-650001c012/sist-en-iec-62127-1-2022)

In the official version, for Bibliography, the following notes have to be added for the standards indicated:

IEC 60500	NOTE Harmonized as EN 60500
IEC 60601-2-5	NOTE Harmonized as EN 60601-2-5
IEC 60601-2-37	NOTE Harmonized as EN 60601-2-37
IEC 60601-2-62	NOTE Harmonized as EN 60601-2-62
IEC 61157	NOTE Harmonized as EN 61157
IEC 61161	NOTE Harmonized as EN 61161
IEC 61828	NOTE Harmonized as EN IEC 61828
IEC 61846	NOTE Harmonized as EN 61846
IEC 61847	NOTE Harmonized as EN 61847
IEC/TS 61949	NOTE Harmonized as CLC/TS 61949
IEC 62359	NOTE Harmonized as EN 62359
IEC 63045	NOTE Harmonized as EN IEC 63045

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 Where an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: www.cenelec.eu.

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60565-1	-	Underwater acoustics - Hydrophones - Calibration of hydrophones - Part 1: Procedures for free-field calibration of hydrophones	EN IEC 60565-1	-
IEC 61689	-	Ultrasonics - Physiotherapy systems - Field specifications and methods of measurement in the frequency range 0,5 MHz to 5 MHz	EN IEC 61689	-
IEC 62127-2	-	Ultrasonics - Hydrophones - Part 2: Calibration for ultrasonic fields up to 40 MHz	EN 62127-2	-
IEC 62127-3	-	Ultrasonics - Hydrophones - Part 3: Properties of hydrophones for ultrasonic fields up to 40 MHz	EN 62127-3	-
IEC 63009	-	Ultrasonics - Physiotherapy systems - Field specifications and methods of measurement in the frequency range 20 kHz to 500 kHz	EN IEC 63009	-
ISO 16269-6	-	Statistical interpretation of data – Part 6: Determination of statistical tolerance intervals	-	-
ISO/IEC Guide 98-3 2008	-	Uncertainty of measurement - Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)	-	-

**iTeh STANDARD
PREVIEW
(standards.iteh.ai)**

SIST EN IEC 62127-1:2022

<https://standards.iteh.ai/catalog/standards/sist/0d399001-f0ad-45a6-ac82-32650061c7d2/sist-en-iec-62127-1-2022>



IEC 62127-1

Edition 2.0 2022-03

INTERNATIONAL STANDARD

NORME INTERNATIONALE

iTeh STANDARD**Ultrasonics – Hydrophones – **PREVIEW**
Part 1: Measurement and characterization of medical ultrasonic fields****(standards.iteh.ai)****Ultrasons – Hydrophones –
Partie 1: Mesurage et caractérisation des champs ultrasoniques médicaux****SIST EN IEC 62127-1:2022**

<https://standards.iteh.ai/catalog/standards/sist/0d399001-f0ad-45a6-ac82-32650061c7d2/sist-en-iec-62127-1-2022>

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 17.140.50

ISBN 978-2-8322-1080-1

**Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

CONTENTS

FOREWORD.....	6
INTRODUCTION.....	8
1 Scope.....	9
2 Normative references	9
3 Terms and definitions	10
4 Symbols	32
5 Measurement requirements	34
5.1 Requirements for hydrophones and amplifiers.....	34
5.1.1 Preface.....	34
5.1.2 General	34
5.1.3 Sensitivity of a hydrophone.....	35
5.1.4 Directional response of a hydrophone.....	35
5.1.5 Effective hydrophone size.....	35
5.1.6 Choice of the size of a hydrophone active element	35
5.1.7 Bandwidth.....	37
5.1.8 Linearity	40
5.1.9 Hydrophone signal amplifier	40
5.1.10 Hydrophone cable length and amplifiers	40
5.2 Requirements for positioning and water baths.....	41
5.2.1 General	41
5.2.2 Positioning systems	41
5.2.3 Water bath.....	42
5.3 Requirements for data acquisition and analysis systems.....	43
5.4 Recommendations for ultrasonic equipment being characterized.....	43
6 Measurement procedure	43
6.1 General.....	43
6.2 Preparation and alignment	44
6.2.1 Preparation.....	44
6.2.2 Aligning an ultrasonic transducer and a hydrophone	44
6.3 Measurement.....	44
6.4 Analysis	44
6.4.1 Corrections for restricted bandwidth and spatial resolution	44
6.4.2 Uncertainties	44
7 Beam characterization	45
7.1 General.....	45
7.2 Primary pressure parameters	46
7.2.1 General	46
7.2.2 Peak-compressional acoustic pressure and peak-rarefactional acoustic pressure	47
7.2.3 Spatial-peak RMS acoustic pressure	47
7.2.4 Local distortion parameter	48
7.3 Intensity parameters derived from acoustic pressure.....	48
7.3.1 General	48
7.3.2 Intensity parameters using pulse-pressure-squared integral	49
8 Requirements for specific ultrasonic fields.....	52

8.1	General.....	52
8.2	Diagnostic fields	52
8.2.1	Simplified procedures and guidelines.....	52
8.2.2	Pulsed wave diagnostic equipment	52
8.2.3	Continuous wave diagnostic equipment	53
8.2.4	Diagnostic equipment with low acoustic output	54
8.3	Therapy fields	54
8.3.1	Physiotherapy equipment.....	54
8.3.2	High intensity therapeutic ultrasonic fields	55
8.3.3	Non-focused and weakly focused pressure pulses	55
8.4	Surgical fields	55
8.4.1	Lithotripters and pressure pulse sources for other therapeutic purposes	55
8.4.2	Low frequency surgical applications.....	56
8.5	Fields from other medical applications	56
9	Conformity statement.....	56
9.1	General.....	56
9.2	Maximum probable values.....	56
9.3	Sampling.....	57
Annex A (informative)	General rationale.....	58
Annex B (informative)	Hydrophones and positioning	60
B.1	General.....	60
B.2	Electrical loading considerations	60
B.3	Hydrophone signal amplifier	60
B.4	Hydrophone cable length and amplifiers.....	60
B.5	Transducer positioning.....	61
B.6	Alignment of hydrophones.....	62
B.7	Water bath lining material.....	62
B.8	Recommendations for ultrasonic equipment being characterized.....	62
B.9	Types of hydrophones.....	63
B.9.1	Ceramic needle hydrophones	63
B.9.2	PVDF needle hydrophones	63
B.9.3	PVDF membrane hydrophones	63
B.9.4	Fibre-optic and optic hydrophones	64
B.9.5	Relative performance of different types.....	65
B.10	Typical specification data for hydrophones.....	65
Annex C (informative)	Acoustic pressure and intensity	66
Annex D (informative)	Voltage to pressure conversion	68
D.1	General.....	68
D.2	Hydrophone deconvolution procedure	69
D.3	Converting the data between double-sided and single-sided spectra	70
D.4	Use of hydrophone calibration data	72
D.4.1	Calibration data interpolation	72
D.4.2	Calibration data extrapolation	72
D.4.3	Regularization filtering	73
D.5	Implication of the hydrophone deconvolution process on measurement duration	74
D.6	Validation of deconvolution implementation.....	75
Annex E (informative)	Correction for spatial averaging.....	76

E.1	Linear and quasilinear fields	76
E.2	Linear fields, quasilinear fields, and broadband nonlinearly distorted waveforms	78
Annex F (informative) Acoustic output parameters for multi-mode medical ultrasonic fields in the absence of scan-frame synchronization		81
F.1	General.....	81
F.2	Current philosophy.....	81
F.3	Need for an alternative approach	82
F.4	Proposed approach.....	82
F.4.1	Alternative philosophy	82
F.4.2	Alternative parameters.....	83
F.5	Measurement methods.....	84
F.5.1	General	84
F.5.2	Peak pressures.....	84
F.5.3	Temporal-average intensity	84
F.5.4	Frequency	85
F.5.5	Power	85
F.6	Discussion	85
F.6.1	Relationship to existing standards	85
F.6.2	Advantages	86
F.6.3	Disadvantages.....	86
Annex G (informative) Propagation medium and degassing.....		87
Annex H (informative) Specific ultrasonic fields.....		88
H.1	Diagnostic fields	88
H.1.1	Useful relationships between acoustical parameters	88
H.1.2	Pulsed wave diagnostic equipment	89
H.1.3	Continuous wave diagnostic equipment	89
H.2	Therapy fields	90
H.2.1	Physiotherapy equipment.....	90
H.2.2	High intensity therapeutic ultrasonic equipment	90
H.2.3	Non-focused and weakly focused pressure pulses	90
H.3	Surgical fields	90
H.3.1	Lithotripters	90
H.3.2	Low frequency surgical applications.....	90
Annex I (informative) Assessment of uncertainty in the acoustic quantities obtained by hydrophone measurements.....		91
I.1	General.....	91
I.2	Overall (expanded) uncertainty	91
I.3	Common sources of uncertainty	91
Annex J (informative) Transducer and hydrophone positioning systems		93
Annex K (informative) Beamwidth midpoint method.....		94
Bibliography.....		95
Figure 1 – Schematic diagram of the different planes and lines in an ultrasonic field		12
Figure 2 – Several apertures and planes for a transducer of unknown geometry		26
Figure 3 – Parameters for describing a focusing transducer of known geometry.....		29
Figure 4 – Schematic diagram of the method of determining pulse duration		46
Figure D.1 – A flow diagram of the hydrophone deconvolution process		70

Figure D.2 – Example of waveform deconvolution	74
Figure J.1 – Schematic diagram of the ultrasonic transducer and hydrophone degrees of freedom	93
Table 1 – Acoustic parameters appropriate to various types of medical ultrasonic equipment.....	45
Table B.1 – Typical specification data for hydrophones, in this case given at 1 MHz [69].....	65
Table C.1 – Properties of distilled or de-ionized water as a function of temperature [71]	67
Table D.1 – Method of conversion from a double- to a single-sided spectrum	71
Table D.2 – Method of conversion from a single- to a double-sided spectrum	71
Table F.1 – Main basic parameters defined in this document or in IEC 61161	82
Table F.2 – List of parameters that are to be used or are to be deleted.....	83
Table K.1 – Decibel beamwidth levels for determining midpoints.....	94

**iTeh STANDARD
PREVIEW
(standards.iteh.ai)**

[SIST EN IEC 62127-1:2022](https://standards.iteh.ai/catalog/standards/sist/0d399001-f0ad-45a6-ac82-32650061c7d2/sist-en-iec-62127-1-2022)

<https://standards.iteh.ai/catalog/standards/sist/0d399001-f0ad-45a6-ac82-32650061c7d2/sist-en-iec-62127-1-2022>

INTERNATIONAL ELECTROTECHNICAL COMMISSION

ULTRASONICS – HYDROPHONES –

Part 1: Measurement and characterization of medical ultrasonic fields

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

IEC 62127-1 has been prepared by IEC technical committee 87: Ultrasonics. It is an International Standard.

This second edition cancels and replaces the first edition published in 2007 and Amendment 1:2013. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition.

- a) The upper frequency limit of 40 MHz has been removed.
- b) Hydrophone sensitivity definitions have been changed to recognize sensitivities as complex-valued quantities.
- c) Procedures and requirements for narrow-band approximation and broadband measurements have been modified; details on waveform deconvolution have been added.
- d) Procedures for spatial averaging correction have been amended.
- e) Annex D, Annex E and bibliography have been updated to support the changes of the normative parts.

The text of this International Standard is based on the following documents:

Draft	Report on voting
87/783/FDIS	87/788/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/standardsdev/publications.

A list of all parts of IEC 62127 series, published under the general title *Ultrasonics – Hydrophones*, can be found on the IEC website.

NOTE Words in **bold** in the text are terms defined in Clause 3.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

PREVIEW
(standards.iteh.ai)
SIST EN IEC 62127-1:2022
<https://standards.iteh.ai/catalog/standards/sist/0d399001-f0ad-45a6-ac82-32650061c7d2/sist-en-iec-62127-1-2022>

INTRODUCTION

The main purpose of this document is to define various acoustic parameters that can be used to specify and characterize ultrasonic fields propagating in liquids, and, in particular, water, using hydrophones. Measurement procedures are outlined that may be used to determine these parameters. Specific device related measurement standards, for example IEC 61689, IEC 61157, IEC 61847 or IEC 62359, can refer to this document for appropriate acoustic parameters. In IEC 62359, some additional measurement methods for attenuated parameters and indices are described addressing the specific needs of acoustic output characterization of ultrasonic diagnostic equipment in accordance with IEC 60601-2-37.

The philosophy behind this document is the specification of the acoustic field in terms of acoustic pressure parameters, acoustic pressure being the primary measurement quantity when hydrophones are used to characterize the field.

Intensity parameters are specified in this document, but these are regarded as derived quantities that are meaningful only under certain assumptions related to the ultrasonic field being measured.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN IEC 62127-1:2022](https://standards.iteh.ai/catalog/standards/sist/0d399001-f0ad-45a6-ac82-32650061c7d2/sist-en-iec-62127-1-2022)

<https://standards.iteh.ai/catalog/standards/sist/0d399001-f0ad-45a6-ac82-32650061c7d2/sist-en-iec-62127-1-2022>

ULTRASONICS – HYDROPHONES –

Part 1: Measurement and characterization of medical ultrasonic fields

1 Scope

This part of IEC 62127 specifies methods of use of calibrated **hydrophones** for the measurement in liquids of acoustic fields generated by ultrasonic medical equipment including **bandwidth** criteria and calibration frequency range requirements in dependence on the spectral content of the fields to be characterized.

This document:

- defines a group of acoustic parameters that can be measured on a physically sound basis;
- defines a second group of parameters that can be derived under certain assumptions from these measurements, and called derived intensity parameters;
- defines a measurement procedure that can be used for the determination of acoustic pressure parameters;
- defines the conditions under which the measurements of acoustic parameters can be made using calibrated **hydrophones**;
- defines procedures for correcting for limitations caused by the use of **hydrophones** with finite **bandwidth** and finite active element size, and for estimating the corresponding **uncertainties**.

NOTE 1 Throughout this document, SI units are used. In the specification of certain parameters, such as **beam areas** and intensities, it can be convenient to use decimal multiples or submultiples. For example, **beam area** is likely to be specified in cm^2 and intensities in W/cm^2 or mW/cm^2 .

NOTE 2 The **hydrophone** as defined can be of a piezoelectric or an optic type.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60565-1, *Underwater acoustics – Hydrophones – Calibration of hydrophones – Part 1: Procedures for free-field calibration of hydrophones*

IEC 61689, *Ultrasonics – Physiotherapy systems – Field specifications and methods of measurement in the frequency range 0,5 MHz to 5 MHz*

IEC 62127-2, *Ultrasonics – Hydrophones – Part 2: Calibration for ultrasonic fields up to 40 MHz*

IEC 62127-3, *Ultrasonics – Hydrophones – Part 3: Properties of hydrophones for ultrasonic fields up to 40 MHz*

IEC 63009, *Ultrasonics – Physiotherapy systems – Field specifications and methods of measurement in the frequency range 20 kHz to 500 kHz*

ISO 16269-6, *Statistical interpretation of data – Part 6: Determination of statistical tolerance intervals*