



Designation: E2653–08 ~~Designation: E 2653 – 09~~

Standard Practice for Conducting an Interlaboratory Study to Determine the Precision of Estimates for a Fire Test Method with Fewer Than Six Participating Laboratories¹

This standard is issued under the fixed designation E 2653; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This practice describes the techniques for planning, conducting, analyzing, and treating results of an interlaboratory study (ILS) for estimating the precision of a fire test method when fewer than six laboratories are available to meet the recommended minimum requirements of Practice E 691. Data obtained from an interlaboratory study are useful in identifying variables that require modifications for improving test method performance and precision.

1.2 Precision estimates developed using this practice will not be statistically equivalent to precision estimates produced by Practice E 691 because a small number of laboratories are used. The smaller number of participating laboratories will seriously reduce the value of precision estimates reported by this practice. However, under circumstances where only a limited number of laboratories are available to participate in an ILS, precision estimates developed by this practice will provide the user with useful information concerning precision for a test method.

1.3 A minimum of three qualified laboratories is required for conducting an ILS using this practice. If six or more laboratories are available to participate in an ILS for a given fire test method, Practice E 691 shall be used for conducting the ILS.

1.4 Since the primary purpose of this practice is the development of the information needed for a precision statement, the experimental design in this practice will not be optimum for evaluating all materials, test methods, or as a tool for individual laboratory analysis.

1.5 Because of the reduced number of participating laboratories a Laboratory Monitor shall be used in the ILS. See Standard Guide E 2335.

1.6 *Field of Application*—This practice is concerned with test methods that yield numerical values or a series of numerical values for different fire-test response properties. The numerical values mentioned above are typically the result of calculations from a set of measurements.

1.7 This practice includes design information suitable for use with the development of interlaboratory studies for test methods that have categorization (go-no-go) allocation test results. However, it does not provide a recommended statistical practice for evaluating the go-no-go data.

1.8

1.8 This fire standard cannot be used to provide quantitative measures.

1.9 This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.

2. Referenced Documents

2.1 *ASTM Standards*:²

E 176 [Terminology of Fire Standards](#)

E 177 [Practice for Use of the Terms Precision and Bias in ASTM Test Methods](#)

E 178 [Practice for Dealing With Outlying Observations](#)

E 456 [Terminology Relating to Quality and Statistics](#)

E 691 [Practice for Conducting an Interlaboratory Study to Determine the Precision of a Test Method](#)

E 1169 [Practice for Conducting Ruggedness Tests](#)

E 2335 [Guide for Laboratory Monitors](#)

¹ This practice is under the jurisdiction of ASTM Committee E05 on Fire Standards and is the direct responsibility of Subcommittee E05.15 on Furnishings and Contents. Current edition approved Oct. 1, 2008. Published November 2008.

Current edition approved April 1, 2009. Published May 2009. Originally approved in 2008. Last previous edition approved in 2008 as E 2653-08.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

3. Terminology

3.1 *Definitions*—For formal definitions of statistical terms, see Terminology E 456. For formal definitions of fire terms, see Terminology E 176.

3.2 *Definitions of Terms Specific to This Standard:*

3.2.1 *test method and protocol*—in this practice, the term “test method” is used both for the actual measurement process and for the written description of the process, while the term “protocol” is used for the directions given to the laboratories for conducting the ILS.

3.2.2 *repeatability and reproducibility*—these terms deal with the variability of test results obtained under specified laboratory conditions. Repeatability concerns the variability between independent test results obtained within a single laboratory in the shortest practical period of time by a single operator with a specific set of test apparatus using test specimens (or test units) taken at random from a single quantity of homogeneous material obtained or prepared for the ILS. Reproducibility deals with the variability between single test results obtained in different laboratories, each of which has applied the test method to test specimens (or test units) taken at random from a single quantity of homogeneous material obtained or prepared for the ILS.

3.3 For further discussion of the terms discussed above, see Practice E 177 and the formal definitions in Terminology E 456.

4. Summary of Practice

4.1 The procedure presented in this practice consists of three basic steps: planning the interlaboratory study, guiding the testing phase of the study, and analyzing the test result data. The analysis evaluates the consistency of the data through the use of numerical estimates of precision of the test method pertaining to both within-laboratory repeatability and between-laboratory reproducibility.

4.2 Planning of the interlaboratory study will include a review of the test procedure to be used in the interlaboratory study. This review will identify portions of the test method that appear to contribute to a loss in precision. Special interlaboratory instructions or modifications to the test method wording are made as needed to clarify these sections and often result in a modification to the test method following the interlaboratory study.

4.3 A manager for the interlaboratory study and an interlaboratory test monitor shall be selected. The same person is allowed to conduct both functions.

4.4 Parties conducting an interlaboratory precision study of a test method will acquire participation agreements with as many laboratories as possible that are willing to take part in the interlaboratory study and have the capability to run the test method of interest. A minimum of three laboratories shall participate in the precision study. Precision results will increase in quality with a larger number of participating laboratories.

4.5 The types of materials and number of test specimens shall be selected for the interlaboratory study. No less than three test specimens shall be selected for the interlaboratory study, and they shall be selected to reflect the range of performance of test specimens normally evaluated by the test method. A minimum of three replicates shall be tested for each test material selected. If a standard reference material is available for the test method, the material shall be included as a specimen in the interlaboratory study. If a standard reference material is not available, a test specimen that consistently produces low variability test results shall be selected as a reference material for the interlaboratory study.

5. Significance and Use

5.1 ASTM regulations require precision statements in all test methods in terms of repeatability and reproducibility. This practice is used when the number of participating laboratories or materials being tested, or both, in a precision study is less than the number specified by Practice E 691. When possible, it is strongly recommended that a full E 691 standard protocol be followed to determine test method precision. Precision results produced by the procedures presented in this standard will not have the same degree of accuracy as results generated by a full E 691 protocol. This procedure will allow for the development of useful precision results when a full compliment of laboratories is not available for interlaboratory testing.

5.2 This practice is based on recommendations for interlaboratory studies and data analysis presented in Practice E 691. This practice does not concern itself with the development of test methods but with a standard means for gathering information and treating the data needed for developing a precision statement for a fire test method when a complete E 691 interlaboratory study and data analysis are not possible.

PLANNING THE ILS

6. Planning

6.1 *Task Group*—Either the task group that developed the test method or a special task group appointed for the purpose must have overall responsibility for the ILS, including funding where appropriate, staffing, the design of the ILS, and decision-making with regard to questionable data. The task group shall decide on the number of laboratories, materials, and test results for each material. In addition, it shall specify any special calibration procedures and the repeatability conditions to be specified in the protocol.

6.2 *ILS Coordinator*—The task group must appoint one individual to act as overall coordinator for conducting the ILS. The coordinator will supervise the distribution of materials and protocols to the laboratories and receive the test result reports from the laboratories. Scanning the reports for gross errors and checking with the laboratories, when such errors are found, will also be the responsibility of the coordinator. The coordinator will consult as needed with a statistician in questionable cases.

6.3 *Laboratory Monitor*—The task group must appoint one individual to act as a laboratory monitor for the ILS. The laboratory monitor will develop an ILS checklist specific to the test method, inspect the test laboratories for equipment conformity and operator training, verify compatibility of the data acquisition system, and based on the Checklist and inspection results report to the sponsoring ASTM Subcommittee. Complete details for the function of a laboratory monitor are located in Guide E 2335.

6.4 *Statistician*—The task group shall obtain the assistance of a person skilled in the use of statistical procedures, the test method being studied, and with the materials being tested in order to ensure that the requirements in this practice are met in an efficient and effective manner. This person will conduct the data analysis using procedures given in this standard and will assist the task group in interpreting results from the data analysis.

7. Basic Design

7.1 Keep the ILS design simple in order to obtain estimates of within-and between-laboratory variability that are free of secondary effects. The basic design is represented by a two-way classification table in which the rows represent the laboratories, the columns represent the materials, and the cell (the intersection of a row and column) contains the test results made by a particular laboratory on a particular material (see Table 1).

7.2 An ILS using this practice shall include enough laboratories to represent a reasonable cross-section of the population of qualified laboratories. A minimum of three laboratories is necessary for carrying out an ILS using this practice.

8. Test Method

8.1 Of prime importance is the existence of a valid, well-written test method that has been developed in one or more competent laboratories, and had been subjected to a ruggedness test prior to the ILS.

8.2 The ruggedness test is a screening procedure for investigating the effects of variations in environmental and other conditions in order to determine how control of such test conditions shall be specified in the written description of the method. Details for ruggedness testing are found in Guide E 1169.

8.3 A written version of the test method must be developed for the ILS (but not necessarily published as a standard method). This draft shall describe the test apparatus and procedure in terms that are easily understood and followed in any properly equipped laboratory by competent personnel with knowledge of the materials and the property to be tested. The method shall contain safety and calibration procedures, details on control related limits that potentially cause test result variability, and specify how test results are to be reported.

(https://standards.iteh.ai)
ILS TESTING
 Document Preview

9. Pilot Run

9.1 Prior to beginning testing for the formal ILS a preliminary laboratory evaluation study shall be carried out using a well characterized test material of known performance. This preliminary study is managed by the ILS Coordinator and Laboratory Monitor and is used to determine if each of the participating laboratories are capable of conducting tests as specified by the written ILS test method. These preliminary tests conducted in the participating laboratories are typically observed by the Laboratory Monitor as a part of the laboratory qualification process.

9.2 The pilot run results give the task group an indication of how well each laboratory will perform in terms of promptness and following the protocol. Laboratories with poor performance are encouraged and helped to take corrective action.

9.3 All steps of the procedures described in this practice shall be followed in detail to ensure that these directions are understood, to disclose any weakness in the protocol or test method.

10. Full Scale Run

10.1 *Materials Preparation and Distribution* :

TABLE 1 Example, ILS Test Result Data

Laboratory	A	B	C	D	E
1	35.3	31.2	38.9	34.0	27.2
	34.0	31.0	35.0	35.5	31.1
	35.5	35.1	50.8	63.1	27.3
2	10.7	12.9	20.6	19.9	15.0
	12.7	15.0	8.0	16.2	8.2
	13.3	12.2	16.2	8.1	12.3
3	36.0	28.0	32.1	32.1	25.1
	36.0	32.1	36.0	32.0	25.0
	29.0	28.0	32.2	32.0	21.2
4	40.9	36.8	32.8	36.7	24.5
	36.7	32.7	28.6	32.7	24.4
	28.6	32.7	32.6	32.7	28.5
5	41.6	37.6	33.2	41.6	29.0
	41.7	25.1	29.2	37.5	29.1
	46.0	29.3	29.0	37.3	29.2