

---

**Kakovost vode - Radij Ra-226 - 1. del: Preskusna metoda s štetjem s tekočinskim scintilatorjem (ISO 13165-1:2013)**

Water quality - Radium-226 - Part 1: Test method using liquid scintillation counting (ISO 13165-1:2013)

Wasserbeschaffenheit - Radium-226 - Teil-1: Verfahren mit dem Flüssigkeitsszintillationszähler (ISO 13165-1:2013)

Qualité de l'eau - Radium 226 - Partie 1: Méthode d'essai par comptage des scintillations en milieu liquide (ISO 13165-1:2013)

**Ta slovenski standard je istoveten z: prEN ISO 13165-1**

SIST EN ISO 13165-1:2020

<http://standards.sist.si/canalog/standard/sist-osist-pr-en-iso-13165-1-2019-06092019>

**ICS:**

13.060.60	Preiskava fizikalnih lastnosti vode	Examination of physical properties of water
17.240	Merjenje sevanja	Radiation measurements

**oSIST prEN ISO 13165-1:2019**

**en,fr,de**



INTERNATIONAL  
STANDARD

ISO  
13165-1

First edition  
2013-04-15

---

---

**Water quality — Radium-226 —**

**Part 1:  
Test method using liquid scintillation  
counting**

*Qualité de l'eau — Radium 226 —*

*Partie 1: Méthode d'essai par comptage des scintillations en milieu  
liquide*

iTeh Standards

(<https://standards.iteh.ai>)

Document Preview

[SIST EN ISO 13165-1:2020](https://standards.iteh.ai/catalog/standards/sist/6d0cf96d-1bbb-40b9-9e27-44ace03de521/sist-en-iso-13165-1-2020)

<https://standards.iteh.ai/catalog/standards/sist/6d0cf96d-1bbb-40b9-9e27-44ace03de521/sist-en-iso-13165-1-2020>



Reference number  
ISO 13165-1:2013(E)

© ISO 2013

ISO 13165-1:2013(E)

iTeh Standards  
(<https://standards.iteh.ai>)  
Document Preview

[SIST EN ISO 13165-1:2020](https://standards.iteh.ai/catalog/standards/sist/6d0cf96d-1bbb-40b9-9e27-44acc03de521/sist-en-iso-13165-1-2020)

<https://standards.iteh.ai/catalog/standards/sist/6d0cf96d-1bbb-40b9-9e27-44acc03de521/sist-en-iso-13165-1-2020>



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2013

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

Page

<b>Foreword</b> .....	<b>iv</b>
<b>Introduction</b> .....	<b>v</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Symbols, definitions and units</b> .....	<b>1</b>
<b>4 Principle</b> .....	<b>2</b>
<b>5 Reagents and equipment</b> .....	<b>2</b>
5.1 Reagents.....	2
5.2 Equipment.....	3
<b>6 Sampling</b> .....	<b>3</b>
<b>7 Instrument set-up and calibration</b> .....	<b>4</b>
7.1 Preparation of calibration sources.....	4
7.2 Optimization of counting conditions.....	4
7.3 Detection efficiency.....	4
7.4 Blank sample preparation and measurement.....	5
<b>8 Procedure</b> .....	<b>5</b>
8.1 Direct counting.....	5
8.2 Thermal preconcentration.....	5
8.3 Sample preparation.....	6
8.4 Sample measurement.....	6
<b>9 Quality control</b> .....	<b>6</b>
<b>10 Expression of results</b> .....	<b>6</b>
10.1 Calculation of massic activity.....	6
10.2 Standard uncertainty.....	6
10.3 Decision threshold.....	7
10.4 Detection limit.....	7
10.5 Confidence limits.....	8
10.6 Calculations using the activity concentration.....	8
<b>11 Interference control</b> .....	<b>8</b>
<b>12 Test report</b> .....	<b>9</b>
<b>Annex A (informative) Set-up parameters and validation data</b> .....	<b>10</b>
<b>Bibliography</b> .....	<b>14</b>

## ISO 13165-1:2013(E)

### Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 13165-1 was prepared by Technical Committee ISO/TC 147, *Water quality*, Subcommittee SC 3, *Radioactivity measurements*.

ISO 13165 consists of the following parts, under the general title *Water quality — Radium-226*:

- *Part 1: Test method using liquid scintillation counting*
- *Part 2: Test method using emanometry*

The following part is under preparation:

- *Part 3: Test method using coprecipitation and gamma-spectrometry*

[SIST EN ISO 13165-1:2020](https://standards.iteh.ai/catalog/standards/sist/6d0cf96d-1bbb-40b9-9e27-44ace03de521/sist-en-iso-13165-1-2020)

<https://standards.iteh.ai/catalog/standards/sist/6d0cf96d-1bbb-40b9-9e27-44ace03de521/sist-en-iso-13165-1-2020>

## Introduction

Radioactivity from several naturally occurring and human-made sources is present throughout the environment. Thus, water bodies (surface waters, groundwaters, sea waters) can contain radionuclides of natural and artificial origin (i.e. human-made).

- a) Natural radionuclides, including potassium-40, and those of the thorium and uranium decay series, in particular radium-226, radium-228, uranium-234, uranium-238, lead-210, can be found in water for natural reasons (e.g. desorption from the soil and wash-off by rain water) or release from technological processes involving naturally occurring radioactive materials (e.g. the mining and processing of mineral sands or phosphate fertilizer production and use).
- b) Human-made radionuclides such as transuranium elements (americium, plutonium, neptunium, curium), tritium, carbon-14, strontium-90 and gamma-emitting radionuclides can also be found in natural waters as they can be authorized to be routinely released into the environment in small quantities in the effluent discharged from nuclear fuel cycle facilities and following their use in unsealed form in medicine or industry. They are also found in water due to fallout from past explosions in the atmosphere of nuclear devices and the accidents at Chernobyl and Fukushima.

Drinking water can thus contain radionuclides at activity concentrations which present a risk to human health. In order to assess the quality of drinking-water (including mineral waters and spring waters) with respect to its radionuclide content and to provide guidance on reducing health risks by taking measures to decrease radionuclide activity concentrations, water resources (groundwater, river, lake, sea, etc.) and drinking water are monitored for their radioactivity content as recommended by the World Health Organization (WHO).

An International Standard on a test method of radium-226 activity concentrations in water samples is justified for test laboratories carrying out these measurements, which are sometimes required by national authorities, as laboratories may have to obtain a specific accreditation for radionuclide measurement in drinking water samples.

Radium-226 activity concentration can vary widely according to local geological and climatic characteristics and ranges from 0,001 Bq l<sup>-1</sup> in surface waters up to 50 Bq l<sup>-1</sup> in natural groundwaters; the guidance level for radium-226 in drinking water as recommended by WHO is 1 Bq l<sup>-1</sup> (Reference [Z]).

**NOTE** The guidance level is the activity concentration with an intake of 2 l day<sup>-1</sup> of drinking water for 1 year that results in an effective dose of 0,1 mSv year<sup>-1</sup> for members of the public, an effective dose that represents a very low level of risk that is not expected to give rise to any detectable adverse health effect.

This International Standard is one of a series on determination of the activity concentration of radionuclides in water samples.





# Water quality — Radium-226 —

## Part 1:

## Test method using liquid scintillation counting

**WARNING** — Persons using this part of ISO 13165 should be familiar with normal laboratory practice. This part of ISO 13165 does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to ensure compliance with any national regulatory conditions.

**IMPORTANT** — It is absolutely essential that tests conducted in accordance with this part of ISO 13165 be carried out by suitably qualified staff.

### 1 Scope

This part of ISO 13165 specifies the determination of radium-226 ( $^{226}\text{Ra}$ ) activity concentration in non-saline water samples by extraction of its daughter radon-222 ( $^{222}\text{Rn}$ ) and its measurement using liquid scintillation counting.

Radium-226 activity concentrations which can be measured by this test method utilizing currently available liquid scintillation counters goes down to 50 mBq l<sup>-1</sup>. This method is not applicable to the measurement of other radium isotopes.

### 2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

ISO 5667-1, *Water quality — Sampling — Part 1: Guidance on the design of sampling programmes and sampling techniques*

ISO 5667-3, *Water quality — Sampling — Part 3: Preservation and handling of water samples*

ISO/IEC 17025, *General requirements for the competence of testing and calibration laboratories*

ISO 80000-10, *Quantities and units — Part 10: Atomic and nuclear physics*

ISO/IEC Guide 98-3:2008, *Uncertainty of measurement — Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

### 3 Symbols, definitions and units

For the purposes of this document, the definitions, symbols and abbreviations given in ISO 80000-10, ISO/IEC Guide 98-3, and the following apply.

- |       |  |
|-------|--|
| $a$   | massic activity of the sample at the measuring time, in becquerels per gram                              |
| $a_S$ | massic activity of the $^{226}\text{Ra}$ standard solution at the measuring time, in becquerels per gram |
| $a^*$ | decision threshold for the massic alpha-activity, in becquerels per gram                                 |

**ISO 13165-1:2013(E)**

$a^{\#}$	detection limit for the massic alpha-activity, in becquerels per gram
$a^{\triangleleft}, a^{\triangleright}$	lower and upper limits of the confidence interval, in becquerels per gram
$c_A$	activity concentration, in becquerels per litre
$m$	mass of the test sample, in grams
$m_1$	mass of initial sample subject to heating or possibly concentration, in grams
$m_2$	mass of heated or concentrated sample, in grams
$m_3$	mass of heated or concentrated sample transferred in the vial, in grams
$m_S$	mass of $^{226}\text{Ra}$ standard solution used for the preparation of the calibration sample, in grams
$r_0$	blank sample count rate in the alpha-window, in reciprocal seconds
$r_g$	sample gross count rate in the alpha-window, in reciprocal seconds
$r_S$	count rate of the calibration sample in the alpha-window, in reciprocal seconds
$t_0$	blank sample counting time, in seconds
$t_g$	sample counting time, in seconds
$t_S$	calibration sample counting time, in seconds
$u(a)$	standard uncertainty associated with the measurement result; in becquerels per gram
$U$	expanded uncertainty, calculated using $U = ku(a)$ , with $k = 1, 2, \dots$ in becquerels per gram
$w$	factor equal to $1/\varepsilon m$
$\varepsilon$	alpha-efficiency
$\rho$	density, in grams per litre

<https://standards.iteh.ai/catalog/standards/sist/6d0cf96d-1bbb-40b9-9e27-44ace03de521/sist-en-iso-13165-1-2020>

**4 Principle**

$^{226}\text{Ra}$  massic activity is determined by liquid scintillation counting of daughter  $^{222}\text{Rn}$  at isotopic equilibrium (99,56 %) reached 30 d after the preparation of the sample. The  $^{222}\text{Rn}$  is extracted from aqueous solution by means of a scintillation cocktail immiscible with water inside the scintillation vial (References [1]–[4]).

The aqueous sample is acidified, heated and, if possible, concentrated by slow evaporation in order to desorb  $^{222}\text{Rn}$  and to achieve a better detection limit. The concentrated aqueous sample is transferred into a radon-tight scintillation vial and a water-immiscible scintillation cocktail is added.

After 30 d, the sample is measured by liquid scintillation counting (LSC) applying alpha and beta discrimination: only alpha-emission of  $^{222}\text{Rn}$  and that of its short lived progeny ( $^{218}\text{Po}$ ,  $^{214}\text{Po}$ ) are considered, as this counting condition ensures a better detection limit.

**5 Reagents and equipment****5.1 Reagents**

All reagents shall be of recognized analytical grade and, except for 5.1.3 and 5.1.4, shall not contain any detectable alpha- and beta-activity.