

---

---

**Biotechnology — Cell counting —**  
**Part 1:**  
**General guidance on cell counting**  
**methods**

*Biotechnologie — Dénombrement des cellules —*

*Partie 1: Lignes directrices générales relatives aux méthodes de  
dénombrement des cellules*

(<https://standards.iteh.ai>)  
Document Preview

ISO 20391-1:2018

<https://standards.iteh.ai/catalog/standards/iso/49a01d17-1cfd-4c65-868e-767fb22aefcd/iso-20391-1-2018>



iTeh Standards  
(<https://standards.iteh.ai>)  
Document Preview

ISO 20391-1:2018

<https://standards.iteh.ai/catalog/standards/iso/49a01d17-1cfd-4c65-868e-767fb22aefcd/iso-20391-1-2018>



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2018

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva, Switzerland  
Tel. +41 22 749 01 11  
Fax +41 22 749 09 47  
[copyright@iso.org](mailto:copyright@iso.org)  
[www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

Page

<b>Foreword</b>	<b>iv</b>
<b>Introduction</b>	<b>v</b>
<b>1 Scope</b>	<b>1</b>
<b>2 Normative references</b>	<b>1</b>
<b>3 Terms and definitions</b>	<b>1</b>
<b>4 General concepts of cell counting</b>	<b>4</b>
4.1 General	4
4.2 Total cell counting	5
4.3 Differential cell counting	5
4.4 Direct cell counting	5
4.5 Indirect cell counting	5
<b>5 Considerations for cell counting measurements</b>	<b>5</b>
5.1 Selection of a cell counting method	5
5.2 Considerations for selecting a cell counting method	6
5.3 Sampling of cells for counting	6
5.4 Preparation of cell samples for counting	7
5.4.1 Environmental factors	7
5.4.2 Procedures	7
5.4.3 Quality and stability of reagents	7
5.5 Performing a measurement	8
<b>6 Qualification, validation, and verification</b>	<b>8</b>
6.1 Instrument qualification	8
6.2 Method validation and verification	8
6.3 Reference materials	9
6.3.1 Certified reference materials	9
6.3.2 In-house reference materials	9
6.3.3 Uses of reference materials	9
<b>7 Data processing, analysis, and reporting</b>	<b>10</b>
7.1 Data processing and analysis	10
7.1.1 General	10
7.1.2 Image processing and analysis	10
7.1.3 Gating	10
7.1.4 Coincidence correction	10
7.2 Reporting	10
<b>Annex A (informative) Description of common cell counting methods</b>	<b>11</b>
<b>Annex B (informative) Common cell counting methods for various measurement purposes</b>	<b>14</b>
<b>Bibliography</b>	<b>15</b>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2. [www.iso.org/directives](http://www.iso.org/directives)

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received. [www.iso.org/patents](http://www.iso.org/patents)

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](http://standards.iteh.ai/Foreword-Supplementary-information)

This document was prepared by ISO/TC 276, *Biotechnology*.

A list of all the parts of ISO 20391 can be found on the ISO website.

ISO 20391-1:2018

<https://standards.iteh.ai/catalog/standards/iso/49a01d17-1cfd-4c65-868e-767fb22aefcd/iso-20391-1-2018>

## Introduction

Cell counting (or cell enumeration) is a fundamental measurement that broadly impacts many aspects of biotechnology, from biomanufacturing to advanced therapy. The cell count (or discrete number of cells) is often expressed as cell concentration (i.e. cell count per volume) when in suspension and area density of cells (i.e. cell count per unit area) when adhered to a surface. Cell count is critical in evaluating potency and efficacy for cell-based therapy. The cell concentration within a bioreactor can serve as a quality assurance metric in cell-based manufacturing processes. Many cell-based bioassays need to be normalized to the respective cell count to allow data inter-comparability. This document (which is Part 1 of a multi-part standard on cell counting) defines terms and provides general guidance for the cell counting measurement process, including method selection, sample preparation, measurement, qualification and validation, and data analysis and reporting.

**iTeh Standards**  
**(<https://standards.iteh.ai>)**  
**Document Preview**

[ISO 20391-1:2018](https://standards.iteh.ai/catalog/standards/iso/49a01d17-1cfd-4c65-868e-767fb22aefcd/iso-20391-1-2018)

<https://standards.iteh.ai/catalog/standards/iso/49a01d17-1cfd-4c65-868e-767fb22aefcd/iso-20391-1-2018>



# Biotechnology — Cell counting —

## Part 1: General guidance on cell counting methods

### 1 Scope

This document defines terms related to cell counting for biotechnology. It describes counting of cells in suspension (generally cell concentration) and cells adhered to a substrate (generally area density of cells). It provides key considerations for general counting methods (including total and differential counting, and direct and indirect counting) as well as for method selection, measurement process, and data analysis and reporting.

This document is applicable to the counting of all cell types – mammalian and non-mammalian (e.g. bacteria, yeast) cells.

This document is not intended for counting of cells while in a tissue section or a biomaterial matrix.

Several sector/application-specific international and national standards for cell counting currently exist. When applicable, the user can consult existing standards when operating within their scope (specific measurement techniques and/or applications).

### 2 Normative references

There are no normative references in this document.

### 3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <https://www.iso.org/obp>

#### 3.1

##### **accuracy**

closeness of agreement between a measured quantity value and a true quantity value of a measurand

Note 1 to entry: The concept of “measurement accuracy” is not a quantity and is not given a numerical quantity value. A measurement is said to be more accurate when it offers a smaller measurement error.

Note 2 to entry: “Measurement accuracy” is sometimes understood as closeness of agreement between measured quantity values that are being attributed to the measurand.

[SOURCE: ISO/IEC Guide 99:2007, 2.13, modified]

#### 3.2

##### **agglomerate**

<cells> two or more cells clustered weakly together and detected as a larger object

Note 1 to entry: Agglomerates of cells can be separated into nominally single cells without causing significant damage to the cell.

### 3.3

#### **aggregate**

<cells> two or more cells clustered together (tightly or loosely) and detected as a larger object

Note 1 to entry: Aggregates of cells are generally more difficult to be separated into single cells.

### 3.4

#### **area density**

<cells> cell count of adherent cells on a surface, typically expressed as number of cells per unit area

### 3.5

#### **attribute**

physical, chemical, biological or microbiological property or characteristic

### 3.6

#### **cell concentration**

cell count per volume

Note 1 to entry: Typically used for cells in suspension.

### 3.7

#### **cell count**

discrete number of cells

Note 1 to entry: Cell count is typically expressed as *cell concentration* (3.6) or *area density* (3.4).

### 3.8

#### **cell counting**

measurement process to determine the cell count

### 3.9

#### **cell suspension**

cells dispersed in a liquid matrix

### 3.10

**debris** <https://standards.iteh.ai/catalog/standards/iso/49a01d17-1cfd-4c65-868e-767fb22aefcd/iso-20391-1-2018>  
<in cell suspensions> fragments of cells and/or particles of biological or non-biological origin

### 3.11

#### **differential cell count**

number of a subset of cells, which have been distinguished from other cell subpopulations by at least one distinct cell attribute identified in the measurement

Note 1 to entry: The concentrations derived from a differential cell count can be expressed in absolute concentration or as a relative measure (i.e. percentage) with respect to the total cell number or another predefined population.

### 3.12

#### **direct cell counting**

counting method in which one signal is (or several signals are) detected for each single event

Note 1 to entry: Each single event should represent a single cell in an idealized measurement.

### 3.13

#### **indirect cell counting**

counting method during which a signal (or a set of signals) is measured from a population of cells and that signal is then related to cell number based on a measurement-specific mathematical model (e.g. calibration curve)



**3.14****limit of quantitation****LoQ**

lowest amount of analyte in a sample that can be quantitatively determined with a suitable precision and accuracy using a specific analytical method

Note 1 to entry: The limit of quantitation describes quantitative assay for low levels of cells in sample matrices.

**3.15****linearity**

ability to elicit test results that are directly, or indirectly by means of well-defined mathematical transformations, proportional to cell count within a given range

**3.16****measurand**

quantity intended to be measured

[SOURCE: ISO/IEC Guide 99:2007, 2.3, modified]

**3.17****precision****measurement precision**

closeness of agreement between indications or measured quantity values obtained by replicate measurements on the same or similar objects under specified conditions

[SOURCE: ISO/IEC Guide 99:2007, 2.15, modified]

**3.18****proportionality**

characteristic exhibited by a collection of measurements in which the ratio of the expected value of the measurement to the value of the input parameter at which the measurements were taken remains constant as the value of the input parameter changes (while all other inputs and measurement conditions are held constant)

Note 1 to entry: When a set of measurements exhibits proportionality over a range of a given input, the expected value of the measurements can be expressed as the input parameter multiplied by a fixed constant, with no bias term.

**3.19****reagent**

substance used in chemical/biochemical analysis or other reactions

**3.20****reference material**

material sufficiently homogeneous and stable with reference to specified properties, which has been established to be fit for its intended use in measurement or in examination of nominal properties

Note 1 to entry: Reference materials with or without assigned quantity values can be used for measurement precision control whereas only reference materials with assigned quantity values can be used for calibration or measurement trueness control.

[SOURCE: ISO/IEC Guide 99:2007, 5.13, modified]

**3.21****reference method**

thoroughly investigated measurement procedure shown to yield values having an uncertainty in measurement commensurate with its intended use, especially in assessing the trueness of other measurement procedures for the same quantity and in characterizing reference material

[SOURCE: ISO 17511:2003, 3.29, modified]

**3.22**

**repeatability**

<results of measurement> measurement precision under defined conditions of measurement

[SOURCE: ISO/IEC Guide 99:2007, 2.21, modified]

**3.23**

**ruggedness**

measure of a method's capacity to remain unaffected by small, but deliberate, variations in method parameters and provides an indication of its reliability during normal usage

[SOURCE: ICH Harmonised Tripartite Guideline, 1994]

**3.24**

**selectivity**

property of a measuring system, used with a specified measurement procedure, whereby it provides measured quantity values for one or more measurands such that the values of each measurand are independent of other measurands or other quantities in the phenomenon, body, or substance being investigated

[SOURCE: ISO/IEC Guide 99:2007, 4.13, modified]

**3.25**

**total cell count**

count of all cells, independent of the attribute(s) of the cell

**3.26**

**uncertainty**

<measurement> non-negative parameter characterizing the dispersion of values attributed to a measurand, based on the information used

[SOURCE: ISO/IEC Guide 99:2007, 2.26, modified]

**3.27**

**validation**

confirmation, through the provision of objective evidence, that the requirements for a specific intended use or application have been fulfilled

[SOURCE: ISO 9000:2015, 3.8.13, modified]

**3.28**

**verification**

confirmation, through the provision of objective evidence, that specified requirements have been fulfilled

[SOURCE: ISO 9000:2015, 3.8.12, modified]

**3.29**

**viable cells**

cells within a sample that have an attribute of being alive (e.g. metabolically active, capable of reproduction, possessed of intact cell membrane, or with the capacity to resume these functions) defined based on the intended use

## **4 General concepts of cell counting**

### **4.1 General**

Various cell counting methods (as described in [Annex A](#)) can be broadly categorized as total or differential cell counting, and direct or indirect cell counting.