

### SLOVENSKI STANDARD SIST EN ISO 28057:2021

01-april-2021

Nadomešča:

SIST EN ISO 28057:2018

Klinična dozimetrija - Dozimetrija s trdnimi termoluminiscenčnimi zaznavali pri fotonskih in elektronskih sevanjih v radioterapiji (ISO 28057:2019)

Clinical dosimetry - Dosimetry with solid thermoluminescence detectors for photon and electron radiations in radiotherapy (ISO 28057:2019)

Dosimetrie mit Festkörper Thermolumineszenzdetektoren für Photonen- und Elektronenstrahlung in der Strahlentherapie (ISO 28057:2019)

(standards.iteh.ai)

Dosimétrie clinique - Dosimétrie avec détecteurs thermoluminescents solides pour les rayonnements de photons et d'électrons en radiothérapie (ISO 28057:2019)

b91197586b72/sist-en-iso-28057-2021

Ta slovenski standard je istoveten z: EN ISO 28057:2021

ICS:

13.280 Varstvo pred sevanjem Radiation protection

SIST EN ISO 28057:2021 en,fr,de

# iTeh STANDARD PREVIEW (standards.iteh.ai)

### EUROPEAN STANDARD NORME EUROPÉENNE EUROPÄISCHE NORM

**EN ISO 28057** 

February 2021

ICS 13.280

Supersedes EN ISO 28057:2018

#### **English Version**

# Clinical dosimetry - Dosimetry with solid thermoluminescence detectors for photon and electron radiations in radiotherapy (ISO 28057:2019)

Dosimétrie clinique - Dosimétrie avec détecteurs thermoluminescents solides pour les rayonnements de photons et d'électrons en radiothérapie (ISO 28057:2019) Dosimetrie mit Festkörper -Thermolumineszenzdetektoren für Photonen- und Elektronenstrahlung in der Strahlentherapie (ISO 28057:2019)

This European Standard was approved by CEN on 18 January 2021.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria; Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Ireland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION COMITÉ EUROPÉEN DE NORMALISATION EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

#### EN ISO 28057:2021 (E)

Contents	Page
European foreword	3

# iTeh STANDARD PREVIEW (standards.iteh.ai)

#### **European foreword**

The text of ISO 28057:2019 has been prepared by Technical Committee ISO/TC 85 "Nuclear energy, nuclear technologies, and radiological protection" of the International Organization for Standardization (ISO) and has been taken over as EN ISO 28057:2021 by Technical Committee CEN/TC 430 "Nuclear energy, nuclear technologies, and radiological protection" the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by August 2021, and conflicting national standards shall be withdrawn at the latest by August 2021.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 28057:2018.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

### (staendorsement notice

The text of ISO 28057:2019 has been approved by CEN as EN ISO 28057:2021 without any modification. https://standards.iteh.a/catalog/standards/sist/0c9f1095-181a-4/88-842e-b91197586b72/sist-en-iso-28057-2021

# iTeh STANDARD PREVIEW (standards.iteh.ai)

# INTERNATIONAL STANDARD

ISO 28057

Second edition 2019-07

# Clinical dosimetry — Dosimetry with solid thermoluminescence detectors for photon and electron radiations in radiotherapy

Dosimétrie clinique — Dosimétrie avec détecteurs thermoluminescents solides pour les rayonnements de photons et The STd/électrons en radiothérapie VIEW

(standards.iteh.ai)



### iTeh STANDARD PREVIEW (standards.iteh.ai)

<u>SIST EN ISO 28057:2021</u> https://standards.iteh.ai/catalog/standards/sist/0c9f1095-181a-4788-842eb91197586b72/sist-en-iso-28057-2021



#### **COPYRIGHT PROTECTED DOCUMENT**

© ISO 2019

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office CP 401 • Ch. de Blandonnet 8 CH-1214 Vernier, Geneva Phone: +41 22 749 01 11 Fax: +41 22 749 09 47 Email: copyright@iso.org Website: www.iso.org

Published in Switzerland

Contents			Page		
For	eword		v		
Intr	oductio	n	vi		
1	Scon	e	1		
	_	-			
2		native references			
3	3 Terms and definitions				
4	Rule	s for the TLD measurement procedure	9		
	4.1	Principle of measurement			
	4.2	Measured quantity			
	4.3	Measurement cycle			
		4.3.1 General requirements			
		4.3.2 Sequence of measurement cycles			
		4.3.3 Common passing of the measurement cycles			
		4.3.4 Handling of TL detectors			
		4.3.5 Pre-irradiation annealing			
		4.3.6 Irradiation			
		4.3.7 Post-irradiation annealing			
	4.4	4.3.8 Reading			
	4.4	Measurement of the absorbed dose to water 4.4.1 Basic formula for the determination of the absorbed dose to water	12		
		4.4.2 Determination of the background value, $M_0$	12		
		4.4.3 Determination of the indicated value, $M_i$ .	12		
		4.4.4 Determination of the individual calibration coefficients, $N_i$			
		4.4.5 Determination of the correction factors, $k_{v}$			
	4.5	Uncertainty of measurement of the absorbed dose	22		
	4.6	Reusabilitytandards.iteh.ai/catalog/standards/sist/0e9f1095-181a-4788-842e-	2.3		
	4.7	Stability check b91197586b72/sist-on-iso-28057-2021	23		
	4.8	Staff			
_	Dogu	irements for the TLD system			
5	5.1	General information			
	3.1	5.1.1 Classification of the requirements			
		5.1.2 Requirements for operation characteristics			
	5.2	Completeness of the TLD system			
	5.2	5.2.1 Technical components			
		5.2.2 Hardware and software components			
		5.2.3 Operating instructions			
		5.2.4 Access to a calibration irradiation device			
	5.3	Requirements for TL detectors			
		5.3.1 Characteristics of TL materials			
		5.3.2 Tailoring of TL materials	26		
		5.3.3 Reusability of TL detectors	27		
		5.3.4 Individual variation	27		
	5.4	Requirements for TL-indicating instruments			
		5.4.1 General remarks			
		5.4.2 Mechanical setup			
		5.4.3 Warm-up time			
		5.4.4 Indication and indication range			
		5.4.5 Background value			
		5.4.6 Overflow indication and effects during evaluation of high doses			
		5.4.7 Test light source			
		5.4.8 Changes in the response			
		5.4.9 Mechanical construction			
		5.4.10 Light shielding	∠9		

		5.4.11	Climatic influences	29
		5.4.12	Electrical requirements	29
		5.4.13		30
		5.4.14	Data output and data backup	32
	5.5	Requirements for auxiliary instruments (pre-irradiation annealing device)		32
		5.5.1	Pre-irradiation annealing	32
		5.5.2	Construction	32
		5.5.3	Electrical requirements	32
		5.5.4	Operation safety	32
		5.5.5	Detection of function failure	33
		5.5.6	Indication of the operating state	33
	5.6	Requir	ements for the entire TLD system	33
		5.6.1	Minimum measuring ranges	33
		5.6.2	Minimum rated ranges of use	33
		5.6.3	Ranges of test parameters	34
	5.7	Requir	ements for the calibration irradiation device	35
	5.8	Requir	ements for the accompanying papers	35
	5.9		ance tests	
		5.9.1	General requirements	35
		5.9.2	Number of TL detectors used	36
		5.9.3	Type of TL detectors used	36
Ribline	oranh	17		37

### iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO 28057:2021

https://standards.iteh.ai/catalog/standards/sist/0c9f1095-181a-4788-842e-b91197586b72/sist-en-iso-28057-2021

#### **Foreword**

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see <a href="www.iso.org/directives">www.iso.org/directives</a>).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see <a href="https://www.iso.org/patents">www.iso.org/patents</a>).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: <a href="https://www.iso.org/iso/foreword.html">www.iso.org/iso/foreword.html</a> ards.iteh.ai

This document was prepared by Technical Committee ISO/TC 85, *Nuclear energy, nuclear technologies, and radiological protection*, Subcommittee SC 2, *Radiological protection*.

This second edition cancels and replaces the first edition (ISO 28057:2014), which has been technically revised.

- The clause on terms and definitions and the clause on rules for TLD measurement procedures, including quality assurance measurements at clinical accelerators, have been complemented and sharpened to ensure the safe application of TL dosimetry in the radiation therapy of cancer.
- Batch dependent changes of the  $k_{\rm Q}$  values have been correlated with the simultaneously occurring mass density variations of TL discs (see <u>4.4.5.5</u>).
- The response of TL materials to the neutrons, occurring within and around photon beams in megavoltage radiotherapy due to the photonuclear effect and eventually generating considerable components of the indicated values, has been dealt with in more detail (see 4.4.5.5).
- It is high-lighted that the  $k_{\rm E}$  values of clinical electron beams are energy independent (see 4.4.5.5).
- Recent experimental results concerning the contribution of "intrinsic effects" to the response of TL detectors have been considered (see <u>4.4.5.5</u>).
- The French title and the numbering of some subclauses of <u>5.4</u> have been corrected; <u>Table 9</u> has been equipped with a heading.

#### Introduction

The thermoluminescence dosimetry (TLD) with lithium fluoride (LiF) detectors has several advantages, in particular:

- small volumes of the detectors;
- applicability to continuous and pulsed radiation;
- fair water equivalency of the detector material;
- few correction factors needed for absorbed dose determinations.

The main disadvantage of thermoluminescence (TL) detectors is that, prior to each dosimetry application, they have to be regenerated by a pre-irradiation annealing procedure. Unfortunately, it is not possible to restore the former response of the detectors perfectly by this annealing. Provided, however, that all detectors of a production batch always undergo the same thermal treatment, one can at least determine the mean alteration of the response of these detectors, with sufficiently small fluctuations of the individually indicated values. From this mean alteration, a correction factor can be derived.

The essential aim of this document is to specify the procedures and to carry out corrections which allow one to achieve

- a) a repeatability of the indicated value within a fraction of a percent<sup>[17]</sup> and thus;
- b) a total uncertainty of measurement (including the calibration steps tracing to the primary standards) of a few percent, as in ionization chamber dosimetry [18][31][25][61][62].

The specifications in this document comprise special terms used in TLD, rules for the measurement technique, and requirements for the measurement system. The defined requirements and the testing techniques can, in whole or in part, serve as a basis for stability checks and acceptance tests. The TLD procedures described in this document can be used for photon radiation within the energy range from 20 keV to 50 MeV, including photon brachytherapy, and for electron radiation within the energy range from 4 MeV to 25 MeV, excluding beta radiation brachytherapy. In order to achieve the repeatability and total uncertainty stated above, this document is applicable in the dose range above 1 mGy. The upper limit of the minimum measuring range is in the order of magnitude of 10 Gy to 100 Gy. In clinical dosimetry, TL detectors are applied taking into account the requirements of high spatial resolution, i.e. in the study of the dose distributions with high gradients occurring in small stereotactic radiation fields and around brachytherapy sources. The other common application is the measurement of dose distributions in large absorbers, e.g. geometrical or tissue equivalent phantoms, either within the radiation field or in its periphery. A further usage is the quality assurance of clinical dosimetry by postal dose intercomparison [1][2][10][12][20][22][26][27][55].

The role of this document is not to anticipate national or international codes of practice in clinical dosimetry, neither for external beam therapy, brachytherapy, whole-body irradiation, mammography, nor dose measurements outside the treatment field or radiation protection of the staff. The authors of this document are well aware of the wide spectrum of the methods of clinical dosimetry, in which TL dosimetry is merely occupying a small sector. But within this framework, this document provides reliable concepts and rules for good practice for the application of TLD methods. The items covered include the terms and definitions, the rules for TLD measurement procedures, and the requirements on the TLD system; this document addresses medical physicists as well as instrument producers. Notably, the numerical examples given are valid for the TL detector materials and products stated in the publications referred to, and tests may be necessary to check whether they apply to TLD materials of other producers. The practical examples given, e.g. for the TL probe calibration conditions and for the numerical values of correction factor,  $k_{\rm Q}$ , accounting for the dependence of the detector response on radiation quality, Q, are not conceived to be pre-emptive in relation to more general standards of the methods of clinical dosimetry or dose intercomparisons. Rather, this document provides access to the reliable application of TLD methods based upon the published results of worldwide development.

The long-standing experience in the clinical usage of TLD, expressed in a set of valuable textbooks, protocols, and recommendations [6][13][25][28][29][42][43][61][62][54], has been accounted for.

### iTeh STANDARD PREVIEW (standards.iteh.ai)

# iTeh STANDARD PREVIEW (standards.iteh.ai)

### Clinical dosimetry — Dosimetry with solid thermoluminescence detectors for photon and electron radiations in radiotherapy

#### 1 Scope

This document describes rules for the procedures, applications, and systems of thermoluminescence dosimetry (TLD) for dose measurements according to the probe method. It is particularly applicable to solid "TL detectors", i.e. rods, chips, and microcubes, made from LiF:Mg,Ti or LiF:Mg,Cu,P in crystalline or polycrystalline form. It is not applicable to LiF powders because their use requires special procedures. The probe method encompasses the arrangement, particularly in a water phantom or in a tissue-equivalent phantom, of single TL detectors or of "TL probes", i.e. sets of TL detectors arranged in thin-walled polymethyl methacrylate (PMMA) casings.

The purpose of these rules is to guarantee the reliability and the accuracy indispensable in clinical dosimetry when applied on or in the patient or phantom. This document applies to dosimetry in teletherapy with both photon radiation from 20 keV to 50 MeV and electron radiation from 4 MeV to 25 MeV, as well as in brachytherapy with photon-emitting radionuclides. These applications are complementary to the use of ionization chambers.

iTeh STANDARD PREVIEW

### Normative references (standards.iteh.ai)

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60601-1, Electromedical equipment — Part 1: General instructions pertaining to safety

IEC 61000-4-2, Electromagnetic compatibility (EMV) — Part 4-2: Test and measurement procedure — Test of immunity against static electric discharges

IEC 61000-4-3, Electromagnetic compatibility (EMC) — Part 4-3: Testing and measurement techniques — Radiated, radio-frequency, electromagnetic field immunity test

IEC 61000-4-4, Electromagnetic compatibility (EMC) — Part 4-4: Testing and measurement techniques — *Electrical fast transient/burst immunity test* 

IEC 61000-4-5, Electromagnetic compatibility (EMC) — Part 4-5: Testing and measurement techniques — Surge immunity test

IEC 61000-4-6, Electromagnetic compatibility (EMC) — Part 4-6: Testing and measurement techniques — Immunity to conducted disturbances, induced by radio-frequency fields

IEC 61000-4-8, Electromagnetic compatibility (EMC) — Part 4-8: Testing and measurement techniques — Power frequency magnetic field immunity test

 ${\tt IEC\,61000-4-11}, \textit{Electromagnetic compatibility (EMC)} -- \textit{Part\,4-11}: \textit{Testing and measurement techniques} --$ Voltage dips, short interruptions and voltage variations immunity tests

IEC 61187, Electrical and electronic measuring equipment — Documentation

#### Terms and definitions

For the purposes of this document, the following terms and definitions apply.