



**SLOVENSKI STANDARD**  
**SIST-TS CEN/TS 17660-2:2025**

**01-januar-2025**

---

**Kakovost zraka - Vrednotenje lastnosti senzorskih sistemov za kakovost zraka - 2.  
del: Delci v zunanjem zraku**

Air quality - Performance evaluation of air quality sensor systems - Part 2: Particulate matter in ambient air

Luftbeschaffenheit - Leistungsbewertung von Luftqualitätssensorsystemen - Teil 2: Partikelförmige Stoffe in der Außenluft

Qualité de l'air - Évaluation des performances des systèmes capteurs de la qualité de l'air - Partie 2 : Particules dans l'air ambiant

**Ta slovenski standard je istoveten z: CEN/TS 17660-2:2024**

SIST-TS CEN/TS 17660-2:2025

<https://standards.sist.net/catalog/standards/sist/13202731/17660-1000/1001/0204/12033/121010/PART-2-SIST-TS-17660-2-2025>

**ICS:**

13.040.20      Kakovost okoljskega zraka      Ambient atmospheres

**SIST-TS CEN/TS 17660-2:2025**

**en,fr,de**



TECHNICAL SPECIFICATION  
SPÉCIFICATION TECHNIQUE  
TECHNISCHE SPEZIFIKATION

**CEN/TS 17660-2**

December 2024

ICS 13.040.20

English Version

**Air quality - Performance evaluation of air quality sensor systems - Part 2: Particulate matter in ambient air**

Qualité de l'air - Évaluation des performances des systèmes capteurs de la qualité de l'air - Partie 2 :  
Particules dans l'air ambiant

Luftbeschaffenheit - Leistungsbewertung von Luftqualitätssensorsystemen - Teil 2: Partikelförmige Stoffe in der Außenluft

This Technical Specification (CEN/TS) was approved by CEN on 6 October 2024 for provisional application.

The period of validity of this CEN/TS is limited initially to three years. After two years the members of CEN will be requested to submit their comments, particularly on the question whether the CEN/TS can be converted into a European Standard.

CEN members are required to announce the existence of this CEN/TS in the same way as for an EN and to make the CEN/TS available promptly at national level in an appropriate form. It is permissible to keep conflicting national standards in force (in parallel to the CEN/TS) until the final decision about the possible conversion of the CEN/TS into an EN is reached.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.

Document Preview

[SIST-TS CEN/TS 17660-2:2025](https://standards.iteh.ai/catalog/standards/sist/45202731-4088-40a1-b2d4-12b9542fb1a4/sist-ts-cen-ts-17660-2-2025)

<https://standards.iteh.ai/catalog/standards/sist/45202731-4088-40a1-b2d4-12b9542fb1a4/sist-ts-cen-ts-17660-2-2025>



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

<b>Contents</b>	<b>Page</b>
<b>European foreword</b> .....	<b>3</b>
<b>Introduction</b> .....	<b>4</b>
<b>1 Scope</b> .....	<b>5</b>
<b>2 Normative references</b> .....	<b>5</b>
<b>3 Terms and definitions</b> .....	<b>5</b>
<b>4 Symbols and abbreviations</b> .....	<b>10</b>
<b>5 Principle of the evaluation</b> .....	<b>13</b>
<b>6 Performance requirements</b> .....	<b>17</b>
<b>7 General requirements for the performance of tests</b> .....	<b>19</b>
<b>8 Laboratory tests of sensor system</b> .....	<b>19</b>
<b>9 Field tests</b> .....	<b>21</b>
<b>10 Classification based on the test results</b> .....	<b>31</b>
<b>11 Test report</b> .....	<b>32</b>
<b>Annex A (informative) Co-location of sensor systems, deployment and management of a network of sensor systems</b> .....	<b>36</b>
<b>A.1 Background</b> .....	<b>36</b>
<b>A.2 Co-location and deployment</b> .....	<b>36</b>
<b>A.2.1 Individual sensor systems</b> .....	<b>36</b>
<b>A.2.2 Sensor networks</b> .....	<b>37</b>
<b>A.3 Calibration and validation of networks</b> .....	<b>39</b>
<b>Annex B (informative) Evaluation of the effect of electromagnetic fields on the sensor system measurements</b> .....	<b>40</b>
<b>Annex C (informative) Air composition in different outdoor type of sites</b> .....	<b>41</b>
<b>Annex D (informative) Selecting the climate for a field trial site</b> .....	<b>45</b>
<b>Annex E (normative) Ordinary least square regression formulae</b> .....	<b>47</b>
<b>(informative) Values for <math>u</math>(bs, RM)</b> .....	<b>49</b>
<b>Annex F (informative) Example of field tests for the classification process</b> .....	<b>52</b>
<b>Annex G (informative) PM testing chamber design for validating coarse particle measurements</b> .....	<b>54</b>
<b>Bibliography</b> .....	<b>55</b>

## European foreword

This document (CEN/TS 17660-2:2024) has been prepared by Technical Committee CEN/TC 264 “Air quality”, the secretariat of which is held by DIN.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

CEN/TS 17660, *Air quality — Performance evaluation of air quality sensor systems* consists of the following parts:

- *Part 1: Gaseous pollutants in ambient air*
- *Part 2: Particulate matter in ambient air*

Any feedback and questions on this document should be directed to the users’ national standards body. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to announce this Technical Specification: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Itih Standards  
(<https://standards.itih.ai>)  
Document Preview

[SIST-TS CEN/TS 17660-2:2025](https://standards.itih.ai/catalog/standards/sist/45202731-4088-40a1-b2d4-12b9542fb1a4/sist-ts-cen-ts-17660-2-2025)

<https://standards.itih.ai/catalog/standards/sist/45202731-4088-40a1-b2d4-12b9542fb1a4/sist-ts-cen-ts-17660-2-2025>

## Introduction

Sensor systems are generally seen as emerging measuring devices for the monitoring of air quality. Sensor systems provide a fast and low-cost complement to the reference and equivalent measurement methods as defined in Directive 2008/50/EC on ambient air quality and cleaner air for Europe [1]. Sensor systems could allow for air pollution monitoring at a lower cost and with a higher spatial density than with the reference and equivalent measurement methods. They also allow for new air pollution applications including monitoring in complex topographies, at traffic junctions, in street canyons, at remote sites, on mobile platforms (pedestrians, cyclists, vehicles, trams) and for citizen science studies, e.g. monitoring around areas of local concern, schools, or parks.

Sensor systems use one or more low-cost sensors that are based on several technologies. However, sensor systems share two common features, portability and low-cost, compared with traditional reference or equivalent measurement methods. Typically, sensor systems can continuously monitor air pollution with fast response times ranging from a few tens of seconds to a few minutes.

Currently, the use of sensor systems for air quality monitoring is limited by the low accuracy of measurements that they can achieve. Additionally, there was no unambiguous protocol for evaluating sensor systems with a structured metrological approach, ensuring traceability from sensor system measurements to national and international standards. A protocol will enable exhaustive and transparent evaluation of sensor systems that can be an important step towards including sensor system measurements into the monitoring of air quality for regulatory and non-regulatory purposes.

The protocol in this document applies to sensor systems and supports the requirements of Directive 2008/50/EC. This procedure evaluates whether the measurement uncertainty of the sensor system meets the data quality objectives defined in Directive 2008/50/EC for indicative measurements. This protocol also allows for a less demanding evaluation of the performance of sensor systems for non-regulatory measurements.

This document defines common procedures and requirements for the evaluation of the performance of sensor systems to facilitate mutual recognition by the relevant bodies or stakeholders and thereby minimise both administrative and cost burdens on manufacturers. It does not describe the roles and responsibilities of manufacturers, test laboratories and relevant bodies under these procedures.

## 1 Scope

This document specifies the general principles, including testing procedures and requirements, for the classification of performance of low-cost sensor systems for the monitoring of particulate matter in ambient air at fixed sites. The classification of sensor systems includes tests that are performed under prescribed conditions. It does not guarantee performance in locations that are different from the tests, variations in meteorological climate from the test programme or account for stability over time, which can only be assessed under ongoing quality control strategies.

The described procedure is applicable to the determination of the mass concentration of particulate matter. The pollutants that are considered in this document are PM<sub>10</sub> and PM<sub>2,5</sub> in the range of concentrations expected in ambient air.

This document provides a classification that is consistent with the requirements for indicative measurements and objective estimation defined in Directive 2008/50/EC. In addition, it provides a classification for applications (non-regulatory measurements) that require more relaxed performance criteria.

This document applies to sensor systems used as individual systems. It does not apply to sensor systems as part of a sensor network. However, for some applications (e.g. in cities) sensor systems are deployed as part of a sensor network. Annex A provides information on the use of sensor systems as nodes in a sensor network.

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 15267-1, *Air quality - Assessment of air quality monitoring equipment - Part 1: General principles of certification*

EN 15267-2, *Air quality - Assessment of air quality monitoring equipment - Part 2: Initial assessment of the manufacturer's quality management system and post certification surveillance for the manufacturing process*

EN 16450, *Ambient air — Automated measuring systems for the measurement of the concentration of particulate matter (PM<sub>10</sub>; PM<sub>2,5</sub>)*

## 3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp/>

### 3.1

#### **ambient air**

outdoor air in the troposphere where provisions concerning health and safety at work apply and to which members of the public do not have regular access

Note 1 to entry: This does not include workplaces as defined by Directive 89/654/EC [2].

[SOURCE: Directive 2008/50/EC] [1]

**CEN/TS 17660-2:2024 (E)****3.2****averaging period**

period of time for which a limit value is associated

Note 1 to entry: For this document, the averaging period is defaulted to 24 h unless otherwise specified (more information can be found in Directive 2008/50/EC).

**3.3****calibration**

operation that, under specified conditions, in a first step, establishes a relation between the quantity values with measurement uncertainties provided by measurement standards and corresponding indications with associated measurement uncertainties and, in a second step, uses this information to establish a relation for obtaining a measurement result from an indication

[SOURCE: JCGM 200:2012, 2.39] [3]

Note 1 to entry: A calibration can be expressed by a statement, calibration function, calibration diagram, calibration curve, or calibration table. In some cases, it can consist of an additive or multiplicative correction of the indication with associated measurement uncertainty. Calibration should not be confused with adjustment of a measuring system, often mistakenly called “self-calibration”, nor with verification of calibration.

Note 2 to entry: This document does not describe the process of calibration of sensor systems.

**3.4****Class 1 sensor system**

measuring device delivering data that are at minimum consistent with the data quality objectives of indicative measurements

Note 1 to entry: The term “indicative measurement” does not refer to the performance of the sensor system.

**3.5****Class 2 sensor system**

measuring device delivering data that are at minimum consistent with the data quality objectives of objective estimations

Note 1 to entry: The term “objective estimation” does not refer to the performance of the sensor system.

**3.6****Class 3 sensor system**

measuring device delivering data that comply with a relaxed target measurement uncertainty, but are not formally associated with any mandatory data quality objective

Note 1 to entry: Relaxed target measurement uncertainties are given in Table 2.

**3.7****combined standard measurement uncertainty**

standard measurement uncertainty that is obtained using the individual standard measurement uncertainties associated with the input quantities in a measurement model

[SOURCE: JCGM 200:2012, 2.31] [3]



### 3.8

#### **coverage factor**

number greater than one by which a combined standard measurement uncertainty is multiplied to obtain an expanded measurement uncertainty

Note 1 to entry: A coverage factor symbol is usually termed  $k$ .

[SOURCE: JCGM 200:2012, 2.38] [3]

### 3.9

#### **drift**

continuous or incremental change over time in measurement, due to changes in properties of a sensor system

Note 1 to entry: Drift is not related either to a change in a quantity being measured or to a change of any recognized influencing quantity.

### 3.10

#### **expanded measurement uncertainty**

expanded uncertainty

product of a combined standard measurement uncertainty and the coverage factor

Note 1 to entry: The factor depends upon the type of probability distribution of the output quantity in a measurement model and on the selected coverage probability. The term “factor” in this definition refers to a coverage factor.

Note 2 to entry: Expanded measurement uncertainty is termed “overall uncertainty” in paragraph 5 of Recommendation INC-1 (1980) (see the GUM) and simply “uncertainty” in IEC documents.

Note 3 to entry: Expanded measurement uncertainty is also referred to as “uncertainty” in this document.

[SOURCE: JCGM 200:2012, 2.35] [3] <https://standards.iteh.ai/catalog/standards/sist/45202731-4088-40a1-b2d4-12b9542fb1a4/sist-ts-cen-ts-17660-2-2025>

### 3.11

#### **equivalent method**

method for the measurement of the concentration of particulate matter ( $PM_{10}$ ;  $PM_{2,5}$ ) meeting the data quality objectives for fixed measurements and tested according to EN 16450

### 3.12

#### **exposure chamber**

test chamber

volume that can be sealed with controlled conditions of temperature, humidity and test aerosol, used for testing the sensor system

### 3.13

#### **fixed measurements**

measurements taken at fixed sites, either continuously or by random sampling, to determine the levels in accordance with the relevant data quality objectives

Note 1 to entry: The data quality objectives are listed in Table 2 (more information can be found in Directive 2008/50/EC).

**CEN/TS 17660-2:2024 (E)****3.14****independent measurement**

measurement that is not influenced by a previous individual measurement, by separating two individual measurements by at least four response times

**3.15****indicative measurements**

measurements which meet data quality objectives that are less strict than those required for fixed measurements

Note 1 to entry: The data quality objectives are listed in Table 2 (more information can be found in Directive 2008/50/EC).

**3.16****PM<sub>10-2,5</sub>**

difference between reported PM<sub>10</sub> and reported PM<sub>2,5</sub> concentrations, also termed coarse particles

**3.17****PM<sub>x</sub>**

particulate matter suspended in air which passes through a size-selective inlet at a constant flow with a 50 % efficiency cut-off at x µm aerodynamic diameter

Note 1 to entry: By convention, the size-selective standard inlet designs prescribed in EN 12341 — used at the prescribed flow rates – possess the required characteristics to sample the relevant PM fraction suspended in ambient air.

Note 2 to entry: The efficiency of the size selectiveness of other inlets used could have a significant effect on the fraction of PM surrounding the cut-off, and, consequently on the mass concentration of PM<sub>x</sub> determined.

[SOURCE: EN 12341:2023] [4]

**3.18****measurement repeatability**

repeatability

measurement precision under a set of repeatability conditions of measurement

[SOURCE: JCGM 200:2012, 2.21] [3]

**3.19****objective estimation**

measurements which meet data quality objectives that are less strict than those required for indicative measurements

Note 1 to entry: The data quality objectives are listed in Table 2 (more information can be found in Directive 2008/50/EC).

**3.20****reference method**

standard gravimetric measurement method for the determination of the PM<sub>10</sub> or PM<sub>2,5</sub> mass concentration of particulate matter according to EN 12341 [4]

**3.21****repeatability condition of measurement**

conditions that include the same measurement procedure, same operators, same measuring system, same operating conditions and same location, and replicate measurements on the same or similar objects over a short period of time

Note 1 to entry: A condition of measurement is a repeatability condition only with respect to a specified set of repeatability conditions.

Note 2 to entry: In chemistry, the term “intra-serial precision condition of measurement” is sometimes used to designate this concept.

[SOURCE: JCGM 200:2012, 2.20] [3]

**3.22****sensor**

individual sensor

physical unit that produces a signal related to the concentration of particulate matter in air

**3.23****sensor system**

single integrated set of hardware that uses one or more sensors to produce a signal related to the concentration of particulate matter in air that can supply real time measurements

Note 1 to entry: The term “instrument” has a very similar definition, but many researchers are typically referring to a reference or equivalent grade device when using the word “instrument”.

Note 2 to entry: All the tests that are intended in this document are designed for sensor systems only.

Note 3 to entry: Sensor systems contain many common components in addition to the basic sensing or analytical element that is used for detection. Common core components and functions can include:

- sensing detector (the actual sensor);
- sampling capability (generally active sampling for PM);
- power systems, which may include batteries;
- analogue to digital conversion;
- signal processing;
- local data storage;
- data transmission;
- enclosure.

**3.24****standard measurement uncertainty**

standard uncertainty

measurement uncertainty expressed as a standard deviation

[SOURCE: JCGM 200:2012, 2.30] [3]

**CEN/TS 17660-2:2024 (E)****3.25****uncertainty of measurement**

measurement uncertainty

non-negative parameter characterizing the dispersion of the quantity values being attributed to a measurand, based on the information used

Note 1 to entry: Measurement uncertainty includes components arising from systematic effects, such as components associated with corrections and the assigned quantity values of measurement standards, as well as the definitional uncertainty. Sometimes estimated systematic effects are not corrected for, but instead, associated measurement uncertainty components are incorporated.

Note 2 to entry: The parameter may be, for example, a standard deviation called standard measurement uncertainty (or a specified multiple of it), or the half-width of an interval, having a stated coverage probability.

Note 3 to entry: Measurement uncertainty comprises, in general, many components. Some of these may be evaluated by Type A evaluation of measurement uncertainty from the statistical distribution of the quantity values from series of measurements and can be characterized by standard deviations. The other components, which may be evaluated by Type B evaluation of measurement uncertainty, can also be characterized by standard deviations, evaluated from probability density functions based on experience or other information.

Note 4 to entry: In general, for a given set of information, it is understood that the measurement uncertainty is associated with a stated quantity value attributed to the measurand. A modification of this value results in a modification of the associated uncertainty.

[SOURCE: JCGM 200:2012, 2.26] [3]

**4 Symbols and abbreviations****4.1 Symbols**

For the purposes of this document, the following symbols apply.

NOTE In the following list, all symbols related to the measurement uncertainty of source contributions are indicated as standard uncertainty, e.g.  $u(X)$ . However, in the text expanded uncertainties of the same source contributions can be used, e.g.  $U(X)$ .

$a$	intercept of the regression line
$b$	slope of the regression line
$c$	intercept of the corrected dataset
$d$	slope of the corrected dataset
$F\left(\frac{FINE}{COARSE}\right)$	factor of the sensor response change from fine to coarse conditions
$k$	coverage factor
$L$	LV, UAT or LAT currently being assessed
$Mdn(sensorratio)_{30-60\%}$	median of <i>Sensorratio</i> with hourly RH% between 30 % and 60 %;
$Mdn(sensorratio)_{85-95\%}$	median of <i>Sensorratio</i> with hourly RH% between 85 % and 95 %.

$n$	is the number of measurement periods where all sensor systems are providing valid data (with identical $n_s$ )
$n_s$	number of replicate sensor systems
$PM_{10-2.5}$	PM coarse, difference between reported $PM_{10}$ and reported $PM_{2.5}$ concentrations
$r$	measurement repeatability; repeatability
$R$	value of the residual sum of squares resulting from the linear regression
$R^2$	coefficient of determination
$RH_{factor}$	factor that expresses the change in sensor response (compared to the equivalent method) when RH changes from low (30 % - 60 %) to high (85 % to 95 %)
$PM_{2.5}(\text{sen, FINE})$	$PM_{2.5}$ sensor response ( $\mu\text{g}/\text{m}^3$ ), averaged over the duration of the test with fine PM fraction
$PM_{2.5}(\text{eq, FINE})$	$PM_{2.5}$ measured by equivalent method ( $\mu\text{g}/\text{m}^3$ ), averaged over the duration of the test with fine PM fraction
$PM_{10-2.5}(\text{sen, COARSE})$	$PM_{10-2.5}$ sensor response ( $\mu\text{g}/\text{m}^3$ ), averaged over the duration of the test with coarse PM fraction
$PM_{10-2.5}(\text{eq, COARSE})$	$PM_{10-2.5}$ measured by equivalent method ( $\mu\text{g}/\text{m}^3$ ), averaged over the duration of the test with coarse PM fraction
$R_{PM_{2.5}(FINE)}$	ratio of sensor to reference method measurements for $PM_{2.5}$ during the test with fine PM fraction
$R_{PM_{10-2.5}(COARSE)}$	ratio of sensor to reference method measurements for $PM_{10-2.5}$ during the test the test with coarse PM fraction
<i>Sensor ratio</i>	ratio of hourly average of the sensor PM measurement to the hourly average of the PM measurements of the equivalent method for $PM_{2.5}$ or $PM_{10}$
$y_{sen}$	hourly average of sensor measurement
$y_{eq}$	hourly average of equivalent method
$u(a)$	standard uncertainty of the intercept of a regression line
$u(b)$	standard uncertainty of the slope of a regression line
$u(\text{bs, RM})$	between reference method standard uncertainty
$u(\text{bs, s})$	between sensor systems standard uncertainty
$U_{\text{field},L}$	expanded uncertainty of the field measurements of the sensor system at the value of 'L =' LV, UAT or LAT currently being assessed, in units of mass concentrations

**CEN/TS 17660-2:2024 (E)**

$U_{\text{field,corr},L}$	expanded uncertainty of the corrected sensor system at the value of 'L' LV, UAT or LAT currently being assessed, in units of mass concentrations
$x$	value of the reference measurement
$x_i$	$i^{\text{th}}$ $x$ -value (reference measurement)
$\bar{x}$	average of the values $x_i$
$y$	value of the sensor system response
$y_i$	$i^{\text{th}}$ $y$ -value (sensor system response)
$y_{i,j}$	value of the sensor system response for period $i$ of sensor system $j$
$\bar{y}$	average of the values $y_i$
$\bar{y}_i$	average for period $i$ of the $n_s$ replicate sensor systems;

**4.2 Abbreviations**

For the purposes of this document, the following abbreviations apply.

AQMS	air quality monitoring station (more information on the measurement methods can be found in Directive 2008/50/EC)
DQO	data quality objective
GPS	global positioning system
LAT	lower assessment threshold
LV	limit value
OLS	ordinary least squares
$R^2$	coefficient of determination
RH	relative humidity
RSS	residual sum of squares resulting from the linear regression
SOP	standard operating procedure
UAT	upper assessment threshold
WLAN	wireless local area network