

# **SLOVENSKI STANDARD**

## **oSIST prEN ISO 13160:2020**

**01-april-2020**

---

**Kakovost vode - Stroncij Sr-90 in stroncij Sr-89 - Preskusne metode s štetjem s tekočinskim scintilatorjem ali proporcionalnim štetjem (ISO/DIS 13160:2020)**

Water quality - Strontium 90 and strontium 89 - Test methods using liquid scintillation counting or proportional counting (ISO/DIS 13160:2020)

Wasserbeschaffenheit - Strontium 90 und Strontium 89 - Verfahren mittels Flüssigszintillationszählung oder Proportionalzählung (ISO/DIS 13160:2020)

Qualité de l'eau - Strontium 90 et strontium 89 - Méthodes d'essai par comptage des scintillations en milieu liquide ou par comptage proportionnel (ISO/DIS 13160:2020)

<https://standards.iteh.ai/catalog/standards/sist/446b89c9-ba07-4542-ab52-76e5118b7880/osist-pr-en-iso-13160-2020>

**Ta slovenski standard je istoveten z: prEN ISO 13160**

---

**ICS:**

13.060.60	Preiskava fizikalnih lastnosti vode	Examination of physical properties of water
17.240	Merjenje sevanja	Radiation measurements

**oSIST prEN ISO 13160:2020**

**en,fr,de**

## **iTeh STANDARD PREVIEW** **(standards.iteh.ai)**

[oSIST prEN ISO 13160:2020](https://standards.iteh.ai/catalog/standards/sist/446b89c9-ba07-4542-ab52-76a5118b7880/osist-pren-iso-13160-2020)

<https://standards.iteh.ai/catalog/standards/sist/446b89c9-ba07-4542-ab52-76a5118b7880/osist-pren-iso-13160-2020>

# DRAFT INTERNATIONAL STANDARD

## ISO/DIS 13160

ISO/TC 147/SC 3

Secretariat: AFNOR

Voting begins on:  
2020-02-24Voting terminates on:  
2020-05-18

## Water quality — Strontium 90 and strontium 89 — Test methods using liquid scintillation counting or proportional counting

*Qualité de l'eau — Strontium 90 et strontium 89 — Méthodes d'essai par comptage des scintillations en milieu liquide ou par comptage proportionnel*

ICS: 13.060.60; 17.240

### iTeh STANDARD PREVIEW (standards.iteh.ai)

[oSIST prEN ISO 13160:2020](https://standards.iteh.ai/catalog/standards/sist/446b89c9-ba07-4542-ab52-76a5118b7880/osist-pren-iso-13160-2020)<https://standards.iteh.ai/catalog/standards/sist/446b89c9-ba07-4542-ab52-76a5118b7880/osist-pren-iso-13160-2020>

THIS DOCUMENT IS A DRAFT CIRCULATED FOR COMMENT AND APPROVAL. IT IS THEREFORE SUBJECT TO CHANGE AND MAY NOT BE REFERRED TO AS AN INTERNATIONAL STANDARD UNTIL PUBLISHED AS SUCH.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

This document is circulated as received from the committee secretariat.

## ISO/CEN PARALLEL PROCESSING



Reference number  
ISO/DIS 13160:2020(E)

© ISO 2020

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[oSIST prEN ISO 13160:2020](https://standards.iteh.ai/catalog/standards/sist/446b89c9-ba07-4542-ab52-76a5118b7880/osist-pren-iso-13160-2020)  
<https://standards.iteh.ai/catalog/standards/sist/446b89c9-ba07-4542-ab52-76a5118b7880/osist-pren-iso-13160-2020>



### **COPYRIGHT PROTECTED DOCUMENT**

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Fax: +41 22 749 09 47  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

Page

<b>Foreword</b>	<b>v</b>
<b>Introduction</b>	<b>vi</b>
<b>1 Scope</b>	<b>1</b>
<b>2 Normative references</b>	<b>1</b>
<b>3 Terms and definitions</b>	<b>1</b>
<b>4 Principle</b>	<b>2</b>
4.1 General	2
4.2 Chemical separation	2
4.3 Detection	3
<b>5 Chemical reagents and equipment</b>	<b>3</b>
<b>6 Procedure</b>	<b>3</b>
6.1 Test sample preparation	3
6.2 Chemical separation	3
6.2.1 General	3
6.2.2 Precipitation techniques	4
6.2.3 Liquid-liquid extraction technique	5
6.2.4 Chromatographic technique	5
6.3 Preparation of the source for test	5
6.3.1 Source preparation for liquid scintillation counter	5
6.3.2 Source preparation for proportional counter	5
6.4 Measurement	6
6.4.1 General	6
6.4.2 Liquid scintillation counter	6
6.4.3 Proportional counter	6
6.4.4 Efficiency calculation	7
6.4.5 Determination of the chemical yield	7
<b>7 Expression of results</b>	<b>8</b>
7.1 Determination of $^{90}\text{Sr}$ in equilibrium with $^{90}\text{Y}$	8
7.1.1 Calculation of the activity concentration	8
7.1.2 Standard uncertainty	8
7.1.3 Decision threshold	9
7.1.4 Detection limit	9
7.2 Determination of $^{90}\text{Sr}$ from separated $^{90}\text{Y}$	9
7.2.1 Calculation of the activity concentration	9
7.2.2 Standard uncertainty	10
7.2.3 Decision threshold	11
7.2.4 Detection limit	11
7.3 Determination of $^{90}\text{Sr}$ in presence of $^{89}\text{Sr}$ when $^{90}\text{Sr}$ is in equilibrium with $^{90}\text{Y}$	11
7.3.1 Calculation of the activity concentration	11
7.3.2 Standard uncertainty	12
7.3.3 Decision threshold	13
7.3.4 Detection limit	14
<b>8 Limits of the coverage intervals</b>	<b>14</b>
8.1 Limits of the of the probabilistically symmetric coverage interval	14
8.2 The shortest coverage interval	15
<b>9 Quality control</b>	<b>15</b>
<b>10 Test report</b>	<b>15</b>
<b>Annex A (informative) Determination of <math>^{89}\text{Sr}</math> and <math>^{90}\text{Sr}</math> by precipitation and proportional counting</b>	<b>17</b>

## ISO/DIS 13160:2020(E)

<b>Annex B (informative) Determination of <math>^{89}\text{Sr}</math> and <math>^{90}\text{Sr}</math> by precipitation and liquid scintillation counting</b>	<b>21</b>
<b>Annex C (informative) Determination of <math>^{90}\text{Sr}</math> from its decay progeny <math>^{90}\text{Y}</math> at equilibrium by organic extraction and liquid scintillation counting</b>	<b>25</b>
<b>Annex D (informative) Determination of <math>^{90}\text{Sr}</math> after ionic exchange separation by proportional counting</b>	<b>28</b>
<b>Annex E (informative) Determination of <math>^{90}\text{Sr}</math> after separation on a crown ether specific resin and liquid scintillation counting</b>	<b>31</b>
<b>Annex F (informative) Determination of <math>^{90}\text{Sr}</math> from its decay progeny <math>^{90}\text{Y}</math> at equilibrium by organic extraction by proportional counting</b>	<b>33</b>
<b>Annex G (informative) Correction factor for Sr-90 purity using proportional counting</b>	<b>37</b>
<b>Bibliography</b>	<b>40</b>

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[oSIST prEN ISO 13160:2020](https://standards.iteh.ai/catalog/standards/sist/446b89c9-ba07-4542-ab52-76a5118b7880/osist-pren-iso-13160-2020)

<https://standards.iteh.ai/catalog/standards/sist/446b89c9-ba07-4542-ab52-76a5118b7880/osist-pren-iso-13160-2020>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 147, *Water quality*, Subcommittee SC 3, *Radioactivity measurements*.

This second edition cancels and replaces the first edition (ISO 13160:2012), which has been technically revised. The main changes compared to the previous edition are as follows:

## Introduction

Radioactivity from several naturally-occurring and anthropogenic sources is present throughout the environment. Thus, water bodies (e.g. surface waters, ground waters, sea waters) can contain radionuclides of natural, human-made, or both origins.

- Natural radionuclides, including  $^{40}\text{K}$ ,  $^3\text{H}$ ,  $^{14}\text{C}$ , and those originating from the thorium and uranium decay series, in particular  $^{226}\text{Ra}$ ,  $^{228}\text{Ra}$ ,  $^{234}\text{U}$ ,  $^{238}\text{U}$ , and  $^{210}\text{Pb}$ , can be found in water for natural reasons (e.g. desorption from the soil and washoff by rain water) or can be released from technological processes involving naturally occurring radioactive materials (e.g. the mining and processing of mineral sands or phosphate fertilizer production and use).
- Human-made radionuclides, such as transuranium elements (americium, plutonium, neptunium, curium),  $^3\text{H}$ ,  $^{14}\text{C}$ ,  $^{90}\text{Sr}$ , and gamma emitting radionuclides can also be found in natural waters. Small quantities of these radionuclides are discharged from nuclear fuel cycle facilities into the environment as the result of authorized routine releases. Some of these radionuclides used for medical and industrial applications are also released into the environment after use. Anthropogenic radionuclides are also found in waters as a result of past fallout contaminations resulting from the explosion in the atmosphere of nuclear devices and accidents such as those that occurred in Chernobyl and Fukushima.

Radionuclide activity concentration in water bodies can vary according to local geological characteristics and climatic conditions and can be locally and temporally enhanced by releases from nuclear installation during planned, existing and emergency exposure situations<sup>[1]</sup>. Drinking water may thus contain radionuclides at activity concentrations which could present a risk to human health.

The radionuclides present in liquid effluents are usually controlled before being discharged into the environment<sup>[2]</sup>. Water bodies and drinking waters are monitored for their radioactivity content as recommended by the World Health Organization (WHO)<sup>[3]</sup> so that proper actions can be taken to ensure that there is no adverse health effect to the public. Following these international recommendations, national regulations usually specify radionuclide authorized concentration limits for liquid effluent discharged to the environment and radionuclide guidance levels for water bodies and drinking waters for planned, existing and emergency exposure situations. Compliance with these limits can be assessed using measurement results with their associated uncertainties as specified by ISO/IEC Guide 98-3 and ISO 5667-20<sup>[4]</sup>.

Depending on the exposure situation, there are different limits and guidance levels that would result in an action to reduce health risk. As an example, during a planned or existing situation, the WHO guidelines for guidance level in drinking water is  $100 \text{ Bq}\cdot\text{l}^{-1}$  for  $^{89}\text{Sr}$  activity concentration and  $10 \text{ Bq}\cdot\text{l}^{-1}$  for  $^{90}\text{Sr}$  activity concentration.

**NOTE 1** The guidance level is the activity concentration with an intake of 2 l/d of drinking water for one year that results in an effective dose of 0,1 mSv/a for members of the public. This is an effective dose that represents a very low level of risk and which is not expected to give rise to any detectable adverse health effects<sup>[3]</sup>.

In the event of a nuclear emergency, the WHO Codex guideline levels<sup>[5]</sup> mentioned that the activity concentration might not be greater than  $1000 \text{ Bq}\cdot\text{l}^{-1}$  for  $^{89}\text{Sr}$  or  $100 \text{ Bq}\cdot\text{l}^{-1}$  for  $^{90}\text{Sr}$  for infant food and  $1000 \text{ Bq}\cdot\text{l}^{-1}$  for  $^{89}\text{Sr}$  or  $100 \text{ Bq}\cdot\text{l}^{-1}$  for  $^{90}\text{Sr}$  for food other than infant food.

**NOTE 2** The Codex guidelines levels (GLs) apply to radionuclides contained in food destined for human consumption and traded internationally, which have been contaminated following a nuclear or radiological emergency. These GLs apply to food after reconstitution or as prepared for consumption, i.e. not to dried or concentrated food, and are based on an intervention exemption level of 1 mSv in a year for members of the public (infant and adult)<sup>[5]</sup>.

Thus, the test method can be adapted so that the characteristic limits, decision threshold, detection limit and uncertainties ensure that the radionuclide activity concentrations test results can be verified to be below the guidance levels required by a national authority for either planned/existing situations or for an emergency situation<sup>[6][7]</sup>.



Usually, the test methods can be adjusted to measure the activity concentration of the radionuclide(s) in either wastewaters before storage or in liquid effluents before discharge to the environment. The test results will enable the plant/installation operator to verify that, before their discharge, wastewaters/liquid effluent radioactive activity concentrations do not exceed authorized limits.

The test method described in this document may be used during planned, existing and emergency exposure situations as well as for wastewaters and liquid effluents with specific modifications that could increase the overall uncertainty, detection limit, and threshold.

The test method may be used for water samples after proper sampling, sample handling, and test sample preparation (see the relevant part of the ISO 5667 series).

This document has been developed to answer the need of test laboratories carrying out these measurements, that are sometimes required by national authorities, as they may have to obtain a specific accreditation for radionuclide measurement in drinking water samples.

This document is one of a set of International Standards on test methods dealing with the measurement of the activity concentration of radionuclides in water samples.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[oSIST prEN ISO 13160:2020](https://standards.iteh.ai/catalog/standards/sist/446b89c9-ba07-4542-ab52-76a5118b7880/osist-pren-iso-13160-2020)

<https://standards.iteh.ai/catalog/standards/sist/446b89c9-ba07-4542-ab52-76a5118b7880/osist-pren-iso-13160-2020>

## **iTeh STANDARD PREVIEW** **(standards.iteh.ai)**

[oSIST prEN ISO 13160:2020](https://standards.iteh.ai/catalog/standards/sist/446b89c9-ba07-4542-ab52-76a5118b7880/osist-pren-iso-13160-2020)

<https://standards.iteh.ai/catalog/standards/sist/446b89c9-ba07-4542-ab52-76a5118b7880/osist-pren-iso-13160-2020>

# Water quality — Strontium 90 and strontium 89 — Test methods using liquid scintillation counting or proportional counting

**WARNING** — Persons using this document should be familiar with normal laboratory practice. This document does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to determine the applicability of any other restrictions.

**IMPORTANT** — It is absolutely essential that tests conducted according to this document be carried out by suitably trained staff.

## 1 Scope

This document provides a selection of test methods and their associated principles for the measurement of the activity of  $^{90}\text{Sr}$  in equilibrium with  $^{90}\text{Y}$ , and  $^{89}\text{Sr}$ , pure beta-emitting radionuclides, in water samples. Different chemical separation methods are presented to produce strontium and yttrium sources, the activity of which are determined using a proportional counter (PC) or liquid scintillation counter (LSC).

The selection of a particular test method depends on the origin of the contamination, the characteristics of the water to be analyzed, the required accuracy of test results and the available resources of the laboratory.

These test methods are used for water monitoring following past or present, accidental or routine, liquid or gaseous discharges. It also covers the monitoring of contamination caused by global fallout.

When fallout occurs immediately following a nuclear accident, the contribution of  $^{89}\text{Sr}$  to the total amount of strontium activity is not negligible. This document provides test methods to determine the activity concentration of  $^{90}\text{Sr}$  in presence of  $^{89}\text{Sr}$ .

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements for this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 11929, *Determination of the characteristic limits (decision threshold, detection limit and limits of the confidence interval) for measurements of ionizing radiation — Fundamentals and application*

ISO 80000-10, *Quantities and units — Part 10: Atomic and nuclear physics*

ISO 19361, *Measurement of radioactivity — Determination of beta emitters activities — Test method using liquid scintillation counting*

## 3 Terms and definitions

For the purposes of this document, the definitions, symbols, and abbreviated terms defined in ISO 11929, ISO 80000-10 and the following apply.

$A_i$	calibration source activity of radionuclide $i$ , at the time of calibration	Bq
$c_{A,i}$	activity concentration of radionuclide $i$	$\text{Bq l}^{-1}$

# ISO/DIS 13160:2020(E)

$c_{A,i}^*$	decision threshold of radionuclide i	Bq l <sup>-1</sup>
$c_{A,i}^\#$	detection limit of radionuclide i	Bq l <sup>-1</sup>
$c_{A,i}^<, c_{A,i}^>$	lower and upper limits of the probabilistically symmetric coverage interval of radionuclide i	Bq l <sup>-1</sup>
$c_{A,i}^{<}, c_{A,i}^{>}$	lower and upper limits of the shortest coverage interval of radionuclide i	Bq l <sup>-1</sup>
$R_{c,i}$	chemical yield of the extraction of radionuclide i	1
$r_0$	background count rate	s <sup>-1</sup>
$r_{0j}$	background count rate for measurement j	s <sup>-1</sup>
$r_g$	gross count rate	s <sup>-1</sup>
$r_{gj}$	gross count rate for measurement j	s <sup>-1</sup>
$r_j$	net count rate for measurement j	s <sup>-1</sup>
$r_s$	calibration source count rate	s <sup>-1</sup>
$t$	time elapsed between separation of <sup>90</sup> Sr/ <sup>90</sup> Y ( $t = 0$ ) and mid-point of counting	s
$t_0$	background counting time	s
$t_d, t_f$	start and finish time respectively of the measurement, referred to $t = 0$	s
$t_g$	sample counting time	s
$t_j$	start time of the measurement j, referred to $t = 0$	s
$t_s$	calibration source counting time	s
$U$	expanded uncertainty, calculated by $U = ku(c_A)$ with $k = 1, 2 \dots$	Bq l <sup>-1</sup>
$u(c_A)$	standard uncertainty associated with the measurement result	Bq l <sup>-1</sup>
$V$	volume of the test sample	l
$\varepsilon_i$	counting efficiency for radionuclide i	1
$\lambda_i$	decay constant of radionuclide i	1

## 4 Principle

### 4.1 General

Strontium-90, <sup>90</sup>Y and <sup>89</sup>Sr are all pure beta-particle emitting radionuclides. Their beta-emission energies and half-lives are given in [Table 1](#).

**Table 1 — Half-lives, maximum energies, and average energies of <sup>90</sup>Sr, <sup>90</sup>Y, and <sup>89</sup>Sr [8]**

Parameter	<sup>90</sup> Sr	<sup>90</sup> Y	<sup>89</sup> Sr
Maximum energy	546,0 keV	2 283,9 keV	1 491,0 keV
Average energy	196,4 keV	935,3 keV	586,3 keV
Half-life	28,80 (7) a	2,6684 (13) d	50,57 (3) d

Strontium-90 can be directly measured or estimated through the measurement of its decay progeny <sup>90</sup>Y. All the test methods are based on a chemical separation step followed by beta-counting of either <sup>90</sup>Sr or <sup>90</sup>Y using PC or LSC. See Table 2.

### 4.2 Chemical separation

Strontium is isolated from the water using precipitation, ion exchange or specific chromatographic separation by crown ether resin (see Reference [9]). Yttrium can then be isolated by precipitation or liquid-liquid extraction.

The method chosen shall be selective with a high chemical yield. When thorium, lead or bismuth radioisotopes are present at high activity levels, they may interfere with  $^{90}\text{Sr}$ ,  $^{90}\text{Y}$  or  $^{89}\text{Sr}$  detection. Other matrix constituents such as alkaline earth metals and in particular calcium for strontium, or transuranic and lanthanide elements for yttrium, reduce the chemical yield of the extraction.

The radiochemical separation yield is calculated using a carrier such as stable strontium or yttrium, or a radioactive tracer such as  $^{85}\text{Sr}$ . Techniques like atomic absorption spectroscopy (AAS), inductively coupled plasma–atomic emission spectroscopy (ICP–AES) or inductively coupled plasma–mass spectrometry (ICP–MS) to measure the carrier, and gamma-spectrometry to measure  $^{85}\text{Sr}$ , are recommended. A carrier can also be measured by gravimetric methods, but the presence of inactive elements, essentially alkaline earth elements, in the leaching solutions can lead to an overestimation of the radiochemical separation yields, particularly for the measurement of strontium.

When stable strontium is added as a carrier, the original strontium concentration in the test sample shall be known to determine the chemical yield.

### 4.3 Detection

The use of LSC, which provides spectra and permits the detection of interference from unwanted radionuclides, is recommended in preference to PC, which does not distinguish between emissions from different beta-emitters. When PC is used, it is recommended that the purity of the precipitate is checked by following the change over an appropriate time of the  $^{90}\text{Y}$  or  $^{89}\text{Sr}$  activity, even though this is time consuming.

Six test methods are presented in [Annexes A, B, C, D, E, and F](#).

## 5 Chemical reagents and equipment

The necessary chemical reagents and equipment for each strontium measurement method are specified in [Annexes A, B, C, D, E, and F](#).

During the analyses, unless otherwise stated, use only reagents of recognized analytical grade and laboratory water such as distilled or demineralized water or water of equivalent purity as specified in ISO 3696[10].

## 6 Procedure

### 6.1 Test sample preparation

Strontium is determined from the water test sample after appropriate sampling procedures[18][19][20].

If filtration is required, it should be done prior to the addition of the tracer or carrier and sufficient time should be allowed to attain chemical equilibrium before starting the test sample preparation.

When stable strontium is added as a carrier, the original strontium concentration in the test sample shall be known to determine the chemical yield. In the case of separation based on extraction chromatography, the total strontium content shall be below the sorption capacity of the resin to avoid saturation of the resin.

### 6.2 Chemical separation

#### 6.2.1 General

There are several routine analyses of  $^{89}\text{Sr}$  and  $^{90}\text{Sr}$  involved in the separation and purification of strontium: precipitation, liquid–liquid extraction or chromatographic techniques (ion exchange or chromatographic extraction). [Annexes A, B, C, D, E, and F](#) describe a test method for each of these techniques.



Two precipitation methods are described: [Annex A](#) employs PC for  $^{89}\text{Sr}$  and  $^{90}\text{Sr}$ ; [Annex B](#) employs LSC for  $^{89}\text{Sr}$  and  $^{90}\text{Sr}$ .

### 6.2.3 Liquid-liquid extraction technique

This technique is based on the extraction of  $^{90}\text{Y}$  in equilibrium with its radioactive parent  $^{90}\text{Sr}$  with an organic solvent. The chemical separation is fast and requires few technical resources. A provisional result may be achieved after 3 d (approximately one yttrium decay period). However, total selectivity of the extraction is not always possible. In the presence of high levels of natural radioactivity, interference may occur, making it difficult to determine very low levels of strontium activity.

Yttrium-90 is extracted from the water test sample fraction using an organic solvent, and then after re-extraction, recovered in the form of an yttrium precipitate. Test methods are presented in [Annexes C](#) and [F](#).

After the source preparation, the  $^{90}\text{Y}$  is measured by PC ([Annex F](#)) or by measuring the Cerenkov radiation from the  $^{90}\text{Y}$  with LSC ([Annex C](#)).

The absence of other interfering beta-emitters is verified during the decay of  $^{90}\text{Y}$  by measuring the decrease in count rate of the  $^{90}\text{Y}$  and once the decay is complete, comparing it with the background level activity.

### 6.2.4 Chromatographic technique

#### 6.2.4.1 Ion exchange resin

This technique is based on Sr(II) exchange on a cationic resin and is used for separation and purification of strontium in large volume samples. A method is presented in [Annex D](#) in which the measurement is carried out with a PC.

#### 6.2.4.2 Crown ether resin

This technique is based on the selective chromatographic separation of strontium using a specific crown ether resin. A method is presented in [Annex E](#) in which the measurement is carried out by LSC.

## 6.3 Preparation of the source for test

### 6.3.1 Source preparation for liquid scintillation counter

The strontium or yttrium precipitate is dissolved and mixed with a liquid scintillation cocktail. When the strontium or yttrium is already in solution, it is mixed directly with the liquid scintillator. The volume of the analyzed aliquot depends on the equipment (vial size) and the specific scintillation cocktail used.

The calibration source shall be prepared from a known activity of tracer ( $^{90}\text{Sr}$ ,  $^{89}\text{Sr}$ ,  $^{90}\text{Sr} + ^{90}\text{Y}$  or  $^{90}\text{Y}$ ) with the same geometry and chemical composition as the source to be measured.

The blank source should be prepared following the method chosen starting with a clean test sample (or water).

### 6.3.2 Source preparation for proportional counter

A strontium or yttrium precipitate is deposited on a filter by filtration or on a stainless steel planchet by direct evaporation.

The filter or planchet size diameter should be similar to the detector size (see [Annex A](#) or [D](#)).

A calibration source shall be prepared from a known amount of tracer ( $^{90}\text{Sr}$ ,  $^{89}\text{Sr}$ ,  $^{90}\text{Sr} + ^{90}\text{Y}$  or  $^{90}\text{Y}$ ) with the same geometry and chemical composition as the source to be measured.