



SLOVENSKI STANDARD SIST EN ISO 13160:2021

01-oktober-2021

Nadomešča:

SIST EN ISO 13160:2016

Kakovost vode - Stroncij Sr-90 in stroncij Sr-89 - Preskusne metode s štetjem s tekočinskim scintilatorjem ali proporcionalnim štetjem (ISO 13160:2021)

Water quality - Strontium 90 and strontium 89 - Test methods using liquid scintillation counting or proportional counting (ISO 13160:2021)

Wasserbeschaffenheit - Strontium-90 und Strontium-89 - Verfahren mittels Flüssigszintillationszählung oder Proportionalzählung (ISO 13160:2021)

Qualité de l'eau - Strontium 90 et strontium 89 - Méthodes d'essai par comptage des scintillations en milieu liquide ou par comptage proportionnel (ISO 13160:2021)

Ta slovenski standard je istoveten z: EN ISO 13160:2021

ICS:

13.060.60	Preiskava fizikalnih lastnosti vode	Examination of physical properties of water
17.240	Merjenje sevanja	Radiation measurements

SIST EN ISO 13160:2021

en,fr,de

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 13160:2021](#)

<https://standards.iteh.ai/catalog/standards/sist/446b89c9-ba07-4542-ab52-76a5118b7880/sist-en-iso-13160-2021>

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN ISO 13160

July 2021

ICS 13.060.60; 17.240

Supersedes EN ISO 13160:2015

English Version

**Water quality - Strontium 90 and strontium 89 - Test
methods using liquid scintillation counting or proportional
counting (ISO 13160:2021)**

Qualité de l'eau - Strontium 90 et strontium 89 -
Méthodes d'essai par comptage des scintillations en
milieu liquide ou par comptage proportionnel (ISO
13160:2021)

Wasserbeschaffenheit - Strontium-90 und Strontium-
89 - Verfahren mittels Flüssigszintillationszählung
oder Proportionalzählung (ISO 13160:2021)

This European Standard was approved by CEN on 10 June 2021.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents	Page
European foreword.....	3

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 13160:2021](https://standards.iteh.ai/catalog/standards/sist/446b89c9-ba07-4542-ab52-76a5118b7880/sist-en-iso-13160-2021)
<https://standards.iteh.ai/catalog/standards/sist/446b89c9-ba07-4542-ab52-76a5118b7880/sist-en-iso-13160-2021>

European foreword

This document (EN ISO 13160:2021) has been prepared by Technical Committee ISO/TC 147 "Water quality" in collaboration with Technical Committee CEN/TC 230 "Water analysis" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by January 2022, and conflicting national standards shall be withdrawn at the latest by January 2022.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 13160:2015.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN websites.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

[https://standards.iteh.ai/catalog/standards/sist/446b89c9-ba07-4542-ab52-](https://standards.iteh.ai/catalog/standards/sist/446b89c9-ba07-4542-ab52-76a5118b7880/sist-en-iso-13160-2021)

[76a5118b7880/sist-en-iso-13160-2021](https://standards.iteh.ai/catalog/standards/sist/446b89c9-ba07-4542-ab52-76a5118b7880/sist-en-iso-13160-2021)

The text of ISO 13160:2021 has been approved by CEN as EN ISO 13160:2021 without any modification.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 13160:2021](#)

<https://standards.iteh.ai/catalog/standards/sist/446b89c9-ba07-4542-ab52-76a5118b7880/sist-en-iso-13160-2021>

INTERNATIONAL
STANDARD

ISO
13160

Second edition
2021-07

**Water quality — Strontium 90 and
strontium 89 — Test methods using
liquid scintillation counting or
proportional counting**

*Qualité de l'eau — Strontium 90 et strontium 89 — Méthodes d'essai
par comptage des scintillations en milieu liquide ou par comptage
proportionnel*

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 13160:2021](https://standards.iteh.ai/catalog/standards/sist/446b89c9-ba07-4542-ab52-76a5118b7880/sist-en-iso-13160-2021)

<https://standards.iteh.ai/catalog/standards/sist/446b89c9-ba07-4542-ab52-76a5118b7880/sist-en-iso-13160-2021>



Reference number
ISO 13160:2021(E)

© ISO 2021

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 13160:2021](https://standards.iteh.ai/catalog/standards/sist/446b89c9-ba07-4542-ab52-76a5118b7880/sist-en-iso-13160-2021)

<https://standards.iteh.ai/catalog/standards/sist/446b89c9-ba07-4542-ab52-76a5118b7880/sist-en-iso-13160-2021>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2021

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Principle	2
4.1 General.....	2
4.2 Chemical separation.....	3
4.3 Detection.....	3
5 Chemical reagents and equipment	3
6 Procedure	4
6.1 Test sample preparation.....	4
6.2 Chemical separation.....	4
6.2.1 General.....	4
6.2.2 Precipitation techniques.....	5
6.2.3 Liquid-liquid extraction technique.....	6
6.2.4 Chromatographic technique.....	6
6.3 Preparation of the source for test.....	6
6.3.1 Source preparation for liquid scintillation counter.....	6
6.3.2 Source preparation for proportional counter.....	6
6.4 Measurement.....	7
6.4.1 General.....	7
6.4.2 Liquid scintillation counter.....	7
6.4.3 Proportional counter.....	7
6.4.4 Efficiency calculation.....	8
6.4.5 Determination of the chemical yield.....	8
7 Expression of results	9
7.1 Determination of ^{90}Sr in equilibrium with ^{90}Y	9
7.1.1 Calculation of the activity concentration.....	9
7.1.2 Standard uncertainty.....	9
7.1.3 Decision threshold.....	10
7.1.4 Detection limit.....	10
7.2 Determination of ^{90}Sr from separated ^{90}Y	10
7.2.1 Calculation of the activity concentration.....	10
7.2.2 Standard uncertainty.....	11
7.2.3 Decision threshold.....	12
7.2.4 Detection limit.....	12
7.3 Determination of ^{90}Sr and ^{89}Sr utilizing $^{90}\text{Sr}/^{90}\text{Y}$ equilibrium.....	12
7.3.1 Calculation of the activity concentration.....	12
7.3.2 Standard uncertainty.....	13
7.3.3 Decision threshold.....	14
7.3.4 Detection limit.....	15
8 Limits of the coverage intervals	15
8.1 Limits of the of the probabilistically symmetric coverage interval.....	15
8.2 Limits of the shortest coverage interval.....	16
9 Quality control	16
10 Test report	16
Annex A (informative) Determination of ^{89}Sr and ^{90}Sr by precipitation and proportional counting	18

ISO 13160:2021(E)

Annex B (informative) Determination of ^{89}Sr and ^{90}Sr by precipitation and liquid scintillation counting	22
Annex C (informative) Determination of ^{90}Sr from its decay progeny ^{90}Y at equilibrium by organic extraction and liquid scintillation counting	26
Annex D (informative) Determination of ^{90}Sr after ionic exchange separation by proportional counting	29
Annex E (informative) Determination of ^{90}Sr after separation on a crown ether specific resin and liquid scintillation counting	32
Annex F (informative) Determination of ^{90}Sr from its decay progeny ^{90}Y at equilibrium by organic extraction and proportional counting	34
Annex G (informative) Correction factor for ^{90}Sr purity using proportional counting	38
Bibliography	41

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 13160:2021](https://standards.iteh.ai/catalog/standards/sist/446b89c9-ba07-4542-ab52-76a5118b7880/sist-en-iso-13160-2021)

<https://standards.iteh.ai/catalog/standards/sist/446b89c9-ba07-4542-ab52-76a5118b7880/sist-en-iso-13160-2021>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 147, *Water quality*, Subcommittee SC 3, *Radioactivity measurements*.

This second edition cancels and replaces the first edition (ISO 13160:2012), which has been technically revised. The main changes compared to the previous edition are as follows:

- The way standard the uncertainty, decision threshold and detection limit are calculated has been updated in conformance with ISO 11929-1:2019.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

ISO 13160:2021(E)

Introduction

Radioactivity from several naturally-occurring and anthropogenic sources is present throughout the environment. Thus, water bodies (e.g. surface waters, ground waters, sea waters) can contain radionuclides of natural, human-made, or both origins.

- Natural radionuclides, including ^{40}K , ^3H , ^{14}C , and those originating from the thorium and uranium decay series, in particular ^{226}Ra , ^{228}Ra , ^{234}U , ^{238}U , and ^{210}Pb , can be found in water for natural reasons (e.g. desorption from the soil and washoff by rain water) or can be released from technological processes involving naturally occurring radioactive materials (e.g. the mining and processing of mineral sands or phosphate fertilizer production and use).
- Human-made radionuclides, such as transuranium elements (americium, plutonium, neptunium, curium), ^3H , ^{14}C , ^{90}Sr , and gamma emitting radionuclides can also be found in natural waters. Small quantities of these radionuclides are discharged from nuclear fuel cycle facilities into the environment as the result of authorized routine releases. Some of these radionuclides used for medical and industrial applications are also released into the environment after use. Anthropogenic radionuclides are also found in waters as a result of past fallout contaminations resulting from the explosion in the atmosphere of nuclear devices and accidents such as those that occurred in Chernobyl and Fukushima.

Radionuclide activity concentration in water bodies can vary according to local geological characteristics and climatic conditions and can be locally and temporally enhanced by releases from nuclear installation during planned, existing and emergency exposure situations^[1]. Drinking water may thus contain radionuclides at activity concentrations which could present a risk to human health.

The radionuclides present in liquid effluents are usually controlled before being discharged into the environment^[2]. Water bodies and drinking waters are monitored for their radioactivity content as recommended by the World Health Organization (WHO)^[3] so that proper actions can be taken to ensure that there is no adverse health effect to the public. Following these international recommendations, national regulations usually specify radionuclide authorized concentration limits for liquid effluent discharged to the environment and radionuclide guidance levels for water bodies and drinking waters for planned, existing and emergency exposure situations. Compliance with these limits can be assessed using measurement results with their associated uncertainties as specified by ISO/IEC Guide 98-3 and ISO 5667-20^[4].

Depending on the exposure situation, there are different limits and guidance levels that would result in an action to reduce health risk. As an example, during a planned or existing situation, the WHO guidelines for guidance level in drinking water is $100 \text{ Bq}\cdot\text{l}^{-1}$ for ^{89}Sr activity concentration and $10 \text{ Bq}\cdot\text{l}^{-1}$ for ^{90}Sr activity concentration^[3].

NOTE 1 The guidance level is the activity concentration with an intake of 2 l d^{-1} of drinking water for one year that results in an effective dose of $0,1 \text{ mSv a}^{-1}$ for members of the public. This is an effective dose that represents a very low level of risk and which is not expected to give rise to any detectable adverse health effects^[3].

In the event of a nuclear emergency, the WHO Codex guideline levels^[5] specifies that the activity concentration might not be greater than $1\,000 \text{ Bq}\cdot\text{kg}^{-1}$ for ^{89}Sr or $100 \text{ Bq}\cdot\text{kg}^{-1}$ for ^{90}Sr for infant food and $1\,000 \text{ Bq}\cdot\text{kg}^{-1}$ for ^{89}Sr or $100 \text{ Bq}\cdot\text{kg}^{-1}$ for ^{90}Sr for food other than infant food.

NOTE 2 The Codex guidelines levels (GLs) apply to radionuclides contained in food destined for human consumption and traded internationally, which have been contaminated following a nuclear or radiological emergency. These GLs apply to food after reconstitution or as prepared for consumption, i.e. not to dried or concentrated food, and are based on an intervention exemption level of 1 mSv in a year for members of the public (infant and adult)^[5].

Thus, the test method can be adapted so that the characteristic limits, decision threshold, detection limit and uncertainties ensure that the radionuclide activity concentrations test results can be verified to be below the guidance levels required by a national authority for either planned/existing situations or for an emergency situation^{[6][7]}.

Usually, the test methods can be adjusted to measure the activity concentration of the radionuclide(s) in either wastewaters before storage or in liquid effluents before discharge to the environment. The test results will enable the plant/installation operator to verify that, before their discharge, wastewaters/liquid effluent radioactive activity concentrations do not exceed authorized limits.

The test method described in this document may be used during planned, existing and emergency exposure situations as well as for wastewaters and liquid effluents with specific modifications that could increase the overall uncertainty, detection limit, and threshold.

The test method may be used for water samples after proper sampling, sample handling, and test sample preparation (see the relevant part of the ISO 5667 series [8],[9],[10]).

This document has been developed to answer the need of test laboratories carrying out these measurements, that are sometimes required by national authorities, as they may have to obtain a specific accreditation for radionuclide measurement in drinking water samples.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 13160:2021](https://standards.iteh.ai/catalog/standards/sist/446b89c9-ba07-4542-ab52-76a5118b7880/sist-en-iso-13160-2021)

<https://standards.iteh.ai/catalog/standards/sist/446b89c9-ba07-4542-ab52-76a5118b7880/sist-en-iso-13160-2021>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 13160:2021](#)

<https://standards.iteh.ai/catalog/standards/sist/446b89c9-ba07-4542-ab52-76a5118b7880/sist-en-iso-13160-2021>

Water quality — Strontium 90 and strontium 89 — Test methods using liquid scintillation counting or proportional counting

WARNING — Persons using this document should be familiar with normal laboratory practice. This document does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to determine the applicability of any other restrictions.

IMPORTANT — It is absolutely essential that tests conducted according to this document be carried out by suitably trained staff.

1 Scope

This document specifies conditions for the determination of ^{90}Sr and ^{89}Sr activity concentration in samples of environmental water using liquid scintillation counting (LSC) or proportional counting (PC).

The method is applicable to test samples of drinking water, rainwater, surface and ground water, marine water, as well as cooling water, industrial water, domestic, and industrial wastewater after proper sampling and handling, and test sample preparation. Filtration of the test sample and a chemical separation are required to separate and purify strontium from a test portion of the sample.

The detection limit depends on the sample volume, the instrument used, the sample count time, the background count rate, the detection efficiency and the chemical yield. The method described in this document, using currently available LSC counters, has a detection limit of approximately 10 mBq l^{-1} and 2 mBq l^{-1} for ^{89}Sr and ^{90}Sr , respectively, which is lower than the WHO criteria for safe consumption of drinking water (100 Bq l^{-1} for ^{89}Sr and 10 Bq l^{-1} for ^{90}Sr). These values can be achieved with a counting time of 1 000 min for a sample volume of 2 l.

The methods described in this document are applicable in the event of an emergency situation. When fallout occurs following a nuclear accident, the contribution of ^{89}Sr to the total amount of radioactive strontium is not negligible. This document provides test methods to determine the activity concentration of ^{90}Sr in presence of ^{89}Sr .

The analysis of ^{90}Sr and ^{89}Sr adsorbed to suspended matter is not covered by this method.

It is the user's responsibility to ensure the validity of this test method selected for the water samples tested.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements for this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO/IEC 17025, *General requirements for the competence of testing and calibration laboratories*

ISO 11929-1, *Determination of the characteristic limits (decision threshold, detection limit and limits of the coverage interval) for measurements of ionizing radiation — Fundamentals and application — Part 1: Elementary applications*

ISO/IEC Guide 98-3, *Uncertainty of measurement — Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

ISO 80000-10, *Quantities and units — Part 10: Atomic and nuclear physics*