

---

---

**Soil quality — Determination of  
particle size distribution in mineral  
soil material — Method by sieving and  
sedimentation**

*Qualité du sol — Détermination de la répartition granulométrique  
de la matière minérale des sols — Méthode par tamisage et  
sédimentation*

iteh standards  
(<https://standards.iteh.ai>)  
Document Preview

[ISO 11277:2020](https://standards.iteh.ai/catalog/standards/iso/8e02daa7-1a38-4ebc-aed9-9d492baddb0d/iso-11277-2020)

<https://standards.iteh.ai/catalog/standards/iso/8e02daa7-1a38-4ebc-aed9-9d492baddb0d/iso-11277-2020>



iTeh Standards  
(<https://standards.iteh.ai>)  
Document Preview

[ISO 11277:2020](https://standards.iteh.ai/catalog/standards/iso/8e02daa7-1a38-4ebe-aed9-9d492baddb0d/iso-11277-2020)

<https://standards.iteh.ai/catalog/standards/iso/8e02daa7-1a38-4ebe-aed9-9d492baddb0d/iso-11277-2020>



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Fax: +41 22 749 09 47  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

	Page
Foreword.....	iv
Introduction.....	v
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms and definitions</b> .....	<b>1</b>
<b>4 Symbols</b> .....	<b>2</b>
<b>5 Principle</b> .....	<b>2</b>
<b>6 Field sampling</b> .....	<b>3</b>
<b>7 Sample preparation</b> .....	<b>4</b>
<b>8 Dry sieving (material &gt;2 mm)</b> .....	<b>4</b>
8.1 General.....	4
8.2 Apparatus.....	4
8.3 Procedure.....	5
8.4 Calculation and expression of results.....	6
<b>9 Wet sieving and sedimentation (material &lt;2 mm)</b> .....	<b>6</b>
9.1 General.....	6
9.2 Apparatus.....	6
9.3 Reagents.....	15
9.4 Calibrations.....	16
9.4.1 Sampling pipette (see <a href="#">Figure 4</a> ).....	16
9.4.2 Dispersing-agent correction.....	16
9.5 Test sample.....	16
9.6 Destruction of organic matter.....	17
9.6.1 General.....	17
9.6.2 Method A.....	18
9.6.3 Method B.....	18
9.7 Removal of soluble salts and gypsum.....	19
9.8 Removal of carbonates.....	19
9.9 Removal of iron oxides.....	20
9.10 Dispersion.....	20
9.11 Wet sieving at 0,063 mm.....	20
9.12 Sedimentation.....	21
9.13 Calculation of results for fractions <2 mm.....	22
<b>10 Test report</b> .....	<b>23</b>
<b>Annex A (normative) Determination of particle size distribution of mineral soil material that is not dried prior to analysis</b> .....	<b>24</b>
<b>Annex B (normative) Determination of particle size distribution of mineral soils by a hydrometer method following destruction of organic matter</b> .....	<b>27</b>
<b>Annex C (informative) Precision of the method</b> .....	<b>36</b>
<b>Bibliography</b> .....	<b>38</b>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 190, *Soil quality*, Subcommittee SC 3, *Chemical and physical characterization*.

This third edition cancels and replaces the second edition (ISO 11277:2009), which has been technically revised. The main changes compared to the previous edition are as follows:

- Alternative digestion methods were added;
- A practical order of preparation steps was added;
- References were updated;
- Document has been editorially revised.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

## Introduction

The physical and chemical behaviour of soils is controlled in part by the amounts of mineral particles of different sizes in the soil. The subject of this document is the quantitative measurement of such amounts (expressed as a proportion or percentage of the total mass of the mineral soil), within stated size classes.

The determination of particle size distribution is affected by organic matter, soluble salts, cementing agents (like iron compounds), relatively insoluble substances such as carbonates and sulfates, or combinations of these. Some soils change their behaviour to such a degree, upon drying, that the particle size distribution of the dried material bears little or no relation to that of the undried material encountered under natural conditions. This is particularly true of soils rich in organic matter, those developed from recent volcanic deposits, some highly weathered tropical soils, and soils often described as “cohesive” (see Reference [4]). Other soils, such as the so-called “sub-plastic” soils of Australia, show little or no tendency to disperse under normal laboratory treatments, despite field evidence of large clay content.

The procedures given in this document recognize these kinds of differences between soils from different environments, and the methodology presented is designed to deal with them in a structured manner. Such differences in soil behaviour can be very important, but awareness of them depends usually on local knowledge. Given that the laboratory is commonly distant from the site of the field operation, the information supplied by field teams becomes crucial to the choice of an appropriate laboratory procedure. This choice can be made only if the laboratory is made fully aware of this background information.

iTeh Standards  
(<https://standards.itih.ai>)  
Document Preview

[ISO 11277:2020](https://standards.itih.ai/catalog/standards/iso/8e02daa7-1a38-4ebc-aed9-9d492baddb0d/iso-11277-2020)

<https://standards.itih.ai/catalog/standards/iso/8e02daa7-1a38-4ebc-aed9-9d492baddb0d/iso-11277-2020>



# Soil quality — Determination of particle size distribution in mineral soil material — Method by sieving and sedimentation

**WARNING** — Persons using this document should be familiar with usual laboratory practice. This document does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to ensure compliance with any national regulatory conditions.

**IMPORTANT** — It is absolutely essential that tests, conducted in accordance with this document, be carried out by suitably qualified staff.

## 1 Scope

This document specifies a basic method of determining the particle size distribution applicable to a wide range of mineral soil materials, including the mineral fraction of organic soils. It also offers procedures to deal with the less common soils mentioned in the introduction. This document has been developed largely for use in the field of environmental science, and its use in geotechnical investigations is something for which professional advice might be required.

A major objective of this document is the determination of enough size fractions to enable the construction of a reliable particle-size-distribution curve.

This document does not apply to the determination of the particle size distribution of the organic components of soil, i.e. the more or less fragile, partially decomposed, remains of plants and animals. It is also realized that the chemical pre-treatments and mechanical handling stages in this document could cause disintegration of weakly cohesive particles that, from field inspection, might be regarded as primary particles, even though such primary particles could be better described as aggregates. If such disintegration is undesirable, then this document is not used for the determination of the particle size distribution of such weakly cohesive materials.

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 565, *Test sieves — Metal wire cloth, perforated metal plate and electroformed sheet — Nominal sizes of openings*

ISO 3310-1, *Test sieves — Technical requirements and testing — Part 1: Test sieves of metal wire cloth*

ISO 3310-2, *Test sieves — Technical requirements and testing — Part 2: Test sieves of perforated metal plate*

ISO 11265, *Soil quality — Determination of the specific electrical conductivity*

ISO 11464, *Soil quality — Pretreatment of samples for physico-chemical analysis*

## 3 Terms and definitions

There are no normative references in this document.

## 4 Symbols

The following symbols are found throughout the text and, where appropriate, units and quantities are as given below (the SI convention is followed for common units, e.g. g = gram; m = metre; mm = millimetre; s = second, etc.).

Mg megagram ( $10^6$  g)

## 5 Principle

The particle size distribution is determined by a combination of sieving and sedimentation, starting from air-dried soil (see Reference [4]). A method for undried soil is given in Annex A. Particles not passing a 2 mm aperture sieve are determined by dry sieving. Particles passing such a sieve, but retained on a 0,063 mm aperture sieve, are determined by a combination of wet and dry sieving, whilst particles passing the latter sieve are determined by sedimentation.

The pipette method is preferred. As an alternative, the hydrometer method is given in Annex B. A combination of sieving and sedimentation enables the construction of a continuous particle-size-distribution curve.

The key points in this procedure are summarized as a flow chart in Figure 2. This document requires that the proportions of fractions separated by sedimentation and sieving be determined from the masses of such fractions obtained by weighing. Other methods of determining the mass of such fractions rely on such things as the interaction of particles with electromagnetic radiation or electrical fields (see Reference [2]). There are often considerable difficulties in relating the values obtained by these different methods for the same sample. It is one of the intentions of this document that close adherence to its details should help minimize interlaboratory variation in the determination of the particle size distribution of mineral soils. Therefore, the proportions of fractions shall be determined only by weighing. If this is not the method used, then conformance with this document cannot be claimed in the test report (see Clause 10).

Both the pipette and hydrometer methods assume that the settling of particles in the sedimentation cylinder is in accordance with Stokes's Law (see References [2],[4], and [7]), and the constraints that this implies, namely:

- a) the particles are rigid, smooth spheres;
- b) the particles settle in laminar flow, i.e. the Reynolds Number is less than about 0,2; this constraint sets an upper equivalent spherical particle diameter (see below) slightly greater than 0,06 mm for Stokesian settling under gravity (Reference [2]);
- c) the suspension of particles is sufficiently dilute to ensure that no particle interferes with the settling of any other particle;
- d) there is no interaction between the particle and fluid;
- e) the diameter of the suspension column is large compared to the diameter of the particle, i.e. the fluid is of "infinite extent";
- f) the particle has reached its terminal velocity;
- g) the particles are of the same relative density.

Thus, the diameter of a particle is defined in terms of the diameter of a sphere whose behaviour in suspension matches that of the particle. This is the concept of *equivalent spherical diameter*. It is the principle upon which the expression of the diameter of particles, as derived from sedimentation, is based in this document.



Stokes's Law can be written, for the purposes of this document, as in [Formula \(1\)](#):

$$t = 18\eta h / [(\rho_s - \rho_w) g d_p^2] \quad (1)$$

where

- $t$  is the settling time, in seconds, of a particle of diameter  $d_p$  (see below);
- $\eta$  is the dynamic viscosity of water at the test temperature (see [Table B.2](#)), in millipascals per second;
- $h$  is the sampling depth, in centimetres;
- $\rho_s$  is the mean particle density, in megagrams per cubic metre (taken as 2,65 Mg/m<sup>3</sup>; see note);
- $\rho_w$  is the density of the liquid containing the soil suspension, in megagrams per cubic metre (taken as 1,00 Mg/m<sup>3</sup>; see note);
- $g$  is the acceleration due to gravity, in centimetres per second squared (taken as 981 cm/s<sup>2</sup>);
- $d_p$  is the equivalent spherical diameter of the particle of interest, in millimetres.

NOTE 1 It is realized that there are considerable differences between the densities of soil particles, but for the purposes of this document it is assumed that the mean particle density is that of quartz, i.e. 2,65 Mg/m<sup>3</sup> (Reference [8]), as this is the commonest mineral in a very wide range of soils. The density of water is 0,998 2 Mg/m<sup>3</sup> and 0,995 6 Mg/m<sup>3</sup> at 20 °C and 30 °C, respectively (see Reference [6]). Given the effect of the addition of a small amount of dispersant (see [9.3.2](#)), the density of water is taken as 1,000 0 Mg/m<sup>3</sup> over the permitted temperature range of this document (see [9.2.2](#)).

Furthermore, for routine use, it is recommended that the sampling times be converted to minutes and/or hours, as appropriate, to lessen the risk of error (see [Table 3](#)).

Particles within particular size ranges or classes are commonly described as cobbles, gravel, coarse sand, silt, etc. The meaning of such trivial names differs between countries, and in some cases there are no exact translations of such words from one language to another; for example, the Dutch word “zavel” has no equivalent in English. The only fraction for which there appears to be common agreement is clay, which is defined as material of less than 0,002 mm equivalent spherical diameter (References [4]). Such trivial names shall not be used in describing the results of particle size determination according to this document. Phrases such as “... passing a 20 mm aperture sieve ...” or “... less than 0,063 mm equivalent spherical diameter ...” shall be used instead. If trivial names shall be used, for example, to cross-reference to another International or National Standard, then the trivial name should be defined explicitly, so as to remove any doubt as to the meaning intended, e.g. silt (0,063 mm to 0,002 mm equivalent spherical diameter) (see Clause 4). Furthermore, it is common to use the word “texture” to describe the results of particle-size-distribution measurements, e.g. “the particle size of this soil is of clay texture”. This is incorrect as the two concepts are different, and the word “texture” shall not be used in the test report (see Clause 10) to describe the results obtained by the use of this document.

## 6 Field sampling

The mass of sample taken in the field shall be representative of the particle size distribution, especially if the amount of the larger particles is to be determined reliably. [Table 1](#) gives recommended minimum masses.

Table 1 — Mass of soil sample to be taken for sieving

Maximum size of material forming > 10 % of the soil (given as test sieve aperture, in mm)	Minimum mass of sample to be taken for sieving kg
63	50
50	35
37,5	15
28	6
20	2
14	1
10	0,5
6,3	0,5
5	0,2
2 or smaller	0,1

## 7 Sample preparation

Samples shall be prepared in accordance with the methods given in ISO 11464.

NOTE For many purposes, particle size distribution is determined only for the fraction of the soil passing a 2 mm aperture sieve. In this case, the test sample (9.5) can be taken either according to the procedures in ISO 11464 or from the material passing a 2 mm aperture sieve according to 8.2.

## 8 Dry sieving (material >2 mm)

### 8.1 General

The procedure specified in this clause applies to material retained on a 2 mm aperture sieve. Table 2 gives the maximum mass which shall be retained on sieves of different diameters and apertures. If more than this amount of material is retained, then it shall be subdivided appropriately and sieved again.

It is common to refer to sieves as having a particular mesh-size or mesh number. These are not the same as the sieve aperture, and the relationship between the various numbers is not immediately obvious. The use of mesh numbers as a measurement of particle size is difficult to justify, and shall not be used in reporting the results of this document.

### 8.2 Apparatus

**8.2.1 Test sieves**, with apertures according to ISO 565, and with well-fitting covers and receivers.

The full range of sieves appropriate to the largest particle(s) present should be used (see Table 1 and 9.2.3). The apertures chosen shall be stated in the test report (Clause 10). The accuracy of the sieves shall be verified monthly against a set of master sieves kept for this purpose, using an accepted method such as particle reference materials, microscopy, etc. (see Reference [2] depending on the sieve aperture. Tolerances shall meet the requirements of ISO 3310-1 and ISO 3310-2. Sieves that do not meet these specifications shall be discarded. A record shall be kept of such testing.

Brass sieves are particularly liable to splitting and distortion, and steel sieves are strongly recommended for the larger apertures.

Special care shall be taken to ensure that covers and receivers do not leak. Sieves shall be inspected weekly when in regular use, and on every occasion if used less often. A record shall be kept of such inspections. Round-hole sieves shall not be used.

**8.2.2 Balance**, capable of weighing to an accuracy of within  $\pm 0,5$  g.

**8.2.3 Mechanical sieve shaker.**

It is usually impracticable to sieve mechanically at sieve apertures much greater than 20 mm, unless very heavy-duty equipment is available. Mechanical sieve shaking is essential to sieve efficiency at smaller apertures.

**8.2.4 A sieve brush and a stiff brush.**

**8.2.5 Ultrasonic bath.**

### 8.3 Procedure

Weigh the dry test sample, prepared in accordance with ISO 11464, to the nearest 0,5 g ( $m_1$ ). Place the weighed material on the 20 mm sieve, and by brushing the material gently over the sieve apertures with the stiff brush (to remove any adhering soil), sieve the material. Take care not to detach any fragments from the primary particles. Sieve the retained material on the nest of sieves of selected apertures (8.2.1) and record the amount retained on each sieve to the nearest 0,5 g. Do not overload the sieves (see Table 1) but sieve the material in portions if necessary.

Weigh the material passing the 20 mm aperture sieve ( $m_2$ ), or a suitable portion of it ( $m_3$ ) (see Table 2) obtained by an appropriate subsampling method (see Clause 6), and place this on a nest of sieves, the lowermost having an aperture of 2 mm. Shake the sieves mechanically until no further material passes any of the sieves (see Note). Record the mass of material retained on each sieve and the mass passing the 2 mm aperture sieve.

The total mass of the fractions should be within 1 % of  $m_2$  or  $m_3$ , as appropriate. If it is not, then check for sieve damage and discard sieves as appropriate (see 8.2.1).

NOTE For practical purposes, it is usual to choose a standard sieve shaking time which gives an acceptable degree of sieving efficiency with a wide range of soil materials. The minimum recommended period is 10 min.

**Table 2 — Maximum mass of material to be retained on each test sieve at the completion of sieving**

Test sieve aperture mm	Maximum mass kg				
	Sieve diameter mm				
	450	400	300	200	100
50	10	8,9	4,5	—	—
37,5	8	7,1	3,5	—	—
28	6	5,3	2,5	—	—
20	4	3,6	2	—	—
14	3	2,7	1,5	—	—
10	2	1,8	1	—	—
6,3	1,5	1,3	0,75	—	—
5	1	0,9	0,5	—	—
3,35	—	—	—	0,3	0,15
2	—	—	—	0,2	0,1
1,18	—	—	—	0,1	0,05
0,6	—	—	—	0,075	0,0375

**Table 2** (continued)

Test sieve aperture	Maximum mass				
	kg				
	Sieve diameter				
mm	450	400	300	200	100
0,425	—		—	0,075	0,0375
0,3	—		—	0,05	0,025
0,212	—		—	0,05	0,025
0,15	—		—	0,04	0,02
0,063	—		—	0,025	0,0125

**8.4 Calculation and expression of results**

For the material retained by the 20 mm and larger aperture sieves, calculate the proportion by mass retained by each sieve as a proportion of  $m_1$ . For example (Formula (2)):

$$\text{Proportion retained on the 20 mm sieve} = [m(20\text{mm})] / m_1 \tag{2}$$

For the material passing the 20 mm sieve, multiply the mass of material passing each sieve by  $m_2/m_3$  and calculate this as a proportion of  $m_1$ . For example (Formula (3)):

$$\text{Proportion retained on the 6,3mm sieve} = m(6,3\text{mm}) [(m_2 / m_3) / m_1] \tag{3}$$

Present the results as a Table showing, to two significant figures, the proportion by mass retained on each sieve and the proportion passing the 2 mm sieve. The data shall also be used to construct a cumulative distribution curve (see Figure 1).

**9 Wet sieving and sedimentation (material < 2 mm)**

**9.1 General**

This clause specifies the procedure (see Figure 2) for the determination of the particle size distribution of the material passing the 2 mm aperture sieve down to <0,002 mm equivalent spherical diameter (see note). In order to ensure that primary particles, rather than loosely bonded aggregates, are measured, organic matter and salts are removed, especially sparingly soluble salts such as gypsum which would otherwise prevent dispersion and/or promote flocculation of the finer soil particles in suspension (see 9.6), and a dispersing agent is added (9.3.2). These procedures are required in this document, and their omission shall invalidate its application. Sometimes iron oxides and carbonates, especially of calcium and/or magnesium, are also removed. Preferred procedures for the removal of these compounds are given in the note in 9.7. The removal of any compound shall be recorded in the test report (see Clause 10).

NOTE 1 Gravitational sedimentation can give a value for the total amount of material <0,002 mm equivalent spherical diameter. However, the method cannot be used to divide this class further with reliability, as particles less than about 0,001 mm equivalent spherical diameter can be kept in suspension almost indefinitely by Brownian motion (see Reference [2]).

**9.2 Apparatus**

The apparatus specified hereafter is sufficient to deal with one sample. Clearly it is more efficient to work in batches. Experience has shown (see Reference [7]) that one operator can process up to 36 samples in a batch at a time, given sufficient apparatus and space, especially if calculations are dealt with by a computer.