
**Molecular in vitro diagnostic
examinations — Specifications for pre-
examination processes for formalin-
fixed and paraffin-embedded (FFPE)
tissue —**

**Part 3:
Isolated DNA**

*Analyses de diagnostic moléculaire in vitro — Spécifications relatives
aux processus préanalytiques pour les tissus fixés au formol et inclus
en paraffine (FFPE) —*

Partie 3: ADN extrait

[ISO 20166-3:2018](https://standards.iteh.ai/catalog/standards/iso/1b80cc28-d488-4517-b997-3ae9709d2c9a/iso-20166-3-2018)

<https://standards.iteh.ai/catalog/standards/iso/1b80cc28-d488-4517-b997-3ae9709d2c9a/iso-20166-3-2018>



iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO 20166-3:2018](https://standards.iteh.ai/catalog/standards/iso/1b80cc28-d488-4517-b997-3ae9709d2c9a/iso-20166-3-2018)

<https://standards.iteh.ai/catalog/standards/iso/1b80cc28-d488-4517-b997-3ae9709d2c9a/iso-20166-3-2018>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2018

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 General considerations	4
5 Outside the laboratory	5
5.1 Specimen collection.....	5
5.1.1 General.....	5
5.1.2 Information about the specimen donor/patient.....	5
5.1.3 Information about the specimen.....	5
5.1.4 Specimen processing.....	6
5.2 Transport requirements.....	6
6 Inside the laboratory	7
6.1 Information about the reception of the specimen.....	7
6.2 Formalin fixation of the specimen or sample(s).....	7
6.3 Evaluation of the pathology of the specimen and selection of the sample(s).....	8
6.4 Post-fixation of frozen samples.....	9
6.5 Decalcification.....	9
6.6 Processing and paraffin embedding.....	9
6.7 Storage requirements.....	10
6.8 Isolation of DNA.....	10
6.8.1 General.....	10
6.8.2 General information for DNA isolation procedures.....	10
6.8.3 Using commercial kits.....	11
6.8.4 Using the laboratories' own protocols.....	11
6.9 Quantity and quality assessment of isolated DNA.....	12
6.10 Storage of isolated DNA.....	12
Annex A (informative) Impact of the storage temperature on DNA Integrity in FFPE blocks of tissue	14
Bibliography	16

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared by Technical ISO/TC 212, *Clinical laboratory testing and in vitro diagnostic test systems*.

A list of all parts in the ISO 20166 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

Molecular in vitro diagnostics, including molecular pathology, has enabled significant progress in medicine. Further progress is expected with new technologies analysing nucleic acids, proteins, and metabolites in human tissues and body fluids. However, the profiles and/or integrity of these molecules can change drastically during specimen collection, transport, storage and processing, thus making the outcome from diagnostics or research unreliable or even impossible because the subsequent examination assay will not determine the situation in the patient but an artificial molecular pattern generated during the pre-examination process. Studies have been undertaken to determine the influencing factors for the DNA examination from formalin-fixed and paraffin-embedded (FFPE) tissue. These studies demonstrated that a standardization of the entire process from specimen collection to the DNA examination is needed. This document draws upon such work to codify and standardize the steps for FFPE tissue with regard to DNA examination in what is referred to as the pre-examination phase.

DNA integrity in tissues can change before, during and after formalin fixation, processing and storage. Chemical modifications introduced into DNA during tissue fixation might lead to fragmentation and sequence alterations, changes in the methylation status or even structural changes which can lead to, for instance, spurious copy number changes in array-CGH profiles. These modifications of the DNA molecules can impact the validity and reliability of the examination test results. Therefore, it is essential to take special measures to minimize the described DNA changes and modifications for subsequent examination.

In this document, the following verbal forms are used:

- "shall" indicates a requirement;
- "should" indicates a recommendation;
- "may" indicates a permission;
- "can" indicates a possibility or a capability.

[ISO 20166-3:2018](https://standards.iteh.ai/catalog/standards/iso/1b80cc28-d488-4517-b997-3ae9709d2c9a/iso-20166-3-2018)

<https://standards.iteh.ai/catalog/standards/iso/1b80cc28-d488-4517-b997-3ae9709d2c9a/iso-20166-3-2018>

Molecular in vitro diagnostic examinations — Specifications for pre-examination processes for formalin- fixed and paraffin-embedded (FFPE) tissue —

Part 3: Isolated DNA

1 Scope

This document gives guidelines on the handling, documentation, storage and processing of formalin-fixed and paraffin-embedded (FFPE) tissue specimens intended for DNA examination during the pre-examination phase before a molecular assay is performed.

This document is applicable to molecular in vitro diagnostic examinations including laboratory developed tests performed by medical laboratories and molecular pathology laboratories. It is also intended to be used by laboratory customers, in vitro diagnostics developers and manufacturers, biobanks, institutions and commercial organizations performing biomedical research, and regulatory authorities.

NOTE International, national or regional regulations or requirements can also apply to specific topics covered in this document.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 15189:2012, *Medical laboratories — Requirements for quality and competence*

ISO 15190, *Medical laboratories — Requirements for safety*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 15189 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

3.1

aliquot

portion of a larger amount of homogeneous material, assumed to be taken with negligible sampling error

Note 1 to entry: The term is usually applied to fluids. Tissues are heterogeneous and therefore cannot be aliquoted.

Note 2 to entry: The definition is derived from References [25], [26], and [27].

**3.2
ambient temperature**

unregulated temperature of the surrounding air

**3.3
analyte**

component represented in the name of a measurable quantity

[SOURCE: ISO 17511:2003, 3.2, modified — EXAMPLE has been removed.]

**3.4
analytical test performance**

accuracy, precision, specificity and sensitivity of a test to measure the *analyte* (3.3) of interest

Note 1 to entry: Other test performance characteristics such as robustness, repeatability can apply as well.

**3.5
cold ischemia**

condition after removal of the tissue from the body until stabilization or fixation

**3.6
diagnosis**

identification of a health or disease state from its signs and/or symptoms, where the diagnostic process can involve *examinations* (3.8) and tests for classification of an individual's condition into separate and distinct categories or subclasses that allow medical decisions about treatment and prognosis to be made

**3.7
DNA**

deoxyribonucleic acid polymer of deoxyribonucleotides occurring in a double-stranded (dsDNA) or single-stranded (ssDNA) form

[SOURCE: ISO 22174:2005, 3.1.2]

**3.8
examination
analytical test**

set of operations having the object of determining the value or characteristics of a property

Note 1 to entry: Processes that start with the isolated analyte and include all kinds of parameter testing or chemical manipulation for quantitative or qualitative examination.

[SOURCE: ISO 15189:2012, 3.7, modified — Notes to entry 1 to 3 have been removed, Note 1 to entry has been added and “analytical test” has been added as a preferred term.]

**3.9
formalin**

saturated aqueous formaldehyde solution which at 100 % contains 37 % formaldehyde by mass (corresponding to 40 % by volume)

**3.10
formalin fixation**

treatment of a *sample* (3.18) with *standard buffered formalin solution* (3.20) for stabilization

**3.11
grossing**

gross examination

inspection of pathology specimens with the bare eye to obtain diagnostic information, while being processed for further microscopic examination

3.12**interfering substances**

endogenous substances of a *specimen* (3.15)/*sample* (3.18) or exogenous substances (e.g. stabilization solution) that can alter an examination result

3.13**paraffin embedding**

process in which a tissue sample is placed in paraffin to generate a hard surrounding matrix so that thin microscopic sections can be cut

3.14**pre-examination process**

pre-analytical phase

pre-analytical workflow

process that starts, in chronological order, from the clinician's request and includes the examination request, preparation and identification of the patient, collection of the *primary sample(s)* (3.15), transportation to and within the medical or pathology laboratory, isolation of *analytes* (3.3), and ends when the analytical examination begins

Note 1 to entry: The pre-examination phase includes preparative processes that influence the outcome of the intended examination.

[SOURCE: ISO 15189:2012, 3.15, modified — “pre-analytical workflow” has been added as a preferred term, Note 1 to entry has been added and the definition has been extended.]

3.15**primary sample specimen**

discrete portion of a body fluid, breath, hair or tissue taken for *examination* (3.8), study or analysis of one or more quantities or properties assumed to apply for the whole

[SOURCE: ISO 15189:2012, 3.16, modified — Notes to entry 1 to 3 have been removed.]

3.16**proficiency test**

evaluation of participant performance against pre-established criteria by means of inter-laboratory comparisons

[SOURCE: ISO 17043:2010, 3.7, modified — Notes to entry 1 and 2 have been removed.]

3.17**room temperature**

for the purposes of this document, temperature in the range of 18 °C to 25 °C

Note 1 to entry: Local or national regulations can have different definitions.

3.18**sample**

one or more parts taken from a *primary sample* (3.15)

[SOURCE: ISO 15189:2012, 3.24, modified — EXAMPLE has been removed.]

3.19**stability**

ability of a sample material, when stored under specified conditions, to maintain a stated property value within specified limits for a specified period of time

Note 1 to entry: The analyte for the purpose of this document is isolated DNA.

[SOURCE: ISO Guide 30:2015, 2.1.15, modified — “reference material” has been replaced by “sample material” and Note 1 to entry has been changed.]

3.20

standard buffered formalin solution neutral buffered formalin

NBF

10 % *formalin* (3.9) solution in water with a mass fraction of 3,7 % (corresponding to a volume fraction of 4 %) formaldehyde, buffered to pH 6,8 to pH 7,2

Note 1 to entry: Standard buffered formalin solutions often contain small amounts of methanol to inhibit oxidation and polymerisation of formaldehyde.

3.21

storage

prolonged interruption of the *pre-analytical workflow* (3.14) of a *sample* (3.18) or *analyte* (3.3) respectively, or of their derivatives, such as stained sections or tissue blocks, under appropriate conditions in order to preserve their properties

Note 1 to entry: Long-term storage typically occurs in laboratory archives or in biobanks.

3.22

tissue processor

automated instrument where tissue fixation, dehydration, clearing and paraffin infiltration occurs

3.23

validation

confirmation, throughout the provision of objective evidence, that the requirements for a specific intended use or application have been fulfilled

Note 1 to entry: "Validated" is used to designate the corresponding status.

[SOURCE: ISO 9000:2015, 3.8.13, modified — Notes to entry 1 and 3 have been removed.]

3.24

warm ischemia

condition before the tissue is removed from the body, but where it is deprived of its normal blood supply

3.25

workflow

series of activities necessary to complete a task

3.26

homogeneous

uniform in structure and composition

4 General considerations

For general statements on medical laboratory quality management systems and in particular on specimen collection and handling (including avoidance of cross contaminations) see ISO 15189:2012, 4.2, 5.4.4, 5.4.6 or ISO/IEC 17020:2012, clause 8 and 7.2. The requirements for laboratory equipment, reagents, and consumables in accordance with ISO 15189:2012, 5.3 shall be followed; ISO 15189:2012, 5.5.1.2 and 5.5.1.3, and ISO/IEC 17020:2012, 6.2 can also apply.

All steps of a diagnostic workflow can influence the final analytical test result. Thus, the entire workflow including biomolecule stability and sample storage conditions shall be verified and validated. Workflow steps which cannot always be controlled (e.g. warm ischemia) shall be documented. A risk assessment of non-controllable workflow steps including their potential impact on the analytical test performance shall be performed and mitigation measures shall be established to enable the required analytical test performance.

In contrast to RNA or proteins, DNA in tissue is relatively stable during warm and cold ischemia. Changes of DNA sequence or copy numbers [e.g. comparative genomic hybridization (CGH) profiles] due