

First edition
2009-03-15

AMENDMENT 1
2019-01

Implants for surgery — Wear of total hip-joint prostheses —

Part 3:

**Loading and displacement parameters
for orbital bearing type wear testing
machines and corresponding
environmental conditions for test**

AMENDMENT 1

Implants chirurgicaux — Usure des prothèses totales de l'articulation de la hanche —

Partie 3: Paramètres de charge et de déplacement pour machines d'essai d'usure du type orbital de maintien et conditions environnementales correspondantes d'essai

AMENDEMENT 1



Reference number
ISO 14242-3:2009/Amd.1:2019(E)

© ISO 2019

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

ISO 14242-3:2009/Amd 1:2019

<https://standards.iteh.ai/catalog/standards/iso/955356c6-43f0-4efe-86db-a90f80ff1c91/iso-14242-3-2009-amd-1-2019>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2019

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 150, *Implants for surgery*, Subcommittee SC 4, *Bone and joint replacements*.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

<https://standards.iteh.ai/catalog/standards/iso/955356c6-43f0-4efe-86db-a90f80ff1c91/iso-14242-3-2009-amd-1-2019>

Implants for surgery — Wear of total hip-joint prostheses —

Part 3:

Loading and displacement parameters for orbital bearing type wear testing machines and corresponding environmental conditions for test

AMENDMENT 1

2 Normative references

Add the following at the end of the Normative references (referenced in 5.1):

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

4 Principle

Replace “in position in their normal configuration” in the first sentence with the following:

in either an “inverted” position, that is, head above liner [Figure 4 a)]; or an “anatomical” position, that is, head below liner [Figure 4 b)]

<https://standards.iteh.ai/catalog/standards/iso/955356c6-43f0-4efe-86db-a90f80ff1c91/iso-14242-3-2009-amd-1-2019>

5.1

Replace

“, calf serum approximately 25 % volume fraction diluted with de-ionized water (balance).

The fluid test medium should be filtered through a filter of pore size 2 µm and have a protein mass concentration of not less than 17 g/l.”

with

“, calf serum diluted with deionized water in accordance with ISO 3696.

The fluid test medium shall be filtered through a 2 µm filter and have a protein mass concentration of 30 g/l ± 2 g/l.”

6.2

Add the following at the end of 6.2 and before the NOTE:

One testing setup, sometimes used for wear testing incorporating 3rd-body particles, is the “inverted position”, that is, head above liner [Figure 4 a)]. Another testing setup is the “anatomical position,” that is, head below liner [Figure 4 b)]. The 23° angle block provides swivel to result in a 46° rocking motion, which is equivalent to an actuation of ±23° in flexion/extension and abduction/adduction (see Figure 2).

6.3

Replace the text (but not the NOTE) with the following:

6.3 Means of aligning and positioning the acetabular and femoral component of the test specimen [for Figure 4 a), the inverted position], so that its centre lies on the axis of rotation [key 7 in Figure 4 a)] of the test machine. This is achieved by calculation based on engineering drawings and setting a suitable thickness of the cement base [key 8 in Figure 4 a)] to achieve the requirement above. The femoral component is assembled to a stem connected to a self-alignment device. The ball bearings in the self-alignment device allow the stem construct to be mobile while aligned with the compressive load.

6.4

Replace the text with the following:

6.4 Means of aligning and positioning the acetabular and femoral component [for Figure 4 b), the anatomic position], to ensure that its centre lies along the axis of rotation [key 7 in Figure 4 b)] of the test machine. This is achieved by calculating and setting the thickness of the femoral fixture base [key 8 in Figure 4 b)] based on engineering drawings to achieve the alignment sought above. The acetabular component is assembled to the self-alignment device and along the axis of compressive load.

Figure 1

Delete the NOTE.

7.4

ISO 14242-3:2009/Amd 1:2019

<https://standards.iteh.ai/catalog/standards/iso/955356c6-43f0-4efe-86db-a90f80ff1c91/iso-14242-3-2009-amd-1-2019>

Replace the text (but not NOTE 1 and NOTE 2) with the following:

“7.4 Assemble the femoral component to the fixture in the test machine for the inverted [Figure 4 a)] and anatomic [Figure 4 b)] positions.”

7.5

Replace the text with the following:

“7.5 Mount the acetabular component in the cup holder for the inverted [Figure 4 a)] and anatomic [Figure 4 b)] positions.”

NOTE For all setup positions, the key is to maintain the centres of the femoral head and acetabular component aligned with the centre of rotation of the machine and axial load line, within the tolerance range of the self-alignment device [Key 1 in both Figure 4 a) and Figure 4 b)].”

Clause 8, c)

Add “whether the “inverted” [Figure 4 a)] or “anatomic” [Figure 4 b)] configuration was used for testing,” between “range of motions and forces, type of systems used for measuring motions and forces,” and “, arrangement for mounting specimen (5.2),” to make c) as follows: