SLOVENSKI STANDARD
oSIST prEN ISO 20647:2020

01-marec-2020

Hrana za dojenčke in prehranska dopolnila za odrasle - Določevanje celotnega joda - Masna spektrometrija z induktivno sklopljeno plazmo (ICP-MS) (ISO 20647:2015)

Infant formula and adult nutritionals -Determination of total iodine - Inductively coupled plasma mass spectrometry (ICP-MS) (ISO 20647:2015)

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Säuglingsanfangsnahrung und Nahrungsergänzungsmittel für Erwachsene - Bestimmung des Gesamtiods - Massenspektrometrie mit induktiv gekoppeltem Plasma (ICP-MS) (ISO 20647:2015)

SIST EN ISO 20647:2020
https://standards.iteh.ai/catalog/standards/sist/f04b8b1b-33ac-4cda-910c-
5211f3a5c3ae4/sist-en-iso-20647-2020

Formules infantiles et produits nutritionnels pour adultes - Détermination de la teneur en iode total - Spectrométrie de masse avec plasma à couplage inductif (ICP-SM) (ISO 20647:2015)

Ta slovenski standard je istoveten z: **prEN ISO 20647**

ICS:

67.050	Splošne preskusne in analizne metode za živilske proizvode	General methods of tests and analysis for food products

oSIST prEN ISO 20647:2020 **en**

2003-01.Slovenski inštitut za standardizacijo. Razmnoževanje celote ali delov tega standarda ni dovoljeno.

oSIST prEN ISO 20647:2020

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 20647:2020
https://standards.iteh.ai/catalog/standards/sist/f04b8b1b-33ae-4cda-910e-82113e5c5ae4/sist-en-iso-20647-2020

oSIST prEN ISO 20647:2020

INTERNATIONAL STANDARD

ISO 20647

IDF 234

First edition
2015-11-01

Infant formula and adult nutritionals — Determination of total iodine — Inductively coupled plasma mass spectrometry (ICP-MS)

Formules infantiles et produits nutritionnels pour adultes — Détermination de la teneur en iode total — Spectrométrie de masse avec plasma à couplage inductif (ICP-SM)

iTeh STANDARD PREVIEW
(standards.iteh.ai)
SIST EN ISO 20647:2020
https://standards.iteh.ai/catalog/standards/sist/f04b8b1b-33ae-4cda-910e-82113e5c5ae4/sist-en-iso-20647-2020

Reference numbers
ISO 20647:2015(E)
IDF 234:2015(E)

© ISO and IDF 2015

oSIST prEN ISO 20647:2020

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 20647:2020
https://standards.iteh.ai/catalog/standards/sist/f04b8b1b-33ae-4cda-910e-82113e5c5ae4/sist-en-iso-20647-2020

COPYRIGHT PROTECTED DOCUMENT

© ISO and IDF 2015, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

International Dairy Federation
Silver Building • Bd Auguste Reyers 70/B • B-1030 Brussels

Tel. + 32 2 325 67 40
Fax + 32 2 325 67 41
info@fil-idf.org
www.fil-idf.org

oSIST prEN ISO 20647:2020

Contents

Page

Forewords ..... **iv**

1 Scope **1**

2 Principle **1**

3 Reagents and materials **1**

4 Apparatus **5**

5 Procedure **6**
- 5.1 Reconstitution 6
- 5.2 Sample preparation 6
- 5.3 Addition of reagents 7
- 5.4 Oven digestion 7
- 5.5 Open vessel microwave digestion 7
- 5.6 Sample filtering 8
- 5.7 Sample dilution 8
- 5.8 Determination 8
- 5.9 Conditioning 9
- 5.10 Calibration 9
- 5.11 Sample analysis 9
- 5.12 Data acceptability 10

6 Calculation **10**

7 Precision **11**
- 7.1 Interlaboratory tests 11
- 7.2 Repeatability 11
- 7.3 Reproducibility 11

Annex A (informative) **Precision data** **12**

Bibliography **13**

iTeh STANDARD PREVIEW
(standards.iteh.ai)
SIST EN ISO 20647:2020
https://standards.iteh.ai/catalog/standards/sist/f04b8b1b-33ae-4cda-910e-
82413e5e5ae4/sist-en-iso-20647-2020

© ISO and IDF 2015 – All rights reserved

oSIST prEN ISO 20647:2020

Forewords

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance wit the editorial rules of ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers Trade (TBT) see the following URL: Foreword - Supplementary information

iTeh STANDARD PREVIEW
(standards.iteh.ai)

The committee responsible for this document is ISO/TC 34, *Food products*, Subcommittee SC 5, *Milk and milk products* and the International Dairy Federation (IDF), in collaboration with AOAC INTERNATIONAL. It is being published jointly by ISO and IDF and separately by AOAC INTERNATIONAL. The method described in this International Standard is equivalent to the AOAC Official Method 2012.15: *Total iodine in infant formula and adult/pediatric nutritional formula – inductively coupled plasma-mass spectrometry*.

SIST EN ISO 20647:2020
https://standards.iteh.ai/catalog/standards/sist/904b8b1b-73ae-4cda-910e-82113e5c5ae4/sist-en-iso-20647-2020

 © ISO and IDF 2015 – All rights reserved

oSIST prEN ISO 20647:2020

IDF (the International Dairy Federation) is a non-profit private sector organization representing the interests of various stakeholders in dairying at the global level. IDF members are organized in National Committees, which are national associations composed of representatives of dairy-related national interest groups including dairy farmers, dairy processing industry, dairy suppliers, academics and governments/food control authorities.

ISO and IDF collaborate closely on all matters of standardization relating to methods of analysis and sampling for milk and milk products. Since 2001, ISO and IDF jointly publish their International Standards using the logos and reference numbers of both organizations.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. IDF shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute and endorsement.

ISO 20647 | IDF 234 was prepared by the IDF Standing Committee on Analytical Methods for Composition and the ISO Technical Committee ISO/TC 34, *Food products*, Subcommittee SC 5, *Milk and milk products*, in collaboration with AOAC INTERNATIONAL. It is being published jointly by ISO and IDF, and separately by AOAC INTERNATIONAL. The method described in this International Standard is equivalent to the AOAC Official Method 2012.15: *Total iodine in infant formula and adult/pediatric nutritional formula – inductively coupled plasma-mass spectrometry*.

All work was carried out by the ISO-IDF Project Group C37 of the Standing Committee on *Analytical Methods for Composition* under the aegis of its project leader, Mr Erik Konings (CH).

iTeh STANDARD PREVIEW
(standards.iteh.ai)
SIST EN ISO 20647:2020
https://standards.iteh.ai/catalog/standards/sist/f04b8b1b-33ae-4cda-910e-82113e5c5ae4/sist-en-iso-20647-2020

© ISO and IDF 2015 – All rights reserved

oSIST prEN ISO 20647:2020

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 20647:2020
https://standards.iteh.ai/catalog/standards/sist/f04b8b1b-33ae-4cda-910e-82113e5c5ae4/sist-en-iso-20647-2020

oSIST prEN ISO 20647:2020

Infant formula and adult nutritionals — Determination of total iodine — Inductively coupled plasma mass spectrometry (ICP-MS)

WARNING — The use of this International Standard can involve hazardous materials, operations and equipment. This International Standard does not purport to address all the safety problems associated with its use. It is the responsibility of the user of this International Standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.

1 Scope

This International Standard specifies a method for the quantitative determination of total iodine in infant formula and adult nutritional formula.[1] The method is applicable to the measurement of total iodine in infant formula and adult nutritional formula from 0,5 µg/100g to 1 500 µg/100g reconstituted final product and for ready-to-feed products from 2,5 µg/100 g to 1 000 µg/100 g using ICP-MS.

Using various infant formula and adult nutritional products, the method was subjected to an interlaboratory study. Levels obtained ranged from 3,47 µg/100 g to 124 µg/100 g. For all precision data related to the interlaboratory study, see Table A.1 located in Annex A.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

2 Principle

Digestion occurs using a potassium hydroxide (KOH) solution in an oven or open-vessel microwave system. Iodine is stabilized with ammonium hydroxide and sodium thiosulfate after digestion. The solution is brought to volume followed by filtration. The filtrate is analysed directly or after dilution by inductively coupled plasma mass spectrometry (ICP-MS).

3 Reagents and materials

During the analysis, unless otherwise stated, use only reagents of recognized analytical grade and distilled or demineralized water or water of equivalent purity. Equivalent chemicals and reagents may be used.

3.1 KOH pellets, certified ACS grade, e.g. Fisher Scientific, Fairlawn, NJ[1)].

NOTE KOH may contribute background levels of iodine.

3.2 Ammonium hydroxide (NH_4OH), 28 % to 30 % (m/m), certified ACS PLUS, Fisher Scientific[1)].

3.3 Sodium thiosulfate ($Na_2S_2O_3$), purity ≥ 99,99 %, metal basis, Fisher Scientific[1)].

3.4 Surfactant, e.g. Triton®X-100, Sigma, St. Louis, MO[1)].

3.5 Concentrated nitric acid (HNO_3), Optima, high purity, Fisher Scientific[1)].

3.6 Perchloric acid ($HClO_4$), 70 % (m/m).

1) This is an example of a suitable product available commercially. This information is given for the convenience of users of this document and does not constitute an endorsement by either ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

© ISO and IDF 2015 – All rights reserved

oSIST prEN ISO 20647:2020

3.7 Purified water, 18 MΩ/cm.

3.8 Reference standards

3.8.1 Iodide standard solution in water, mass concentration ρ = 1 000 µg/ml, SPEX CertiPrep[1)].

3.8.2 Iodide standard solution in 1 % triethylamine (TEA), ρ = 1 000 µg/ml, Inorganic Ventures[1)].

3.8.3 Standard Reference Material (SRM), National Institute of Standards and Technology (NIST) SRM 1849a, Infant/Adult Nutritional Formula.

Either stock iodide reference solutions may be used for intermediate and working standard solutions preparation. The remaining source may be used as a continuing calibration verification (CCV) standard. Equivalent reference standards may be substituted. 'Iodide' may be referred to as 'iodine' throughout this International Standard.

3.9 Internal standards

3.9.1 Praseodymium (Pr) standard solution in 5 % HNO_3, ρ = 10 µg/ml, Inorganic Ventures[1)].

Individual values of iodine are reported for each test sample using praseodymium as the internal standard. Equivalent stock internal standard solutions may be substituted.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

3.10 Preparation of reagent solutions

3.10.1 General

Prepare all reagent solutions as recommended by either mass per volume or volume per volume. Adjusting for purity and/or concentration is not required.

SIST EN ISO 20647:2020
https://standards.iteh.ai/catalog/standards/sist/f04b8b1b-33ae-4cda-910e-82113e5c5ae4/sist-en-iso-20647-2020

3.10.2 KOH solution, ρ = 50 g/l

Dissolve 25 g of KOH pellets in an appropriate amount of purified water, then dilute to 500 ml with purified water. This solution may be added using a re-pipet volumetric bottle top dispenser. Store this solution at room temperature. Reagent expires 6 months after preparation date.

3.10.3 Stabilizer concentrate

Dissolve 5 g of $Na_2S_2O_3$ in an appropriate amount of purified water, add 50 ml of NH_4OH, then dilute to 500 ml with purified water. The resulting concentration is 10 % NH_4OH and 1 % $Na_2S_2O_3$ in purified water. Store this solution at room temperature. Reagent expires 6 months after preparation date.

3.10.4 Wash solution (rinse)

Dissolve 2 g of surfactant (3.4) in an appropriate amount of purified water, add 20 ml of NH_4OH, then dilute to 2 000 ml with purified water. The resulting concentration is 1 % NH_4OH and 0,1 % surfactant in purified water. This solution may be added using a re-pipet volumetric bottle top dispenser. Store this solution at room temperature. Reagent expires 6 months after preparation date.

3.10.5 Diluent

Dissolve 10 g of KOH pellets and 0,4 g $Na_2S_2O_3$ in an appropriate amount of purified water, add 4 ml NH_4OH, then dilute to 2 000 ml with purified water. Store this solution at room temperature. Reagent expires 6 months after preparation date.

 © ISO and IDF 2015 – All rights reserved

oSIST prEN ISO 20647:2020

Alternatively, for a smaller volume, dilute 50 ml of 5 % KOH and 10 ml of stabilizer concentrate (3.10.3) to 500 ml with purified water. The resulting concentrations for both preparations are 0,5 % KOH, 0,2 % NH_4OH, and 0,02 % $Na_2S_2O_3$ in purified water. Store this solution at room temperature. Reagent expires 6 months after preparation date.

3.10.6 Conditioning solution

Prepare by aliquoting 25 ml of KOH solution (3.10.2), then dilute to 250 ml with purified water. Use this solution to prepare the instrument for analysis. The resulting concentration is 5 g/l KOH. Store this solution at room temperature. Reagent expires 6 months after preparation date.

3.10.7 Carrier solution

This solution is equivalent to the wash solution (3.10.4). The carrier solution is used to deliver the sample solution to the nebulizer through the ICP-MS autosampler introduction system. The carrier solution is introduced via a peristaltic pump using 0,76 mm inside diameter (id) two-stop polyvinyl chloride pump tubing. Store this solution at room temperature. Reagent expires 6 months after preparation date.

3.11 Preparation of standard solutions

3.11.1 General

Stock solutions are stable until the end date indicated on the certificate of analysis. Intermediate stock standard, calibration standard, continuing calibration verification and internal standard solutions are stable at room temperature until the earliest expiration date of all components used to prepare the solution. These solutions are ready to use for analysis. Do not subject them to any of the various procedures used to prepare actual samples (i.e. infant formulas and adult nutritionals) for analysis.

iTeh STANDARD PREVIEW
(standards.iteh.ai)
SIST EN ISO 20647:2020
https://standards.iteh.ai/catalog/standards/sist/f0cb8b1b-32ae-4cda-910e-82113e5c5ae4/sist-en-iso-20647-2020

3.11.2 Stock iodine and praseodymium solutions

Purchase of stock iodine and praseodymium standard solutions with accompanying certificates of analysis is recommended.

3.11.3 Intermediate stock standard (ISS) iodine solutions

Prepare the intermediate stock standard iodine solutions according to Table 1.

These ISS solutions are used for calibration standard preparation and are typically prepared according to Table 1. The ISS concentrations presented are nominal. Using the stock iodine concentration found on the certificate of analysis, determine the exact concentration of each ISS. An electronic adjustable volume pipet, capable of delivering 100 µl to 5 000 µl should be used.

Table 1 — Preparation of intermediate stock standard (ISS) iodine solutions

Iodine standard solution ID	ID of solution used for preparation	Initial iodine concentration ng/ml	Aliquot volume ml	Final volume ml	Final iodine concentration ng/ml
10 000 (ISS)	Stock	1 000 000	0,5	50	10 000
1 000 (ISS)	10 000 (ISS)	10 000	5	50	1 000
10,0 (ISS)	1 000 (ISS)	1 000	0,5	50	10,0
Aliquot the appropriate amount of iodine standard solution into a single use 50 ml tube (4.1) and add 5 ml of stabilizer concentrate (3.10.3), fill to the 50 ml mark on the tube with water, cap the tube and then mix thoroughly. The resulting matrix concentration is 1 % NH_4OH and 0,1 % $Na_2S_2O_3$ in water.					

3.11.4 Calibration standard (CS) iodine solutions

Prepare the intermediate stock standard iodine solutions according to Table 2.

© ISO and IDF 2015 – All rights reserved