

ISO/TC 168

Secretariat: DIN

Voting begins on:
2016-02-11

Voting terminates on:
2016-04-11

Prosthetics — Structural testing of lower-limb prostheses — Requirements and test methods

Prothèses — Essais portant sur la structure des prothèses de membres inférieurs — Exigences et méthodes d'essai

iTeh STANDARD PREVIEW
(standards.iteh.ai)
Full standard:
<https://standards.iteh.ai/catalog/standards/sist/1d8235c-a762-4b4a-8cb5-fcd594ad445b/iso-10328-2016>

Please see the administrative notes on page iii

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.



Reference number
ISO/FDIS 10328:2016(E)

ISO/CEN PARALLEL PROCESSING

This final draft has been developed within the International Organization for Standardization (ISO), and processed under the **ISO-lead** mode of collaboration as defined in the Vienna Agreement. The final draft was established on the basis of comments received during a parallel enquiry on the draft.

This final draft is hereby submitted to the ISO member bodies and to the CEN member bodies for a parallel two-month approval vote in ISO and formal vote in CEN.

Positive votes shall not be accompanied by comments.

Negative votes shall be accompanied by the relevant technical reasons.

iTeh STANDARD PREVIEW
(standards.iteh.ai)
Full standard:
<https://standards.iteh.ai/catalog/standards/sist/1d88357c-a762-4b4a-8cb5-fcd594ad445b/iso-10328-2016>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2016, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

	Page
Foreword	vii
Introduction	ix
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Designations and symbols of test forces and moments	2
5 Strength and related performance requirements and conditions of use	3
6 Coordinate systems and test configurations	5
6.1 General.....	5
6.2 Axes of coordinate systems.....	5
6.3 Reference planes.....	5
6.3.1 General.....	5
6.3.2 Top reference plane, T.....	5
6.3.3 Knee reference plane, K.....	5
6.3.4 Ankle reference plane, A.....	6
6.3.5 Bottom reference plane, B.....	6
6.4 Reference points.....	7
6.5 Test force.....	8
6.6 Load line.....	8
6.7 Longitudinal axis of the foot and effective joint centres and centrelines.....	8
6.7.1 General.....	8
6.7.2 Longitudinal axis of the foot.....	8
6.7.3 Effective ankle-joint centre.....	8
6.7.4 Effective ankle-joint centreline.....	10
6.7.5 Effective knee-joint centreline.....	10
6.7.6 Effective knee-joint centre.....	11
6.8 Reference distances.....	11
6.8.1 Offsets.....	11
6.8.2 Combined offsets.....	11
6.8.3 Effective lever arms L_A and L_K	11
6.8.4 Distance L_{BT}	11
7 Test loading conditions and test loading levels	11
7.1 Test loading conditions.....	11
7.1.1 General.....	11
7.1.2 Test loading conditions of principal structural tests.....	12
7.1.3 Test loading conditions of separate structural tests.....	12
7.2 Test loading levels.....	12
8 Values of test loads, dimensions and cycles	13
9 Compliance	21
9.1 General.....	21
9.2 Selection of tests required to claim compliance with this International Standard.....	22
9.3 Arrangements for tests on samples of prosthetic structures including ankle-foot devices or foot units, required to claim compliance with this International Standard.....	22
9.3.1 General.....	22
9.3.2 Particular arrangements concerning the ankle-foot device or foot unit.....	22
9.3.3 Particular arrangements and requirements concerning the part required to connect the ankle-foot device or foot unit to the remainder of the prosthetic structure.....	22
9.4 Number of tests and test samples required to claim compliance with this International Standard.....	23
9.5 Multiple use of test samples.....	23

9.5.1	General.....	23
9.5.2	Restriction.....	23
9.6	Testing at particular test loading levels not specified in this International Standard.....	24
10	Test samples.....	26
10.1	Selection of test samples.....	26
10.1.1	General.....	26
10.1.2	Selection of ankle-foot devices and foot units of appropriate size of foot.....	27
10.2	Types of test samples.....	27
10.2.1	Complete structure.....	27
10.2.2	Partial structure.....	30
10.2.3	Any other structure.....	30
10.3	Preparation of test samples.....	30
10.4	Identification of test samples.....	31
10.5	Alignment of test samples.....	32
10.5.1	Test samples for principal tests and separate tests on knee locks.....	32
10.5.2	Test samples for separate tests on ankle-foot devices and foot units.....	32
10.5.3	Test samples for separate static ultimate strength tests in maximum knee flexion for knee joints and associated parts.....	32
10.5.4	Test samples for separate tests on knee locks.....	33
10.6	Worst-case alignment position of test samples.....	33
11	Responsibility for test preparation.....	34
12	Test submission document.....	35
12.1	General requirements.....	35
12.2	Information required for test samples.....	36
12.2.1	All test samples.....	36
12.2.2	Test samples for tests on ankle-foot devices and foot units.....	36
12.2.3	Test samples for static ultimate strength tests in maximum knee flexion for knee joints and associated parts.....	36
12.3	Information required for tests.....	36
12.3.1	General.....	36
12.3.2	For all tests.....	36
12.3.3	For static tests in torsion and on ankle-foot devices and foot units.....	37
12.3.4	For static ultimate strength tests.....	37
12.3.5	For cyclic tests.....	37
12.3.6	For tests in torsion.....	37
12.3.7	For tests on ankle-foot devices and foot units.....	37
13	Equipment.....	37
13.1	General.....	37
13.2	Equipment for the principal tests specified in 16.2 and 16.3.....	38
13.2.1	End attachments.....	38
13.2.2	Jig (optional).....	40
13.2.3	Test equipment.....	40
13.3	Equipment for the separate static test in torsion specified in 17.1.....	42
13.3.1	Test equipment.....	42
13.4	Equipment for the separate tests on ankle-foot devices and foot units specified in 17.2.....	42
13.4.1	Test equipment.....	42
13.5	Equipment for the separate static ultimate strength test in maximum knee flexion for knee joints and associated parts specified in 17.3.....	46
13.5.1	Extension pieces.....	46
13.5.2	Test equipment to perform static compression loading – (Compression testing machine or other equipment).....	46
13.6	Equipment for the separate tests on knee locks specified in 17.4.....	46
13.6.1	End attachments.....	46
13.6.2	Jig (optional).....	46
13.6.3	Test equipment.....	46
14	Accuracy.....	47

14.1	General	47
14.2	Accuracy of equipment	47
14.3	Accuracy of procedure	47
15	Test principles	48
15.1	General	48
15.2	Static test procedure	48
15.3	Cyclic test procedure	48
16	Test procedures – Principal structural tests	48
16.1	Test loading requirements	48
16.1.1	Preparation for test loading	48
16.1.2	Application of test loading	48
16.2	Principal static test procedure	50
16.2.1	Principal static proof test	50
16.2.2	Principal static ultimate strength test	55
16.3	Principal cyclic test procedure	59
16.3.1	General requirements	59
16.3.2	Test method	59
16.3.3	Performance requirements	63
16.3.4	Compliance conditions	64
17	Test procedures – Separate structural tests	68
17.1	Separate static test in torsion	68
17.1.1	General	68
17.1.2	Purpose of test	68
17.1.3	Test method	68
17.1.4	Performance requirements	70
17.1.5	Compliance conditions	70
17.2	Separate tests on ankle-foot devices and foot units	72
17.2.1	General	72
17.2.2	Purpose of tests	72
17.2.3	Separate static proof test for ankle-foot devices and foot units	72
17.2.4	Separate static ultimate strength test for ankle-foot devices and foot units	76
17.2.5	Separate cyclic test for ankle-foot devices and foot units	81
17.3	Separate static ultimate strength test in maximum knee flexion for knee joints and associated parts	85
17.3.1	General	85
17.3.2	Purpose of test	86
17.3.3	Applicability of the test to specific test samples	86
17.3.4	Test method	86
17.3.5	Performance requirement	87
17.3.6	Compliance conditions	87
17.4	Separate tests on knee locks	88
17.4.1	General	88
17.4.2	Purpose of tests	89
17.4.3	Separate static proof test for knee locks	89
17.4.4	Separate static ultimate strength test for knee locks	93
17.4.5	Separate cyclic test for knee locks	95
18	Test laboratory/facility log	104
18.1	General requirements	104
18.2	Specific requirements	104
19	Test report	104
19.1	General requirements	104
19.2	Specific requirements	105
19.3	Options	105
20	Classification and designation	105
20.1	General	105

20.2	Examples of classification and designation.....	105
21	Labelling.....	106
21.1	General.....	106
21.2	Use of mark “*”) and warning symbol.....	107
21.3	Examples of label layout.....	107
21.4	Label placement.....	107
Annex A	(informative) Description of internal loads and their effects	109
Annex B	(informative) Reference data for the specification of test loading conditions and test loading levels of principal cyclic tests.....	113
Annex C	(informative) Guidance on the application of an alternative static ultimate strength test.....	117
Annex D	(normative) Guidance on the application of an additional test loading levels P6, P7 and P8	118
Annex E	(informative) Summary of the records to be entered in the test laboratory/facility log	121
Annex F	(informative) Background information on the loading profiles generated by test equipment according to 13.4.1.2 for separate cyclic tests for ankle-foot devices and foot units according to 17.2.5.1.....	136
Annex G	(informative) Reference to the essential principles of safety and performance of medical devices according to ISO/TR 16142	138
Annex ZA	(informative) Relationship between this European Standard and the essential requirements of Directive 93/42/EEC [OJ L 169] aimed to be covered	139
Bibliography	141

iTech STANDARD PREVIEW

(standards.iteh.ai)

Full standard:

https://standards.iteh.ai/catalog/standards/sist/1d88457c-a762-4b4a-8cb5-fcd594ad445b/iso-10328-2016

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: Foreword - Supplementary information

The committee responsible for this document is ISO/TC 168, *Prosthetics and orthotics*.

This second edition cancels and replaces the first edition ISO 10328:2006 which has been technically revised with the following changes:

- a) Test loading levels P7 and P8 have been introduced in [Table B.1](#), [Table B.2](#), [Table B.3](#), [Table 4.1](#), [Table D.1](#), [Table D.2](#), [Table D.3](#) and the clauses pointing at these tables have been updated. Additional information on P7 and P8 is given in Annex B.1;
- b) [Table 9](#) has been revised;
- c) [Annex D](#) has changed from informative to normative.

European Foreword

By agreement between ISO and CEN, this European Foreword is included in the FDIS but will not appear in the published ISO document.

The following referenced documents are indispensable for the application of this document. For undated references, the latest edition of the referenced document (including any amendments) applies. For dated references, only the edition cited applies. However, for any use of this standard ‘within the meaning of [Annex ZA](#)’, the user should always check that any referenced document has not been superseded and that its relevant contents can still be considered the generally acknowledged state-of-art.

When an IEC or ISO standard is referred to in the ISO standard text, this shall be understood as a normative reference to the corresponding EN standard, if available, and otherwise to the dated version of the ISO or IEC standard, as listed below.

NOTE 1 The way in which these referenced documents are cited in normative requirements determines the extent (in whole or in part) to which they apply.

Table 1 — Correlation between normative references and dated EN and ISO standards

Normative references as listed in Clause 2 of the ISO standard	Equivalent dated standard	
	EN	ISO or IEC
ISO 8549-1:1989	—	ISO 8549-1:1989
ISO/TR 16142:1999	—	ISO/TR 16142:1999
ISO 22523:2006	EN ISO 22523:2006	ISO 22523:2006
ISO 22675:2016	EN ISO 22675:2016	ISO 22675:2016

Introduction

Throughout this International Standard, the term prosthesis means an externally applied device used to replace wholly, or in part, an absent or deficient limb segment.

As a result of concern in the international community about the need to provide prostheses that are safe in use, and also because of an awareness that test standards would assist the development of better prostheses, a series of meetings was held under the aegis of the International Society for Prosthetics and Orthotics (ISPO). The final one was held in Philadelphia, PA, USA in 1977 at which a preliminary consensus was reached on methods of testing and the required load values. From 1979 onwards this work was continued by ISO Technical Committee 168 leading to the development of ISO 10328:1996. The test procedures may not be applicable to prostheses of mechanical characteristics different from those used in the consensus.

During use, a prosthesis is subjected to a series of load actions, each varying individually with time. The test methods specified in this International Standard use static and cyclic strength tests which typically produce compound loadings by the application of a single test force.

The static tests relate to the worst loads generated in any activity. The cyclic tests relate to normal walking activities where loads occur regularly with each step. This International Standard specifies fatigue testing of structural components. The tests specified do not provide sufficient data to predict actual service life.

The evaluation of lower-limb prostheses and their components requires controlled field trials in addition to the laboratory tests specified in this International Standard.

The laboratory tests and field trials should be repeated when significant design changes are made to a load-bearing part of a prosthesis.

Ideally, additional laboratory tests should be carried out to deal with function, wear and tear, new material developments, environmental influences and user activities as part of the evaluation procedure. There are no standards for such tests, so appropriate procedures will need to be determined.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Full standard:
<https://standards.iteh.ai/catalog/standards/sist/1d88357c-a762-4b4a-8cb5-fcd594ad445b/iso-10328-2016>

Prosthetics — Structural testing of lower-limb prostheses — Requirements and test methods

1 Scope

IMPORTANT — This International Standard is *suitable* for the assessment of the conformity of lower limb prosthetic devices/structures with the strength requirements specified in 4.4 of ISO 22523:2006 (see NOTE 1). Prosthetic ankle-foot devices and foot units on the market, which have demonstrated their compliance with the strength requirements specified in 4.4 of ISO 22523:2006 through submission to the relevant tests of ISO 10328:2006, need not be retested to ISO 22675:2016.

WARNING — This International Standard is *not suitable* to serve as a guide for the selection of a specific lower limb prosthetic device/structure in the prescription of an individual lower limb prosthesis! Any disregard of this warning can result in a safety risk for amputees.

This International Standard specifies procedures for static and cyclic strength tests on lower-limb prostheses (see NOTE 2) which typically produce compound loadings by the application of a single test force. The compound loads in the test sample relate to the peak values of the components of loading which normally occur at different instants during the stance phase of walking.

The tests described in this International Standard comprise

- principal static and cyclic tests for all components;
- a separate static test in torsion for all components;
- separate static and cyclic tests on ankle-foot devices and foot units for all ankle-foot devices as single components including ankle units or ankle attachments and all foot units as single components;
- a separate static ultimate strength test in maximum knee flexion on knee joints and associated parts for all knee units or knee-shin-assemblies and adjacent components that normally provide the flexion stop on a complete prosthesis;
- separate static and cyclic tests on knee locks for all mechanisms which lock the knee joint in the extended position of the knee unit or knee-shin-assembly.

The tests described in this International Standard apply to specific types of ankle-disarticulation prostheses (see NOTE 3), to transtibial (below-knee), knee-disarticulation and transfemoral (above-knee) prostheses and to the distal (lower) part of hip-disarticulation and hemi-pelvectomy prostheses (see NOTE 4).

NOTE 1 The tests can be performed on complete structures, on part structures or on individual components.

NOTE 2 The tests only apply to ankle-disarticulation prostheses which include (foot) components of prosthetic ankle-foot devices taken from the normal production line.

NOTE 3 The distal part comprises the knee unit, the ankle-foot device and all parts between. Tests on hip units are described in ISO 15032.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 8549-1:1989, *Prosthetics and orthotics — Vocabulary — Part 1: General terms for external limb prostheses and external orthoses*

ISO/TR 16142:1999, *Medical devices — Guidance on the selection of standards in support of recognized essential principles of safety and performance of medical devices*

ISO 22523:2006, *External limb prostheses and external orthoses — Requirements and test methods*

ISO 22675:2016, *Prosthetics — Testing of ankle-foot devices and foot units — Requirements and test methods*

3 Terms and definitions

For the purposes of this document, the terms and definitions in ISO 8549-1:1989 and the following definitions apply.

3.1 proof strength

static load representing an occasional severe event, which can be sustained by the prosthetic device/structure and still allow it to function as intended

3.2 ultimate strength

static load representing a gross single event, which can be sustained by the prosthetic device/structure but which could render it thereafter unusable

3.3 fatigue strength

cyclic load which can be sustained by the prosthetic device/structure for a given number of cycles

3.4 batch

set of test samples of a prosthetic device/structure submitted together to a test laboratory/facility to undertake tests to demonstrate compliance with one or more requirements of this International Standard

4 Designations and symbols of test forces and moments

The designations and symbols of all relevant test forces and moments are listed in [Table 1](#).

Table 1 — Designations and symbols of test forces and moments

Designation	Symbol
Test forces; twisting moments	$F, F_1, F_2; M_u$
Proof test force of end attachments	F_{pa}
Stabilizing test force	F_{stab}
Settling test force	F_{set}
Static proof test force	F_{sp}
Static proof test force on heel/forefoot	F_{1sp}, F_{2sp}
Static ultimate test force	F_{su}
Static ultimate test force on heel/forefoot	F_{1su}, F_{2su}
NOTE Further details of the test forces and moments listed are given in Table 3 .	

Table 1 (continued)

Designation	Symbol
Minimum test force	F_{cmin}
Maximum test force	F_{cmax}
Range of pulsating test force	F_{cr}
Mean test force	F_{cmean}
Amplitude of pulsating test force	F_{ca}
Pulsating test force	$F_c(t)$
Final static test force	F_{fin}
Minimum test force on heel/forefoot	F_{1cmin}, F_{2cmin}
Maximum test force on heel/forefoot	F_{1cmax}, F_{2cmax}
Range of pulsating test force on heel/forefoot	F_{1cr}, F_{2cr}
Mean test force on heel/forefoot	F_{1cmean}, F_{2cmean}
Amplitude of pulsating test force on heel/forefoot	F_{1ca}, F_{2ca}
Pulsating test force on heel/forefoot	$F_{1c}(t), F_{2c}(t)$
Final static test force on heel/forefoot	F_{1fin}, F_{2fin}
Stabilizing twisting moment	M_{u-stab}
Settling twisting moment	M_{u-set}
Maximum twisting moment	M_{u-max}
NOTE Further details of the test forces and moments listed are given in Table 3 .	

5 Strength and related performance requirements and conditions of use

5.1 According to ISO 22523:2006, 4.4.1, a lower limb prosthetic device/structure “... shall have the strength to sustain the loads occurring during use by amputees [...] in the manner intended by the manufacturer for that device according to his written instructions on its intended use”.

For the assessment of the conformity of lower limb prosthetic devices/structures with the above requirement (see also Scope), this International Standard provides a means of determining the three categories of strength defined in [3.1](#) to [3.3](#) and, in addition, the static strength in torsion and the security against slipping of clamped components.

All of these are listed in [Table 2](#), together with the related performance requirements and the test methods for their verification.

5.2 In order to satisfy the general requirement in [5.1](#) for a specific lower limb prosthetic device/structure, the following safety concept shall apply.

The device/structure shall

- a) comply with the requirements of this International Standard (see [9.1](#), [9.2](#) and [9.3](#)) for a specific test loading level (see [7.2](#))

and

- b) be used in accordance with the body mass limit specified by the manufacturer in consideration of the intended use of that device (see NOTE).