

ISO/TC 85/SC 2

Secretariat: AFNOR

Voting begins on:
2020-04-08

Voting terminates on:
2020-06-03

Radiation protection — Sealed sources — Leakage test methods

Radioprotection — Sources scellées — Méthodes d'essai d'étanchéité

iTeh STANDARD PREVIEW
(standards.iteh.ai)
Full standard:
<https://standards.iteh.ai/catalog/standards/sist/0ff56b1f-d75e-4d82-946c-5108917fd04d/iso-fdis-9978>

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.



Reference number
ISO/FDIS 9978:2020(E)

iTeh STANDARD PREVIEW
(standards.iteh.ai)
Full standard:
<https://standards.iteh.ai/catalog/standards/sist/0ff56b1f-d75e-4d82-946c-5108917fd04d/iso-fdis-9978>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

	Page
Foreword	iv
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Requirements	3
5 Test methods by radioactive means	5
5.1 Immersion tests	5
5.1.1 Immersion test (hot liquid)	5
5.1.2 Immersion test (boiling liquid)	5
5.1.3 Immersion test with a liquid scintillator	5
5.1.4 Immersion test at room temperature	5
5.1.5 Acceptance criteria	6
5.2 Gaseous emanation tests	6
5.2.1 Gaseous emanation test by absorption (for radium-226 sealed sources)	6
5.2.2 Gaseous emanation test by immersion with a scintillator (for radium-226 sealed sources)	6
5.2.3 Gaseous emanation test (for krypton-85 sealed sources)	6
5.2.4 Other gaseous emanation tests	6
5.2.5 Acceptance criteria	6
5.3 Wipe tests	6
5.3.1 Wet wipe test	6
5.3.2 Dry wipe test	7
5.3.3 Acceptance criteria	7
6 Test methods by volumetric means	7
6.1 Helium mass spectrometer leakage tests	7
6.1.1 Helium test [equivalent to leak test type B6 in ISO 20485]	7
6.1.2 Helium pressurization test [equivalent to leak test type B5 in ISO 20485]	8
6.1.3 Acceptance criteria	8
6.2 Bubble leakage tests	8
6.2.1 Vacuum bubble test [equivalent to immersion technique using vacuum in EN 1593 ^[6]	8
6.2.2 Hot-liquid bubble test [equivalent to immersion technique using liquid at elevated temperature in EN 1593 ^[6]	9
6.2.3 Gas pressurization bubble test [equivalent to immersion technique using pressurisation of the object in EN 1593 ^[6]	9
6.2.4 Liquid nitrogen bubble test	9
6.2.5 Acceptance criteria	9
6.3 Water pressurization test	9
Annex A (informative) Guidance for the choice of the tests to be carried out according to purpose and sealed source type	10
Bibliography	12

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 85, *Nuclear energy, nuclear technologies, and radiological protection*, Subcommittee SC 2, *Radiological protection*.

This second edition cancels and replaces the first edition (ISO 9978:1992), which has been technically revised. The main changes compared to the previous edition are as follows:

- [Clause 4](#): Revised to add text specifying factors to be considered in designing an effective leak testing regime for a particular type of sealed source;
- [Clause 4](#): Requirement added that personnel performing leak tests be appropriately trained and qualified, informative reference to ISO 9712 added;
- [Clause 4](#): Requirement added that measurement uncertainty shall be considered in sentencing non-binary test results;
- [Table 1](#) — “Threshold detection values and limiting values for different test methods” has been revised for clarity;
- [5.1](#): Informative reference to suitable assay techniques for immersion test liquid samples added: ISO 19361 and ISO 19581;
- [5.1.1](#), [5.1.2](#), [5.1.4](#): Composition of suitable immersion test liquids clarified;
- [5.3](#): Informative reference to suitable wipe testing techniques (ISO 7503-2) added and clarification that acceptance criteria is absolute without correction for wiping efficiency required;
- [6.1](#): Normative reference to ISO 20485 added for methods of helium leak testing and calculation of acceptance limits;
- [6.2](#): Cautionary text added to state that efficacy of tests assume ideal conditions for vision of bubbles;
- [6.2.1](#): Cautionary text added regarding bubble testing of self-heated sources;

- [A.1](#): Text expanded to clarify which tests to use under given circumstances.

iTeh STANDARD PREVIEW
(standards.iteh.ai)
Full standard:
<https://standards.iteh.ai/catalog/standards/sist/0ff56b1f-d75e-4d82-946c-5108917fd04d/iso-fdis-9978>

Introduction

The use of sealed sources has become so widespread that standards to guide the user, manufacturer and regulatory agencies are necessary. When establishing these standards, radiation protection is the prime consideration.

The purpose of this document, in conjunction with ISO 2919, is to minimise the risk to the public caused by leakage of radioactive material into the general environment.

Leakage test methods for sealed sources were standardised in the first edition of this document. The experience acquired since this date has necessitated the revision of this document.

iTeh STANDARD PREVIEW
(standards.iteh.ai)
Full standard:
<https://standards.iteh.ai/catalog/standards/sist/0ff56b1f-d75e-4d82-946c-5108917fd04d/iso-fdis-9978>

Radiation protection — Sealed sources — Leakage test methods

1 Scope

This document specifies the different leakage test methods for sealed sources. It gives a comprehensive set of procedures using radioactive and non-radioactive means.

This document applies to the following situations:

- leakage testing of test sources following design classification testing in accordance with ISO 2919^[1];
- production quality control testing of sealed sources;
- periodic inspections of the sealed sources performed at regular intervals, during the working life.

[Annex A](#) of this document gives guidance to the user in the choice of the most suitable method(s) according to situation and source type.

It is recognized that there can be circumstances where special tests, not described in this document, are required.

It is emphasized, however, that insofar as production, use, storage and transport of sealed radioactive sources are concerned, compliance with this document is no substitute for complying with the requirements of the relevant IAEA regulations and other relevant national regulations. It is also recognized that countries can enact statutory regulations which specify exemptions for tests, according to sealed source type, design, working environment, and activity (e.g., for very low activity reference sources where the total activity is less than the leakage test limit).

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 20485:2017, *Non-destructive testing — Leak testing — Tracer gas method*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

3.1

capsule

protective envelope, used to prevent leakage of radioactive material

3.2

dummy sealed source

facsimile of a sealed source, the capsule of which has the same construction and is made with exactly the same materials as those of the sealed source that it represents, but containing, in place of the radioactive material, a substance resembling it as closely as is practical in physical and chemical properties

3.3

leachable

soluble in water, yielding quantities greater than 0,1 mg/g in 100 ml of still water maintained at 50 °C for 4 h

3.4

leakage

transfer of contained radioactive material from the sealed source to the environment

3.5

leaktight

term applied to sealed sources which, after undergoing leakage testing, meet the acceptance criteria

Note 1 to entry: The acceptance criteria are given in [Table 1](#).

3.6

model designation

manufacturer's unique term (number, code or a combination of these) which is used to identify a specific design of sealed source

3.7

non-destructive test

test used to detect internal, surface and concealed defects or imperfections in materials, using techniques that do not damage or destroy the items being tested

3.8

non-leachable

insoluble in water, yielding quantities less than 0,1 mg/g in 100 ml of still water maintained at 50 °C for 4 h

3.9

sealed source

radioactive material sealed in a capsule or associated with a material to which it is closely bonded, this capsule or bonding material being strong enough to maintain leaktightness of the sealed source under the conditions of use and wear for which it was designed

3.10

simulated sealed source

facsimile of a sealed source, the capsule of which has the same construction and is made with exactly the same materials as those of the sealed source that it represents but it contains, in place of the radioactive material, a substance resembling it as closely as possible in physical and chemical properties and trace quantities of radioactive material

Note 1 to entry: The tracer is in a form soluble in a solvent which does not attack the capsule and has the maximum activity compatible with its use in a containment enclosure.

3.11

standard helium leakage rate

helium leakage rate at an upstream pressure of $10^5 \text{ Pa} \pm 5 \times 10^3 \text{ Pa}$ and a downstream pressure of 10^3 Pa or less at a temperature of $296 \text{ K} \pm 7 \text{ K}$ ($23 \text{ °C} \pm 7 \text{ °C}$)

Note 1 to entry: In this document, the unit Pascal cubic meter per second is used¹⁾.

1) $[1 \times 10^{-6} \text{ Pa} \cdot \text{m}^3 \cdot \text{s}^{-1} = 1 \mu\text{Pa} \cdot \text{m}^3 \cdot \text{s}^{-1} \approx 10^{-5} \text{ atm} \cdot \text{cm}^3 \cdot \text{s}^{-1} \approx 1 \times 10^{-5} \text{ mbar} \cdot \text{l} \cdot \text{s}^{-1} \approx 7,5 \times 10^{-3} \text{ lusec.}]$

3.12**test source**

sample used in the performance tests, having the same material and construction as sealed sources of the model for which classification is being established

Note 1 to entry: A test source may be a simulated sealed source, a dummy sealed source or production source.

Note 2 to entry: The performance tests are described in ISO 2919.

4 Requirements

The tests described in this document are all designed to test and verify that the sealed source is leaktight. However not all tests are applicable in all circumstances. Correct application and choice of test method and testing media is critically important in designing an effective leak test programme. Factors to be considered include:

- the chemical form of the active material if leak test is by radioactive means;
- the type of test liquid used in immersion tests;
- the number of encapsulations;
- the internal void volume when tests are carried out by volumetric means;
- the temperature of the sealed source;
- the suitability of the test method for the environment in which it is being performed;
- the reason for performing the test (integrity testing of a test source, production leakage tests, routine in service testing);
- the required sensitivity and acceptance criteria.

The test programme for test and production sealed sources should be considered as part of the design process and validated or justified as appropriate to demonstrate its effectiveness and sensitivity. This process may include the analysis of historic data.

The tests described in this document shall be designed, validated and carried out by competent and qualified persons who have demonstrable appropriate training in the applied test methods. For test methods by radioactive means, the persons shall also have appropriate training in radiation protection and measurement.

NOTE 1 Qualification and certification methods for non-destructive testing personnel can be found in ISO 9712^[2].

An evaluation should be made of uncertainty in the case of non-binary test results (e.g. radiation measurements on immersion test samples) and taken account of in sentencing the result.

Guidance for choosing suitable tests are specified in [Annex A](#).

According to the test type and the sealed source type, at least one of each of the tests described in [Clauses 5](#) and [6](#) should be carried out [see [Annex A](#) for the choice of the test(s)].

It should be noted that it is best practice to carry out more than one type of leakage test and also to perform a final wipe as a contamination check.

The tests described in this document do not form an exhaustive list, and other test methods may be developed. However, in the case where a special test, which is not described in this document, is carried out (see [Clause 1](#)), the organisation shall validate that the applied method is at least as effective as the corresponding method(s) given in this document in order to be able to claim compliance with this document.

At the conclusion of the performed test(s), the sealed source shall be considered to be leaktight if it complies with the acceptance criteria specified in [Table 1](#).

It has been asserted that there is correspondence between the acceptance criteria for volumetric and radioactive leak tests. Whilst there is no universally accepted basis for this assertion, experience has shown that sources meeting the acceptance criteria shown in [Table 1](#) have not subsequently been found to leak.

NOTE 2 A leakage rate of $10 \mu\text{Pa} \cdot \text{m}^3 \cdot \text{s}^{-1}$ for non-leachable solid contents and a rate of $0,1 \mu\text{Pa} \cdot \text{m}^3 \cdot \text{s}^{-1}$ for leachable solids and liquids was historically considered to be equivalent to the activity release limit of 2 000 Bq ($\approx 50 \text{ nCi}$)^[18].

NOTE 3 A further confirmation of the volumetric acceptance threshold is given by Reference [8]. A leakage rate of $10^{-7} \text{ atm} \cdot \text{cm}^3 \cdot \text{s}^{-1}$ or more based on dry air at 298 K (25 °C) and for a pressure difference of 1 atm against a vacuum of 10^{-2} atm (equivalent to or less) is considered to represent a loss of leaktightness, irrespective of the physical nature of the content.

Table 1 — Threshold detection values and limiting values for different test methods

Test method	Subclause	Threshold of detection ^a	Acceptance criteria	
			Non-leachable content	Leachable or gaseous content
Radioactive methods				
Immersion test (hot liquid)	5.1.1	(10 to 1) Bq	<200 Bq	<200 Bq
Immersion test (boiling liquid)	5.1.2	(10 to 1) Bq	<200 Bq	<200 Bq
Immersion test with a liquid scintillator	5.1.3	(10 to 1) Bq	<200 Bq	<200 Bq
Gaseous emanation test	5.2.1	(4 to 0,4) Bq	<i>Unsuitable</i>	<200 Bq (²²² Rn/12h)
Emanation test with a liquid scintillator	5.2.2	(0,4 to 0,004) Bq	<i>Unsuitable</i>	<200 Bq (²²² Rn/12h)
Wet wipe test	5.3.1	(10 to 1) Bq	<200 Bq	<200 Bq
Dry wipe test	5.3.2	(10 to 1) Bq	<200 Bq	<200 Bq
Non-radioactive methods – Helium tests Standard helium leakage rate				
Helium test (<i>He filling before sealing</i>)	6.1.1	(10^{-2} to 10^{-4}) $\mu\text{Pa} \cdot \text{m}^3 \cdot \text{s}^{-1}$	$<1 \mu\text{Pa} \cdot \text{m}^3 \cdot \text{s}^{-1}$	$<0,01 \mu\text{Pa} \cdot \text{m}^3 \cdot \text{s}^{-1}$
Helium pressurization test (<i>He bombing after sealing</i>)	6.1.2	(1 to 10^{-2}) $\mu\text{Pa} \cdot \text{m}^3 \cdot \text{s}^{-1}$	$<1 \mu\text{Pa} \cdot \text{m}^3 \cdot \text{s}^{-1}$	$<0,01 \mu\text{Pa} \cdot \text{m}^3 \cdot \text{s}^{-1}$
Non-radioactive methods – Bubble tests Corresponding standard helium leakage rate				
Vacuum bubble test	6.2.1	(10 to 1) $\mu\text{Pa} \cdot \text{m}^3 \cdot \text{s}^{-1b}$	No bubbles observed	<i>Not sensitive enough</i>
Hot-liquid bubble test	6.2.2	(50 to 5) $\mu\text{Pa} \cdot \text{m}^3 \cdot \text{s}^{-1b}$	<i>Not sensitive enough</i>	<i>Not sensitive enough</i>
Gas pressurization bubble test	6.2.3	(10 to 1) $\mu\text{Pa} \cdot \text{m}^3 \cdot \text{s}^{-1b}$	No bubbles observed	<i>Not sensitive enough</i>
Liquid nitrogen bubble test	6.2.4	(10^{-1} to 10^{-2}) $\mu\text{Pa} \cdot \text{m}^3 \cdot \text{s}^{-1b}$	No bubbles observed	No bubbles observed
Non-radioactive methods – Mass gain Mass gain of water [μg]				
Water pressurization test	6.3	10	Mass gain < 50	<i>Not sensitive enough</i>

^a The threshold of detection is expressed as a range; its upper end defines the smallest detectable leak under typical, well controlled industrial leak testing conditions and its lower end indicates the smallest detectable leak under excellent (ideal) industrial leak testing conditions. Smaller leaks than those indicated can be detected under laboratory conditions.

^b Threshold values shown for bubble tests are rough approximations of the corresponding standard helium leakage rates and are applicable only to single leaks under favourable visual conditions.