



Designation: E887 – 88 (Reapproved 2009)

Standard Test Method for Silica in Refuse-Derived Fuel (RDF) and RDF Ash¹

This standard is issued under the fixed designation E887; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This test method covers the determination of silica in RDF, RDF ash, fly ash, bottom ash, or slag.

1.2 The test method is an acid dehydration gravimetric procedure and is independent of interferences.

1.3 The values stated in SI units are to be regarded as the standard. The values given in parentheses are for information only.

1.4 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.* For hazard statement, see Section 6.

2. Referenced Documents

2.1 *ASTM Standards*:²

D1193 Specification for Reagent Water

E791 Test Method for Calculating Refuse-Derived Fuel Analysis Data from As-Determined to Different Bases

E829 Practice for Preparing Refuse-Derived Fuel (RDF) Laboratory Samples for Analysis (Withdrawn 2002)³

3. Summary of Test Method

3.1 Silicon compounds in RDF ash, fly ash, bottom ash, or slag are dissolved by alkali fusion and dehydrated with hydrochloric acid (HCl). Dehydration is completed by ignition, and the silica is volatilized as silicon tetrafluoride.

4. Apparatus

4.1 *Analytical Balance*, capable of weighing to 0.0001 g.

¹ This test method is under the jurisdiction of ASTM Committee D34 on Waste Management and is the direct responsibility of Subcommittee D34.03 on Treatment, Recovery and Reuse.

Current edition approved July 1, 2009. Published August 2009. Originally approved in 1982. Last previous edition approved in 2004 as E 887 – 88 (2004). DOI: 10.1520/E0887-88R09.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

³ The last approved version of this historical standard is referenced on www.astm.org.

4.2 *Muffle Furnace*—The furnace shall have an operating temperature of up to 1200°C.

4.3 *Hot Plate or Steam Bath*.

4.4 *Platinum Crucibles*, 35 to 85-mL capacity.

4.5 *Graphite Crucibles*, 35 to 85-mL capacity.

4.6 *Fused Quartz Dishes*, 35 to 85-mL capacity.

5. Reagents and Materials

5.1 *Purity of Reagents*—Reagent grade chemicals shall be used in this test. Unless otherwise indicated, it is intended that all reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society where such specifications are available.⁴ Other grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

5.2 *Purity of Water*—Unless otherwise indicated, reference to water shall be understood to mean at least Type III reagent water conforming to Specification D1193.

5.3 *Sodium Carbonate* (Na₂CO₃), anhydrous powder.

5.4 *Hydrochloric Acid* (HCl), concentrated, sp gr 1.19.

5.5 *Hydrochloric Acid* (1 + 3), Mix 1 volume of concentrated HCl with 3 volumes of water.

5.6 *Hydrochloric Acid* (1 + 1)—Mix 1 volume of concentrated HCl with 1 volume of water.

5.7 *Hydrochloric Acid* (1 + 99)—Mix 1 volume of concentrated HCl with 99 volumes of water.

5.8 *Sulfuric Acid* (1 + 1)—Mix 1 volume of concentrated sulfuric acid (H₂SO₄, sp gr 1.84) with 1 volume of water.

5.9 *Hydrofluoric Acid* (HF), concentrated 48 to 51 %.

6. Hazards

6.1 Due to the origins of RDF in municipal waste, common sense dictates that precautions should be observed when

⁴ *Reagent Chemicals, American Chemical Society Specifications*, American Chemical Society, Washington, DC. For Suggestions on the testing of reagents not listed by the American Chemical Society, see *Analar Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmacopeial Convention, Inc. (USPC), Rockville, MD.