
**Optics and photonics — Test method
for refractive index of optical
glasses —**

**Part 1:
Minimum deviation method**

*Optique et photonique — Méthode d'essai pour déterminer l'indice de
réfraction des verres optiques —*

Partie 1: Méthode de la déviation minimale

Document Preview

ISO 21395-1:2020

<https://standards.iteh.ai/catalog/standards/iso/255a1f8b-6b16-487a-918b-accdb02dbb25/iso-21395-1-2020>



iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO 21395-1:2020](https://standards.iteh.ai/catalog/standards/iso/255a1f8b-6b16-487a-918b-accdb02dbb25/iso-21395-1-2020)

<https://standards.iteh.ai/catalog/standards/iso/255a1f8b-6b16-487a-918b-accdb02dbb25/iso-21395-1-2020>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Principle	1
5 Measuring apparatus	2
5.1 General construction	2
5.2 Goniometer	3
5.3 Light source	3
5.4 Detector	4
6 Specimen prism	4
6.1 General	4
6.2 Dimensions	5
6.3 Apex angle	5
6.4 Flatness	5
7 Environmental condition of measurement	5
7.1 Temperature	5
7.2 Humidity	5
7.3 Atmospheric pressure	5
8 Measurement	6
8.1 Adjustment of the measurement specimen prism	6
8.2 Measurement of the apex angle, α	6
8.3 Measurement of the angle of minimum deviation, δ_{\min}	7
9 Indication	8
10 Test report	8
Annex A (informative) Calculation of principal dispersion, Abbe number, partial dispersion and relative partial dispersion	9
Annex B (informative) Dispersion formulae for calculation of refractive index at arbitrary wavelength	12
Annex C (informative) Correction of refractive index for temperature, humidity and atmospheric pressure	14
Annex D (informative) Other measurement methods of the apex angle	16
Annex E (informative) Other measurement methods of the angle of minimum deviation	18
Bibliography	21

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 172, *Optics and photonics*, Subcommittee SC 3, *Optical materials and components*.

A list of all parts in the ISO 21395 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

The refractive index of optical glasses has been measured by various methods, but up to now, an International Standard for the measurement has not been available. The refractive index of optical glasses is the most important characteristic for the optical elements to be manufactured from them. This document defines a suitable method for measuring the refractive index of optical glasses accurately and also helps to improve communication between raw optical glass suppliers and optical element manufacturers.

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO 21395-1:2020](https://standards.iteh.ai/catalog/standards/iso/255a1f8b-6b16-487a-918b-accdb02dbb25/iso-21395-1-2020)

<https://standards.iteh.ai/catalog/standards/iso/255a1f8b-6b16-487a-918b-accdb02dbb25/iso-21395-1-2020>

Optics and photonics — Test method for refractive index of optical glasses —

Part 1: Minimum deviation method

1 Scope

This document specifies the measuring method for the refractive index of optical glasses with the accuracy within 1×10^{-5} used in the spectral range from 365 nm to 2 400 nm.

Additional information on how to apply the refractive index in the dispersion and the various dispersion formulae of optical glasses is given in [Annex A](#) and [Annex B](#).

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

There are no normative references in this document.

3 Terms and definitions

No terms and definitions are listed in this document.

ISO and IEC maintain terminological databases for use in standardization at the following addresses.

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

4 Principle

As shown in [Figure 1](#), when the monochromatic light beam is refracted by the specimen prism at the angle of minimum deviation, the relative refractive index of the specimen prism to the air at the wavelength of the monochromatic light beam is described by the following [Formula \(1\)](#):

$$n_{\text{rel}} = \frac{\sin \frac{\alpha + \delta_{\text{min}}}{2}}{\sin \frac{\alpha}{2}} \quad (1)$$

where

- n_{rel} is the relative refractive index ($n_{\text{rel}} = n_{\text{abs}}/n_{\text{air}}$);
- n_{abs} is the absolute refractive index of the specimen prism;
- n_{air} is the refractive index of air;
- α is the apex angle;
- δ_{min} is the angle of minimum deviation.

As shown in [Figure 1](#), light enters the plane AB of the specimen prism at an angle of incidence (θ) and exits from the plane AC at an exit angle (θ'). The incident and the exiting light ray form the deviation angle (δ). With the deviation angle minimized, the incident and the exiting angles are equal. The smallest angle of deviation is called the angle of minimum deviation. The angle of minimum deviation (δ_{min}) and the apex angle (α) of the specimen prism are measured, and the refractive index is calculated using those angles. Formulae for the calculation of principal dispersion, Abbe number, partial dispersion and relative partial dispersion are given in [Annex A](#). The dispersion formulae that calculate the refractive index at the wavelength different from the measured wavelengths are given in [Annex B](#). The correction of the refractive index of optical glasses for temperature, humidity and atmospheric pressure of optical glasses are given in [Annex C](#).

NOTE When measuring the refractive index with this method, it is necessary to consider temperature, pressure, humidity and measurement errors. Expressions for the relations of these errors are described in ISO 17328. The dependence of the refractive index of air on temperature and pressure can be found in ISO 12123:2018, A.3.





Figure 1 — Principle of minimum deviation method

5 Measuring apparatus

5.1 General construction

The measuring apparatus is shown in [Figure 2](#).

**Key**

- | | |
|--|---|
| 1 light source | 4 rotating stage coupled with prism (5) |
| 2 collimator | 5 specimen prism |
| 3 goniometer coupled with telescope (6) and detector (7) | 6 telescope |
| | 7 detector |

Figure 2 — Schematic of minimum deviation method**5.2 Goniometer**

The goniometer shall provide the capability of reading the angle within ± 1 arc sec.

5.3 Light source

The light source should be a mercury, hydrogen, helium, rubidium, cesium or cadmium lamp, also He-Ne laser or Nd:YAG laser defined in ISO 7944. The spectral lines and their associated wavelengths are shown in [Table 1](#).

Light sources and their corresponding wavelengths not defined in [Table 1](#) are also applicable for measurement, but the spectral bandwidth of the light source and the accuracy/certainty of emission line (for example D line (589,3 nm)) should be checked before use.

Table 1 — Wavelength and spectral line of light source

Wavelength/nm	Spectral line	Light source
365,01	i	Mercury lamp
404,66	h	Mercury lamp
435,83	g	Mercury lamp
479,99	F'	Cadmium lamp
486,13	F	Hydrogen lamp
543,5	—	He-Ne laser
546,07	e	Mercury lamp
587,56	d	Helium lamp
632,8	—	He-Ne laser
643,85	C'	Cadmium lamp
656,27	C	Hydrogen lamp
706,52	r	Helium lamp

Table 1 (continued)

Wavelength/nm	Spectral line	Light source
780,00	—	Rubidium lamp
852,11	s	Cesium lamp
1 013,98	t	Mercury lamp
1 064,1	—	Nd:YAG laser
1 128,7	—	Mercury lamp
1 395,1	—	Mercury lamp
1 529,6	—	Mercury lamp
1 813,1	—	Mercury lamp
1 970,1	—	Mercury lamp
2 325,4	—	Mercury lamp


5.4 Detector

A general-type detector that is capable of detecting each wavelength or easily exchangeable for different spectral ranges should be used.

6 Specimen prism

6.1 General

An example of the shape of specimen prism is shown in Figure 3.



Key

- l length
- t thickness
- α apex angle

Figure 3 — Shape of the specimen prism