



SLOVENSKI STANDARD
oSIST prEN ISO 23298:2022

01-februar-2022

Zobozdravstvo - Preskusne metode za natančnost obdelave računalniško podprtih strojev za rezkanje (ISO/DIS 23298:2021)

Dentistry - Test methods for machining accuracy of computer-aided milling machines (ISO/DIS 23298:2021)

Zahnheilkunde - Prüfverfahren zur Bewertung der Genauigkeit von computergesteuerten Fräsmaschinen (ISO/DIS 23298:2021)

Médecine bucco-dentaire - Méthodes d'essai pour l'exactitude d'usinage des fraiseuses à commande numérique (ISO/DIS 23298:2021)

**Ta slovenski standard je istoveten z: oSIST prEN ISO 23298:2022
prEN ISO 23298**
<https://standards.iteh.ai/catalog/standards/sist/bf86f59b-cc88-42dd-9fb9-695b6498a563/osist-pren-iso-23298-2022>

ICS:

11.060.01 Zobozdravstvo na splošno Dentistry in general

oSIST prEN ISO 23298:2022

en,fr,de

**iTeh STANDARD
PREVIEW
(standards.iteh.ai)**

oSIST prEN ISO 23298:2022

<https://standards.iteh.ai/catalog/standards/sist/bf86f59b-cc88-42dd-9fb9-695b6498a563/osist-pren-iso-23298-2022>

DRAFT INTERNATIONAL STANDARD

ISO/DIS 23298

ISO/TC 106/SC 9

Secretariat: JISC

Voting begins on:
2021-12-10Voting terminates on:
2022-03-04

Dentistry — Test methods for machining accuracy of computer-aided milling machines

Médecine bucco-dentaire — Méthodes d'essai pour l'exactitude d'usinage des fraiseuses à commande numérique

ICS: 11.060.01

iTeh STANDARD PREVIEW (standards.iteh.ai)

[oSIST prEN ISO 23298:2022](https://standards.iteh.ai/catalog/standards/sist/bf86f59b-cc88-42dd-9fb9-695b6498a563/osist-pren-iso-23298-2022)

<https://standards.iteh.ai/catalog/standards/sist/bf86f59b-cc88-42dd-9fb9-695b6498a563/osist-pren-iso-23298-2022>

This document is circulated as received from the committee secretariat.

THIS DOCUMENT IS A DRAFT CIRCULATED FOR COMMENT AND APPROVAL. IT IS THEREFORE SUBJECT TO CHANGE AND MAY NOT BE REFERRED TO AS AN INTERNATIONAL STANDARD UNTIL PUBLISHED AS SUCH.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

ISO/CEN PARALLEL PROCESSING



Reference number
ISO/DIS 23298:2021(E)

© ISO 2021

**iTeh STANDARD
PREVIEW
(standards.iteh.ai)**

oSIST prEN ISO 23298:2022

<https://standards.iteh.ai/catalog/standards/sist/bf86f59b-cc88-42dd-9fb9-695b6498a563/osist-pren-iso-23298-2022>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2021

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

	Page
Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 General	2
5 Test methods	2
5.1 Metal die method.....	2
5.1.1 Target restorations.....	2
5.1.2 Apparatus.....	2
5.1.3 Measurement of metal dies.....	5
5.1.4 Preparation of three-dimensional data.....	5
5.1.5 Machining of restorations.....	7
5.1.6 Evaluation of accuracy.....	8
5.2 Test methods for software method.....	14
5.2.1 General.....	14
5.2.2 Test object.....	15
5.2.3 Equipment and apparatus.....	18
5.2.4 Machining of specimens.....	19
5.2.5 Measurement.....	20
5.2.6 Data alignment procedures.....	21
5.2.7 Data analysis procedure.....	22
5.2.8 Calculation of total errors.....	25
6 Test report	26
6.1 General information.....	26
6.2 Specific information.....	26
6.2.1 Die method.....	26
6.2.2 Software method.....	27
6.3 Averaged characteristic accuracy values.....	27
6.3.1 Die method.....	27
6.3.2 Software method.....	27
Annex A (informative) Flow chart of test method	28
Annex B (normative) Measurement of die set(s) and preparation of CAD data of target restoration(s)	30
Annex C (informative) Contents of test report	40
Bibliography	44

ISO/DIS 23298:2021(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 9, *Dental CAD/CAD systems*.

This document cancels and replaces ISO/TR 18845:2017 which have been technically revised.

The main changes compared to ISO/TR 18845:2017 are as follows:

- changing the type of document from Technical Report (TR) to International Standard (IS);
- specifying two test methods using metal dies and software as the normative test methods;
- clarification of selection guidance of test methods;
- revision of the detail procedures of both test methods based on the inter-laboratory test.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

Dental CAD/CAM systems have been successfully used for the fabrication of indirect dental restorations such as inlays, crowns and bridges. The accuracy of these restorations is one of the most important factors for their clinical success. This document provides standardized test methods to evaluate the machining accuracy of computer-aided milling machines which are used as a part of dental CAD/CAM systems and the information to be provided by the manufacturer. A flow chart of test method is shown in [Annex A](#).

There are two methods using metal dies or software to evaluate machining accuracy of the target restoration(s). Either or both test methods should be selected to evaluate the machining accuracy.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[oSIST prEN ISO 23298:2022](https://standards.iteh.ai/catalog/standards/sist/bf86f59b-cc88-42dd-9fb9-695b6498a563/osist-pren-iso-23298-2022)

<https://standards.iteh.ai/catalog/standards/sist/bf86f59b-cc88-42dd-9fb9-695b6498a563/osist-pren-iso-23298-2022>

**iTeh STANDARD
PREVIEW
(standards.iteh.ai)**

oSIST prEN ISO 23298:2022

<https://standards.iteh.ai/catalog/standards/sist/bf86f59b-cc88-42dd-9fb9-695b6498a563/osist-pren-iso-23298-2022>

Dentistry — Test methods for machining accuracy of computer-aided milling machines

1 Scope

This document specifies the test methods to evaluate the machining accuracy of computer-aided milling machines as a part of dental CAD/CAM systems, which fabricate dental restorations, e.g. inlays, crowns and bridges

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 1942, *Dentistry — Vocabulary*

ISO 18739, *Dentistry — Vocabulary of process chain for CAD/CAM systems*

ISO 25178-2, *Geometrical product specifications (GPS) — Surface texture: Areal — Part 2: Terms, definitions and surface texture parameters*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 1942, ISO 18739 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

3.1

computer-aided milling machine

computer-aided machining device designed for subtractive manufacturing of dental prostheses using rotary instruments for cutting and grinding

3.2

blank

material for machining on a *computer-aided milling machine* (3.1). It may be a *block* (3.3) or a *disk* (3.4)

3.3

block

cuboidal material with holding device for machining on a *computer-aided milling machine* (3.1)

3.4

disc

disc-shaped material for machining on a *computer-aided milling machine* (3.1)

3.5

stock material

material *blanks* (3.2) that are in stock for machining on a *computer-aided milling machine* (3.1)

ISO/DIS 23298:2021(E)

4 General

There are two methods to evaluate accuracy of the target restoration(s). The accuracy of target restoration(s) shall be evaluated using one or both of the test methods described in [Clause 5](#). The test method(s) selected and corresponding results shall be provided in the instruction for use, the technical manual or other means. When the machining accuracy is affected by the material, appropriate material(s) shall be tested. Testing shall be performed on each material type that the manufacturer indicates for use by the device. The metal die method ([5.1](#)) is a measurement method based on the marginal adaptability of a machined restoration to a master die. Measurements obtained using this method can be used to assess adaptability at restoration margins. The software method ([5.2](#)) is a measurement method based on a comparison of the scanned file of milled restoration to a master manufacturing file using reverse engineering software. Measurements obtained using this method can be used to assess restoration margin, intaglio and external surface accuracy.

5 Test methods

5.1 Metal die method

5.1.1 Target restorations

Three types of restorations, a) Class II inlay, b) crown and c) four-unit/element bridge, are the targets of this test method. Choose the restoration type(s) specified in the manufacturer's instructions for use; and technical manual. If any of the restoration types are not specified by the manufacturer's technical manual for the equipment being tested, this restoration type shall be eliminated from the test procedure.

NOTE This test method is designed adopting the same principle as the examination method of clinical marginal adaptation. The clinical adaptation is examined checking the discrepancy between the restoration and the cavity margin or between it and the shoulder margin of abutment.

<https://standards.iteh.ai/catalog/standards/sist/bf86f59b-cc88-42dd-9fb9-695b6498a563/osist-pren-iso-23298-2022>

5.1.2 Apparatus

<https://standards.iteh.ai/catalog/standards/sist/bf86f59b-cc88-42dd-9fb9-695b6498a563/osist-pren-iso-23298-2022>

5.1.2.1 Metal dies

2022

Two types of metal dies given in [Figure 1](#) (Class II inlay) and [Figure 2](#) (crown and four-unit/element bridge dies) are used both for the preparation of three-dimensional data (manufacturing data set) and the evaluation of the accuracy of restorations. Dies shall be constructed based on the drawings in [Figure 1](#) and [Figure 2](#), these dies consist of a non-malleable metal die and one or more removable structure(s) used for the evaluation of accuracy.

The diameter of the removable occlusal part shall be not less than the diameter of abutment and the difference of diameter shall be not more than 10 µm.


The surface roughness (Sa) of the die, excepting the surfaces which do not come in contact with the test specimens/machined restorations, shall be less than 2 µm. Refer to ISO 25178-2 and other parts for test methods.

If a mark for reference point is necessary, either a groove or a ridge, or both, may be placed on the part, but shall be placed so as to not influence the evaluation of the results.

The removal occlusal part and removable shoulder are used for preparation of three-dimensional data, but not used for evaluation of accuracy.

NOTE An example of the machining device to fabricate the dies is Vertical Center NEXUS 410B.

Dimensions in millimetres



oSIST prEN ISO 23298:2022


<https://standards.iteh.ai/catalog/standards/sist/bf86f59b-cc88-42dd-9fb9-695b6498a563/osist-pren-iso-23298-2022>

Key

- 1 metal die
- 2 removable part
- 3 positioning pin
- 4 fixing screw

Figure 1 — Die for class II inlay specimen

Dimensions in millimetres



Key

- 1 metal die
- 2 abutment
- 3 removable occlusal part
- 4 removable shoulder
- h_s height of removable shoulder

NOTE The recommended size of the height of removable shoulder (h_s) is $(3,6 \pm 0,05)$ mm.

Figure 2 — Die for crown and bridge specimen

5.1.2.2 Measuring devices to be used for metal dies

Measuring devices with accuracy of $\leq 2 \mu\text{m}$ shall be used for measurement of metal dies (5.1.3). Coordinate measuring machine (CMM) can be useful to measure the size of die.

NOTE An example of a CMM is America Strato-Apex 574.

5.1.2.3 Measuring devices to be used for discrepancy measurement

Measuring devices with accuracy of $\leq 5 \mu\text{m}$ shall be used for discrepancy measurement in 5.1.4. Three-dimensional measuring microscopes, displacement meters and digital micrometers can be used.

5.1.3 Measurement of metal dies

Each die shall be measured using a measuring device specified in 5.1.2.2 to confirm the shape and dimensions specified in Figure 1 or Figure 2. The specified dimensions of constructed die necessary to prepare CAD data shall be measured in accordance with Annex B. The measured data are used to prepare the three-dimensional data (see 5.1.4).

In case of the metal die for crown and bridge specimen, the height of removable shoulder (h_s in Figure 2), and the height from the upper surface of the removable shoulder (4 in Figure 2) to the upper surface of the removable occlusal part (3 in Figure 2) shall be measured.

5.1.4 Preparation of three-dimensional data

5.1.4.1 General

The surface to be in contact with the metal die of each specimen type is determined by the measurements of the dies made in 5.1.3. The external surfaces of each specimen type are determined by the sections 5.1.4.2 and 5.1.4.3.

5.1.4.2 Class II inlay

The shapes and sizes of test specimen of class II inlay shall conform to the cavity of metal die (see Figure 1). The occlusal and proximal surfaces shall be the same planes with the corresponding surface of the metal die.

5.1.4.3 Crown and bridges

The shapes and sizes of test specimen of the crown and the bridge shall conform to Figure 3 (crown) and Figure 4 (bridge). A mark to distinguish direction when placing the restoration on the metal die shall be made on the top surface of crown. In case of bridges, the mark shall be made on either crown.


5.1.4.4 Preparation of CAD data (STL data)

To fabricate the target restorations, a CAD data (STL data) for each of the restorations specified in 5.1.4.2 and 5.1.4.3 shall be prepared in accordance with Annex B. This CAD data shall then be processed by the CAM software to prepare the manufacturing data set. The CAD data shall not be modified by the CAM software.

The dimensions of any surfaces in contact with the die surfaces are obtained from the measuring process in 5.1.3. Other dimensions are determined from Figure 3 and Figure 4.

The CAD data shall be prepared to ensure that the restoration meets the die without an allowance for cement space.


Dimensions in millimetres



Key

- 1 metal die
- 2 abutment
- 3 removable occlusal part
- 4 removable shoulder
- 5 test specimen
- 6 mark to distinguish direction

Figure 3 — Test specimen of crown

**Key**

- 1 metal die
 - 2 abutment
 - 3 test specimen
 - 4 mark to distinguish direction
- <https://standards.iteh.ai/catalog/standards/sist/bf86f59b-9f88-421d-9fb9-695b6498a563/osist-pren-iso-23298-2022>

Figure 4 — Test specimen of bridge**5.1.5 Machining of restorations**

The prepared manufacturing data set shall be input into the computer-aided milling machine following the manufacturer's instruction. The CAM software shall use the same configuration and parameters as is usually delivered. The target restoration shall be machined using the material specimen (blank) following the manufacturer's instruction.

NOTE 1 A manufacturer means a natural or legal person actually manufacturing a computer-aided milling machine or a natural or legal person supplying necessary information to use the computer-aided milling machine.

The target restoration shall be the same size of the prepared manufacturing data set. CAM software contains a scaling factor to compensate for shrinkage of material during an additional process such as sintering. The CAM software scaling factor used in this test shall be 1,00.

This test is carried out using a computer-aided milling machine maintained in accordance with the manufacturer's instruction.