

---

**In vitro diagnostični preskusni sistemi - Informacije proizvajalca (označevanje) - 5. del: Diagnostični instrumenti in vitro za samopreskušanje (ISO/DIS 18113-5:2021)**

In vitro diagnostic medical devices - Information supplied by the manufacturer (labelling) - Part 5: In vitro diagnostic instruments for self-testing (ISO/DIS 18113-5:2021)

In-vitro-Diagnostika - Bereitstellung von Informationen durch den Hersteller - Teil 5: Geräte für in-vitro-diagnostische Untersuchungen zur Eigenanwendung (ISO/DIS 18113-5:2021)

**(standards.iteh.ai)**

Dispositifs médicaux de diagnostic in vitro - Informations fournies par le fabricant (étiquetage) - Partie 5: Instruments de diagnostic in vitro pour auto-tests (ISO/DIS 18113-5:2021)

<https://standards.iteh.ai/catalog/standards/sist/04df3406-6bc2-4da3-a58d-8dd47cc10112/osist-pren-iso-18113-5-2021>

**Ta slovenski standard je istoveten z: prEN ISO 18113-5**

---

**ICS:**

11.100.10	Diagnostični preskusni sistemi in vitro	In vitro diagnostic test systems
-----------	---	----------------------------------

**oSIST prEN ISO 18113-5:2021**

**en,fr,de**

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[oSIST prEN ISO 18113-5:2021](https://standards.iteh.ai/catalog/standards/sist/04df3406-6bc2-4da3-a58d-8dd47cc10112/osist-pren-iso-18113-5-2021)

<https://standards.iteh.ai/catalog/standards/sist/04df3406-6bc2-4da3-a58d-8dd47cc10112/osist-pren-iso-18113-5-2021>

# DRAFT INTERNATIONAL STANDARD

## ISO/DIS 18113-5

ISO/TC 212

Secretariat: ANSI

Voting begins on:  
2021-08-02Voting terminates on:  
2021-10-25

---

---

## In vitro diagnostic medical devices — Information supplied by the manufacturer (labelling) —

Part 5:

### In vitro diagnostic instruments for self-testing

*Dispositifs médicaux de diagnostic in vitro — Informations fournies par le fabricant (étiquetage) —  
Partie 5: Instruments de diagnostic in vitro pour auto-tests*

ICS: 11.100.10

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[oSIST prEN ISO 18113-5:2021](https://standards.iteh.ai/catalog/standards/sist/04df3406-6bc2-4da3-a58d-8dd47cc10112/osist-pren-iso-18113-5-2021)<https://standards.iteh.ai/catalog/standards/sist/04df3406-6bc2-4da3-a58d-8dd47cc10112/osist-pren-iso-18113-5-2021>

THIS DOCUMENT IS A DRAFT CIRCULATED FOR COMMENT AND APPROVAL. IT IS THEREFORE SUBJECT TO CHANGE AND MAY NOT BE REFERRED TO AS AN INTERNATIONAL STANDARD UNTIL PUBLISHED AS SUCH.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

This document is circulated as received from the committee secretariat.

**ISO/CEN PARALLEL PROCESSING**



Reference number  
ISO/DIS 18113-5:2021(E)

© ISO 2021

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[oSIST prEN ISO 18113-5:2021](https://standards.iteh.ai/catalog/standards/sist/04df3406-6bc2-4da3-a58d-8dd47cc10112/osist-pren-iso-18113-5-2021)  
<https://standards.iteh.ai/catalog/standards/sist/04df3406-6bc2-4da3-a58d-8dd47cc10112/osist-pren-iso-18113-5-2021>



### **COPYRIGHT PROTECTED DOCUMENT**

© ISO 2021

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

Page

<b>Foreword</b> .....	<b>iv</b>
<b>Introduction</b> .....	<b>v</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms and definitions</b> .....	<b>1</b>
<b>4 Essential requirements</b> .....	<b>2</b>
<b>5 Labels and marking</b> .....	<b>2</b>
5.1 General.....	2
5.2 Identification of the IVD instrument.....	2
5.2.1 IVD instrument name.....	2
5.2.2 Serial number.....	2
5.2.3 <i>in vitro</i> diagnostic use.....	3
5.2.4 Unique device identifier/identification (UDI).....	3
<b>6 Elements of the instructions for use</b> .....	<b>3</b>
<b>7 Content of the instructions for use</b> .....	<b>4</b>
7.1 Manufacturer.....	4
7.2 Identification of the IVD instrument.....	4
7.2.1 IVD instrument name.....	4
7.2.2 Module and software identification.....	5
7.3 Intended use/Intended purpose.....	5
7.4 Storage and handling.....	5
7.5 Warnings and precautions and/or measures to be taken and limitations of use regarding the instrument.....	5
7.6 Instrument installation.....	6
7.6.1 General.....	6
7.6.2 Action upon delivery.....	6
7.6.3 Bringing into operation.....	6
7.7 Principles of measurement.....	6
7.8 Performance of the IVD instrument.....	6
7.9 Limitations of use.....	6
7.10 Preparation prior to operation.....	7
7.11 Operating procedure.....	7
7.12 Control procedure.....	7
7.13 Reading of examination results.....	7
7.14 Special functions.....	7
7.15 Shut-down procedure.....	8
7.16 Disposal information.....	8
7.17 Maintenance.....	8
7.18 Troubleshooting.....	8
7.19 Follow-up action.....	8
7.20 Document control.....	8
<b>Bibliography</b> .....	<b>9</b>

## ISO/DIS 18113-5:2021(E)

### Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC [or ISO/PC] 212, *Clinical laboratory testing and in vitro diagnostic test systems*.

This second edition cancels and replaces the first edition (ISO 18113:2009), which has been technically revised.

The main changes compared to the previous edition are as follows:

- Updated text to reflect changes in regulations and provide examples for clarity
- Added information pertaining to UDI (Unique Device Identification)
- Updated the bibliography

In this document, the following verbal forms are used: — “shall” indicates a requirement; — “should” indicates a recommendation; — “may” indicates a permission; — “can” indicates a possibility or a capability. Information marked as “NOTE” is for guidance in understanding or clarifying the associated requirement.

A list of all parts in the ISO 18113 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

## Introduction

Manufacturers of *in vitro* diagnostic (IVD) instruments for self-testing supply users with information to enable the safe use and expected performance of their devices. Adequate instructions for use are essential for the safe and proper operation of IVD instruments. The type and level of detail varies according to the intended uses and country-specific regulations.

The International Medical Device Regulators Forum (IMDRF) encourages convergence of the evolution of regulatory systems for medical devices at the global level. Eliminating differences among regulatory jurisdictions could allow patients earlier access to new technologies and treatments. This document provides a basis for harmonization of labelling requirements for IVD instruments for self-testing.

This document is concerned solely with information supplied with IVD instruments and equipment intended for self-testing. It is intended to be used in conjunction with ISO 18113-1, which contains the general requirements for information supplied by the manufacturer and definitions of This document is intended to support the essential labelling requirements of all the IMDRF [7] partners, as well as other countries that have enacted or plan to enact labelling regulations for IVD medical devices.

For IVD instruments that are intended to be used as a system with reagents provided by the same manufacturer, document is also intended to be used together with ISO 18113-1 and ISO 18113-4.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[oSIST prEN ISO 18113-5:2021](https://standards.iteh.ai/catalog/standards/sist/04df3406-6bc2-4da3-a58d-8dd47cc10112/osist-pren-iso-18113-5-2021)

<https://standards.iteh.ai/catalog/standards/sist/04df3406-6bc2-4da3-a58d-8dd47cc10112/osist-pren-iso-18113-5-2021>

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[oSIST prEN ISO 18113-5:2021](https://standards.iteh.ai/catalog/standards/sist/04df3406-6bc2-4da3-a58d-8dd47cc10112/osist-pren-iso-18113-5-2021)

<https://standards.iteh.ai/catalog/standards/sist/04df3406-6bc2-4da3-a58d-8dd47cc10112/osist-pren-iso-18113-5-2021>



# In vitro diagnostic medical devices — Information supplied by the manufacturer (labelling) —

## Part 5: In vitro diagnostic instruments for self-testing

### 1 Scope

This document specifies requirements for information supplied by the manufacturer of IVD instruments intended for self-testing.

This document also applies to apparatus and equipment intended to be used with *in vitro* diagnostic (IVD) instruments for self-testing.

This document can also be applied to accessories.

This document does not apply to

- a) instructions for instrument servicing or repair;
- b) IVD reagents, including calibrators and control materials for use in control of the reagent;
- c) IVD instruments for professional use.

### 2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 14971, *Medical devices — Application of risk management to medical devices*

ISO 15223-1, *Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied — Part 1: General requirements*

ISO 18113-1, *In vitro diagnostic medical devices — Information supplied by the manufacturer (labelling) — Part 1: Terms, definitions and general requirements*

IEC 61010-1, *Safety requirements for electrical equipment for measurement, control and laboratory use — Part 1: General requirements*

IEC 61010-2-101, *Safety requirements for electrical equipment for measurement, control and laboratory use — Part 2-101: Particular requirements for in vitro diagnostic (IVD) medical equipment*

IEC 61326-2-6, *Electrical equipment for measurement, control and laboratory use — EMC requirements — Part 2-6: Particular requirements — In vitro diagnostic (IVD) medical equipment*

IEC 62366-1, *Medical devices — Part 1: Application of usability engineering to medical devices*

### 3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 18113-1 apply.

## ISO/DIS 18113-5:2021(E)

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

### 4 Essential requirements

The requirements of ISO 18113-1 apply.

ISO standards for specific IVD medical devices may also contain requirements for information supplied by the manufacturer.

EXAMPLES ISO 15197; ISO 17593.

### 5 Labels and marking

#### 5.1 General

The requirements of IEC 61010-1, IEC 61010-2-101 and IEC 61326-2-6 concerning labels and markings apply.

For the use of symbols, the requirements of ISO 15223-1 apply.

The labelling shall not contain any language regarding the manufacturer's liability in the case of damage or injury resulting from any use or malfunction of the medical device or IVD medical device that contradicts the laws or regulations in the jurisdiction of use.

The labelling shall not contain any disclaimers related to the safety and performance of the medical device or IVD medical device for its intended purpose that are incompatible with the laws or regulations in the jurisdiction of use, or the obligations of the manufacturer to design and manufacture a product that is safe and performs as intended throughout its expected lifetime.

#### 5.2 Identification of the IVD instrument

##### 5.2.1 IVD instrument name

The name or trade name of the IVD instrument shall be given.

When the name does not uniquely identify the IVD instrument, an additional means of identification shall also be given.

EXAMPLES Catalogue number, commodity number.

##### 5.2.2 Serial number

A unique serial number shall be given for IVD instruments.

All instruments covered by IEC 61010 require serial numbers.

Where serial numbers are not practical for apparatus, equipment or accessories intended to be used with IVD instruments, a batch code may be used instead.

EXAMPLE A primary sample receptacle would be assigned a batch code.