



**SLOVENSKI STANDARD**  
**oSIST prEN ISO 15004-2:2023**

**01-december-2023**

**Nadomešča:**

**SIST EN ISO 15004-2:2007**

---

**Oftalmični instrumenti - Osnovne zahteve in preskusne metode - 2. del: Zaščita pred nevarno svetlobo (ISO/DIS 15004-2:2023)**

Ophthalmic instruments - Fundamental requirements and test methods - Part 2: Light hazard protection (ISO/DIS 15004-2:2023)

Ophthalmische Instrumente - Grundlegende Anforderungen und Prüfverfahren - Teil 2: Schutz gegen Gefährdung durch Licht (ISO/DIS 15004-2:2023)

Instruments ophtalmiques - Exigences fondamentales et méthodes d'essai - Partie 2: Protection contre les dangers de la lumière (ISO/DIS 15004-2:2023)

**Ta slovenski standard je istoveten z: prEN ISO 15004-2**

<https://standards.iteh.ai/catalog/standards/sist/24f22144-1683-47ef-807a-2a4777b56e8d/osist-pren-iso-15004-2-2023>

---

**ICS:**

11.040.70      Oftalmološka oprema      Ophthalmic equipment

**oSIST prEN ISO 15004-2:2023**      **en,fr,de**



# DRAFT INTERNATIONAL STANDARD

## ISO/DIS 15004-2

ISO/TC 172/SC 7

Secretariat: DIN

Voting begins on:  
2023-09-11Voting terminates on:  
2023-12-04

---

---

## Ophthalmic instruments — Fundamental requirements and test methods —

### Part 2: Light hazard protection

*Instruments ophtalmiques — Exigences fondamentales et méthodes d'essai —  
Partie 2: Protection contre les dangers de la lumière*

ICS: 11.040.70

## iTeh Standards (<https://standards.iteh.ai>) Document Preview

[oSIST prEN ISO 15004-2:2023](https://standards.iteh.ai/catalog/standards/sist/24f22144-f683-47ef-807a-2a4777b56e8d/osist-pren-iso-15004-2-2023)<https://standards.iteh.ai/catalog/standards/sist/24f22144-f683-47ef-807a-2a4777b56e8d/osist-pren-iso-15004-2-2023>

This document is circulated as received from the committee secretariat.

THIS DOCUMENT IS A DRAFT CIRCULATED FOR COMMENT AND APPROVAL. IT IS THEREFORE SUBJECT TO CHANGE AND MAY NOT BE REFERRED TO AS AN INTERNATIONAL STANDARD UNTIL PUBLISHED AS SUCH.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

### ISO/CEN PARALLEL PROCESSING



Reference number  
ISO/DIS 15004-2:2023(E)

© ISO 2023

iTeh Standards  
(<https://standards.iteh.ai>)  
Document Preview

[oSIST prEN ISO 15004-2:2023](https://standards.iteh.ai/catalog/standards/sist/24f22144-f683-47ef-807a-2a4777b56e8d/osist-pren-iso-15004-2-2023)

<https://standards.iteh.ai/catalog/standards/sist/24f22144-f683-47ef-807a-2a4777b56e8d/osist-pren-iso-15004-2-2023>



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2023

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

	Page
Foreword.....	v
Introduction.....	vi
<b>1 Scope.....</b>	<b>1</b>
<b>2 Normative references.....</b>	<b>1</b>
<b>3 Terms and definitions.....</b>	<b>1</b>
<b>4 Classification principles.....</b>	<b>7</b>
<b>5 Requirements for Classification.....</b>	<b>9</b>
5.1 General.....	9
5.2 Requirements for classification as a Group 1 instrument.....	9
5.3 Requirements for Classification as a Group 2 instrument.....	10
5.4 Emission limits for Group 1 classification.....	11
5.4.1 Continuous wave instruments.....	11
5.4.2 Pulsed instruments.....	11
5.4.3 Time-limited instruments.....	11
5.4.4 Dose limited instruments.....	11
5.4.5 Scanning instruments.....	11
5.4.6 Instruments with multiple sources.....	12
5.4.7 Instruments for long-term repetitive daily use.....	12
5.5 Recommended Maximum Exposure (RME) values for Group 2 instruments.....	12
5.5.1 Continuous wave instruments.....	12
5.5.2 Pulsed instruments.....	12
5.5.3 Time-limited instruments.....	13
5.5.4 Scanning instruments.....	13
5.5.5 Instruments with multiple sources.....	13
5.5.6 Instruments for long-term repetitive daily use.....	13
<b>6 Test methods.....</b>	<b>19</b>
6.1 General.....	19
6.2 Classification of instruments into Group 1 or Group 2.....	19
6.3 Spectral Measurements.....	19
6.4 Group 2 instruments: Determination of time and number of pulses to reach recommended maximum exposure.....	20
6.4.1 Determination of time $t_{\max}$ to reach the recommended maximum exposure for weighted corneal and lenticular ultraviolet radiation radiant exposure, $H_{S-CL}$ .....	20
6.4.2 Determination of time $t_{\max}$ to reach the recommended maximum exposure for unweighted corneal and lenticular ultraviolet radiation radiant exposure, $H_{UV-CL}$ .....	20
6.4.3 Determination of time $t_{\max}$ to reach the recommended maximum exposure for photochemical aphakic retinal radiant exposure.....	20
6.4.4 Determination of the number of pulses necessary to reach the recommended maximum exposure for photochemical aphakic retinal exposure, $n_{\max}$ (for pulsed instruments).....	21
<b>7 Information supplied by the manufacturer.....</b>	<b>21</b>
7.1 Information provided on request.....	21
7.2 Accompanying documents.....	21
7.2.1 General.....	21
7.2.2 Dose-limited instruments.....	21
7.2.3 Cautionary Statements.....	21
7.3 Marking.....	24
7.3.1 Radiation output.....	24
7.3.2 Consult accompanying documents.....	24
7.3.3 Safety information for optical radiation.....	24

**ISO/DIS 15004-2:2023(E)**

<b>Annex A (normative) Spectral weighting functions</b> .....	<b>26</b>
<b>Annex B (informative) Ophthalmic instruments with product-related International Standards that contain a specific light hazard section</b> .....	<b>33</b>
<b>Annex D (informative) Measurement methods for radiance/irradiance</b> .....	<b>37</b>
<b>Annex E (informative) Guidance on the direct measurement of irradiance</b> .....	<b>42</b>
<b>Annex G (informative) Safety signs</b> .....	<b>55</b>
<b>Annex ZA (informative) Relationship between this European Standard and the General Safety and Performance Requirements of Regulation (EU) 2017/745 aimed to be covered</b> .....	<b>56</b>
<b>Bibliography</b> .....	<b>58</b>

iTeh Standards  
(<https://standards.iteh.ai>)  
Document Preview

[oSIST prEN ISO 15004-2:2023](https://standards.iteh.ai/catalog/standards/sist/24f22144-f683-47ef-807a-2a4777b56e8d/osist-pren-iso-15004-2-2023)

<https://standards.iteh.ai/catalog/standards/sist/24f22144-f683-47ef-807a-2a4777b56e8d/osist-pren-iso-15004-2-2023>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 172, *Optics and photonics, Subcommittee SC 7, Ophthalmic optics and instruments*.

This second edition cancels and replaces the first edition (ISO 15004-2:2007), which has been technically revised.

The main changes are as follows:

- The terms and definitions include dose limited instruments and time limited instruments;
- The safe exposure limits have been reorganized into a single 4-part table ([Table 2a-2d](#)), and the associated measurement conditions have been reorganized into a companion table ([Table 3](#));
- A number of Group 1 exposure limits and Group 2 recommended maximum exposures (RMEs) have been updated to conform to recent research and relevant standards;
- The language has been clarified and simplified throughout the document, and a flowchart has been added as a guide to make the standard more accessible to first-time users;
- Clauses and associated exposure limits have been added for long-term repetitive exposures, such as may apply to extensive use of head-mounted displays by people with visual impairments;
- Provisions have been added to ensure that the exposure limits and RMEs applicable to specific devices are easily accessible to end users.

A list of all parts in the ISO 15004 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

## ISO/DIS 15004-2:2023(E)

### Introduction

Ophthalmic instruments are classified into two groups to distinguish instruments that can present a potential hazard from those that cannot. Those that present a potential light hazard need to undergo significantly greater levels of testing than those that do not. The two groups are named **Group 1** and **Group 2**. They are defined as follows:

**Group 1 instruments: ophthalmic instruments for which no potential light hazard exists when used as intended (see [Clause 4](#)).**

**Group 2 instruments: ophthalmic instruments for which a potential light hazard exists (see [Clause 4](#)).**

Limits and guidelines for optical radiation exposure of the eye during ophthalmic examination can differ from those of non-ophthalmic applications. They can be more restrictive because of pupils dilation or retinal image stabilization, or less restrictive based on benefit/risk ratios. Furthermore, interruptions of exposure during surgical procedures mitigate the risk.

All group 2 instruments pose a potential risk of injury at the upper emission values of the instrument. This is true for both photochemical (where time is critical) and thermal, where transmission and absorption values can vary. Clinical judgement of individual susceptibility is also required.

NOTE 1 The basic limits and guidelines in this document are based on the International Commission on Non-Ionizing Radiation Protection (ICNIRP) guidelines for human exposure to optical radiation. They also cover cases where pupils are dilated, and/or the image is stabilized on the retina during ophthalmic examinations.

NOTE 2 This document provides exposure limits for ocular tissues. The exposures can be calculated based upon the measured instrument emissions.

The flow chart in [Figure 1](#) at the beginning of [clause 4](#) provides guidance on how to apply this document to any device to be tested and/or designed for light hazards compliance.

## Document Preview

[oSIST prEN ISO 15004-2:2023](#)

<https://standards.iteh.ai/catalog/standards/sist/24f22144-f683-47ef-807a-2a4777b56e8d/osist-pren-iso-15004-2-2023>



# Ophthalmic instruments — Fundamental requirements and test methods —

## Part 2: Light hazard protection

### 1 Scope

This document specifies fundamental requirements for optical radiation safety for ophthalmic instruments and is applicable to all ophthalmic instruments that direct optical radiation into or at the eye. It is also applicable to all new and emerging ophthalmic instruments that direct optical radiation into or at the eye, as well as to those portions of therapeutic or surgical systems that direct optical radiation into or at the eye for diagnostic, illumination, measurement, imaging or alignment purposes.

NOTE For the purpose of this document, optical radiation relates to the wavelength range of 250 nm – 2 500 nm.

This document does not apply to therapeutic radiation. However, in the case of the treatment beams of therapeutic devices, when conducting risk assessments for non-target tissues, the limits given in this document may be applied to those parts of the treatment beam that strike non-target tissue.

Where vertical (instrument-specific) International Standards contain specific light hazard requirements different from those given in ISO 15004-2, then those in the vertical International Standard shall take precedence.

This document classifies ophthalmic instruments into either Group 1 or Group 2 to distinguish instruments that are non-hazardous from those that are potentially hazardous.

### 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 15004-1, *Ophthalmic instruments — Fundamental requirements and test methods — Part 1: General requirements applicable to all ophthalmic instruments*

### 3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 15004-1 and the following apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

## ISO/DIS 15004-2:2023(E)

### 3.1 aperture

aperture stop

opening that defines the area over which average optical emission is measured

Note 1 to entry: For spectral irradiance measurements this opening is usually the entrance of a small integrating sphere placed in front of the radiometer/spectroradiometer entrance slit.

### 3.2 continuous exposure

CW

radiant exposure equal to or greater than 0,25 s duration

### 3.3 continuous wave instrument

CW instrument

ophthalmic instrument that is designed to expose the eye to one or more continuous wave radiation sources

### 3.4 continuous wave radiation source

CW radiation source

radiation source that is, or can be, operated with a continuous output for a time that can be equal or greater than 0,25 s (i.e., a non-pulsed radiation source)

### 3.5 dose-limited instrument

ophthalmic instrument whose emission could exceed the Group 1 exposure limits, but through its design and construction and accounting for multiple exposures within a 24-hour period, cannot under reasonably foreseeable conditions expose any given eye to radiation that exceeds the Group 1 cumulative exposure limits up to 30 000 s given in [Table 2a](#)

Note 1 to entry: This instrument would otherwise be a Group 2 instrument; for example, some UV-Fluorescent diagnostic instruments.

### 3.6 effective aperture

portion of the aperture that limits the amount of light delivered to the retina

Note 1 to entry: For an obscured or noncircular aperture, the effective aperture is defined as the uniformly illuminated, unobscured circular aperture that transmits the same radiant flux.

### 3.7 emission limit

maximum permitted value of optical radiation output to which the eye potentially may be exposed

### 3.8 exposure limit

maximum value of optical radiation exposed to which an ocular tissue may be potentially exposed

### 3.9 endoilluminator

instrument consisting of a light source and an associated fibre optic light guide that is intended for insertion into the eye to illuminate any portion of the interior of the eye

### 3.10 field-of-view

conical solid angle as “seen” by the detector, such as the eye or the radiometer/ spectroradiometer, over which the detector receives radiation

Note 1 to entry: This is also referred to as acceptance angle.

Note 2 to entry: The field-of-view denotes the angle over which radiance is averaged (sampled) and should not be confused with the angular subtense of the source  $\alpha$ , which denotes source size.

Note 3 to entry: In this document, a plane angle is used to describe a circular symmetric solid angle field of view.

### 3.11

#### **Group 1 instrument**

ophthalmic instrument for which no potential optical radiation hazard exists and that can be shown to fulfil Group 1 requirements

### 3.12

#### **Group 2 instrument**

ophthalmic instrument for which a potential optical radiation hazard exists and that does not fulfil Group 1 requirements but does fulfil the Group 2 requirements

Note 1 to entry: Since Group 2 instruments do not fulfil the requirements of Group 1 instruments, special precautions are necessary.

### 3.13

#### **immobilized eye**

an eye that is physically fixed or whose movements are compensated so that a retinal image does not move

Note 1 to entry: For the purpose of this document, this term also refers to a retinal image stabilized by eye tracking technology. It does not refer to an eye that is maintaining voluntary fixation, e.g., during an ophthalmic examination.

### 3.14

#### **irradiance**

$E$

at a point on a surface, quotient of the radiant power  $d\Phi$  incident on an element of a surface containing the point, by the area  $dA$  of that element, i.e. given in [equation \(1\)](#),

$$E = \frac{d\Phi}{dA} \quad (1)$$

Note 1 to entry: Irradiance is expressed in units of watts per square centimetre, W/cm<sup>2</sup>.

### 3.15

#### **manufacturer**

natural or legal person with responsibility for design and/or manufacture of an ophthalmic instrument with the intention of making the ophthalmic instrument available for use, under his/her name, whether or not such an ophthalmic instrument is designed and/or manufactured by that person himself/herself or on his/her behalf by another person(s)

[SOURCE: ISO 13485:2016, 3.10, modified — The word "medical device" has been replaced by "ophthalmic instrument", neutered.]

### 3.16

#### **maximum intensity**

highest optical radiation intensity the instrument can deliver under normal conditions

### 3.17

#### **operation microscope**

stereo microscope used for observation of surgical and other medical procedures, consisting of an illumination system and an observation system

### 3.18

#### **optical radiation hazard**

hazard related to the risk of damage to the eye by exposure to optical radiant energy

## ISO/DIS 15004-2:2023(E)

### 3.19

#### photoretinitis

photochemically-induced injury to the retina resulting from retinal radiant exposure

Note 1 to entry: The term photic maculopathy is also used to describe photoretinitis in the macular area of the retina.

### 3.20

#### pulse duration

The time interval equal to the full width at half maximum of the pulse.

### 3.21

#### pulsed instrument

ophthalmic instrument that emits radiation in the form of a single exposure of known duration of less than 0,25 s or a train of pulses where each pulse in that train has a duration of less than 0,25 s

Note 1 to entry: A continuous train of pulses or modulated radiant energy where the peak radiated power is at least ten times the minimum radiated power is considered to be pulsed radiation.

### 3.22

#### radiance

$L$

electromagnetic radiation emitted in a given direction at a given point of a real or imaginary surface defined by [equation \(2\)](#):

$$L = \frac{d\Phi}{dA \times \cos\theta \times d\Omega} \quad (2)$$

where

$d\Phi$  is the radiant power transmitted by an elementary beam passing through the given point and propagating in the solid angle  $d\Omega$  containing the given direction;

$dA$  is the area of a section of that beam containing the given point;

$\theta$  is the angle between the normal to that section and the direction of the beam.

Note 1 to entry: The same definition holds for the time-integrated radiance  $L_t$  if, in the equation for  $L$ , the radiant power  $d\Phi$  is replaced by the radiant energy  $dQ$ .

Note 2 to entry: Radiance is expressed in watts per steradian square centimetre,  $W/(sr \cdot cm^2)$ ; time-integrated radiance is expressed in Joules per steradian square centimetre,  $J/(sr \cdot cm^2)$ .

### 3.23

#### radiant exposure

$H$

at a point of a surface, for a given duration quotient of the radiant energy,  $dQ$ , incident on an element of a surface containing the point over the given duration by unit area  $dA$  of that element as expressed in [equation \(3\)](#):

$$H = \frac{dQ}{dA} \quad (3)$$

Equivalently, the radiant exposure is defined as the integral of the irradiance,  $E$ , at a given point over duration  $\Delta t$  as expressed in [equation \(4\)](#):

$$H = \int_{\Delta t} E \times dt \quad (4)$$

Note 1 to entry: Radiant exposure is expressed in Joules per square centimetre,  $J/cm^2$ .

### 3.24 recommended maximum exposure RME

Group 2 radiant exposure value above which a significant risk of injury exists

### 3.25 scanning instrument

instrument that emits radiation having a time-varying direction, origin or pattern of propagation with regard to a stationary frame of reference

### 3.26 slit-lamp microscope

instrument consisting of a microscope and a swiveling illumination system used in ocular examinations that illuminates the eye with a variable-width beam of light

### 3.27 spectral irradiance

$E_\lambda$   
quotient of the spectral radiant power  $d\Phi(\lambda)$  in a wavelength interval  $d\lambda$ , incident on an element of a surface, by the area  $dA$  of that element and by the wavelength interval  $d\lambda$ , as expressed in [equation \(5\)](#):

$$E_\lambda = \frac{d^2\Phi(\lambda)}{dA \times d\lambda} \quad (5)$$

Note 1 to entry: Spectral irradiance is expressed in watts per square centimetre nanometre,  $W/(cm^2 \cdot nm)$ .

### 3.28 spectral radiance

$L_\lambda$   
for a wavelength interval  $d\lambda$ , in a given direction at a given point ratio of the spectral radiant power  $d\Phi(\lambda)$  passing through that point and propagating within the solid angle  $d\Omega$  in the given direction, to the product of the wavelength interval  $d\lambda$  and the areas of a section of that beam on a plane perpendicular to this direction ( $\cos \theta dA$ ) containing the given point and to the solid angle  $d\Omega$ , as expressed in [equation \(6\)](#):

$$L_\lambda = \frac{d^2\Phi(\lambda)}{dA \times \cos \theta \times d\Omega \times d\lambda} \quad (6)$$

Note 1 to entry: Spectral radiance is expressed in watts per steradian square centimetre nanometre,  $W/(sr \cdot cm^2 \cdot nm)$ .

### 3.29 spectral radiant exposure

$H_\lambda$   
quotient of the spectral radiant exposure  $dQ(\lambda)$  in a wavelength interval  $d\lambda$ , incident on an element of a surface, by the area  $dA$  of that element and by the wavelength interval  $d\lambda$ , given in [equation \(7\)](#):

$$H_\lambda = \frac{d^2Q(\lambda)}{dA \times d\lambda} \quad (7)$$

Note 1 to entry: Spectral radiant exposure is expressed in joules per square centimetre nanometre,  $J/(cm^2 \cdot nm)$ .

## ISO/DIS 15004-2:2023(E)

## 3.30

**spot size**

the arithmetic average of the full-width half-maximum values of the minimum and maximum dimensions of an irradiated area where the values are limited to no more than 1,7 mm and no less than 0,03 mm

Note 1 to entry: For the irradiated area with a non-circular cross-section, the value of spot size shall be determined by averaging the maximum and minimum cross-section lengths.

## 3.30.1

**small spot**

spot with a spot size  $\leq 0,03$  mm

## 3.30.2

**intermediate spot**

spot with a spot size  $> 0,03$  mm and  $< 1,7$  mm

## 3.30.3

**large spot**

spot with a spot size  $\geq 1,7$  mm

## 3.31

**time-limited instrument**

ophthalmic instrument for which the maximum exposure duration for each subject per 24-hour period is limited by the manufacturer for each intended use and specified in the instructions for use

**Table 1 — Symbols, Quantities and Units**

Symbol	Quantity	Unit
$E$	irradiance (at a point on a surface)	W/cm <sup>2</sup>
$E_{\lambda}$	spectral irradiance	W/(cm <sup>2</sup> ·nm)
$L$	radiance (in a given direction at a given point of a real or imaginary surface)	W/(sr·cm <sup>2</sup> )
$L_{\lambda}$	spectral radiance (for a wavelength interval $d\lambda$ , in a given direction at a given point)	W/(sr·cm <sup>2</sup> ·nm)
$L_i$	time-integrated radiance	J/(sr·cm <sup>2</sup> )
$H$	radiant exposure (at a point of a surface, for a given duration)	J/cm <sup>2</sup>
$H_{\lambda}$	spectral radiant exposure; the time-integral of spectral irradiance	J/(cm <sup>2</sup> ·nm)
$E_{S-CL}$	$S(\lambda)$ weighted corneal and lenticular ultraviolet radiation irradiance	W/cm <sup>2</sup>
$E_{UV-CL}$	unweighted corneal and lenticular ultraviolet radiation irradiance	W/cm <sup>2</sup>
$E_{A-R}$	$A(\lambda)$ weighted retinal irradiance	W/cm <sup>2</sup>
$E_{B-R}$	$B(\lambda)$ weighted retinal irradiance	W/cm <sup>2</sup>
$E_{IR-CL}$	unweighted corneal and lenticular infrared radiation irradiance	W/cm <sup>2</sup>
$E_{VIR-AS}$	unweighted anterior segment visible and infrared radiation irradiance	W/cm <sup>2</sup>
$E_{VIR-I}$	$R(\lambda)$ weighted iris visible and infrared radiation thermal irradiance	W/cm <sup>2</sup>
$E_{VIR-R}$	$R(\lambda)$ weighted retinal visible and infrared radiation thermal irradiance	W/cm <sup>2</sup>
$L_{A-R}$	$A(\lambda)$ weighted retinal radiance	W/(sr·cm <sup>2</sup> )
$L_{i,A-R}$	$A(\lambda)$ weighted retinal time-integrated radiance	J/(sr·cm <sup>2</sup> )
$L_{i,VIR-R}$	$R(\lambda)$ weighted, retinal visible and infrared radiation time-integrated radiance	J/(sr·cm <sup>2</sup> )
$L_V$	Luminance (in a given direction point of a real or imaginary surface)	cd/cm <sup>2</sup>
$L_{VIR-R}$	$R(\lambda)$ weighted retinal visible and infrared radiation radiance	W/(sr·cm <sup>2</sup> )

\*  $R(\lambda)$  refers to  $R_A(\lambda)$  or  $R_P(\lambda)$  as appropriate.

NOTE: For simplicity, " $\tau(\lambda)$ " is assumed to be 1,0 and " $\tau_p(\lambda)$ " is the re-normalized crystalline lens transmittance as specified in [Table A.1](#).