# INTERNATIONAL STANDARD

ISO 6320

Fifth edition 2017-02

## Animal and vegetable fats and oils — Determination of refractive index

Corps gras d'origines animale et végétale — Détermination de l'indice de réfraction

### iTeh Standards (https://standards.iteh.ai) Document Preview

ISO 6320:2017

https://standards.iteh.ai/catalog/standards/iso/b96cc3e7-1a74-42a5-9943-530819a7t6t9/iso-6320-2017



### iTeh Standards (https://standards.iteh.ai) Document Preview

ISO 6320:2017

https://standards.iteh.ai/catalog/standards/iso/b96cc3e7-1a74-42a5-9943-5308f9a7f6f9/iso-6320-2017



#### COPYRIGHT PROTECTED DOCUMENT

© ISO 2017, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office Ch. de Blandonnet 8 • CP 401 CH-1214 Vernier, Geneva, Switzerland Tel. +41 22 749 01 11 Fax +41 22 749 09 47 copyright@iso.org www.iso.org

Contents		Page
Fore	eword	iv
1	Scope	1
2	Normative reference	1
3	Terms and definitions	1
4	Principle	1
5	Reagents	2
6	Apparatus	2
7	Sampling	2
8	Preparation of test sample	2
9	Procedure	2
	9.1 Calibration of the instrument	
	9.2 Determination	
10	Calculation	3
11	Precision	4
	11.1 Interlaboratory test	4
	11.2 Repeatability	4
	11.3 Reproducibility Scandards	4
12	Test report	
Anne	ex A (normative) Results of an interlaboratory test	5
Ribliography Dogument Droy, our		6

#### ISO 6320:2017

011/10/https://standards.iteh.ai/catalog/standards/iso/b96cc3e7-1a74-42a5-9943-5308t9a7t6t9/iso-6320

#### **Foreword**

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see <a href="www.iso.org/directives">www.iso.org/directives</a>).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see <a href="https://www.iso.org/patents">www.iso.org/patents</a>).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: <a href="https://www.iso.org/iso/foreword.html">www.iso.org/iso/foreword.html</a>.

This document was prepared by Technical Committee ISO/TC 34, *Food products*, Subcommittee SC 11, *Animal and vegetable fats and oils*.

This fifth edition cancels and replaces the fourth edition (ISO 6320:2000), of which it constitutes a minor revision by the addition of an exclusion for fat coming from milk and milk products.

https://standards.iteh.ai/catalog/standards/iso/b96cc3e7-1a74-42a5-9943-5308f9a7f6f9/iso-6320-2017

## Animal and vegetable fats and oils — Determination of refractive index

#### 1 Scope

This document specifies a method for the determination of the refractive index of animal and vegetable fats and oils.

Milk and milk products (or fat coming from milk and milk products) are excluded from the scope of this document.

#### 2 Normative reference

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 661, Animal and vegetable fats and oils — Preparation of test sample

ISO 5725-1, Accuracy (trueness and precision) of measurement methods and results — Part 1: General principles and definition

ISO 5725-2, Accuracy (trueness and precision) of measurement methods and results — Part 2: Basic method for the determination of repeatability and reproducibility of a standard measurement method

#### 3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <a href="http://www.electropedia.org/">http://www.electropedia.org/</a>
- ISO Online browsing platform: available at <a href="http://www.iso.org/obp">http://www.iso.org/obp</a>

#### 3.1

#### refractive index (of a medium)

ratio of the velocity of light of a definite wavelength in a vacuum to its velocity in the medium

Note 1 to entry: In practice, the velocity of light in air is used in place of that in a vacuum and, unless otherwise specified, the selected wavelength is the mean wavelength of the sodium D lines (589,6 nm).

Note 2 to entry: The refractive index of a given substance varies with the wavelength of the incident light and with temperature. The notation used is  $n_{\rm D}^t$ , where t is the temperature in degrees Celsius.

#### 4 Principle

By means of a suitable refractometer, the refractive index of a liquid sample is measured at a specified temperature.

#### 5 Reagents

Use only reagents of recognized analytical grade and distilled or demineralized water or water of equivalent purity.

- **5.1 Ethyl laurate**, of quality suitable for refractometry, and of known refractive index.
- **5.2 Hexane**, or other suitable solvents, such as light petroleum, acetone or toluene, for cleaning the prism of the refractometer.

#### 6 Apparatus

Usual laboratory equipment and, in particular, the following.

Modern digital refractometers may be used. The manufacturer's instructions should be followed in this case.

**6.1 Refractometer**, for example of the Abbé type, suitable for measurements of refractive index to within  $\pm 0,000$  1 over the range  $n_D$  = 1,300 to  $n_D$  = 1,700.

#### 6.2 Light source: sodium vapour lamp

White light can also be used if the refractometer is fitted with an achromatic compensation system.

- **6.3 Glass plate**, of known refractive index.
- **6.4 Water bath**, thermostatically controlled, with a circulation pump, and capable of being maintained to the nearest  $\pm 0.1$  °C.
- **6.5 Water bath**, capable of being maintained at the temperature at which the measurements are to be made (in the case of solid samples).

#### 7 Sampling

It is important that the laboratory receive a sample which is truly representative and has not been damaged or changed during transportation or storage.

Sampling is not part of the method specified in this document. A recommended sampling method is given in ISO 5555.

#### 8 Preparation of test sample

Prepare the test sample in accordance with ISO 661.

The refractive index shall be determined on dried and filtered fats and oils.

In the case of a solid sample, transfer the sample prepared in accordance with ISO 661 to a suitable container and place it in the water bath  $(\underline{6.5})$ , set at the temperature at which the measurements are to be made. Allow sufficient time for the temperature of the sample to stabilize.

#### 9 Procedure

NOTE If it is required to check whether the repeatability requirement (see  $\underline{11.2}$ ) is met, carry out two single determinations in accordance with  $\underline{9.1}$  and  $\underline{9.2}$ .