
Oprema, ki se uporablja za pripravo pitne vode v stavbah - Naprave z živosrebrnimi nizkotlačnimi ultravijoličnimi sevalniki - Zahteve za delovanje, varnost in preskušanje

Water conditioning equipment inside buildings - Devices using mercury low-pressure ultraviolet radiators - Requirements for performance, safety and testing

Anlagen zur Behandlung von Trinkwasser innerhalb von Gebäuden - Geräte mit Quecksilberdampf-Niederdruckstrahlern - Anforderungen an Ausführung, Sicherheit und Prüfung

Équipements de traitement de l'eau à l'intérieur des bâtiments - Lampes à rayonnement UV à mercure et basse pression - Exigences relatives aux performances, à la sécurité et aux essais

Ta slovenski standard je istoveten z: EN 14897:2006

ICS:

13.060.20	Pitna voda	Drinking water
91.140.60	Sistemi za oskrbo z vodo	Water supply systems

SIST EN 14897:2006 en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 14897:2006

<https://standards.iteh.ai/catalog/standards/sist/f39eabae-2d3b-4745-baad-329d67e4ad0d/sist-en-14897-2006>

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN 14897

June 2006

ICS 91.140.60

English Version

**Water conditioning equipment inside buildings - Devices using
mercury low-pressure ultraviolet radiators - Requirements for
performance, safety and testing**

Équipements de conditionnement de l'eau à l'intérieur des
bâtiments - Dispositifs utilisant des radiateurs à mercure et
basse pression de rayonnement UV - Exigences relatives
aux performances, à la sécurité et aux essais

Anlagen zur Behandlung von Trinkwasser innerhalb von
Gebäuden - Geräte mit Quecksilberdampf-
Niederdruckstrahlern - Anforderungen an Ausführung,
Sicherheit und Prüfung

This European Standard was approved by CEN on 10 May 2006.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: rue de Stassart, 36 B-1050 Brussels

Contents

Page

Foreword.....	3
1 Scope	4
2 Normative references	4
3 Definitions	4
4 Requirements	7
4.1 General.....	7
4.2 Radiation chamber.....	7
4.3 Low-pressure mercury UV lamps.....	7
4.4 Electrical	7
4.4.1 General.....	7
4.4.2 UV disinfection devices	7
4.4.3 UV bactericidal treatment devices	8
4.5 Performance.....	8
4.6 Labelling	8
4.7 Manual.....	9
5 Testing	9
5.1 General.....	9
5.2 Test rig and installation	10
5.3 Biosimetric measurements	10
5.4 Performance test procedure.....	11
5.4.1 UV disinfection devices	11
5.4.2 Processing of the data	13
5.4.3 UV bactericidal treatment devices.....	15
5.4.4 Processing of the data	15
Annex A (normative) Requirements for the device sensor.....	17
A.1 Calibration	17
A.2 Selectivity	17
A.3 Measuring range linearity	17
A.4 Measuring range and resolution	17
A.5 Temperature drift	18
A.6 Stability over time	18
Annex B (normative) Biosimulator calibration.....	19
Annex C (informative) Monitoring window	22
Annex D (normative) Manufacturer's information for the type test	23
Annex E (normative) Manufacturer's information for the UV device user	25
Bibliography	26

Foreword

This document (EN 14897:2006) has been prepared by Technical Committee CEN/TC 164 "Water supply", the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by December 2006, and conflicting national standards shall be withdrawn at the latest by December 2006.

With respect to potential adverse effects on the quality of water intended for human consumption/caused by the product covered by this standard, the following is pointed out to the user of the standard.

- 1) This standard provides no information as to whether the product may be used without restriction in any of the Member States.
- 2) It should be noted that, while awaiting the adoption of verifiable European criteria, existing national regulations concerning the use and/or characteristics of this product remain in force.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom. (standards.iteh.ai)

[SIST EN 14897:2006](https://standards.iteh.ai/catalog/standards/sist/f39eabae-2d3b-4745-baad-329d67e4ad0d/sist-en-14897-2006)

<https://standards.iteh.ai/catalog/standards/sist/f39eabae-2d3b-4745-baad-329d67e4ad0d/sist-en-14897-2006>

EN 14897:2006 (E)

1 Scope

This document specifies definitions, principles of construction, requirements and methods for testing the performance of UV devices for drinking water installations inside buildings which are permanently connected to the mains supply at the point of entry into a building or within the water distribution system inside the building.

UV devices in the sense of this standard are UV bactericidal treatment devices or UV disinfection devices.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 1717, *Protection against pollution of potable water in water installations and general requirements of devices to prevent pollution by backflow*

3 Definitions

For the purposes of this document, the following terms and definitions apply.

3.1 absorption

decrease of the incident irradiance of a light beam including transformation into other forms suffered by radiant energy passing through a material substance (e. g. heat)

3.2 irradiance

measure of the UV light flux divided by the area that intercepts the radiation, in W/m^2

NOTE The irradiance measured in UV disinfection devices by the UV device sensor is mainly influenced by the lamp output, the transmittance of the water, and scaling/fouling of the protective quartz sleeves and the position of the lamps in the radiation chamber.

3.3 disinfection

action of killing or inactivating all types of pathogenic bacteria to a specified degree of at least 99,999 % and all types of pathogenic viruses to a degree of at least 99,99 % using a UV disinfection device

3.4 bactericidal treatment

action of inactivating or killing bacteria present in water to an unspecified degree using a UV bactericidal treatment device

3.5 fluence

dose
product of irradiance in W/m^2 and exposure time in s, in J/m^2

NOTE Fluence is the correct term from a strictly scientific point of view.

3.6 germicidal fluence

fluence weighted with the germicidal UV sensitivity, in J/m^2

3.7**reduction equivalent fluence (REF)**

average germicidal fluence measured by the biodosimeter in accordance with 5.3 in the radiation chamber, in J/m^2

3.8**radiation chamber**

part of the device that comprises the radiation zone and the connecting pipes

3.9**radiation zone**

part of the radiation chamber whose volume is used for the calculation of the fluence

3.10**exposure time**

time interval during which a specific volume of water within the radiation zone is exposed to the radiation, in s

3.11**microbiological dosimeter;**

biodosimeter

test organism used to determine the equivalent fluence, whose UV inactivation behaviour has been determined in a standard collimated beam apparatus (see Annex B), e. g. *Bacillus subtilis* spores

3.12**minimum irradiance**

value determined in the type test that ensures the required reduction equivalent fluence at a defined water flow rate and at a defined UV transmittance value, in W/m^2

3.13**flow rate (Q)**

volume of water per unit time flowing through the UV device, in l/min or m^3/h

3.14**maximum flow rate (Q_{max})**

highest flow at which, at a defined UV transmittance of the water and a defined irradiance, the required reduction equivalent fluence can be guaranteed, in l/min or m^3/h

3.15**permissible operation range**

those limit values for the operation parameters (irradiance at the sensor or UV transmittance of the water) and flow rate where adequate bactericidal treatment or disinfection is assured

3.16**sensor**

system for the measurement of the irradiance in UV disinfection devices

3.16.1**reference sensor**

sensor used to countercheck the signal of the device sensor where national regulations apply.

NOTE The reference sensor should comply to national standards where existent, e.g. [1], [2].

3.16.2**selectivity**

percentage of the sensor signal that is produced by radiation with a wavelength of 254 nm

3.16.3**device sensor**

calibrated sensor monitoring device used for continuous measurement of the irradiance

EN 14897:2006 (E)

3.17

attenuation

absorption and diffraction of radiation passing through an optical medium in a specific direction

3.18

lamp service life

service life of a UV radiator after which the irradiance that is necessary to guarantee the minimum fluence can no longer be reached under the mode of operation given by the manufacturer, and the lamp has to be replaced, in h

3.19

UV lamp

radiator which produces UV light

3.20

UV device

general expression for products using UV light to irradiate water flow through, with the purpose of inactivating microorganisms being present in the water

3.21

UV disinfection device

device designed to disinfect water

3.22

UV bactericidal treatment device

device designed for bactericidal water treatment

3.23

turbidity

reduction of optical transmittance of a liquid caused by the presence of undissolved matter

3.24

UV transmittance (% T_{100})

spectral transmission rate at a wavelength of 254 nm at an optical path length in the medium of 100 mm, in %

NOTE In general, the UV transmittance includes the influence of attenuation and absorption of the through passing medium. The UV transmittance is measured in the unfiltered sample in quartz cuvettes of at least 40 mm at a wavelength of 254 nm in a spectrophotometer and is given in %.

3.25

UV radiation (UV)

electromagnetic radiation according to Table 1

Table 1 — UV radiation

Type	Wavelength nm
UV-C	$> 100 \leq 280$
UV-B	$> 280 \leq 315$
UV-A	$> 315 \leq 400$
NOTE For bactericidal and disinfection purposes, part of the UV-C range is used.	

4 Requirements

4.1 General

The treatment of water with the UV device shall yield a reduction equivalent fluence of at least 400 J/m² (40 mJ/cm²), at a wavelength of 254 nm in the defined operational range.

4.2 Radiation chamber

The radiation chamber is made of corrosion resistant and UV resistant materials. Materials and substances used shall be suitable for contact with drinking water.

NOTE Product intended for use in water supply systems should comply, when existing, with national regulations and testing arrangements that ensure fitness for contact with drinking water. The Member states, relevant regulators and the EC Commission agreed on the principles of a future unique European Acceptance Scheme (EAS), which would provide a common testing and approval arrangement at European level. If and when the EAS is adopted, European Product Standards will be amended by the addition of an Annex Z/EAS under Mandate M/136, which will contain formal references to the testing, certification and product marking requirements of the EAS.

The mechanical design and the construction shall comply with the relevant requirements for the pressure present. Direct or indirect leaking of radiation from the radiation chamber to the environment with a wavelength below 400 nm shall be avoided.

Radiation chambers shall be constructed in a way that they are easily serviceable.

For UV disinfection devices, the radiation chamber shall be provided with a sensor for a representative irradiation measurement. The location of the sensor shall be designed so that the irradiation measurement is not disturbed by gas bubbles or sediment deposits.

4.3 Low-pressure mercury UV lamps

<https://standards.iteh.ai/catalog/standards/sist/b39eabae-2d3b-4745-baad-329d67e4ad0d/sist-en-14897-2006>

In order not to produce ozone, only lamps with a radiation range above 240 nm shall be used. At the line of 254 nm shall be 85 % of the total radiation intensity in the UV C range at the mercury resonance.

Lamp(s) shall be marked with a designation of type. Only those lamps used for the type test shall be used in the UV device. The UV lamp(s) shall be approved for the device by the manufacturer or be equivalent to the approved type used at the type test.

4.4 Electrical

4.4.1 General

For the electrotechnical design of UV devices, the relevant EC Directives and CE marking requirements shall be accommodated. Compliance with these EC Directives is a requirement of this standard.(see [3])

4.4.2 UV disinfection devices

4.4.2.1 Controller

The UV disinfection device shall be equipped with a controller, which provides the following functions:

- when switching on the device, the signal for the waterflow shall be delayed until the minimum irradiance is reached;
- operation and failure of the electrical function of each lamp shall be indicated;
- operation beyond the permissible limits of operation shall be indicated and a signal shall be provided which allows the waterflow to be stopped;

EN 14897:2006 (E)

- general malfunction signal shall be provided;
- when shutting down the device or in case of a breakdown of the electric power supply, a signal shall be provided which allows the stop of waterflow.

The following functions shall be displayed:

- device in function;
- failure signal for each lamp;
- irradiance, in W/m^2 ;
- service time of the UV lamps;
- flow-related alarm point(s).

4.4.2.2 Sensor

For the measurement of the irradiation, a sensor shall be provided to ensure disinfection under consideration of possible changes in water UV transmittance and lamp performance. Requirements for the device sensor are given in Annex A. Where national regulations apply, a sensor and monitoring window may have to fulfil certain requirements.

NOTE An example for a monitoring window is shown in Annex C.

4.4.3 UV bactericidal treatment devices

The UV bactericidal treatment devices shall be equipped with a controller, which provides the following functions:

- operation and failure of the electrical function of the lamp(s) shall be indicated;
- general malfunction signal shall be provided.

The following functions shall be displayed:

- device in function;
- service time of the UV lamps.

4.5 Performance

The UV device shall apply a reduction equivalent fluence of at least $400 J/m^2$ at the end of the lamp service life for the specified flow rates and UV transmittance values.

The performance is evaluated with a type test for which the manufacturer shall provide information in accordance with Annex D.

4.6 Labelling

The information to be given on a nameplate, which shall be permanently fixed to the UV device and be legible when installed, shall be in accordance with Annex E.

4.7 Manual

The manual shall describe operation, control, cleaning and service measures.

The manual shall also contain at least the following information:

- operating diagram: transmittance vs. maximum admissible flow;
- water resulting from sampling (rinsing) shall be adequately disposed of to the provisions of EN 1717;
- replacement intervals for UV lamp(s), sensor (only for UV disinfection devices).

5 Testing

5.1 General

The manufacturer shall provide the details and documentation described in Annex D. The device to be tested is checked for conformity to the documentation.

The purpose of the type test is to verify that the UV fluence delivered by the device under test meets a reduction equivalent fluence of 400 J/m^2 at the end of the lamp life, at the specified flow rates and transmittance values.

Parameters to be changed during the test are the flow rate of the water and the UV transmittance of the water for the test of UV bactericidal treatment devices and the flow rate, the transmittance and the lamp-output, for the test of UV disinfection devices.

For the type test of UV disinfection devices the UV device shall be equipped with a sensor or with a specified monitoring window and a specified sensor if national regulation for these apply. The testing in the test rig comprises of five steps:

- checking the compliance of the device to be tested with the specifications;
- data collection during the test (flow rate, water temperature, electrical power consumption,);
- radiation physics tests, i.e. determination of the irradiance (only for UV disinfection devices)
- microbiological test with the biosimulator;
- evaluation of the UV device and specification of the operating range.

The manufacturer shall provide data (flow rate versus UV transmittance), that give the testing points at which the UV fluence (400 J/m^2) is reached at the end of the lamp service time and the percentage of UV output at the end of lamp service life (e.g. 70 %). The permissible operational range is determined by measurements of at least three test points which should cover the whole operational range.

The UV device to be tested shall have new lamps that have been in service for 100 h. The manufacturer shall provide an appropriate method to vary the output of the UV lamps.