
**Biotechnology — Biobanking —
Process and quality requirements
for establishment, maintenance and
characterization of mammalian cell
lines**

*Biotechnologie — Biobanking — Exigences de processus et de qualité
pour la génération, le maintien et la caractérisation des lignées
cellulaires de mammifères*

(<https://standards.iteh.ai>)
Document Preview

ISO 21709:2020

<https://standards.iteh.ai/catalog/standards/iso/ffd6e501-779d-4cba-a17b-f918f3f4b81a/iso-21709-2020>



iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

ISO 21709:2020

<https://standards.iteh.ai/catalog/standards/iso/ffd6e501-779d-4cba-a17b-f918f3f4b81a/iso-21709-2020>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Requirements	3
4.1 General	3
4.2 Legal and ethical requirements	3
4.3 Facilities	4
4.3.1 General	4
4.3.2 Cell culture facility	4
4.4 Equipment	4
4.4.1 General	4
4.4.2 Equipment inspection	5
4.5 Reagents	6
4.6 Informed consent	6
4.6.1 General	6
4.6.2 Procedure of obtaining informed consent	6
4.6.3 Special circumstances for informed consent	7
4.6.4 Information provided to the donor/patient or the donor's/patient's legally designated or nominated representative	7
4.6.5 Informed consent signature	8
4.7 Personnel	8
4.7.1 General	8
4.7.2 Personnel competence	9
4.7.3 Personnel training	9
4.7.4 Biorisk and biosafety of personnel	9
5 Process requirements	9
5.1 Establishment of cell lines within the biobank	9
5.1.1 General	9
5.1.2 Isolation and purification of primary cells	10
5.1.3 Primary cultures	10
5.1.4 Cell lines	11
5.2 Reception of established cell lines	11
5.2.1 General	11
5.2.2 Review of the deposit request	11
5.2.3 Decision on the deposit request	12
5.2.4 Transfer of material and associated data	12
5.2.5 Characterization and authentication of the cell line	13
5.2.6 Assignment of accession number	13
5.3 Cell line management	14
5.3.1 Planning for cell banking	14
5.3.2 Accession	14
5.3.3 Propagation of cell lines	14
5.3.4 Preservation and storage	15
5.3.5 Inventory management	15
5.3.6 Disposal management	15
5.4 Distribution	16
5.4.1 General	16
5.4.2 Acceptance of order request	16
5.4.3 Distribution review	16
5.4.4 Transport	16

5.5	Quality control, validation and verification	17
5.5.1	Quality control	17
5.5.2	Validation and verification	17
6	Performance evaluation and improvement	17
Annex A (informative)	Recommended units for biobanking of cell lines	19
Bibliography		20

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

ISO 21709:2020

<https://standards.iteh.ai/catalog/standards/iso/ffd6e501-779d-4cba-a17b-f918f3f4b81a/iso-21709-2020>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 276, *Biotechnology*.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

ISO 21709:2020

<https://standards.iteh.ai/catalog/standards/iso/ffd6e501-779d-4cba-a17b-f918f3f4b81a/iso-21709-2020>

Introduction

Scientific research using cell lines has contributed greatly to the understanding of human health. Cell cultures are increasingly used to complement studies using animal models. Although cell lines are important research tools, potential problems have recently been identified.

Cell lines have unique characteristics and behaviour that can change as they continue to be passaged. The original phenotype (e.g. expression of specific biomarkers) can be lost or new characteristics or behaviour (e.g. development of tumorigenicity) may develop. It is important to minimize passaging to retain the original characteristics that were present when the cell line was first established.

Other problems such as contamination, either with microorganisms or another cell line, and misidentification can also arise. Cultures can become contaminated during cell line establishment or later when cultures are passaged. These problems are often not visible by eye and require specific testing to be detected.

In order to help address these issues, the research community has called for an international effort to create standards for biobanks. ISO 20387 was published to provide an overarching standard for biobanks. This document provides additional technical specifications for biobanks that handle mammalian cell lines. Such biobanks can demonstrate their competence in biobanking by complying with the specifications within this document, in addition to the requirements prescribed in ISO 20387.

In this document, the following verbal forms are used:

- “shall” indicates a requirement;
- “should” indicates a recommendation;
- “may” indicates a permission;
- “can” indicates a possibility or a capability.

Further details can be found in the ISO/IEC Directives, Part 2.

[ISO 21709:2020](https://standards.iteh.ai/catalog/standards/iso/ffd6e501-779d-4cba-a17b-f918f3f4b81a/iso-21709-2020)

<https://standards.iteh.ai/catalog/standards/iso/ffd6e501-779d-4cba-a17b-f918f3f4b81a/iso-21709-2020>

Biotechnology — Biobanking — Process and quality requirements for establishment, maintenance and characterization of mammalian cell lines

1 Scope

This document specifies process and quality requirements for the biobanking of mammalian (including human) cell lines. It describes requirements for the fundamental procedures of the biobank handling cell lines, such as establishment, reception, identification, propagation, preservation, storage, quality control, and distribution of cell lines.

This document can be used by organizations performing biobanking activities with mammalian cell lines used for research and development, biobank users, organizations and schemes using peer-assessment and accreditation bodies.

This document does not apply to biological material intended for therapeutic use.

NOTE International, national or regional regulations or requirements can also apply to specific topics covered in this document.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 20387:2018, *Biotechnology — Biobanking — General requirements for biobanking*

ISO 20391-1, *Biotechnology — Cell counting — Part 1: General guidance on cell counting methods* 2020

ISO 20391-2, *Biotechnology — Cell counting — Part 2: Experimental design and statistical analysis to quantify counting method performance*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

3.1

cell culture

growth of cells dissociated from the parent tissue by spontaneous migration or mechanical or enzymatic dispersal for propagation and consecutive passages in vitro

Note 1 to entry: Additional information can be found in Reference [6].

3.2

cell line

progeny of a *primary culture* (3.16) after it has been passaged beyond crises or beyond senescence either spontaneously or after introduction of immortalizing factors

Note 1 to entry: A cell line is continuous for proliferation.

Note 2 to entry: Additional information can be found in Reference [7].

3.3

cell morphology

form and structure of the cell

Note 1 to entry: Morphology can be represented by a single parameter or a combination of two or more parameters.

3.4

cell strain

progeny of a *primary culture* (3.16) before it has been passaged beyond crises or beyond senescence or immortalized by introduction of immortalizing factors

Note 1 to entry: Not all the cell strains will continue proliferation to reach the cell lines.

3.5

cell type

classification used to distinguish among distinct cell forms

3.6

cryopreservation

process by which cells are maintained in a low temperature (e.g. -70°C to -80°C) at an inactive state so they can be revived later

3.7

derivative material

biological material that was derived from the original donation

EXAMPLE Derivative materials can be cell lines.

3.8

doubling time

time taken for cultured cell count to double

3.9

establishment of cell line

process of producing a cell line capable of indefinite proliferation

3.10

identity verification

part of the process of verifying authenticity of a cell line in which cell origin is genetically confirmed

3.11

informed consent

process by which an individual or its designated legal representative or nominated representative voluntarily confirms willingness to donate biological material for research purposes, after having been informed of all aspects of the potential research that are relevant for the decision to donate

3.12

microbe

microorganism

unicellular living cells which cannot be observed with naked eye but only under a microscope

3.13**passage number**

number of subculturing that occurred

3.14**population doubling level****PDL**

total number of population doublings of a *cell line* (3.2) or *cell strain* (3.4) since its initiation *in vitro*

Note 1 to entry: Additional information can be found in Reference [7].

3.15**primary cells**

cells isolated directly from tissue or organs taken directly from an organism, using enzymatic or mechanical methods

3.16**primary culture**

culture started from cells, tissues, or organs taken directly from an organism, and before the first subculture, propagation and consecutive passages *in vitro*

Note 1 to entry: Additional information can be found in Reference [6].

3.17**viability**

attribute of being alive (e.g. metabolically active, capable of reproducing, have intact cell membrane, or have the capacity to resume these functions) as defined based on the intended use

4 Requirements**4.1 General**

During the whole cell culture workflow, precautions shall be taken to avoid cross contamination between different samples/specimens, e.g. by using disposable material whenever feasible or appropriate cleaning procedures between processing of different specimens/samples.

The biobank shall establish, implement, and maintain a quality management system as described in ISO 20387:2018, Clause 8.

4.2 Legal and ethical requirements

ISO 20387:2018, 4.3 shall be followed.

The biobank shall consult an ethics review committee, which is responsible for the investigation and evaluation of any related ethical principles/requirements.

The biobank shall comply with ISO 20387:2018, 4.1.6. The biological material, which is the source of the cell line, shall be collected, transported, and handled according to internationally accepted procedures.

The biobank shall be aware of and able to demonstrate compliance with relevant national, regional and international approved ethics, laws and regulations relating to the biological material held in the biobank.

Further ethical requirements related to the informed consent are included in 4.6.

The biobank shall meet the organizational structure, resource, process, and quality requirements described in ISO 20387:2018, Clauses 5, 6 and 7, in addition to those in this document.

4.3 Facilities

4.3.1 General

ISO 20387:2018, 6.3 shall be followed.

The processing facility shall be constructed and operated to minimize the introduction, generation, and retention of particles and microorganisms.

The biobank shall define a protocol to test the quality of the biobanking environment(s).

Where appropriate, test methods proposed in the ISO 14644 series can be considered.

4.3.2 Cell culture facility

The biobank culturing and preserving cell lines shall establish a cell culture facility.

The biobank shall assess the biological safety level of the cell lines that will be handled. The biobank shall assess the risk of asphyxiation based on the number and condition of liquid nitrogen tanks and storage methods and install an oxygen monitor accordingly.

The biobank shall meet biological safety and security requirements for its facilities to minimize the risk to personnel and surroundings from handling the cell lines.

The biobank shall take necessary measures to control airflow.

The biobank should make efforts to secure a separate space for storage of duplicated cell lines in order to prevent the loss of collections in emergency situations. If the biobank cannot secure a separate alternate space for storage, the biobank should duplicate cell lines and store them in separate liquid nitrogen tanks. If the biobank cannot duplicate all stored cell lines, the biobank should make duplicates of the relevant cell lines or those that can be required by contractual agreements.

The biobank shall designate a separate area for handling deposited cell lines and/or tissues until they are determined to be free of microbial contamination.

<https://standards.iteh.ai/catalog/standards/iso/ffd6e501-779d-4cba-a17b-f918f3f4b81a/iso-21709-2020>

4.4 Equipment

4.4.1 General

ISO 20387:2018, 6.5 shall be followed.

Equipment used in biobanking should be selected under quality specific criteria defined by the biobank, and for managing risks that can affect or impact performed assays, tests or their validity.

All incubators used, for working in accordance with this document, shall be monitored.

The biobank shall be furnished with or have controlled access to the following equipment as minimum requirements for cell culturing and ensure quality and safety requirements regarding the equipment.

- a) Biological safety cabinet (Class II, Class III): physical structure protecting cell cultures from contaminants and personnel from potentially harmful biological substances. The cabinet can require appropriate biosafety measures, depending upon the biosafety level of the biological materials being handled.
- b) Cell counting equipment: an automated cell counting feature is recommended, manual cell counting slide (e.g. hemocytometer slide) may also be used.
- c) CO₂ incubator: a self-cleaning feature is recommended;
- d) Freezer: data logger, alarming system and backup system are recommended.