
**Določevanje plamenišča - Metoda z zaprto posodo po Pensky-Martensu -
Dopolnilo 1: Korekcija termometrov (ISO 2719:2016/DAM 1:2020)**

Determination of flash point - Pensky-Martens closed cup method - Amendment 1:
Thermometers correction (ISO 2719:2016/DAM 1:2020)

Bestimmung des Flammpunktes - Verfahren nach Pensky-Martens mit geschlossenem
Tiegel - ÄNDERUNG 1: Korrektur des Thermometers (ISO 2719:2016/DAM 1:2020)

Détermination du point d'éclair - Méthode Pensky-Martens en vase clos - Amendement
1: Correction des thermomètres (ISO 2719:2016/DAM 1:2020)

<https://standards.iteh.ai/catalog/standards/sist/3d0b7418-e558-4930-b204-1a324e92421a/sist-en-iso-2719-2016/oprA1:2020>

Ta slovenski standard je istoveten z: EN ISO 2719:2016/prA1

ICS:

75.080	Naftni proizvodi na splošno	Petroleum products in general
--------	-----------------------------	----------------------------------

SIST EN ISO 2719:2016/oprA1:2020 **en,fr,de**

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO 2719:2016/oprA1:2020

<https://standards.iteh.ai/catalog/standards/sist/3d0b7418-e558-4930-b204-1a32dc92421a/sist-en-iso-2719-2016-opra1-2020>

DRAFT AMENDMENT

ISO 2719:2016/DAM 1

ISO/TC 28

Secretariat: NEN

Voting begins on:
2020-04-14Voting terminates on:
2020-07-07

Determination of flash point — Pensky-Martens closed cup method

AMENDMENT 1: Thermometers correction

Détermination du point d'éclair — Méthode Pensky-Martens en vase clos
AMENDEMENT 1

ICS: 75.080

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 2719:2016/oprA1:2020](https://standards.iteh.ai/catalog/standards/sist/3d0b7418-e558-4930-b204-1a32dc92421a/sist-en-iso-2719-2016-opra1-2020)
<https://standards.iteh.ai/catalog/standards/sist/3d0b7418-e558-4930-b204-1a32dc92421a/sist-en-iso-2719-2016-opra1-2020>

THIS DOCUMENT IS A DRAFT CIRCULATED FOR COMMENT AND APPROVAL. IT IS THEREFORE SUBJECT TO CHANGE AND MAY NOT BE REFERRED TO AS AN INTERNATIONAL STANDARD UNTIL PUBLISHED AS SUCH.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

This document is circulated as received from the committee secretariat.

ISO/CEN PARALLEL PROCESSING



Reference number
ISO 2719:2016/DAM 1:2020(E)

© ISO 2020

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO 2719:2016/oprA1:2020

<https://standards.iteh.ai/catalog/standards/sist/3d0b7418-e558-4930-b204-1a32dc92421a/sist-en-iso-2719-2016-opra1-2020>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

The committee responsible for this document is ISO/TC 28, *Petroleum products and related products of synthetic or biological origin*, Working Group 9, who developed the document in conjunction with ISO/TC 35, *Paints and varnishes*, CEN/TC 19, *Gaseous and liquid fuels*, lubricants and related products of petroleum, synthetic and biological origin, and CEN/TC 139, *Paints and varnishes*.

This fourth edition cancels and replaces the third edition (ISO 2719:2002). This Amendment is to address an important ambiguity regarding which types of thermometers are suitable for use following study results from the Energy Institute (UK).

iTeh STANDARD PREVIEW **(standards.iteh.ai)**

SIST EN ISO 2719:2016/oprA1:2020

<https://standards.iteh.ai/catalog/standards/sist/3d0b7418-e558-4930-b204-1a32dc92421a/sist-en-iso-2719-2016-opra1-2020>

Determination of flash point — Pensky-Martens closed cup method

AMENDMENT 1: Thermometers correction

9.2 Repeatability, r

Replace the 1st paragraph:

"The difference between two test results obtained by the same operator with the same apparatus under constant operating conditions on identical test material would in the long run, in the normal and correct operation of the test method, exceed the values given in Tables 1, 2 and 3 in only one case in twenty."

by:

"The difference between two independent results obtained in the normal and correct operation of the test method by the same operator in a given laboratory applying the same method, for test material considered to be the same, within a short interval of time, with the same apparatus under constant operating test conditions, would exceed the values in Tables 1, 2 and 3, with an approximate probability of 5 %."

[SIST EN ISO 2719:2016/oprA1:2020](https://standards.iteh.ai/catalog/standards/sist/3d0b7418-e558-4930-b204-1a32dc92421a/sist-en-iso-2719-2016-opra1-2020)

[https://standards.iteh.ai/catalog/standards/sist/3d0b7418-e558-4930-](https://standards.iteh.ai/catalog/standards/sist/3d0b7418-e558-4930-b204-1a32dc92421a/sist-en-iso-2719-2016-opra1-2020)

[b204-1a32dc92421a/sist-en-iso-2719-2016-opra1-2020](https://standards.iteh.ai/catalog/standards/sist/3d0b7418-e558-4930-b204-1a32dc92421a/sist-en-iso-2719-2016-opra1-2020)

9.3 Reproducibility, R

Replace the 1st paragraph:

"The difference between two single and independent test results obtained by different operators working in different laboratories on identical test material would in the long run, in the normal and correct operation of the test method, exceed the values given in Tables 4, 5 and 6 in only one case in twenty."

by:

"The difference between two independent results obtained by different operators in the normal and correct operation of the same method, for test material considered to be the same, in different laboratories using different apparatus, would exceed the values given in Tables 4, 5 and 6, with an approximate probability of 5 %."

C.2 Liquid-in-glass thermometers

Add the following two paragraphs between the first paragraph and Table C.1:

"Some alternative low hazard precision liquids can have significantly higher coefficients of expansion than mercury, making them unsuitable for these applications due to stem correction requirements. Liquid-in-glass thermometers using a Gallium based liquid do not have this issue and should be used."

ISO 2719:2016/DAM 1:2020(E)

Previously available mercury thermometers IP 15C/ASTM 9C, IP 16C/ASTM 10C, IP 101C and ASTM 88C may be used as well."

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO 2719:2016/oprA1:2020

<https://standards.iteh.ai/catalog/standards/sist/3d0b7418-e558-4930-b204-1a32dc92421a/sist-en-iso-2719-2016-opra1-2020>