



**SLOVENSKI STANDARD**  
**oSIST prEN ISO 8130-2:2020**

**01-november-2020**

---

**Praškasti premazi - 2. del: Določanje gostote s piknometrom za primerjavo plinov (referenčna metoda) (ISO/DIS 8130-2:2020)**

Coating powders - Part 2: Determination of density by gas comparison pyknometer (referee method) (ISO/DIS 8130-2:2020)

Pulverlacke - Teil 2: Bestimmung der Dichte mit einem Gasvergleichspyknometer (Schiedsverfahren) (ISO/DIS 8130-2:2020)

Poudres pour revêtement - Partie 2: Détermination de la masse volumique à l'aide d'un pycnomètre à gaz (méthode de référence) (ISO/DIS 8130-2:2020)

<https://standards.iteh.ai/catalog/standards/sist/63e53131-296b-4f64-8fc0-c6859a76317/osist-pr-en-iso-8130-2-2020>

**Ta slovenski standard je istoveten z: prEN ISO 8130-2**

---

**ICS:**

87.040

Barve in laki

Paints and varnishes

**oSIST prEN ISO 8130-2:2020**

**en,fr,de**

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[oSIST prEN ISO 8130-2:2020](#)

<https://standards.iteh.ai/catalog/standards/sist/63e53131-296b-4f64-8fc0-ee6859a76317/osist-pren-iso-8130-2-2020>

# DRAFT INTERNATIONAL STANDARD

## ISO/DIS 8130-2

ISO/TC 35/SC 9

Secretariat: BSI

Voting begins on:  
2020-09-22Voting terminates on:  
2020-12-15

---

---

## Coating powders —

### Part 2: Determination of density by gas comparison pycnometer (referee method)

*Poudres pour revêtement —**Partie 2: Détermination de la masse volumique à l'aide d'un pycnomètre à gaz (méthode de référence)*

ICS: 87.040

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[oSIST prEN ISO 8130-2:2020](https://standards.iteh.ai/catalog/standards/sist/63e53131-296b-4f64-8fc0-ee6859a76317/osist-pren-iso-8130-2-2020)<https://standards.iteh.ai/catalog/standards/sist/63e53131-296b-4f64-8fc0-ee6859a76317/osist-pren-iso-8130-2-2020>

THIS DOCUMENT IS A DRAFT CIRCULATED FOR COMMENT AND APPROVAL. IT IS THEREFORE SUBJECT TO CHANGE AND MAY NOT BE REFERRED TO AS AN INTERNATIONAL STANDARD UNTIL PUBLISHED AS SUCH.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

This document is circulated as received from the committee secretariat.



Reference number  
ISO/DIS 8130-2:2020(E)

© ISO 2020

## iTeh STANDARD PREVIEW (standards.iteh.ai)

<https://standards.iteh.ai/catalog/standards/sist/63e53131-296b-4f64-8fc0-ee6859a76317/osist-pren-iso-8130-2-2020>



### **COPYRIGHT PROTECTED DOCUMENT**

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

	Page
Foreword .....	iv
<b>1 Scope .....</b>	<b>1</b>
<b>2 Normative references .....</b>	<b>1</b>
<b>3 Terms and definitions .....</b>	<b>1</b>
<b>4 Principle .....</b>	<b>1</b>
<b>5 Materials .....</b>	<b>1</b>
<b>6 Apparatus .....</b>	<b>1</b>
<b>7 Sampling .....</b>	<b>2</b>
<b>8 Procedure .....</b>	<b>2</b>
<b>9 Expression of results .....</b>	<b>3</b>
<b>10 Precision .....</b>	<b>3</b>
<b>11 Test report .....</b>	<b>3</b>
<b>Bibliography .....</b>	<b>5</b>

## iTeh STANDARD PREVIEW (standards.iteh.ai)

<https://standards.iteh.ai/catalog/standards/sist/63e53131-296b-4f64-8fc0-ee6859a76317/osist-pren-iso-8130-2-2020>

## ISO/DIS 8130-2:2020(E)

### Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 35, *Paints and varnishes*, Subcommittee SC 9, *General test methods for paints and varnishes*.

This second edition cancels and replaces the first edition (ISO 8130-2:1992), which has been technically revised. The main changes compared to the previous edition are as follows:

- a) the scope has been revised editorially;
- b) a terms and definitions clause has been added;
- c) the gas has been changed from air or helium to helium or nitrogen;
- d) the procedure has been aligned with actual practice;
- e) the text has been editorially revised and the normative references have been updated.

A list of all parts in the ISO 8130, *Coating powders*, series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

# Coating powders —

## Part 2:

# Determination of density by gas comparison pycnometer (referee method)

## 1 Scope

This document specifies a method for the determination of density for all types of coating powders using a gas comparison pycnometer.

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 8130-14, *Coating powders — Part 14: Vocabulary*

ISO 15528, *Paints, varnishes and raw materials for paints and varnishes — Sampling*

## 3 Terms and definitions

oSIST prEN ISO 8130-2:2020

For the purposes of this document, the terms and definitions given in ISO 8130-14 apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

## 4 Principle

The density is calculated from the mass and the volume of the test portion. The volume of a weighed test portion is determined by measuring the volume of gas displaced within a receptacle when the test portion is introduced. This is achieved by measuring the pressure difference which arises due to the displacement of the gas.

## 5 Materials

### 5.1 Helium or nitrogen, minimum Grade 4,8, in a steel cylinder.

Other high purity gases may be used provided that the product under test is not affected and this deviation from the method is noted in the test report.

## 6 Apparatus

Ordinary laboratory apparatus, together with the following:


## ISO/DIS 8130-2:2020(E)

**6.1 Gas comparison pycnometer** for the automatic or manual determination of density.

The essential design of a typical gas comparison pycnometer employs a displacement gas as the medium, outlined in [Figure 1](#). It consists of two chambers, sample and reference, of equal dimensions. The chambers are connected by a valve and a pressure difference meter. The test portion is placed in the sample chamber. The displacement gas is allowed to flow from one chamber into the other chamber and the resulting pressure difference is measured.

NOTE Automated gas pycnometers could help to minimise errors versus manual pycnometers.

Manual pycnometers may be used and due care is needed during testing and measurement.

**Key**

- 1 reference chamber
- 2 pressure difference meter
- 3 flushing valve
- 4 connecting valve
- 5 sample container
- 6 sample chamber

**Figure 1 — Measuring principle of a gas comparison pycnometer**

**7 Sampling**

Take a representative sample of the product to be tested as described in ISO 15528.

**8 Procedure**

Carry out the determination in duplicate at  $(23 \pm 1) ^\circ\text{C}$  and, where applicable, at a relative humidity of  $(50 \pm 5) \%$ .

The procedure depends on the type of pycnometer and the instructions of the apparatus shall be followed.

Calibration of the pycnometer may be performed with a known volume standard or similar test specimens previously measured.

Adjust the apparatus, in particular to allow for the volume of the test sample container.

The measured pressure difference shall not be greater than 1 %.

For manual pycnometers, the swept volumes of the cylinders measured should be equal to or within 0,5 % (relative).

Weigh the sample container to the nearest 1 mg and fill it with the product under test, making sure that no powder is on the outside walls of the container. Weigh the container plus contents to the nearest 1 mg to determine the mass ( $m$ ) of the test portion.

Place the container and contents in the apparatus and determine the volume of the test portion. Repeat the volume measurement in order to ensure that the reading is consistent and calculate the mean ( $V$ ) of the two volumes.

NOTE Depending on the grade of gas, a change in density can be observed due to the drying of the test portion.

## 9 Expression of results

Calculate the density  $\rho_p$ , in grams per cubic centimetre, of the coating powder at  $(23 \pm 1) ^\circ\text{C}$  following [Formula \(1\)](#):

$$\rho_p = \frac{m}{V} \quad (1)$$

where

$m$  is the mass, in grams, of the test portion;

$V$  is the volume, in cubic centimetres, of the test portion.

If the two results differ by more than 2 %, repeat the procedure described in [Clause 8](#).

Calculate the mean of two valid determinations and report the result to the nearest  $0,01 \text{ g/cm}^3$ .

## 10 Precision

No precision data are currently available.

## 11 Test report

The test report shall contain at least the following information:

- a) all details necessary to identify the product tested;
- b) a reference to this document, i.e. ISO 8130-2:—;
- c) the type of gas comparison pycnometer, manual or automated;
- d) the grade and type of the gas used;
- e) the result of the test as indicated in [Clause 9](#);
- f) any deviation from the test method specified;
- g) any unusual features (anomalies) observed during the test.

**ISO/DIS 8130-2:2020(E)**

h) the date of the test.

**iTeh STANDARD PREVIEW  
(standards.iteh.ai)**

[oSIST prEN ISO 8130-2:2020](https://standards.iteh.ai/catalog/standards/sist/63e53131-296b-4f64-8fc0-ee6859a76317/osist-pren-iso-8130-2-2020)  
<https://standards.iteh.ai/catalog/standards/sist/63e53131-296b-4f64-8fc0-ee6859a76317/osist-pren-iso-8130-2-2020>