

ISO/TC 147/SC 2

Secretariat: DIN

Voting begins on:
2020-03-10

Voting terminates on:
2020-05-05

Water quality — Determination of total organic carbon (TOC), dissolved organic carbon (DOC), total bound nitrogen (TN_b), dissolved bound nitrogen (DN_b), total bound phosphorus (TP_b) and dissolved bound phosphorus (DP_b) after wet chemical catalysed ozone hydroxyl radical oxidation (COHR)

Qualité de l'eau — Détermination du carbone organique total (COT), du carbone organique dissous (COD), de l'azote total lié (TNb), de l'azote dissous lié (DNb), du phosphore total lié et du phosphore dissous lié (DPb) après oxydation par l'ozone avec des radicaux hydroxyles et catalyseur en milieux aqueux

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.



Reference number
ISO/FDIS 21793:2020(E)

© ISO 2020

iTeh STANDARD PREVIEW
(standards.iteh.ai)
Full standard:
<https://standards.iteh.ai/catalog/standards/sist/3b912b79-a49e-417c-a88a-2d642de6224e/iso-21793-2020>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

	Page
Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Principle	2
5 Interferences	3
5.1 General	3
5.2 TOC	3
5.3 TN	3
5.4 TP	4
6 Reagents	4
7 Solution preparations	5
7.1 Blank solution	5
7.2 Sample oxidation solutions	5
7.2.1 Sulfuric acid TIC and catalyst solution	5
7.2.2 Base oxidation solution	5
7.2.3 TP solution	5
7.2.4 TP hydrolysis solution	5
7.3 Calibration stock solutions	6
7.4 Individual calibration standard solutions	6
7.4.1 TOC	6
7.4.2 TN	6
7.4.3 TP	7
7.4.4 Combined calibration standard solutions	7
7.5 System check stock solutions	7
7.5.1 TOC and TN stock solution	7
7.5.2 TP stock solution	7
7.6 Particle processing control solution	7
7.7 Gases	8
8 Apparatus	8
9 Quality requirements for the analytical system	8
9.1 System check	8
9.2 Recovery and variation of replicate determinations for particle processing control for TOC, TN _b , and TP _b	9
10 Sampling and sample preparation	9
10.1 Sampling and sample injection	9
11 Procedure	9
11.1 General	9
11.2 Calibration	10
11.3 Validity check of the calibration function	10
11.4 Measurement	10
11.4.1 General	10
11.4.2 Determination	10
12 Evaluation	11
13 Expression of results	11
14 Test report	11

Annex A (normative) Determination of total organic carbon that includes purgeable organic carbon by difference	13
Annex B (informative) Performance data TOC/TN	15
Annex C (informative) Informative validation data, system check recovery and particle check recovery for TOC/TN/TP	17
Bibliography	31

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Full standard:
<https://standards.iteh.ai/catalog/standards/sist/3b912b79-a49e-417c-a88a-2d642de6224e/iso-21793-2020>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 147, *Water quality*, Subcommittee SC 2, *Physical, chemical and biochemical methods*.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

Total organic carbon (TOC), dissolved organic carbon (DOC) total bound nitrogen (TN_b), dissolved bound nitrogen (DN_b), total bound phosphorus (TP_b) and dissolved bound phosphorus (DP_b) are an analytical convention, the respective result of which is a parameter used for water quality control purposes. These parameters represent the sum of organically bound carbon, the sum of inorganic and organic nitrogen, and the sum of inorganic and organic phosphorus. These parameters can be dissolved in water or bonded to dissolved or suspended matter under specified conditions. If the sample is not filtered the parameter is associated with suspended matter. This document does not give information on the nature of the substances.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Full standard:
<https://standards.iteh.ai/catalog/standards/sist/3b912b79-a49e-417c-a88a-2d642de6224e/iso-21793-2020>

Water quality — Determination of total organic carbon (TOC), dissolved organic carbon (DOC), total bound nitrogen (TN_b), dissolved bound nitrogen (DN_b), total bound phosphorus (TP_b) and dissolved bound phosphorus (DP_b) after wet chemical catalysed ozone hydroxyl radical oxidation (COHR)

WARNING — Persons using this document should be familiar with normal laboratory practices. This document does not purport to address all safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices.

IMPORTANT — It is essential that tests conducted in accordance with this document be carried out by suitably qualified staff.

1 Scope

This document specifies a multi-parameter method for the determination of total organic carbon (TOC), total nitrogen (TN_b) and total phosphorus (TP) in drinking water, raw water, ground water, surface water, sea water, saline water, process water, domestic and industrial wastewater, after a chemical oxidation process. It is applicable to both dissolved and bound suspended materials.

The method allows for determination of TOC, TN and TP. The lower and upper working ranges for these parameters are dependent upon instrument conditions (for example sample volume, reaction chemistry amounts) and can be adjusted for a wider range. Typical ranges are shown in Figures C.6 to C.8.

The analysis procedure is carried out instrumentally by a single oxidation process.

Dissolved nitrogen gas is not included in the TN_b measurement in this method. When present in the sample, elemental carbon, cyanide and cyanate will be included in the TOC result.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 5725-2, *Accuracy (trueness and precision) of measurement methods and results — Part 2: Basic method for the determination of repeatability and reproducibility of a standard measurement method*

ISO 8245, *Water quality — Guidelines for the determination of total organic carbon (TOC) and dissolved organic carbon (DOC)*

ISO 8466-1, *Water quality — Calibration and evaluation of analytical methods and estimation of performance characteristics — Part 1: Statistical evaluation of the linear calibration function*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

— ISO Online browsing platform: available at <https://www.iso.org/obp>

— IEC Electropedia: available at <http://www.electropedia.org/>

3.1
total carbon

TC
sum of organically and inorganically bound carbon present in water

3.2
total inorganic carbon

TIC
sum of inorganic carbon present in water measured under the conditions of this method

Note 1 to entry: TIC is measured as CO₂ originating only from carbonates, hydrogen carbonates and dissolved carbon dioxide.

3.3
total organic carbon

TOC
sum of organically bound carbon present in water, bonded to dissolved or suspended matter, including cyanate, thiocyanate, and elemental carbon measured

3.4
purgeable organic carbon

POC
organic carbon present in water which can be purged under the conditions of this method

3.5
non purgeable organic carbon

NPOC
organic carbon present in water which is not purged

3.6
total nitrogen

TN
sum of organically and inorganically bound nitrogen present in water or suspended

3.7
total phosphorus

TP
sum of organically and inorganically bound phosphorus present in water and suspended matter measured under the conditions of this method

4 Principle

Wet chemical oxidation of the sample by hydroxyl radicals, and catalysed ozone for the measurement of TOC, TN_b, and TP_b.

Organic carbon (TOC) is oxidized to carbon dioxide (CO₂). For the determination of total organic carbon that includes purgeable organic carbon by difference, the procedure shall be as specified in [Annex A](#). Detection is by nondispersive infrared (NDIR) spectrometry.

Detection is by photometric analysis in the ultraviolet wavelength range 200 nm to 220 nm. Detection is by colorimetric analysis using a photometer under the visible spectra between 380 nm and 470 nm wavelengths.

This document can be applied for the determination of TOC, TN_b and TP_b separately, or for simultaneous analysis, which consists of a non-dispersive infrared carbon dioxide analyser and a visible and UV photometer.

Quality control is necessary to verify the validity of the calibration function. Replicate determinations can be necessary, depending on the matrix. The method of standard addition can be used if matrix interferences are expected.

Inorganic carbon is removed by acidification and purging with a carrier gas (see [11.4](#)).

Performance results from the interlaboratory trial is provided in [Annex B](#). Supplemental single laboratory performance data is provided in [Annex C](#).

5 Interferences

5.1 General

Depending on concentration and analysis range, interferences with the determination of TOC, TN_b and TP_b can arise from possible sample carry over effects. In some cases, replicate injections can be necessary. When carryover is suspected, insert a reagent blank in the analytical run, immediately after the suspect sample.

Interferences can arise from memory effects. Replicate injections maybe necessary ([11.4.1](#)).

Samples with high pH values, highly buffered samples and samples with high chloride content can cause interference.

Large Suspended particles can lead to a loss of quality of the analytical result. If a homogenized sample containing suspended particles produces results (obtained from replicate measurements) which deviate by more than 10 %, an accurate TOC TN_b and TP_b result cannot be obtained on the sample.

Seek advice from the manufacturer to resolve these interferences. Particles in general can cause interference with sample injection. To mitigate particle interference, samples can be homogenized or filtered using filters and screen meshes. When the sample is filtered through a 0,45 μm membrane filter, the TOC result represents dissolved organic carbon (DOC) and dissolved TN & TP as specified in ISO 8245, results from samples analysed without filtration are reported as TOC, TN_b and TP_b .

If a homogenized sample containing large suspended material produces results (obtained from replicate measurements) which deviate by more than 10 % from each other, an accurate and precise TOC, TN_b , or TP_b result can not be quantified.

5.2 TOC

Inorganic carbon (for example CO_2 or ions of carbonic acid) present in the sample can interfere with the determination of TOC or DOC. Inorganic carbon is removed by acidification and sparging prior to sample oxidation.

NOTE 1 Purgeable organic carbon (POC) compounds, such as benzene, toluene, cyclohexane and chloroform, can partly escape during the acidification and sparging of the TIC. In the presence of these substances, the TOC concentration can be determined by applying the TC-TIC method (see [Annex A](#)).

Interference from chloride can occur at chloride concentrations greater than 3 %. If chloride interference is suspected, dilute the sample and re-analyse.

5.3 TN

Depending on sample matrix conditions, interferences may occur with the measurement of TN_b . If interferences are suspected, perform a suitable sample dilution, or by applying standard addition techniques.

Interference from chloride can occur at chloride concentrations greater than 3 %. If chloride interference is suspected, dilute the sample and re-analyse.

5.4 TP

Depending on sample matrix conditions, interferences can occur with the measurement of TP_b . If interferences are suspected, perform a suitable sample dilution, or by applying standard addition techniques.

Interference from chloride can occur at chloride concentrations greater than 3 %. If chloride interference is suspected, dilute the sample and re-analyse.

6 Reagents

Use only reagents of analytical grade.

Dry all solid reagents for at least 1 h at $(105 \pm 5) ^\circ\text{C}$. Store the dried solid in a desiccator before weighing.

NOTE There is no need to dry cellulose before usage.

Prepare concentrations and volumes of solutions as described in the manufacturers' instrument manuals. Alternatively, use commercially available stock solutions at the required concentration.

6.1 Water.

The contents of bound nitrogen, phosphorus, and carbon in water being used for the preparation of samples and solutions shall be sufficiently low to be negligible in comparison with the lowest TOC, TN_b and TP_b concentration to be determined.

6.2 Sulfuric acid, $\rho = 1,84 \text{ g/ml}$.

6.3 Manganese sulfate monohydrate, $\text{MnSO}_4 \cdot \text{H}_2\text{O}$, $\geq 99 \%$.

6.4 Sodium hydroxide, NaOH , $\geq 97 \%$.

6.5 Hydrochloric acid, $\omega(\text{HCl}) = 37 \%$.

6.6 Ammonium heptamolybdate tetrahydrate, $[(\text{NH}_4)_6\text{Mo}_7\text{O}_{24} \cdot 4\text{H}_2\text{O}]$, $\geq 99 \%$.

6.7 Ammonium metavanadate, NH_4VO_3 , $\geq 99 \%$.

6.8 Potassium hydrogen phthalate, $(\text{C}_8\text{H}_5\text{KO}_4)$, $\geq 99,7 \%$.

6.9 Sodium nitrate, NaNO_3 , $\geq 99 \%$.

6.10 Potassium dihydrogen phosphate, KH_2PO_4 , $\geq 99 \%$.

6.11 Nicotinic acid, $\text{C}_6\text{H}_5\text{NO}_2$, $> 99,5 \%$.

6.12 Triethyl phosphate.

6.13 Cellulose, $(\text{C}_6\text{H}_{10}\text{O}_5)_n$, microcrystalline, of particle size ranging from 0,02 mm to 0,1 mm.

7 Solution preparations

7.1 Blank solution

Use water (6.1) as the blank solution.

7.2 Sample oxidation solutions

7.2.1 Sulfuric acid TIC and catalyst solution

Prepare the sulfuric acid TIC and catalyst by slowly adding 49,9 ml of sulfuric acid (6.2) to 800 ml of water (6.1) in a 1 000 ml volumetric flask. Dilute to volume with water (6.1) and cool to room temperature. Add 0,04 g of manganese sulfate (6.3) to the sulfuric acid/water mixture. Determine the acid normality and adjust to $(1,80 \text{ N} \pm 0,01\text{N}) \text{ H}_2\text{SO}_4$.

NOTE For higher sulfuric acid TIC and catalyst oxidation concentration, maintain an acid to base concentration (7.2.2) ratio of 3:2, keeping the amount of manganese sulfate at 0,04 g.

The range of sulfuric acid normality should be 1,8 N to 2,5 N.

7.2.2 Base oxidation solution

Prepare the sodium hydroxide solution by slowly adding 48 g of NaOH (6.4) to a 1 000 ml volumetric flask containing approximately 500 ml of water (6.1). Mix the solution until the NaOH has dissolved, then stopper the flask and allow the solution to come to room temperature. Bring solution to volume with water (6.1).

For higher sodium hydroxide concentrations, maintain an acid (7.2.1) to base concentration ratio of 3:2.

The range of sodium hydroxide molarity should be 1,2 M to 1,7 M.

7.2.3 TP solution

7.2.3.1 Prepare the hydrochloric acid (A) solution by adding 249 ml of concentrated HCl (6.5) to an empty 1 000 ml volumetric flask. Slowly add 600 ml water (6.1) to the flask. Mix gently and allow the solution to come to room temperature. Bring the solution to volume with water (6.1), then stopper the flask.

7.2.3.2 Prepare the ammonium heptamolybdate tetrahydrate (B) solution by adding 25 g of ammonium heptamolybdate tetrahydrate (6.6) to 300 ml of water (6.1) in a 500 ml beaker. Stir to dissolve.

7.2.3.3 Prepare the ammonium metavanadate solution (C) by adding 2,5 g of ammonium metavanadate (6.7) to 300 ml of water (6.1) in a 500 ml beaker. Dissolve by bringing the solution to a boil. Cool to room temperature.

7.2.3.4 While stirring, slowly add the contents of solution A (7.2.3.1) to solution B (7.2.3.2) in a 1 000 ml volumetric flask. Cool to room temperature.

7.2.3.5 Finally while stirring, slowly pour the content of solution C (7.2.3.3) into the volumetric flask that contains the mixture of solution A and B (7.2.3.4) and bring to volume with water (6.1).

7.2.4 TP hydrolysis solution

Prepare the hydrochloric acid solution by adding 249 ml of concentrated HCl (6.5) to an empty 1 000 ml volumetric flask. Slowly add 600 ml water (6.1) to the flask. Mix gently and allow the solution to come to room temperature. Bring the solution to volume with water (6.1), then stopper the flask.