



SLOVENSKI STANDARD
oSIST prEN IEC 62990-1:2022
01-februar-2022

Zrak na delovnem mestu - 1. del: Detektorji plina - Zahteve za delovanje detektorjev strupenih plinov

Workplace Atmospheres - Part 1: Gas detectors - Performance requirements of detectors for toxic gases

**iTeh STANDARD
PREVIEW**

Atmosphères des lieux de travail - Partie 1: Détecteurs de gaz - Exigences d'aptitude à la fonction des détecteurs de gaz toxiques

(standards.iteh.ai)

Ta slovenski standard je istoveten z: EN IEC 62990-1

<https://standards.iteh.ai/catalog/standards/sist/d6b0874d-4a42-4612-b941-e53240fd6378/osist-pren-iec-62990-1-2022>

ICS:

- | | | |
|-----------|----------------------------------|---------------------------|
| 13.040.30 | Kakovost zraka na delovnem mestu | Workplace atmospheres |
| 13.320 | Alarmni in opozorilni sistemi | Alarm and warning systems |

oSIST prEN IEC 62990-1:2022

en,fr,de

**iTeh STANDARD
PREVIEW
(standards.iteh.ai)**

oSIST prEN IEC 62990-1:2022

<https://standards.iteh.ai/catalog/standards/sist/d6b0874d-4a42-4bf2-b941-e53240fd6378/osist-pren-iec-62990-1-2022>

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

DRAFT
prEN IEC 62990-1

December 2021

ICS 29.260.20

English Version

Workplace Atmospheres - Part 1: Gas detectors - Performance requirements of detectors for toxic gases (IEC 62990-1:2019 + COR1:2019)

Atmosphères des lieux de travail - Partie 1: Détecteurs de gaz - Exigences d'aptitude à la fonction des détecteurs de gaz toxiques
(IEC 62990-1:2019 + COR1:2019)

To be completed
(IEC 62990-1:2019 + COR1:2019)

This draft European Standard is submitted to CENELEC members for enquiry.
Deadline for CENELEC: 2022-03-04.

The text of this draft consists of the text of IEC 62990-1:2019 (31/1463/FDIS) + COR1:2019.

If this draft becomes a European Standard, CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

This draft European Standard was established by CENELEC in three official versions (English, French, German).
A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Recipients of this draft are invited to submit, with their comments, notification of any relevant patent rights of which they are aware and to provide supporting documentation.

Warning : This document is not a European Standard. It is distributed for review and comments. It is subject to change without notice and shall not be referred to as a European Standard.



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

prEN IEC 62990-1:2021 (E)**European foreword**

This document (prEN IEC 62990-1:2021) consists of the text of IEC 62990-1:2019 and IEC 62990-1:2019/COR1:2019 prepared by IEC/TC 31 "Equipment for explosive atmospheres" in cooperation with ISO/TC 146 "Air quality, sub-committee 2: Workplace atmospheres".

This document is currently submitted to the Enquiry.

The following dates are proposed:

- latest date by which the existence of this document (doa) dor + 6 months
has to be announced at national level
- latest date by which this document has to be (dop) dor + 12 months
implemented at national level by publication of an
identical national standard or by endorsement
- latest date by which the national standards (dow) dor + 36 months
conflicting with this document have to be withdrawn (to be confirmed or
modified when voting)

**iTeh STANDARD
PREVIEW
(standards.iteh.ai)**

[oSIST prEN IEC 62990-1:2022](https://standards.iteh.ai/catalog/standards/sist/d6b0874d-4a42-4bf2-b941-e53240fd6378/osist-pren-iec-62990-1-2022)

<https://standards.iteh.ai/catalog/standards/sist/d6b0874d-4a42-4bf2-b941-e53240fd6378/osist-pren-iec-62990-1-2022>



IEC 62990-1

Edition 1.0 2019-06

INTERNATIONAL STANDARD

NORME INTERNATIONALE



iTeh STANDARD

Workplace atmospheres – **PREVIEW**
Part 1: Gas detectors – Performance requirements of detectors for toxic gases
(standards.iteh.ai)

Atmosphères des lieux de travail –
Partie 1: Détecteurs de gaz – Exigences d'aptitude à la fonction des détecteurs
de gaz toxiques

<https://standards.iteh.ai/catalog/standards/sist/d6b0874d-4a42-4bf2-b941-e53240fd6378/osist-pren-iec-62990-1-2022>

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 29.260.20

ISBN 978-2-8322-7080-6

**Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	6
1 Scope.....	7
2 Normative references	8
3 Terms and definitions	8
4 General requirements	16
4.1 Overview.....	16
4.2 Design	17
4.2.1 General	17
4.2.2 Indicating devices	17
4.2.3 Alarm signal	19
4.2.4 Fault signals	19
4.2.5 Adjustments.....	20
4.2.6 Battery-powered equipment	20
4.2.7 Gas detection transmitter for use with separate gas detection control units	20
4.2.8 Separate gas detection control units for use with gas detection transmitter(s).....	21
4.2.9 Software-controlled equipment	21
4.3 Labelling and marking.....	22
4.4 Instruction manual	23
5 Test methods.....	25
5.1 General.....	25
5.2 General requirements for tests.....	26
5.2.1 Samples and sequence of tests.....	26
5.2.2 Preparation of equipment before each test.....	27
5.2.3 Mask for calibration and test.....	28
5.3 Normal conditions for test	28
5.3.1 General	28
5.3.2 Test gas(es)	28
5.3.3 Flow rate for test gases	28
5.3.4 Power supply.....	29
5.3.5 Temperature.....	29
5.3.6 Pressure.....	29
5.3.7 Humidity	29
5.3.8 Acclimation time	29
5.3.9 Orientation.....	29
5.3.10 Communications options.....	29
5.3.11 Gas detection equipment as part of systems.....	30
5.4 Tests	30
5.4.1 General	30
5.4.2 Unpowered storage	30
5.4.3 Measurement of deviations	31
5.4.4 Mechanical tests.....	31
5.4.5 Environmental tests	33
5.4.6 Performance tests	36
5.4.7 Electrical tests.....	41

5.4.8	Stability	44
5.4.9	Fault signal tests	45
5.4.10	Software controlled equipment.....	46
5.4.11	Protection against water	46
6	Uncertainty of measurement and lower limit of measurement for type HM equipment	47
6.1	General.....	47
6.2	Method of calculation of uncertainty of measurement	47
6.2.1	Basic concept.....	47
6.2.2	Sources of uncertainty.....	48
6.2.3	Calculation of relative expanded uncertainty.....	53
6.3	Method of calculation of lower limit of measurement	54
6.4	Acceptance criteria	55
6.4.1	Uncertainty of measurement.....	55
6.4.2	Lower limit of measurement.....	55
6.5	Relation between uncertainty and accuracy	55
Annex A	(normative) Gas specific performance requirements	56
Annex B	(informative) Determination of time of response and time of recovery	57
B.1	Aspirated equipment.....	57
B.1.1	Test rig.....	57
B.1.2	Equipment without internal pump.....	57
B.1.3	Equipment with internal pump.....	57
B.2	Equipment that samples by diffusion.....	58
B.2.1	Calibration mask method	58
B.2.2	Diffusion or flow methods.....	58
Bibliography	60
Figure 1	– Relationship between indication range and measuring range	14
Figure 2	– Warm-up time in clean air (typical).....	15
Figure 3	– Warm-up time in standard test gas (typical)	15
Figure 4	– Example of zero uncertainty.....	16
Figure B.1	– Schematic example of test rig for use with aspirated equipment.....	58
Figure B.2	– Schematic example of test chamber for diffusion method.....	59
Figure B.3	– Schematic example of test chamber for flow method.....	59
Table 1	– Factors to be considered in the expanded uncertainty estimate	48
Table A.1	– Gas specific performance requirements	56

INTERNATIONAL ELECTROTECHNICAL COMMISSION

WORKPLACE ATMOSPHERES –

**Part 1: Gas detectors –
Performance requirements of detectors for toxic gases**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as “IEC Publication(s)”). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 62990-1 has been prepared by Joint Working Group (JWG) 45 of IEC technical committee 31: Equipment for explosive atmospheres in cooperation with ISO technical committee 146: Air quality, sub-committee 2: Workplace atmospheres.

It is published as a dual logo standard.

The text of this International Standard is based on the following documents:

FDIS	Report on voting
31/1463/FDIS	31/1480/RVD

Full information on the voting for the approval of this International Standard can be found in the report on voting indicated in the above table.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

A list of all parts in the IEC 62990 series, published under the general title *Workplace atmospheres*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "http://webstore.iec.ch" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

The contents of the corrigendum of December 2019 have been included in this copy.

IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

**ITeH STANDARD
PREVIEW
(standards.iteh.ai)**

oSIST prEN IEC 62990-1:2022

<https://standards.iteh.ai/catalog/standards/sist/d6b0874d-4a42-4bf2-b941-e53240fd6378/osist-pren-iec-62990-1-2022>

INTRODUCTION

This part of IEC 62990 specifies general requirements for construction, testing and performance of equipment intended to measure the concentration of toxic gas and vapour in workplace atmospheres and other industrial and commercial applications. The performance requirements are intended to apply under environmental conditions present at the site of operation. However, because a wide range of environmental conditions are encountered in practise, this document specifies requirements that have to be fulfilled by equipment when tested under prescribed laboratory conditions.

This document applies to the following types of equipment: Health Monitoring (HM) and Safety Monitoring (SM). For a given measurement task of Type HM equipment the range over which the requirements must be met depends on the occupational exposure limit value. However, for most toxic gases and vapours the occupational exposure limit values have not been harmonized at the international level. Therefore, it was decided to use a reference value instead of the occupational exposure limit value for the performance tests. The list of reference values is given in Annex A. The reference values chosen are equal to or close to the occupational exposure limit values used in different countries but are intended to be used only for type testing equipment without any legal implications.

Electrical equipment used for the direct detection and direct concentration measurement of toxic gases and vapours generate readings in clean air (nominally zero), which vary with environmental conditions and time. This document therefore includes test methods and requirements for acceptable variations in measured values at application of zero gas and of defined test gases.

For gas detection equipment including additional function for detecting flammable gas and/or oxygen, consideration should be given to the relevant standards.

General requirements for construction, testing and performance of gas detectors for flammable gases and vapours are set out in IEC 60079-29-1, *Explosive atmospheres – Part 29-1: Gas detectors – Performance requirements of detectors for flammable gases*.

General requirements for construction, testing and performance of open path detectors for flammable gases are set out in IEC 60079-29-4, *Explosive atmospheres – Part 29-4: Gas detectors – Performance requirements of open path detectors for flammable gases*.

Guidance for the selection, installation, use and maintenance of gas detecting equipment is set out in IEC 60079-29-2, *Explosive atmospheres – Part 29-2, Gas detectors – Selection, installation, use and maintenance of detectors for flammable gases and oxygen*.

Guidance for functional safety of fixed gas detection systems is set out in IEC 60079-29-3, *Explosive atmospheres – Part 29-3, Gas detectors – Guidance on functional safety of fixed gas detection systems*.

WORKPLACE ATMOSPHERES –

Part 1: Gas detectors – Performance requirements of detectors for toxic gases

1 Scope

This part of IEC 62990 specifies general requirements for design, function and performance, and describes the test methods that apply to portable, transportable, and fixed equipment for the detection and concentration measurement of toxic gases and vapours in workplace atmospheres and other industrial and commercial applications.

This document is applicable to continuously sensing equipment whose primary purpose is to provide an indication, alarm and/or other output function the purpose of which is to indicate the presence of a toxic gas or vapour in the atmosphere and in some cases to initiate automatic or manual protective action(s). It is applicable to equipment in which the sensor generates an electrical signal when gas is present.

This document applies to two types of equipment:

- Type HM (Health Monitoring) 'occupational exposure' equipment:
For occupational exposure measurement, the performance requirements are focused on uncertainty of measurement of gas concentrations in the region of Occupational Exposure Limit Values (OELV). The upper limit of measurement will be defined by the manufacturer in accordance with 4.2.1.
- Type SM (Safety Monitoring) 'general gas detection' equipment:
For general gas detection applications (e.g. safety warning, leak detection), the performance requirements are focused on alarm signalling. The upper limit of measurement will be defined by the manufacturer according to the intended use of the equipment.

In general, the requirements for accuracy will be higher for Type HM equipment than for Type SM equipment. The same equipment may meet the requirements of both Type HM and Type SM.

For equipment used for sensing the presence of multiple gases this document applies only to the detection of toxic gas or vapour.

This document is not applicable to equipment:

- with samplers and concentrators such as sorbents or paper tape having an irreversible indication;
- used for the measurement of gases and vapours related to the risk of explosion;
- used for the measurement of oxygen;
- used only in laboratories for analysis or measurement;
- used only for process measurement purposes;
- used in the domestic environment;
- used in environmental air pollution monitoring;
- used for open-path (line of sight) area gas measurement;
- used for ventilation control in car parks or tunnels.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60068-2-6, *Environmental testing – Part 2-6: Tests – Test Fc: Vibration (sinusoidal)*

IEC 60079-0, *Explosive atmospheres – Part 0: Equipment – General requirements*

IEC 60529, *Degrees of protection provided by enclosures (IP Code)*

IEC 61000-4-29, *Electromagnetic compatibility (EMC) – Part 4-29: Testing and measurement techniques – Voltage dips, short interruptions and voltage variations on d.c. input power port immunity tests*

IEC 61000-6-3, *Electromagnetic compatibility (EMC) – Part 6-3: Generic standards – Emission standard for residential, commercial and light-industrial environments*

IEC 61000-6-4, *Electromagnetic compatibility (EMC) – Part 6-4: Generic standards – Emission standard for industrial environments*

IEC 61326-1, *Electrical equipment for measurement, control and laboratory use – EMC requirements – Part 1: General requirements*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

3.1

gas properties

3.1.1

ambient air

normal atmosphere surrounding the equipment

3.1.2

clean air

air that is free of gases or vapours which the sensor is sensitive to or which influence the performance of the sensor

3.1.3

occupational exposure limit value

OELV

limit of the time-weighted average of the concentration of a chemical agent in the air within the breathing zone of a worker in relation to a specified reference period

Note 1 to entry: The term “limit value” is often used as a synonym for “occupational exposure limit value”, but the term “occupational exposure limit value” is preferred because there is more than one limit value (e.g., biological limit value and occupational exposure limit value).

Note 2 to entry: Occupational exposure limit values (OELVs) are often set for reference periods of 8 h, but can also be set for shorter periods or concentration excursions.

[SOURCE: ISO 18158:2016, 2.1.5.4, modified (Note 2 to entry is shortened)]

3.1.4

reference value

<for type HM equipment> value used as the basis to specify requirements such as measuring range, resolution and alarm set point

Note 1 to entry: The term reference value is used in this document because for most toxic gases and vapours the occupational exposure limit values have not been harmonized at the International level.

3.1.5

poisons

<for sensing elements> substances that lead to temporary or permanent change of performance, particularly loss of sensitivity of the sensing element

3.1.6

toxic gas

gas or vapour that can be harmful to human health and/or the performance of persons due to its physical or physico-chemical properties

Note 1 to entry: For the purpose of this document, the term “toxic gas” includes “toxic vapours”.

3.1.7

volume fraction

quotient of the volume of a specified component and the sum of the volumes of all components of a gas mixture before mixing, all volumes referring to the pressure and the temperature of the gas mixture

Note 1 to entry: The volume fraction and volume concentration take the same value if, at the same state conditions, the sum of the component volumes before mixing and the volume of the mixture are equal. However, because the mixing of two or more gases at the same state conditions is usually accompanied by a slight contraction or, less frequently, a slight expansion, this is not generally the case.

3.1.8

zero gas

gas recommended by the manufacturer, which is free of toxic gases and interfering and contaminating substances, the purpose of which is calibration or adjustment of the equipment zero

3.1.9

standard test gas

test gas with a composition specified for each item of equipment and gas and/or vapour to be used for all tests unless otherwise stated

3.1.10

time-weighted average concentration

TWA concentration

concentration of gas in air averaged over a reference period

3.1.11

reference period

specified period of time for which the occupational exposure limit value of a chemical agent applies

Note 1 to entry: The reference period is usually 8 h for long-term measurements and 15 min for short-term measurements.

[SOURCE: ISO 18158:2016, 2.1.5.7, modified (term “biological agent” deleted as not relevant to the current document)]

3.2 types of equipment

3.2.1

alarm-only equipment

equipment with an alarm but not having an indication of measured value

3.2.2

aspirated equipment

equipment that samples the gas by drawing it to the gas sensor

Note 1 to entry: A hand operated or electric pump is often used to draw gas to the sensor.

3.2.3

automatically aspirated equipment

aspirated equipment with an integral pump or separate pump, which is connected directly to the equipment

3.2.4

continuous duty equipment

equipment that is powered for long periods of time, but may have either continuous or intermittent sensing

Note 1 to entry: Within this document, all equipment is regarded as continuous duty.

3.2.5

diffusion equipment

equipment in which the transfer of gas from the atmosphere to the sensor takes place without aspirated flow

3.2.6

fixed equipment

equipment fastened to a support, or otherwise secured in a specific location when energized

3.2.7

portable equipment

equipment intended to be carried by a person during its operation

Note 1 to entry: Portable equipment is battery powered and includes, but is not limited to

- a) hand-held equipment, typically less than 1 kg, which requires use of only one hand to operate,
- b) personal monitors, similar in size and mass to the hand-held equipment, that are continuously operating while they are attached to the user, and
- c) larger equipment that can be operated by the user while it is carried either by hand, by a shoulder strap or carrying harness and which may or may not have a hand directed probe.

3.2.8

transportable equipment

equipment not intended to be carried by a person during operation, nor intended for fixed installation

3.2.9

gas detection transmitter

fixed gas detection equipment that provide a conditioned electronic signal or output indication to a generally accepted industry standard (such as 4-20 mA), intended to be utilized with separate gas detection control units or signal processing data acquisition, central monitoring and similar systems, which typically process information from various locations and sources including, but not limited to gas detection equipment

3.2.10**gas detection control unit**

equipment intended to provide display indication, alarm functions, output contacts or alarm signal outputs or any combination when operated with remote sensor(s)

3.2.11**separate gas detection control unit**

equipment intended to provide display indication, alarm functions, output contacts or alarm signal outputs or any combination when operated with gas detection transmitter(s)

3.2.12**equipment with integral sensor(s)**

equipment that provides display indication, alarm functions, output contacts and/or alarm signal outputs using a sensor which is within or directly assembled to the equipment housing

3.2.13**accessory**

component which can be fitted to the equipment for special purpose

EXAMPLE: External gas pump, sampling probe, hoses, collecting cone, weather protection device.

3.3**sensors****3.3.1****sensing element**

part of the sensor which is sensitive to the gas/vapour to be measured

3.3.2**sensor**

assembly in which the sensing element is housed and that may also contain associated circuit components

3.3.3**integral sensor**

sensor which is within or directly assembled to the equipment housing

3.3.4**remote sensor**

sensor which is installed separately, but is connected to a gas detection control unit, gas detection transmitter, or to transportable or portable equipment

3.4**supply of gas to equipment****3.4.1****sample line**

means by which the gas being sampled is conveyed to the sensor

Note 1 to entry: Accessories such as filter or water trap are often included in the sample line.

3.4.2**sampling probe**

separate accessory sample line which is optionally attached to the equipment

Note 1 to entry: It is usually short (e.g. of the order of 1 m) and rigid, although it can be telescopic. In some cases, it is connected by a flexible tube to the equipment.

iTeh STANDARD
PREVIEW
(standards.iteh.ai)

<https://standards.iteh.ai/catalog/standards/sist/d6b0874d-4a42-4bf2-b941-e53240fd6378/osist-pren-iec-62990-1-2022>