

~~ISO/TC 194~~

~~Date: 2022-09-01~~

~~ISO/DTR 10993-55:2022(E)~~

ISO/TC 194/WG 5

Secretariat: DIN

**Biological evaluation of medical devices — Part 55: Interlaboratory study on cytotoxicity**

iTeh STANDARD PREVIEW

Copyright notice

~~This~~ First edition

~~ISO/TR 10993-55:2023~~ Date: 2022-09-30

<https://standards.iteh.ai/catalog/standards/sist/c0db1318-fd0c-4f75-bf11-b675fb6852af/iso-tr-10993-55-2023>

© ISO document is a working draft 2022

All rights reserved. Unless otherwise specified, or committee draft and is copyright protected by ISO. While required in the reproduction context of working drafts or committee drafts in any form for use by participants in the ISO standards development process is permitted without prior permission from ISO, neither its implementation, no part of this document nor any extract from its publication may be reproduced, stored or utilized otherwise in any form or transmitted in any form for any other purpose by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission from ISO. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO Copyright Office

CP 401 • CH-1214 Vernier, Geneva

Phone: + 41 22 749 01 11

Email: [copyright@iso.org](mailto:copyright@iso.org)

Website: [www.iso.org](http://www.iso.org) Requests for permission to reproduce this document for the purpose of selling it should be addressed as shown below or to ISO's member body in the country of the requester.

[Indicate the full address, telephone number, fax number, telex number, and electronic mail address, as appropriate, of the Copyright Manager of the ISO member body responsible for the secretariat of the TC or SC within the framework of which the working document has been prepared.]

Reproduction for sales purposes may be subject to royalty payments or a licensing agreement.

Violators may be prosecuted.

<https://standards.iteh.ai/catalog/standards/sist/c0db1318-fd0c-4f75-bf11-b675fb6852af/iso-tr-10993-55-2022>  
Published in Switzerland.

<b>Contents</b>	<b>Page</b>
Foreword.....	iii
Introduction .....	iv
1 Scope.....	1
2 Normative references .....	1
3 Terms and definitions.....	1
4 Participants .....	1
5 Materials and Sample preparation .....	2
6 Test procedures .....	2
7 Results.....	2
7.1 Neutral Red Uptake.....	2
7.1.1 Sodium Lauryl Sulfate as positive control .....	2
7.1.2 Test samples.....	3
7.2 Colony Formation Assay.....	3
7.2.1 Negative reference material .....	4
7.2.2 Positive reference materials.....	5
8 Assessment of results.....	7
Annex A (informative) Interlaboratory study protocol for the neutral red uptake (NRU) cytotoxicity test.....	9
Annex B (informative) Interlaboratory study protocol for the colony formation cytotoxicity test.....	17
Bibliography.....	24

<https://standards.iteh.ai/catalog/standards/sist/c0db1318-fd0c-4f75-bf11-b675fb6852af/iso-tr-10993-55-2023>

ISO/DTR 10993-55:2022(E)

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 194, *Biological and clinical evaluation of medical devices*.

A list of all parts in the ISO 10993 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

## Introduction

The first ~~version~~edition of ISO 10993-5, published in 1992, allowed several different ways to assess cytotoxicity of medical devices and gave an imprecise description of how to perform the tests. Qualitative assays were accepted and only a small amount of guidance was given for the interpretation of the results. Not surprisingly, the first interlaboratory study ~~by the members of working group 5 (WG 5) of ISO/TC 194~~ in 2000 resulted in quite low reproducibility of results. ~~It was therefore the consensus of WG 5 to include~~Therefore, detailed protocols were included into the standard and ~~to evaluate~~ in another study the practicability of the protocols and reference materials ~~proposed by the working group members were~~ evaluated. The results of this second interlaboratory study mainly influenced the revision of ~~the standard~~ISO 10993-5, which was published in 2009.

This ~~technical report is~~document provides the historical report of the second interlaboratory study, conducted in 2006.

iTeh STANDARD PREVIEW  
(standards.iteh.ai)

ISO/TR 10993-55:2023

<https://standards.iteh.ai/catalog/standards/sist/c0db1318-fd0c-4f75-bf11-b675fb6852af/iso-tr-10993-55-2023>



## Biological evaluation of medical devices — Part 55: Interlaboratory study on cytotoxicity

### 1 Scope

This document describes the results of an international interlaboratory study conducted in 2006 to evaluate the performance of two different test protocols in terms of the cytotoxic effects in the biological evaluation of medical devices. The results of these tests were used for the revision of ISO 10993-5-~~[1,2]~~. Furthermore, the results of these tests were used to estimate the accuracy of these test systems with living cells to define a threshold what is considered a cytotoxic effect.

NOTE The determination of cytotoxic effects has a high relevance in the biological evaluation of medical devices; according to ISO 10993-1<sup>[1]</sup> it is one of the very few tests which are proposed to be performed for every kind of device.

### 2 Normative references

There are no normative references in this document.

### 3 Terms and definitions

No terms and definitions are listed in this document.

ISO and IEC maintain ~~terminological~~terminology databases for use in standardization at the following addresses:

— ISO Online browsing platform: available at <https://www.iso.org/obp>

— IEC Electropedia: available at <https://www.electropedia.org/>

### 4 Participants

Twelve laboratories participated in this study<sup>1</sup> ~~mainly from entities represented in ISO/TC 194, which is responsible for this document~~. Eleven reports on a neutral red uptake (NRU) assay ~~were received~~ and ~~ten~~10 reports on a colony formation (CF) assay ~~were received~~. Four participants were commercial test laboratories, four participants were internal test laboratories of medical device manufacturers and four laboratories were in research institutes.

The laboratories were located in six different countries.

<sup>1</sup> The participating laboratories include: Deutsche Institute für Textil- und Faserforschung, Germany; Hatano Research Institute, Food and Drug Safety Center, Japan; Medical University Vienna, Austria; National Institute of Health Sciences, Japan; Envigo CRS GmbH, Germany; Terumo Corporation R&D, Japan; BD Technologies, United States; NAMSA, United States; Gambro BCT, United States and three other laboratories.

one each in Austria, France and the Netherlands, and three each in Germany, Japan and the United States.

## 5 Materials and sample preparation

The following materials were used for the study:

- a) ~~reference material-C~~ [RM-C; ~~Hatano Research Institute~~ (HRI)]: high density polyethylene sheet;
- b) ~~RM-A~~ (HRI): segmented polyurethane film containing 0,1 % zinc diethyldithiocarbamate (ZDEC);
- c) ~~RM-B~~ (HRI): segmented polyurethane film containing 0,25 % zinc dibutyldithiocarbamate (ZDBC).

RM-C (HRI), RM-A (HRI) and RM-B (HRI) have been widely used as reference materials for cytotoxicity tests of medical devices. The Food and Drug Safety Center of ~~the Hatano Research Institute~~ (HRI) (~~Ochiai 729-5, Hadanoshi, Kanagawa 257-8523, Japan~~) has certified these materials ~~and offers them for sale~~. HRI agreed to provide them for the interlaboratory study. Test samples were cut (2 mm ~~x~~ 15 mm) and sterilized with ethylene oxide (EO) and were distributed from HRI to the participants. Extraction was then performed in the participating laboratories according to the protocols.

## 6 Test procedures

Two test protocols were chosen by the working group developing the tests for cytotoxicity in vitro: ~~neutral red uptake~~ (NRU) and ~~colony formation~~ (CF). The NRU assay protocol is based on the protocol, which was used in a validation study of Interagency Coordinating Committee on the Validation of Alternative Methods (ICCVAM).<sup>[4]</sup> The CF assay protocol is based on the cytotoxicity test of the Japanese guidelines for basic biological tests of medical materials and devices.<sup>[5]</sup> These original protocols were modified to meet the requirements of this specific study (see ~~Annexes A and B~~). The protocols were sent to the participants together with the test materials.

## 7 Results

### 7.1 Neutral red uptake

#### 7.1.1 General

Eleven laboratories participated in this study. All test samples were extracted once as described in ~~Annex A~~. Each concentration of the dilution series was tested in six replicates. The mean values were used to calculate the concentration producing a 50 % inhibition of cell viability ( $IC_{50}$ ) values.

#### 7.1.2 Sodium lauryl sulfate as positive control

The laboratories ~~were using~~ different internal reference materials as positive controls. It was therefore decided that all participants use the same common chemical substance as positive control and sodium lauryl sulfate (SLS, CAS #~~Registry Number~~<sup>®</sup> 151-21-3<sup>2</sup>) was selected for this purpose. ~~is a summary of Table 1 summarizes~~ the results.

Table 1 —  $IC_{50}$ -values of SLS in the NRU assay

<sup>2</sup> CAS Registry Number<sup>®</sup> is a trademark of CAS corporation. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO of the product named. Equivalent products may be used if they can be shown to lead to the same results.



Laboratory		Laboratory $IC_{50}$ µg/ml										
1		34.0										
2		83.0										
	1	2	3	4	5	6	7	8	9	10	11	
$IC_{50}$ µg/ml	34,0	83,0	62,4	77,0	85,9	75,6	67,8	47,8	49,8	62,1	22,0	
5		85.9										
6		75.6										
7		67.8										
8		47.8										
9		49.8										
10		62.1										
11		22.0										

- Deleted Cells
- Deleted Cells
- Deleted Cells
- Deleted Cells
- Deleted Cells
- Deleted Cells
- Deleted Cells
- Deleted Cells
- Deleted Cells
- Deleted Cells

The variation of the  $IC_{50}$  was from 22,0 µg/ml to 85,9 µg/ml, the mean  $IC_{50}$  was  $(60,7 \pm 20,4)$  µg/ml.

In Annex A, an  $IC_{50}$ -value between 70 µg/ml and 116 µg/ml was requested as acceptance criterion. This was an error in the 2009 version of ISO 10993-5:2009 and will be removed in the next version edition. Historical  $IC_{50}$ -values are typical for a specific laboratory but cannot be compared between laboratories.

7.1.27.1.3 Test samples

The three different test samples RM-A, RM-B and RM-C were extracted as described in Annex A and the extracts were diluted as defined. The cell viabilities at different extract concentrations were determined as described in Annex A. The  $IC_{50}$  was determined from the concentration-response. This was done by using validated software, which is available in public, see Reference [6]. The results of the eleven participants are summarized in Table 2 and illustrated in Figure 1. Initially, the testing was conducted using the following concentrations of the RM-A extract: 0,25 %, 0,5 %, 1,0 %, 2,0 %, 3,0 % and 4,0 %. Unexpectedly,  $IC_{50}$ -values were higher than expected from the colony formation assay, because the neutral red assay is less sensitive probably due to the shorter exposure time. Therefore, the laboratories repeated the test with the following concentrations of the RM-A extract: 5 %, 10 %, 20 %, 30 %, 40 % and 50 %.

Table 2 —  $IC_{50}$ -values of sample extracts in the NRU assay


Laboratory		Laboratory $IC_{50}$ %										
		RM-A	RM-B	RM-C								
1		16.5	32.0	=								
2		26.4	54.0	=								
	1	2	3	4	5	6	7	8	9	10	11	
RM-A	16,5%	26,4%	18,5%	15,3%	20,5%	15,6%	24,3%	15,1%	11,7%	6,7%	16,7%	
	4%	5%	4%	6%	6%	=	%	%	%			

- Merged Cells
- Deleted Cells
- Deleted Cells
- Deleted Cells
- Deleted Cells
- Deleted Cells
- Deleted Cells
- Deleted Cells
- Deleted Cells
- Deleted Cells
- Deleted Cells
- Deleted Cells
- Deleted Cells
- Deleted Cells

	5	20.6	79.6	=										
RM-B	32,0%	54,0%	93,2%	56,6%	79,15,6%	45,6%	93,3%	43,3%	39,2%	89,4%	33,6%			
	RM-CZ	none	2	none	9	none	none	none	none	none	none	none	none	
	8	15.1	43.3	=										
	9	11.7	38.2	=										
	10	6.7	89.4	=										
	11	16.7	33.6	=										

- Deleted Cells
- Deleted Cells
- Deleted Cells
- Deleted Cells
- Deleted Cells
- Deleted Cells
- Deleted Cells
- Deleted Cells
- Deleted Cells
- Deleted Cells
- Deleted Cells
- Deleted Cells
- Deleted Cells
- Deleted Cells
- Deleted Cells
- Deleted Cells
- Deleted Cells
- Deleted Cells
- Deleted Cells
- Deleted Cells

Results for RM-A varied from 6,7 % to 26,4 %, mean  $IC_{50}$  was  $(17,0 \pm 5,5)$  %. Results for RM-B varied from 32,0 % to 93,3%, the mean  $IC_{50}$  was  $(59,9 \pm 24,4)$  %.



**Key**  
 X — Laboratory number  
 Y —  $IC_{50}$   
 ■ — Reference material A  
 ■ — Reference material B

X laboratory number  
 Y IC<sub>50</sub>

reference material A  
 reference material B

Figure 1 — Comparison of IC<sub>50</sub>-values of sample extracts in the NRU assay

7.2 Colony formation assay

7.2.1 General

The test samples RM-A, RM-B and RM-C were those materials, which were already recommended to be used as reference materials in the Japanese Guidelines for the colony formation assay. For the study only these materials were used by the laboratories, to assess the differences between individual laboratories. The test samples were extracted as described in Annex B and the extracts were diluted as proposed. The cell viabilities at different extract concentrations were determined by counting the colonies formed [plating efficiency (PE)]. IC<sub>50</sub> was calculated as the dose with 50 % PE which was calculated from the line which passed through a dose with higher PE and a dose with lower PE than 50 %.

Ten of the twelve participants in the NRU study also participated in the CF assay and communicated their results. All test samples were extracted once as described in Annex B. Each concentration of the dilution series was tested in triplicate. The mean values were used to calculate IC<sub>50</sub>-values.

The plating efficiency of the controls in the different laboratories is listed in Table 3.

Table 3 — Control PE of the CF assay in the participating laboratories

Laboratory								Plating efficiency	
								%	
1								68,0	
2								62,8	10
3								75,7	
4								101,5	
5								92,2	
7								71,2	
8								108,7	
9								85,0	
10	68,0%	62,8%	75,7%	101,5%	92,2%	71,2%	108,7%	85,0%	106,3%
12								75,0%	
									106,3

The PE in the controls varied from 62,8 % to 108,7 %, mean value was (84,6 ± 15,8) %.

7.2.1.2.2 Negative reference material

The test sample RM-C is certified not to give any positive response in the test. Nevertheless, the extract of RM-C was used in this study to detect the variation of results in this biological system. The results of the 10 participants are summarized in Table 4 and illustrated in Figure 2.

Table 4 — Plating efficiencies of RM-C in the CF assay

Inserted Cells

Deleted Cells

Deleted Cells

Deleted Cells

Deleted Cells

Deleted Cells

Deleted Cells

Deleted Cells

Deleted Cells

Deleted Cells

Deleted Cells

Deleted Cells

Deleted Cells

Deleted Cells


Deleted Cells

Deleted Cells

Concentration of RM-C %	Laboratory Plating efficiency %									
	Lab 1	Lab 2	Lab 3	Lab 4	Lab 5	Lab 7	Lab 8	Lab 9	Lab 10	Lab 12
0	100,0%	100,0%	100,0%	100,0%	100,0%	100,0%	100,0%	100,0%	100,0%	100,0%
25	112,0%	96,8%	100,0%	100,2%	93,3%	97,8%	85,3%	102,4%	98,7%	101,9%
50	94,0%	87,1%	81,5%	100,5%	103,1%	92,6%	101,8%	94,9%	79,6%	99,4%
75	110,0%	92,9%	96,0%	97,5%	92,9%	98,2%	96,6%	98,5%	85,3%	98,4%
100	106,0%	94,8%	110,1%	97,2%	95,1%	91,2%	92,0%	96,1%	77,8%	98,1%



ISO/TR 10993-55:2023  
<https://standards.iteh.ai/catalog/standards/sist/c0db1318-fd0c-4f75-bf11-b675fb6852af/iso-tr-10993-55-2023>



**Key**  
 X — extract concentration in %  
 Y — plating efficiency in % of the control  
 Lab 1  
 Lab 2  
 Lab 3  
 Lab 4  
 Lab 5