



Designation: E1428 – 99(Reapproved 2009)

## Standard Test Method for Evaluating the Performance of Antimicrobials in or on Polymeric Solids Against Staining by *Streptover-* *reticulum* (A Pink Stain Organism)<sup>1</sup>

This standard is issued under the fixed designation E1428; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

### 1. Scope

1.1 This test method is used for determining the performance of antimicrobial agents used in or on synthetic polymeric solids against staining by the actinomycete *Streptover-*  
*reticulum*. This organism has been chosen as an indicator organism, although other organisms have been known to cause undesirable staining in polymeric solids.

1.2 This test method is not suitable for evaluating dark-pigmented test samples.

1.3 A knowledge of microbiological techniques is recommended for these procedures.

1.4 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.5 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

### 2. Summary of Test Method

2.1 Test specimens are placed on an agar surface inoculated with *Streptover-*  
*reticulum* and incubated. After incubation, test specimens are rated visually by percentage of sample area stained.

### 3. Significance and Use

3.1 This test method provides a technique for evaluating antimicrobials in or on polymeric solids against staining by *Streptover-*  
*reticulum*, and should assist in the prediction of performance of treated articles under actual field conditions.

3.2 Conditioning of the specimens, such as exposure to leaching, weathering, and heat treatment, may have significant effects on performance of antimicrobials against staining. Determination of these effects is not included in this test method.

### 4. Interferences

4.1 An interference may be caused by contamination of plates and agar by unwanted organisms that settle from the environment.

4.2 Dark pigments mask observation of the pink stain.

### 5. Apparatus

5.1 *Petri dishes*, 100 mm diameter.

NOTE 1—Presterilized and disposable plastic petri dishes are available from most laboratory supply houses.

5.2 *Cotton swabs*, sterile.

5.3 *Incubator*—Incubating equipment for this test method shall maintain a temperature of  $29 \pm 1^\circ\text{C}$ .

5.4 *Autoclave*.

5.5 *Sterilizer*, ethylene oxide (optional).

### 6. Reagents and Materials

6.1 *Yeast Malt Extract Agar (ISP Medium 2)*<sup>2</sup>—Prepare this medium according to manufacturer's directions.

6.2 *Inoculum*—*Streptover-*  
*reticulum* ATCC 25607.<sup>3</sup> Maintain stock cultures on yeast malt extract agar. The stock may be kept for not more than 12 months at approximately 3 to 10°C. Subcultures, incubated at  $29 \pm 1^\circ\text{C}$  for 7 to 14 days, shall be used for inoculation.

<sup>1</sup> This test method is under the jurisdiction of ASTM Committee E35 on Pesticides and Alternative Control Agents and is the direct responsibility of Subcommittee E35.15 on Antimicrobial Agents.

Current edition approved Oct. 1, 2009. Published November 2009. Originally approved in 1991. Last previous edition approved in 2004 as E1428 – 99(2004). DOI: 10.1520/E1428-99R09.

<sup>2</sup> The sole source of supply of Bacto-Yeast Malt Extract Agar, Stock No. 0770-01 known to the committee at this time is Difco Labs, P.O. Box 1058A, Detroit, MI 48232. If you are aware of alternative suppliers, please provide this information to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee,<sup>1</sup> which you may attend.

<sup>3</sup> American Type Culture Collection, 10801 University Blvd., Manassas, VA 20110.