
Medical suction equipment —
Part 1:
Electrically powered suction
equipment

AMENDMENT 1: Changes to
requirements for operating at extremes
of temperature

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Appareils d'aspiration médicale —

Partie 1: Appareils électriques d'aspiration

*AMENDEMENT 1: Modifications des exigences de fonctionnement à
des valeurs extrêmes de température*

<https://standards.iteh.ai>

[ISO 10079-1:2015/Amd 1:2018](https://standards.iteh.ai)

[47ad0cd93f5d/iso-10079-1-2015-amd-1-2018](https://standards.iteh.ai)



iTeh STANDARD PREVIEW
(standards.iteh.ai)

<https://standards.iteh.ai/catalog/standards/sist/2bc27243-99ba-4066-8a84-47ad0cd93f6d/iso-10079-1-2015-amd-1-2018>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2018

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 8, *Suction devices for hospital and emergency care use*.

ISO 10079-1:2015/Amd.1:2018
<https://standards.iteh.ai/catalog/standards/sist/2bc27243-99ba-4066-8a84-47ad0cd93f6d/iso-10079-1-2015-amd-1-2018>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO 10079-1:2015/Amd 1:2018

<https://standards.iteh.ai/catalog/standards/sist/2bc27243-99ba-4066-8a84-47ad0cd93f6d/iso-10079-1-2015-amd-1-2018>

Medical suction equipment —

Part 1:

Electrically powered suction equipment

AMENDMENT 1: Changes to requirements for operating at extremes of temperature

10.1

Replace subclause 10.1 with:

10.1 Operating conditions

10.1.1 Suction equipment intended for field and/or transport use shall meet the requirements of Clauses 7 and 9, as appropriate, after being subjected to temperatures of -18 °C and $+50\text{ °C}$.

Check compliance by the tests given in A.18.2.1 and A.18.2.2.

10.1.2 Manufacturers of suction equipment intended for field and/or transport use that cannot meet the requirements of Clauses 7 and 9 at the extremes of temperatures specified in 10.1.1 shall declare, in their instructions for use, [see 11.3 k)] the maximum and minimum temperatures at which the suction equipment can meet the requirements of Clauses 7 and 9.

Check compliance by inspection of the instructions for use.

ISO 10079-1:2015/Amd.1:2018
<https://standards.iteh.ai/catalog/standards/sist/2bc27243-99ba-4066-8a84-47ad0cd93f6d/iso-10079-1-2015-amd-1-2018>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[ISO 10079-1:2015/Amd 1:2018](https://standards.iteh.ai/catalog/standards/sist/2bc27243-99ba-4066-8a84-47ad0cd93f6d/iso-10079-1-2015-amd-1-2018)
<https://standards.iteh.ai/catalog/standards/sist/2bc27243-99ba-4066-8a84-47ad0cd93f6d/iso-10079-1-2015-amd-1-2018>