
Radiološka zaščita - Referenčno sevanje z rentgenskimi in gama žarki za kalibracijo dozimetров in merilnikov doze sevanja ter za ugotavljanje njihovega odzivanja kot funkcije fotonske energije - 1. del: Značilnosti sevanja in proizvodne metode (ISO 4037-1:2019)

Radiological protection - X and gamma reference radiation for calibrating dosimeters and dose rate meters and for determining their response as a function of photon energy - Part 1: Radiation characteristics and production methods (ISO 4037-1:2019)

Strahlenschutz - Röntgen- und Gamma-Referenzstrahlungsfelder zur Kalibrierung von Dosimetern und Dosisleistungsmessgeräten und zur Bestimmung ihres Ansprechvermögens als Funktion der Photonenenergie - Teil 1: Strahlungseigenschaften und Erzeugungsmethoden

Radioprotection - Rayonnements X et gamma de référence pour l'étalonnage des dosimètres et des débitmètres, et pour la détermination de leur réponse en fonction de l'énergie des photons - Partie 1: Caractéristiques des rayonnements et méthodes de production (ISO 4037-1:2019)

Ta slovenski standard je istoveten z: prEN ISO 4037-1

ICS:

17.240	Merjenje sevanja	Radiation measurements
--------	------------------	------------------------

oSIST prEN ISO 4037-1:2020

en,fr,de

INTERNATIONAL STANDARD

ISO
4037-1

Second edition
2019-01

Radiological protection — X and gamma reference radiation for calibrating dosimeters and doserate meters and for determining their response as a function of photon energy —

Part 1: Radiation characteristics and production methods

*Radioprotection — Rayonnements X et gamma de référence
pour l'étalonnage des dosimètres et des débitmètres, et pour la
détermination de leur réponse en fonction de l'énergie des photons —*

*Partie 1: Caractéristiques des rayonnements et méthodes de
production*



Reference number
ISO 4037-1:2019(E)

© ISO 2019

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

SIST EN ISO 4037-1:2021

<https://standards.iteh.ai/catalog/standards/sist/e65560ba-99cd-4cab-a471-d60b610e05f0/sist-en-iso-4037-1-2021>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2019

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	3
3 Terms and definitions	3
4 Continuous reference filtered X radiation	7
4.1 General	7
4.1.1 Realisation of reference radiation fields	7
4.1.2 Basis of conversion coefficients	7
4.1.3 Radiation quality	8
4.1.4 Choice of reference radiation	8
4.2 Conditions and methods for producing reference X radiation	13
4.2.1 Characteristics of the high voltage generator	13
4.2.2 Tube potential and protective resistor	14
4.2.3 Filtration	15
4.2.4 Limitations concerning matched fields	19
4.2.5 X radiation shutter	20
4.2.6 Beam aperture	20
4.3 Field uniformity and scattered radiation	20
4.3.1 Field diameter	20
4.3.2 Field uniformity	20
4.3.3 Scattered radiation	20
4.4 Summary of the requirements for reference X radiation fields	21
4.5 Validation of reference X radiation	21
4.5.1 General	21
4.5.2 Criteria for validation by HVL determination	22
4.5.3 Apparatus for HVL measurement	23
4.5.4 HVL measurement procedure	24
4.5.5 Criteria for validation by dosimetry	24
4.5.6 Criteria for validation by spectrometry	24
5 Gamma radiation emitted by radionuclides	25
5.1 General	25
5.2 Radionuclides used for the production of gamma radiation	25
5.3 Specification of radiation sources	25
5.3.1 Sources	25
5.3.2 Encapsulation	26
5.4 Irradiation facility and influence of scattered radiation	26
5.4.1 General requirements	26
5.4.2 Collimated geometry installation	26
5.4.3 Variation of air kerma rate by means of lead attenuators	27
5.5 Checking installation conformity	27
6 Photon radiation with energies between 4 MeV and 9 MeV	28
6.1 General	28
6.2 Production of reference radiation	28
6.2.1 General	28
6.2.2 Photon reference radiation from de-excitation of ^{16}O in the $^{19}\text{F}(p, \alpha\gamma)^{16}\text{O}$ reaction	28
6.2.3 Photon reference radiation from de-excitation of ^{12}C	30
6.3 Beam diameter and uniformity of radiation field	31
6.4 Contamination of photon reference radiation	31
6.4.1 General	31

ISO 4037-1:2019(E)

6.4.2	Contamination of reference radiation common to all methods of production of reference radiation.....	32
6.4.3	Additional contamination of accelerator produced reference radiation from de-excitation of ^{16}O	32
Annex A (informative) Fluorescence X radiation with not enough information for matched or characterized fields.....		33
Annex B (informative) Gamma radiation emitted by ^{241}Am radionuclide with not enough information for matched or characterized fields.....		40
Annex C (informative) Continuous filtered X radiation based on the quality index.....		42
Bibliography		45

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

SIST EN ISO 4037-1:2021

<https://standards.iteh.ai/catalog/standards/sist/e65560ba-99cd-4cab-a471-d60b610e05f0/sist-en-iso-4037-1-2021>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 85, *Nuclear energy, nuclear technologies and radiological protection*, Subcommittee SC 2, *Radiological protection*.

This second edition cancels and replaces the first edition (ISO 4037-1:1996), which has been technically revised. The main changes are:

- introduction of two types of reference fields, matched reference fields and characterized reference fields;
- introduction of validation for matched reference fields;
- introduction of limits for the allowed deviation of parameters like high voltage, filter purity and filter thickness from their nominal values. These limits now depend on the definition depth of the phantom related quantity. This is done to achieve an overall uncertainty ($k = 2$) of about 6 % to 10 % for the phantom related operational quantities.

A list of all the parts in the ISO 4037 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

ISO 4037-1:2019(E)**Introduction**

This maintenance release of this document incorporates the improvements to high voltage generators from 1996 to 2017 (e.g., the use of high frequency switching supplies providing nearly constant potential), and the spectral measurements at irradiation facilities equipped with such generators (e.g., the catalogue of X-ray spectra by Ankerhold^[4]). It also incorporates all published information with the aim to adjust the requirements for the technical parameters of the reference fields to the targeted overall uncertainty of about 6 % to 10 % for the phantom related operational quantities of the International Commission on Radiation Units and Measurements (ICRU)^[5]. It does not change the general concept of the existing ISO 4037.

ISO 4037 focusing on photon reference radiation fields is divided into four parts. ISO 4037-1 gives the methods of production and characterization of reference radiation fields in terms of the quantities spectral photon fluence and air kerma free-in-air. ISO 4037-2 describes the dosimetry of the reference radiation qualities in terms of air kerma and in terms of the phantom related operational quantities of the International Commission on Radiation Units and Measurements (ICRU)^[5]. ISO 4037-3 describes the methods for calibrating and determining the response of dosimeters and doserate meters in terms of the phantom related operational quantities of the ICRU^[5]. ISO 4037-4 gives special considerations and additional requirements for calibration of area and personal dosimeters in low energy X reference radiation fields, which are reference fields with generating potential lower or equal to 30 kV.

The general procedures described in ISO 29661 are used as far as possible in this document. Also, the symbols used are in line with ISO 29661.

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[SIST EN ISO 4037-1:2021](https://standards.iteh.ai/catalog/standards/sist/e65560ba-99cd-4cab-a471-d60b610e05f0/sist-en-iso-4037-1-2021)

<https://standards.iteh.ai/catalog/standards/sist/e65560ba-99cd-4cab-a471-d60b610e05f0/sist-en-iso-4037-1-2021>

Radiological protection — X and gamma reference radiation for calibrating dosimeters and doserate meters and for determining their response as a function of photon energy —

Part 1: Radiation characteristics and production methods

1 Scope

This document specifies the characteristics and production methods of X and gamma reference radiation for calibrating protection-level dosimeters and doserate meters with respect to the phantom related operational quantities of the International Commission on Radiation Units and Measurements (ICRU)[5]. The lowest air kerma rate for which this standard is applicable is $1 \mu\text{Gy h}^{-1}$. Below this air kerma rate the (natural) background radiation needs special consideration and this is not included in this document.

For the radiation qualities specified in [Clauses 4 to 6](#), sufficient published information is available to specify the requirements for all relevant parameters of the matched or characterized reference fields in order to achieve the targeted overall uncertainty ($k = 2$) of about 6 % to 10 % for the phantom related operational quantities. The X ray radiation fields described in the informative [Annexes A to C](#) are not designated as reference X-radiation fields.

NOTE The first edition of ISO 4037-1, issued in 1996, included some additional radiation qualities for which such published information is not available. These are fluorescent radiations, the gamma radiation of the radionuclide ^{241}Am , S-Am, and the high energy photon radiations R-Ti and R-Ni, which have been removed from the main part of this document. The most widely used radiations, the fluorescent radiations and the gamma radiation of the radionuclide ^{241}Am , S-Am, are included nearly unchanged in the informative [Annexes A and B](#). The informative [Annex C](#) gives additional X radiation fields, which are specified by the quality index.

The methods for producing a group of reference radiations for a particular photon-energy range are described in [Clauses 4 to 6](#), which define the characteristics of these radiations. The three groups of reference radiation are:

- a) in the energy range from about 8 keV to 330 keV, continuous filtered X radiation;
- b) in the energy range 600 keV to 1,3 MeV, gamma radiation emitted by radionuclides;
- c) in the energy range 4 MeV to 9 MeV, photon radiation produced by accelerators.

The reference radiation field most suitable for the intended application can be selected from [Table 1](#), which gives an overview of all reference radiation qualities specified in [Clauses 4 to 6](#). It does not include the radiations specified in the [Annexes A, B and C](#).

The requirements and methods given in [Clauses 4 to 6](#) are targeted at an overall uncertainty ($k = 2$) of the dose(rate) value of about 6 % to 10 % for the phantom related operational quantities in the reference fields. To achieve this, two production methods are proposed:

The first one is to produce “*matched reference fields*”, whose properties are sufficiently well-characterized so as to allow the use of the conversion coefficients recommended in ISO 4037-3. The existence of only a small difference in the spectral distribution of the “matched reference field” compared to the nominal reference field is validated by procedures, which are given and described in detail in ISO 4037-2. For matched reference radiation fields, recommended conversion coefficients are given in ISO 4037-3 only for specified distances between source and dosimeter, e.g., 1,0 m and 2,5 m.

ISO 4037-1:2019(E)

For other distances, the user has to decide if these conversion coefficients can be used. If both values are very similar, e.g., differ only by 2 % or less, then a linear interpolation may be used.

The second method is to produce “characterized reference fields”. Either this is done by determining the conversion coefficients using spectrometry, or the required value is measured directly using secondary standard dosimeters. This method applies to any radiation quality, for any measuring quantity and, if applicable, for any phantom and angle of radiation incidence. In addition, the requirements on the parameters specifying the reference radiations depend on the definition depth in the phantom, i.e., 0,07 mm, 3 mm and 10 mm, therefore, the requirements are different for the different depths. Thus, a given radiation field can be a “matched reference field” for the depth of 0,07 mm but not for the depth of 10 mm, for which it can then be a “characterized reference field”. The conversion coefficients can be determined for any distance, provided the air kerma rate is not below 1 µGy/h.

Both methods need charged particle equilibrium for the reference field. However, this is not always established in the workplace field for which the dosimeter is calibrated. This is especially true at photon energies without inherent charged particle equilibrium at the reference depth d , which depends on the actual combination of energy and reference depth d . Electrons of energies above 65 keV, 0,75 MeV and 2,1 MeV can just penetrate 0,07 mm, 3 mm and 10 mm of ICRU tissue, respectively, and the radiation qualities with photon energies above these values are considered as radiation qualities without inherent charged particle equilibrium for the quantities defined at these depths.

To determine the dose(rate) value and the associated overall uncertainty of it, a calibration of all measuring instruments used for the determination of the quantity value is needed which is traceable to national standards.

This document does not specify pulsed reference radiation fields.

Table 1 — List of X and gamma reference radiation, their mean energies, $\bar{E}(\Phi)$, for 1 m distance and their short names

Radiation quality	$\bar{E}(\Phi)$ keV	Radiation quality	$\bar{E}(\Phi)$ keV	Radiation quality	$\bar{E}(\Phi)$ keV	Radiation quality	$\bar{E}(\Phi)$ keV
L-10	9,0	N-10	8,5	W-30	22,9	H-10	8,0
L-20	17,3	N-15	12,4	W-40	29,8	H-20	13,1
L-30	26,7	N-20	16,3	W-60	44,8	H-30	19,7
L-35	30,4	N-25	20,3	W-80	56,5	H-40	25,4
L-55	47,8	N-30	24,6	W-110	79,1	H-60	38,0
L-70	60,6	N-40	33,3	W-150	104	H-80	48,8
L-100	86,8	N-60	47,9	W-200	138	H-100	57,3
L-125	109	N-80	65,2	W-250	172	H-150	78,0
L-170	149	N-100	83,3	W-300	205	H-200	99,3
L-210	185	N-120	100			H-250	122
L-240	211	N-150	118			H-280	145
		N-200	165			H-300	143
		N-250	207			H-350	167
		N-300	248			H-400	190
		N-350	288				
		N-400	328				

Table 1 (continued)

Radionuclides			High energy photon radiations		
Radiation quality	Radionuclide	$\bar{E}(\Phi)$ keV	Radiation quality	Reaction	$\bar{E}(\Phi)$, $\bar{E}[H^*(10)]_a$ MeV
S-Cs	^{137}Cs	662	R-C	$^{12}\text{C}(\text{p,p}'\gamma)^{12}\text{C}$	4,2; 4,4
S-Co	^{60}Co	1250	R-F	$^{19}\text{F}(\text{p},\alpha\gamma)^{16}\text{O}$	4,4; 6,5
NOTE In the informative Annexes A to C, further radiation qualities are given. These cover the mean photon energies from 8 keV up to 270 keV.					
^a Mean photon energy weighted by distribution of ambient dose equivalent, $H^*(10)$, with respect to photon energy E .					

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 2919, *Radiological protection — Sealed radioactive sources — General requirements and classification*

ISO 4037-2:2018, *Radiological protection — X and gamma reference radiation for calibrating dosimeters and doserate meters and for determining their response as a function of photon energy — Part 2: Dosimetry for radiological protection over the energy ranges 8 keV to 1,3 MeV and 4 MeV to 9 MeV*

ISO 4037-3, *Radiological protection — X and gamma reference radiation for calibrating dosimeters and doserate meters and for determining their response as a function of photon energy — Part 3: Calibration of area and personal dosimeters and the measurement of their response as a function of energy and angle of incidence*

ISO 29661, *Reference radiation fields for radiation protection — Definitions and fundamental concepts*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 29661 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

3.1

air kerma-to-dose-equivalent conversion coefficient

h_K

quotient of the dose equivalent, H , and the air kerma free-in-air, K_a , at a point in the photon radiation field

$$h_K = \frac{H}{K_a}$$

Note 1 to entry: The unit of the air kerma-to-dose-equivalent conversion coefficient is sievert per gray (Sv·Gy⁻¹).

Note 2 to entry: This definition differs from the one given by ISO 29661:2012, 3.2.4, as it uses the air kerma instead of the air collision kerma. See also 4.1.2.

ISO 4037-1:2019(E)

Note 3 to entry: The full specification of an air kerma-to-dose-equivalent conversion coefficient includes the specification of the type of dose equivalent, e.g. ambient, directional or personal. The conversion coefficient, h_K , depends on the energy and, for $H_p(10)$, $H_p(3)$, $H_p(0,07)$, $H'(3, \overline{\Gamma})$ and $H'(0,07, \overline{\Gamma})$, also on the directional distribution of the incident radiation. It is, therefore, useful to consider the conversion coefficient as a function, $h_K(E, \alpha)$, of the energy, E , of monoenergetic photons at several angles of incidence α .

Note 4 to entry: The conversion coefficients from the air kerma free-in-air, K_a , to $H'(0,07)$, to $H'(3)$, to $H^*(10)$, to $H_p(10)$, to $H_p(3)$ or to $H_p(0,07)$ for the radiation quality U and the angle of incidence α are indicated as $h'_K(0,07; U, \alpha)$, $h'_K(3; U, \alpha)$, $h^*_K(10; U)$, $h_{pK}(10; U, \alpha)$, $h_{pK}(3; U, \alpha)$, and $h_{pK}(0,07; U, \alpha)$, respectively.

3.2 characterized reference radiation field

reference radiation field whose properties are not sufficiently well-characterized so as to allow the use of recommended conversion coefficients but the mean energy of which is close enough to the nominal value to be used as a reference radiation field with the given designation

Note 1 to entry: Either this is done by determining the conversion coefficients using spectrometry, or the required value is measured directly using secondary standard dosimeters.

3.3 effective energy (of radiation comprised of X-rays with a range of energies)

E_{eff}
energy of the monoenergetic photons which have the same first HVL

3.4 generating potential

U_{gen}
potential difference between positive and negative output of the high voltage generator

3.5 half-value thickness half-value layer

HVL

thickness of the attenuating layer that reduces the quantity of interest of a unidirectional beam of infinitesimal width to half of its initial value

[SOURCE: ISO 80000-10:—, 10.53]

Note 1 to entry: For this document, the quantity of interest is the air kerma.

Note 2 to entry: In this definition, the contribution of all scattered radiation, other than any which might be present initially in the beam concerned, is deemed to be excluded.

3.6 homogeneity coefficient

h
ratio of the first *half-value layer* (3.5) to the second half-value layer (air kerma)

$$h = \frac{1^{\text{st}} \text{ HVL}}{2^{\text{nd}} \text{ HVL}}$$

3.7 matched reference radiation field

reference radiation field whose properties are sufficiently well-characterized so as to allow the use of the recommended conversion coefficients

3.8**mean photon energy****mean energy**

$$\bar{E}(\Phi)$$

ratio defined by the formula:

$$\bar{E}(\Phi) = \frac{\int_0^{E_{\max}} \Phi_E E \, dE}{\int_0^{E_{\max}} \Phi_E \, dE}$$

where Φ_E is the *spectral fluence* (3.14)

3.9**monitor**

instrument used to monitor the stability of the air kerma rate during irradiation or to compare values of air kerma after successive irradiations

3.10**primary radiation****primary beam**

radiation or beam emitted by the X-ray tube or the radionuclide or the target of the accelerator including scattered radiation inherently present in the beam, which cannot be removed from the beam by any means

3.11**pulse height spectrum**

distribution of number of pulses N with respect to charge Q generated in the detector, dN/dQ

3.12**relative tube potential deviation**

$$\Delta U_{\text{rel}}$$

ratio defined for a given nominal tube potential by the formula:

$$\Delta U_{\text{rel}} = \left| \frac{U_{\text{tube, meas}} - U_{\text{tube, nom}}}{U_{\text{tube, nom}}} \right|$$

where

$U_{\text{tube, meas}}$ is the measured value

$U_{\text{tube, nom}}$ is the nominal value of the tube potential

3.13**spectral air kerma**

distribution of air kerma K_a with respect to photon energy E

$$K_a(E) = \frac{dK_a}{dE}$$

3.14**spectral fluence**

distribution of fluence Φ with respect to photon energy E

$$\Phi_E = \frac{d\Phi}{dE}$$

ISO 4037-1:2019(E)

3.15

spectral resolution
resolution
(full width at half maximum)

R_E
ratio defined by the formula:

$$R_E = \frac{\Delta E}{E}$$

where ΔE is the width of the spectrum at half maximum

Note 1 to entry: In the case where fluorescence radiation is present in the spectrum, the spectrum width measured is based upon the continuum only.

3.16

tube potential

U_{tube}
potential difference between cathode and anode of the X-ray tube

3.17

unfolding

determination of the spectral fluence ([3.14](#)), Φ_E , from the (measured) *pulse height spectrum* ([3.11](#)), dN/dQ

3.18

value of peak-to-peak voltage
ripple

ratio defined for a fixed current value by the formula:

$$\frac{U_{\text{max}} - U_{\text{min}}}{U_{\text{max}}}$$

where

U_{max} is the maximum value

U_{min} the minimum value between which the voltage oscillates

3.19

X-ray tube

vacuum tube designed to produce X-rays by bombardment of the anode by a beam of electrons accelerated through a potential difference

3.20

X-ray tube shielding

fixed or mobile panel intended to reduce the contribution of scattered X-radiation to the primary beam

3.21

X-ray unit

assembly comprising a high-voltage supply, an *X-ray tube* ([3.19](#)) with its protective housing and high-voltage electrical connections