

---

**Radiološka zaščita - Referenčno sevanje z rentgenskimi in gama žarki za kalibracijo dozimetrov in merilnikov doze sevanja ter za ugotavljanje njihovega odzivanja kot funkcije fotonske energije - 1. del: Značilnosti sevanja in proizvodne metode (ISO 4037-1:2019)**

Radiological protection - X and gamma reference radiation for calibrating dosimeters and doserate meters and for determining their response as a function of photon energy - Part 1: Radiation characteristics and production methods (ISO 4037-1:2019)

**iTeh STANDARD PREVIEW**

Strahlenschutz - Röntgen- und Gamma-Referenzstrahlungsfelder zur Kalibrierung von Dosimetern und Dosisleistungsmessgeräten und zur Bestimmung ihres Ansprechvermögens als Funktion der Photonenenergie - Teil 1: Strahlungseigenschaften und Erzeugungsmethoden

<https://standards.iteh.ai/catalog/standards/sist/e65560ba-99cd-4cab-a471-d60b610e05f0/sist-en-iso-4037-1-2021>

Radioprotection - Rayonnements X et gamma de référence pour l'étalonnage des dosimètres et des débitmètres, et pour la détermination de leur réponse en fonction de l'énergie des photons - Partie 1: Caractéristiques des rayonnements et méthodes de production (ISO 4037-1:2019)

**Ta slovenski standard je istoveten z: EN ISO 4037-1:2021**

---

**ICS:**

13.280	Varstvo pred sevanjem	Radiation protection
17.240	Merjenje sevanja	Radiation measurements

**SIST EN ISO 4037-1:2021**

**en,fr,de**

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN ISO 4037-1:2021](https://standards.iteh.ai/catalog/standards/sist/e65560ba-99cd-4cab-a471-d60b610e05f0/sist-en-iso-4037-1-2021)

<https://standards.iteh.ai/catalog/standards/sist/e65560ba-99cd-4cab-a471-d60b610e05f0/sist-en-iso-4037-1-2021>

EUROPEAN STANDARD

EN ISO 4037-1

NORME EUROPÉENNE

EUROPÄISCHE NORM

February 2021

ICS 17.240

English Version

Radiological protection - X and gamma reference radiation  
for calibrating dosimeters and doserate meters and for  
determining their response as a function of photon energy  
- Part 1: Radiation characteristics and production methods  
(ISO 4037-1:2019)

Radioprotection - Rayonnements X et gamma de  
référence pour l'étalonnage des dosimètres et des  
débitmètres, et pour la détermination de leur réponse  
en fonction de l'énergie des photons - Partie 1:  
Caractéristiques des rayonnements et méthodes de  
production (ISO 4037-1:2019)

Strahlenschutz - Röntgen- und Gamma-  
Referenzstrahlungsfelder zur Kalibrierung von  
Dosimetern und Dosisleistungsmessgeräten und zur  
Bestimmung ihres Ansprechvermögens als Funktion  
der Photonenenergie - Teil 1: Strahlungseigenschaften  
und Erzeugungsmethoden (ISO 4037-1:2019)

**iTeh STANDARD PREVIEW**  
(standards.iteh.ai)

This European Standard was approved by CEN on 18 January 2021.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

<https://standards.iteh.ai/catalog/standards/sist/e65560ba-99cd-4cab-a471-d60b610e05f0/sist-en-iso-4037-1-2021>

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

<b>Contents</b>	<b>Page</b>
<b>European foreword.....</b>	<b>3</b>

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN ISO 4037-1:2021](https://standards.iteh.ai/catalog/standards/sist/e65560ba-99cd-4cab-a471-d60b610e05f0/sist-en-iso-4037-1-2021)  
<https://standards.iteh.ai/catalog/standards/sist/e65560ba-99cd-4cab-a471-d60b610e05f0/sist-en-iso-4037-1-2021>

## European foreword

The text of ISO 4037-1:2019 has been prepared by Technical Committee ISO/TC 85 "Nuclear energy, nuclear technologies, and radiological protection" of the International Organization for Standardization (ISO) and has been taken over as EN ISO 4037-1:2021 by Technical Committee CEN/TC 430 "Nuclear energy, nuclear technologies, and radiological protection" the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by August 2021, and conflicting national standards shall be withdrawn at the latest by August 2021.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

## iTeh STANDARD PREVIEW Endorsement notice (standards.iteh.ai)

The text of ISO 4037-1:2019 has been approved by CEN as EN ISO 4037-1:2021 without any modification.

[SIST EN ISO 4037-1:2021](https://standards.iteh.ai/catalog/standards/sist/e65560ba-99cd-4cab-a471-d60b610e05f0/sist-en-iso-4037-1-2021)

<https://standards.iteh.ai/catalog/standards/sist/e65560ba-99cd-4cab-a471-d60b610e05f0/sist-en-iso-4037-1-2021>

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN ISO 4037-1:2021](https://standards.iteh.ai/catalog/standards/sist/e65560ba-99cd-4cab-a471-d60b610e05f0/sist-en-iso-4037-1-2021)

<https://standards.iteh.ai/catalog/standards/sist/e65560ba-99cd-4cab-a471-d60b610e05f0/sist-en-iso-4037-1-2021>

INTERNATIONAL  
STANDARD

ISO  
4037-1

Second edition  
2019-01

---

---

**Radiological protection — X and  
gamma reference radiation for  
calibrating dosimeters and dose rate  
meters and for determining their  
response as a function of photon  
energy —**

iTeh STANDARD PREVIEW

(standards.iteh.ai)

**Part 1:  
Radiation characteristics and  
production methods**

[https://standards.iteh.ai/catalog/standards/sist/e65560ba-99cd-4cab-a471-](https://standards.iteh.ai/catalog/standards/sist/e65560ba-99cd-4cab-a471-d601-4037-1)

*601-4037-1*  
**Radioprotection — Rayonnements X et gamma de référence  
pour l'étalonnage des dosimètres et des débitmètres, et pour la  
détermination de leur réponse en fonction de l'énergie des photons —**

*Partie 1: Caractéristiques des rayonnements et méthodes de  
production*



Reference number  
ISO 4037-1:2019(E)

© ISO 2019

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 4037-1:2021](https://standards.iteh.ai/catalog/standards/sist/e65560ba-99cd-4cab-a471-d60b610e05f0/sist-en-iso-4037-1-2021)

<https://standards.iteh.ai/catalog/standards/sist/e65560ba-99cd-4cab-a471-d60b610e05f0/sist-en-iso-4037-1-2021>



### **COPYRIGHT PROTECTED DOCUMENT**

© ISO 2019

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Fax: +41 22 749 09 47  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland



# Contents

Page

<b>Foreword</b> .....	<b>v</b>
<b>Introduction</b> .....	<b>vi</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>3</b>
<b>3 Terms and definitions</b> .....	<b>3</b>
<b>4 Continuous reference filtered X radiation</b> .....	<b>7</b>
4.1 General.....	7
4.1.1 Realisation of reference radiation fields.....	7
4.1.2 Basis of conversion coefficients.....	7
4.1.3 Radiation quality.....	8
4.1.4 Choice of reference radiation.....	8
4.2 Conditions and methods for producing reference X radiation.....	13
4.2.1 Characteristics of the high voltage generator.....	13
4.2.2 Tube potential and protective resistor.....	14
4.2.3 Filtration.....	15
4.2.4 Limitations concerning matched fields.....	19
4.2.5 X radiation shutter.....	20
4.2.6 Beam aperture.....	20
4.3 Field uniformity and scattered radiation.....	20
4.3.1 Field diameter.....	20
4.3.2 Field uniformity.....	20
4.3.3 Scattered radiation.....	20
4.4 Summary of the requirements for reference X radiation fields.....	21
4.5 Validation of reference X radiation.....	21
4.5.1 General.....	21
4.5.2 Criteria for validation by HVL determination.....	22
4.5.3 Apparatus for HVL measurement.....	23
4.5.4 HVL measurement procedure.....	24
4.5.5 Criteria for validation by dosimetry.....	24
4.5.6 Criteria for validation by spectrometry.....	24
<b>5 Gamma radiation emitted by radionuclides</b> .....	<b>25</b>
5.1 General.....	25
5.2 Radionuclides used for the production of gamma radiation.....	25
5.3 Specification of radiation sources.....	25
5.3.1 Sources.....	25
5.3.2 Encapsulation.....	26
5.4 Irradiation facility and influence of scattered radiation.....	26
5.4.1 General requirements.....	26
5.4.2 Collimated geometry installation.....	26
5.4.3 Variation of air kerma rate by means of lead attenuators.....	27
5.5 Checking installation conformity.....	27
<b>6 Photon radiation with energies between 4 MeV and 9 MeV</b> .....	<b>28</b>
6.1 General.....	28
6.2 Production of reference radiation.....	28
6.2.1 General.....	28
6.2.2 Photon reference radiation from de-excitation of $^{16}\text{O}$ in the $^{19}\text{F}(p, \alpha\gamma)^{16}\text{O}$ reaction.....	28
6.2.3 Photon reference radiation from de-excitation of $^{12}\text{C}$ .....	30
6.3 Beam diameter and uniformity of radiation field.....	31
6.4 Contamination of photon reference radiation.....	31
6.4.1 General.....	31

## ISO 4037-1:2019(E)

6.4.2	Contamination of reference radiation common to all methods of production of reference radiation.....	32
6.4.3	Additional contamination of accelerator produced reference radiation from de-excitation of $^{16}\text{O}$ .....	32
<b>Annex A (informative) Fluorescence X radiation with not enough information for matched or characterized fields.....</b>		<b>33</b>
<b>Annex B (informative) Gamma radiation emitted by <math>^{241}\text{Am}</math> radionuclide with not enough information for matched or characterized fields.....</b>		<b>40</b>
<b>Annex C (informative) Continuous filtered X radiation based on the quality index.....</b>		<b>42</b>
<b>Bibliography.....</b>		<b>45</b>

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 4037-1:2021](https://standards.iteh.ai/catalog/standards/sist/e65560ba-99cd-4cab-a471-d60b610e05f0/sist-en-iso-4037-1-2021)

<https://standards.iteh.ai/catalog/standards/sist/e65560ba-99cd-4cab-a471-d60b610e05f0/sist-en-iso-4037-1-2021>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 85, *Nuclear energy, nuclear technologies and radiological protection*, Subcommittee SC 2, *Radiological protection*.

This second edition cancels and replaces the first edition (ISO 4037-1:1996), which has been technically revised. The main changes are:

- introduction of two types of reference fields, matched reference fields and characterized reference fields;
- introduction of validation for matched reference fields;
- introduction of limits for the allowed deviation of parameters like high voltage, filter purity and filter thickness from their nominal values. These limits now depend on the definition depth of the phantom related quantity. This is done to achieve an overall uncertainty ( $k = 2$ ) of about 6 % to 10 % for the phantom related operational quantities.

A list of all the parts in the ISO 4037 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

**ISO 4037-1:2019(E)****Introduction**

This maintenance release of this document incorporates the improvements to high voltage generators from 1996 to 2017 (e.g., the use of high frequency switching supplies providing nearly constant potential), and the spectral measurements at irradiation facilities equipped with such generators (e.g., the catalogue of X-ray spectra by Ankerhold<sup>[4]</sup>). It also incorporates all published information with the aim to adjust the requirements for the technical parameters of the reference fields to the targeted overall uncertainty of about 6 % to 10 % for the phantom related operational quantities of the International Commission on Radiation Units and Measurements (ICRU)<sup>[5]</sup>. It does not change the general concept of the existing ISO 4037.

ISO 4037 focusing on photon reference radiation fields is divided into four parts. ISO 4037-1 gives the methods of production and characterization of reference radiation fields in terms of the quantities spectral photon fluence and air kerma free-in-air. ISO 4037-2 describes the dosimetry of the reference radiation qualities in terms of air kerma and in terms of the phantom related operational quantities of the International Commission on Radiation Units and Measurements (ICRU)<sup>[5]</sup>. ISO 4037-3 describes the methods for calibrating and determining the response of dosimeters and doserate meters in terms of the phantom related operational quantities of the ICRU<sup>[5]</sup>. ISO 4037-4 gives special considerations and additional requirements for calibration of area and personal dosimeters in low energy X reference radiation fields, which are reference fields with generating potential lower or equal to 30 kV.

The general procedures described in ISO 29661 are used as far as possible in this document. Also, the symbols used are in line with ISO 29661.

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN ISO 4037-1:2021](https://standards.iteh.ai/catalog/standards/sist/e65560ba-99cd-4cab-a471-d60b610e05f0/sist-en-iso-4037-1-2021)

<https://standards.iteh.ai/catalog/standards/sist/e65560ba-99cd-4cab-a471-d60b610e05f0/sist-en-iso-4037-1-2021>

# Radiological protection — X and gamma reference radiation for calibrating dosimeters and doserate meters and for determining their response as a function of photon energy —

## Part 1: Radiation characteristics and production methods

### 1 Scope

This document specifies the characteristics and production methods of X and gamma reference radiation for calibrating protection-level dosimeters and doserate meters with respect to the phantom related operational quantities of the International Commission on Radiation Units and Measurements (ICRU)<sup>[5]</sup>. The lowest air kerma rate for which this standard is applicable is 1  $\mu\text{Gy h}^{-1}$ . Below this air kerma rate the (natural) background radiation needs special consideration and this is not included in this document.

For the radiation qualities specified in [Clauses 4 to 6](#), sufficient published information is available to specify the requirements for all relevant parameters of the matched or characterized reference fields in order to achieve the targeted overall uncertainty ( $k = 2$ ) of about 6 % to 10 % for the phantom related operational quantities. The X ray radiation fields described in the informative [Annexes A to C](#) are not designated as reference X-radiation fields.

NOTE The first edition of ISO 4037-1, issued in 1996, included some additional radiation qualities for which such published information is not available. These are fluorescent radiations, the gamma radiation of the radionuclide  $^{241}\text{Am}$ , S-Am, and the high energy photon radiations R-Ti and R-Ni, which have been removed from the main part of this document. The most widely used radiations, the fluorescent radiations and the gamma radiation of the radionuclide  $^{241}\text{Am}$ , S-Am, are included nearly unchanged in the informative [Annexes A and B](#). The informative [Annex C](#) gives additional X radiation fields, which are specified by the quality index.

The methods for producing a group of reference radiations for a particular photon-energy range are described in [Clauses 4 to 6](#), which define the characteristics of these radiations. The three groups of reference radiation are:

- in the energy range from about 8 keV to 330 keV, continuous filtered X radiation;
- in the energy range 600 keV to 1,3 MeV, gamma radiation emitted by radionuclides;
- in the energy range 4 MeV to 9 MeV, photon radiation produced by accelerators.

The reference radiation field most suitable for the intended application can be selected from [Table 1](#), which gives an overview of all reference radiation qualities specified in [Clauses 4 to 6](#). It does not include the radiations specified in the [Annexes A, B and C](#).

The requirements and methods given in [Clauses 4 to 6](#) are targeted at an overall uncertainty ( $k = 2$ ) of the dose(rate) value of about 6 % to 10 % for the phantom related operational quantities in the reference fields. To achieve this, two production methods are proposed:

The first one is to produce “*matched reference fields*”, whose properties are sufficiently well-characterized so as to allow the use of the conversion coefficients recommended in ISO 4037-3. The existence of only a small difference in the spectral distribution of the “*matched reference field*” compared to the nominal reference field is validated by procedures, which are given and described in detail in ISO 4037-2. For matched reference radiation fields, recommended conversion coefficients are given in ISO 4037-3 only for specified distances between source and dosimeter, e.g., 1,0 m and 2,5 m.

## ISO 4037-1:2019(E)

For other distances, the user has to decide if these conversion coefficients can be used. If both values are very similar, e.g., differ only by 2 % or less, then a linear interpolation may be used.

The second method is to produce “characterized reference fields”. Either this is done by determining the conversion coefficients using spectrometry, or the required value is measured directly using secondary standard dosimeters. This method applies to any radiation quality, for any measuring quantity and, if applicable, for any phantom and angle of radiation incidence. In addition, the requirements on the parameters specifying the reference radiations depend on the definition depth in the phantom, i.e., 0,07 mm, 3 mm and 10 mm, therefore, the requirements are different for the different depths. Thus, a given radiation field can be a “matched reference field” for the depth of 0,07 mm but not for the depth of 10 mm, for which it can then be a “characterized reference field”. The conversion coefficients can be determined for any distance, provided the air kerma rate is not below 1  $\mu\text{Gy/h}$ .

Both methods need charged particle equilibrium for the reference field. However, this is not always established in the workplace field for which the dosimeter is calibrated. This is especially true at photon energies without inherent charged particle equilibrium at the reference depth  $d$ , which depends on the actual combination of energy and reference depth  $d$ . Electrons of energies above 65 keV, 0,75 MeV and 2,1 MeV can just penetrate 0,07 mm, 3 mm and 10 mm of ICRU tissue, respectively, and the radiation qualities with photon energies above these values are considered as radiation qualities without inherent charged particle equilibrium for the quantities defined at these depths.

To determine the dose(rate) value and the associated overall uncertainty of it, a calibration of all measuring instruments used for the determination of the quantity value is needed which is traceable to national standards.

This document does not specify pulsed reference radiation fields.

STANDARD PREVIEW  
(standards.iteh.ai)

**Table 1 — List of X and gamma reference radiation, their mean energies,  $\bar{E}(\Phi)$ , for 1 m distance and their short names**

Radiation quality	$\bar{E}(\Phi)$ keV	Radiation quality	$\bar{E}(\Phi)$ keV	Radiation quality	$\bar{E}(\Phi)$ keV	Radiation quality	$\bar{E}(\Phi)$ keV
L-10	9,0	N-10	8,5	W-30	22,9	H-10	8,0
L-20	17,3	N-15	12,4	W-40	29,8	H-20	13,1
L-30	26,7	N-20	16,3	W-60	44,8	H-30	19,7
L-35	30,4	N-25	20,3	W-80	56,5	H-40	25,4
L-55	47,8	N-30	24,6	W-110	79,1	H-60	38,0
L-70	60,6	N-40	33,3	W-150	104	H-80	48,8
L-100	86,8	N-60	47,9	W-200	138	H-100	57,3
L-125	109	N-80	65,2	W-250	172	H-150	78,0
L-170	149	N-100	83,3	W-300	205	H-200	99,3
L-210	185	N-120	100			H-250	122
L-240	211	N-150	118			H-280	145
		N-200	165			H-300	143
		N-250	207			H-350	167
		N-300	248			H-400	190
		N-350	288				
		N-400	328				

Table 1 (continued)

Radionuclides			High energy photon radiations		
Radiation quality	Radionuclide	$\bar{E}(\Phi)$ keV	Radiation quality	Reaction	$\bar{E}(\Phi)$ ; $\bar{E}[H^*(10)]_a$ MeV
S-Cs	$^{137}\text{Cs}$	662	R-C	$^{12}\text{C}(\text{p,p}'\gamma)^{12}\text{C}$	4,2; 4,4
S-Co	$^{60}\text{Co}$	1250	R-F	$^{19}\text{F}(\text{p},\alpha\gamma)^{16}\text{O}$	4,4; 6,5

NOTE In the informative Annexes A to C, further radiation qualities are given. These cover the mean photon energies from 8 keV up to 270 keV.

<sup>a</sup> Mean photon energy weighted by distribution of ambient dose equivalent,  $H^*(10)$ , with respect to photon energy  $E$ .

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 2919, *Radiological protection — Sealed radioactive sources — General requirements and classification*

ISO 4037-2:2018, *Radiological protection — X and gamma reference radiation for calibrating dosimeters and doserate meters and for determining their response as a function of photon energy — Part 2: Dosimetry for radiological protection over the energy ranges 8 keV to 1,3 MeV and 4 MeV to 9 MeV*

ISO 4037-3, *Radiological protection — X and gamma reference radiation for calibrating dosimeters and doserate meters and for determining their response as a function of photon energy — Part 3: Calibration of area and personal dosimeters and the measurement of their response as a function of energy and angle of incidence*

<https://standards.iteh.ai/catalog/standards/sist/e65560ba-99cd-4cab-a471-466b376c9394/sist-en-iso-4037-1-2021>

ISO 29661, *Reference radiation fields for radiation protection — Definitions and fundamental concepts*

## 3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 29661 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

### 3.1

#### air kerma-to-dose-equivalent conversion coefficient

$h_K$

quotient of the dose equivalent,  $H$ , and the air kerma free-in-air,  $K_a$ , at a point in the photon radiation field

$$h_K = \frac{H}{K_a}$$

Note 1 to entry: The unit of the air kerma-to-dose-equivalent conversion coefficient is sievert per gray (Sv·Gy<sup>-1</sup>).

Note 2 to entry: This definition differs from the one given by ISO 29661:2012, 3.2.4, as it uses the air kerma instead of the air collision kerma. See also 4.1.2.