

SLOVENSKI STANDARD oSIST prEN ISO 28399:2020

01-december-2020

Zobozdravstvo - Proizvodi za zunanje beljenje zob (ISO/FDIS 28399:2020)

Dentistry - External tooth bleaching products (ISO/FDIS 28399:2020)

Zahnheilkunde - Äußere Zahnbleichmittel (ISO/FDIS 28399:2020)

Médecine bucco-dentaire - Produits d'éclaircissement dentaire par voie externe (ISO/FDIS 28399:2020)

Ta slovenski standard je istoveten z: prEN ISO 28399

ICS:		
11.060.01	Zobozdravstvo na splošno	Dentistry in general
71.100.70	Kozmetika. Toaletni pripomočki	Cosmetics. Toiletries

oSIST prEN ISO 28399:2020

en,fr,de

oSIST prEN ISO 28399:2020

iTeh Standards (https://standards.iteh.ai) Document Preview

SIST EN ISO 28399:2021

https://standards.iteh.ai/catalog/standards/sist/1a9cacc0-bf0b-424d-8fdd-31b54ede2aa1/sist-en-iso-28399-2021

FINAL DRAFT

INTERNATIONAL STANDARD

ISO/FDIS 28399

ISO/TC 106/SC 7

Secretariat: JISC

Voting begins on: **2020-10-28**

Voting terminates on: 2020-01-20

Dentistry — External tooth bleaching products

Médecine bucco-dentaire — Produits d'éclaircissement dentaire par voie externe

iTeh Standards (https://standards.iteh.ai) Document Preview

SIST EN ISO 28399:2021

https://standards.iteh.ai/catalog/standards/sist/1a9cacc0-bf0b-424d-8fdd-31b54ede2aa1/sist-en-iso-28399-2021

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNO-LOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STAN-DARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

ISO/CEN PARALLEL PROCESSING



Reference number ISO/FDIS 28399:2020(E)

iTeh Standards (https://standards.iteh.ai) Document Preview

SIST EN ISO 28399:2021

https://standards.iteh.ai/catalog/standards/sist/1a9cacc0-bf0b-424d-8fdd-31b54ede2aa1/sist-en-iso-28399-2021



COPYRIGHT PROTECTED DOCUMENT

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office CP 401 • Ch. de Blandonnet 8 CH-1214 Vernier, Geneva Phone: +41 22 749 01 11 Email: copyright@iso.org Website: www.iso.org

Published in Switzerland

Contents

Page

Forew	vord	iv
Intro	duction	v
1	Scope	1
2	Normative references	1
3	Terms and definitions	2
4	Classification4.1General4.2Products for professional application4.3Products for consumer application	2
5	Requirements5.1Concentration of active ingredients for bleaching5.2Surface microhardness5.3Surface erosion	
6	Measurements and test methods6.1Preparation of tooth specimens6.2Preparation and application of tooth bleaching product6.3Surface microhardness	
7	Packaging, marking and information to be supplied by the manufacturer7.1General7.2Packaging7.3Marking and instructions for use	
Annez	x A (informative) Test method for the measurement of hydrogen peroxide concentra	tion 5
Annex	x B (informative) Light microscopy method for measuring erosion of enamel and dentine caused by external tooth bleaching products	6
Annex Biblio	x C (informative) Test method for laboratory assessment of tooth bleaching efficacy. Ography	

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 7, *Oral care products*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 55, *Dentistry*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

<u>SIST EN ISO 28399:2021</u>

This third edition cancels and replaces the second edition (ISO 28399:2020), which has been editorially revised. The changes compared to the previous edition are as follows:

- "TEM" has been removed from "TEM grid" throughout this third edition;
- <u>B.10.2.6</u> EXAMPLE has been rewritten;
- "<u>B.9</u> A)" has been replaced with "<u>B.9</u>";
- "See <u>Figure B.9</u> B" has been removed;
- "In Figure B.9 C, the value is 51,916 3 (i.e. 51,9 μm)" has been removed and Figure B.9, k), "number of " has been replaced with "width in";
- In <u>C.3.3.3</u>, "delta E*ab" has been added in the space between two commas.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at <u>www.iso.org/members.html</u>.

Introduction

External tooth bleaching products are used in dentistry for changing the colour of natural teeth towards a lighter or whiter shade. They are applied in the oral cavity directly on the outer surfaces of teeth.

Specific qualitative and quantitative requirements for freedom from biological hazard are not included in this document. It is recommended, that reference should be made to ISO 10993-1 and ISO 7405 when assessing possible biological or toxicological hazards.

iTeh Standards (https://standards.iteh.ai) Document Preview

SIST EN ISO 28399:2021

https://standards.iteh.ai/catalog/standards/sist/1a9cacc0-bf0b-424d-8fdd-31b54ede2aa1/sist-en-iso-28399-2021

oSIST prEN ISO 28399:2020

iTeh Standards (https://standards.iteh.ai) Document Preview

SIST EN ISO 28399:2021

https://standards.iteh.ai/catalog/standards/sist/1a9cacc0-bf0b-424d-8fdd-31b54ede2aa1/sist-en-iso-28399-2021

FINAL DRAFT INTERNATIONAL STANDARD

Dentistry — External tooth bleaching products

1 Scope

This document specifies requirements and test methods for external tooth bleaching products. These products are intended for use in the oral cavity, either by professional application (in-office tooth bleaching products) or consumer application (professional or non-professional home use of tooth bleaching products), or both. It also specifies requirements for their packaging, labelling and manufacturer's instructions for use.

This document is not applicable to tooth bleaching products:

— specified in ISO 11609;

- intended to change colour perception of natural teeth by mechanical methods (e.g. stain removal) or using restorative approaches, such as veneers or crowns;
- auxiliary or supplementary materials (e.g. tray materials) and instruments or devices (e.g. lights) that are used in conjunction with the bleaching products.

This document does not specify biological safety aspects of tooth bleaching products.

NOTE Maximum concentration of a bleaching agent for professional or non-professional use is subject to each country's regulatory body.

2 Normative references_{ocument} Preview

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 1042, Laboratory glassware — One-mark volumetric flasks

ISO 1942, Dentistry — Vocabulary

ISO 3696, Water for analytical laboratory use — Specification and test methods

ISO 6344-1, Coated abrasives — Grain size analysis — Part 1: Grain size distribution test

ISO 7405, Dentistry — Evaluation of biocompatibility of medical devices used in dentistry

ISO 8601-1, Date and time — Representations for information interchange — Part 1: Basic rules

ISO 8601-2, Date and time — Representations for information interchange — Part 2: Extensions

ISO 10993-1, Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process

ISO/CIE 11664-1, Colorimetry — Part 1: CIE standard colorimetric observers

ISO 11664-2, Colorimetry — Part 2: CIE standard illuminants

ANSI/ADA Specification No 41: Recommended Standard Practices for Biological Evaluation of Dental Materials

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 1942 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at https://www.iso.org/obp
- IEC Electropedia: available at http://www.electropedia.org/

3.1

bleaching

<natural teeth> removal of intrinsic or acquired discolorations of natural teeth, or changing their colour towards a lighter or whiter shade using chemicals, sometimes in combination with the application of auxiliary means

Note 1 to entry: Changing the colour of natural teeth towards a lighter or whiter shade is not limited to a discoloration.

Note 2 to entry: Auxiliary means for bleaching other than the application of external energy are also conceivable.

Note 3 to entry: [SOURCE: ISO 1942:2009, 2.28, modified — Note 1 to entry and Note 2 to entry have been added.]

3.2

professional home use

cyroduct> use prescribed by a professional and for use at home under the repeated supervision of the dentist

3.3

erosion

<tooth surface> progressive loss of calcified tissue by chemical processes that do not involve bacterial action

Note 1 to entry: [SOURCE: ISO 1942:2009, 2.292, modified — terms have been modified and a domain has been added.] <u>SIST EN ISO 28399:2021</u>

nttps://standards.iteh.ai/catalog/standards/sist/1a9cacc0-bf0b-424d-8fdd-31b54ede2aa1/sist-en-iso-28399-2021

4 Classification

4.1 General

External tooth bleaching products can be classified for either:

- a) professional application; or
- b) consumer application.

NOTE External tooth bleaching products can be used alone or in conjunction with auxiliary means of application.

4.2 Products for professional application

These products are tooth bleaching products that are intended by the manufacturer to be applied only by dental professionals (in office tooth bleaching products).

4.3 **Products for consumer application**

These products are tooth bleaching products that are intended by the manufacturer to be applied by the consumer (for professional home use or for non-professional home use).

NOTE Such external bleaching products can be prescribed by a dental professional or directly available to consumers.