

SLOVENSKI STANDARD SIST EN 12177:2023

01-januar-2023

Nadomešča: SIST EN 12177:1999 SIST EN 12177:1999/AC:2001

Tekoči naftni proizvodi - Neosvinčeni motorni bencini - Določevanje benzena s plinsko kromatografijo

Liquid petroleum products - Unleaded petrol - Determination of benzene content by gas chromatography

Flüssige Mineralölerzeugnisse - Unverbleite Ottokraftstoffe - Bestimmung des Benzolgehaltes mittels Gaschromatographie

https://standards.iteh.ai/catalog/standards/sist/cd894475-52e7-4463-b34d-Produits pétroliers liquides - Essence sans plomb - Détermination de la teneur en benzène par chromatographie en phase gazeuse

Ta slovenski standard je istoveten z: EN 12177:2022

<u>ICS:</u>

75.160.20 Tekoča goriva

Liquid fuels

SIST EN 12177:2023

en,fr,de



iTeh STANDARD PREVIEW (standards.iteh.ai)

<u>SIST EN 12177:2023</u> https://standards.iteh.ai/catalog/standards/sist/cd894475-52e7-4463-b34d-560c6d601be4/sist-en-12177-2023

SIST EN 12177:2023

EUROPEAN STANDARD NORME EUROPÉENNE EUROPÄISCHE NORM

EN 12177

November 2022

ICS 75.160.20

Supersedes EN 12177:1998

English Version

Liquid petroleum products - Unleaded petrol -Determination of benzene content by gas chromatography

Produits pétroliers liquides - Essence sans plomb -Détermination de la teneur en benzène par chromatographie en phase gazeuse Flüssige Mineralölerzeugnisse - Unverbleiter Ottokraftstoff - Bestimmung des Benzolgehaltes mittels Gaschromatographie

This European Standard was approved by CEN on 16 October 2022.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.

https://standards.iteh.ai/catalog/standards/sist/cd8944/5-52e/-4463-b34d-560c6d601be4/sist-en-12177-2023



EUROPEAN COMMITTEE FOR STANDARDIZATION COMITÉ EUROPÉEN DE NORMALISATION EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

SIST EN 12177:2023

EN 12177:2022 (E)

Contents

Europ	European foreword		
1	Scope	4	
2	Normative references	4	
3	Terms and definitions	4	
4	Principle	4	
5	Reagents and materials	5	
6 6.1	Apparatus Gas chromatographic assembly	5 5	
7	Sampling	6	
8 8.1 8.1.1 8.1.2 9.2	Procedure Setup of the apparatus General Carrier gas	7 7 7 7	
8.2 8.3	Determination of density	7 7	
8.4	Preparation of test sample 12002005 Iten 21	7	
8.5 8.6	Introduction of test portion Examination of the chromatogram	8 8	
9 9.1	Calculation of mass of benzene in the test sample	8 8	
9.2	Calculation of volume of benzene in the test sample	8	
9.5 9.4	Calculation of volume of test sample	0 9	
10	Expression of results	9	
11	Precision	9	
11.1	General	9	
11.2 11 3	Repeatability, r	9 9	
12	Test report	0	
Annex	Annex A (informative) Guidance on the column switching technique		
A.1	Introduction	1	
A.2	General apparatus parameters for the flow switching technique1	2	
A.3	Gas chromatogram1	2	
Bibliog	Bibliography		

European foreword

This document (EN 12177:2022) has been prepared by Technical Committee CEN/TC 19 "Gaseous and liquid fuels, lubricants and related products of petroleum, synthetic and biological origin", the secretariat of which is held by NEN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by May 2023, and conflicting national standards shall be withdrawn at the latest by May 2023.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 12177:1998.

At the time of the development of the test method, the ethanol content of gasoline was limited to 5 % (*V*/*V*) which was reflected in the scope of the test method. Since several years, fuels containing up to 10 % of ethanol (3,7 % (m/m)) are common in the European market. This revision is meant to extend the scope of the test methods to those fuels, thus reflecting market needs.

Any feedback and questions on this document should be directed to the users' national standards body. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

560c6d601be4/sist-en-12177-2023

1 Scope

This document specifies a column switching gas chromatographic method for the quantitative determination of benzene content in the range 0,05 % (V/V) to 6 % (V/V) in unleaded petrol having a final boiling point not greater than 220 °C.

The method described in this document is suitable for determining benzene in petrol, including petrol containing oxygenates up to E10 (up to 3,7 % (m/m) oxygen content), in line with the relevant EC Directives [1].

NOTE For the purposes of this document, the terms "% (V/V)" and "% (m/m)" are used to represent respectively the volume fraction and the mass fraction.

WARNING — Use of this document might involve hazardous materials, operations and equipment. This document does not purport to address all of the safety problems associated with its use. It is the responsibility of the user of this document to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN ISO 3170, Petroleum liquids — Manual sampling (ISO 3170)

EN ISO 3171, Petroleum liquids — Automatic pipeline sampling (ISO 3171)

EN ISO 3675, Crude petroleum and liquid petroleum products — Laboratory determination of density — Hydrometer method (ISO 3675)

EN ISO 3838, Crude petroleum and liquid or solid petroleum products — Determination of density or relative density — Capillary-stoppered pyknometer and graduated bicapillary pyknometer methods (ISO 3838)

EN ISO 12185, Crude petroleum and petroleum products — Determination of density — Oscillating U-tube method (ISO 12185)

3 Terms and definitions

No terms and definitions are listed in this document.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at https://www.electropedia.org/
- ISO Online browsing platform: available at <u>https://www.iso.org/obp/</u>

4 Principle

The benzene-containing fraction is isolated from the sample using a capillary column and, in a second capillary column, the benzene is separated and detected using a flame ionization detector.

NOTE 1 Some oxygenates are known to interfere with the determination of benzene using a single column gas chromatographic method.

NOTE 2 Guidance on the column switching technique is given in Annex A.

5 Reagents and materials

Use only chemicals and reagents of recognized analytical grade.

5.1 Gases

5.1.1 Carrier gas, helium, nitrogen or hydrogen, of at least 99,999 % (*V*/*V*) purity. Any oxygen present is removed by a chemical resin filter.

5.1.2 Hydrogen, grade suitable for flame ionization detectors.

5.1.3 Compressed air regulated for flame ionization detectors.

WARNING - Hydrogen is explosive in mixtures with air at concentrations of hydrogen ranging approximately from 4 % (V/V) to 75 % (V/V). All joints and lines carrying hydrogen shall be made gas tight to prevent leakage of hydrogen into a confined space.

5.2 Reagents for the preparation of calibration samples

NOTE Calibration samples will normally be combinations of benzene, solvent and an internal standard.

5.2.1 Benzene, not less than 99,0 % (*m*/*m*) pure.

WARNING - Benzene is toxic and carcinogenic.

5.2.2 Solvent, not containing benzene or internal standard, e.g. heptane.

5.3 Internal standard, reagent for which it has been established that it is not present in the sample.

NOTE Isobutyl methyl ketone is preferred as the internal standard.

https://standards.iteh.ai/catalog/standards/sist/cd894475-52e7-4463-b34d-

6 Apparatus 560c6d601be4/sist-en-12177-2023

Usual laboratory apparatus and glassware, together with the following:

6.1 Gas chromatographic assembly

6.1.1 Gas chromatograph, provided with a means for column switching, equipped with a programmable oven temperature controller, or controllers in the case of dual oven gas chromatographs, a split injector and two flame ionization detectors (FID). See Annex A for typical operating conditions.

It is recommended that a system constructed entirely of glass from the sample injection port up to the detector system is used since petrol may contain oxygenates which can give rise to corrosion and changes of retention times in systems constructed partly using metals.

6.1.2 Two capillary columns, of appropriate dimensions, each internally coated with a substance of different polarity, such that the resolution between the benzene peak and the matrix is at least 1 after elution from the second column.

The resolution, *R*, between peaks A and B (see Figure 1) is calculated using Formula 1:

$$R = 1.18 \frac{t'_{\rm B} - t'_{\rm A}}{w_{\rm A} + w_{\rm B}}$$
(1)

where

EN 12177:2022 (E)

- $t'_{\rm A}$ is the reduced retention time of component A;
- $t'_{\rm B}$ is the reduced retention time of component B;
- w_A is the peak width at half-height of component A;
- w_B is the peak width at half-height of component B.



NOTE t_N is the hold-up time zero of the column, i.e. the time taken for an inert component, such as methane, to travel through the column without chromatography taking place.

6.1.3 Device for the control of the flow of carrier gas, the chromatograph shall be able to deliver a constant carrier gas flow over the whole range of the analysis

6.2 Test sample container, normally with a capacity of between 10 ml and 100 ml, fitted with a self-sealing rubber septum coated with polytetrafluoroethylene (PTFE).

6.3 Balance, capable of weighing to the nearest 0,1 mg.

7 Sampling

Unless otherwise specified in the commodity specification, samples shall be taken as described in EN ISO 3170 or EN ISO 3171 and/or in accordance with the requirements of national standards or regulations for the sampling of petrol.

8 Procedure

8.1 Setup of the apparatus

8.1.1 General

Prepare the equipment and set the test conditions in accordance with the manufacturer's instructions.

8.1.2 Carrier gas

Adjust the pressure and flow rate of the carrier gas (5.1.1) to levels such that the resolution is in accordance with 6.1.2.

8.2 Calibration

Prepare the calibration sample by combining known masses of benzene (5.2.1) and internal standard (5.3) in a suitable solvent (5.2.2).

Inject a suitable quantity of the prepared calibration sample into the gas chromatograph such that the capacity of the columns and other components is not exceeded and the linearity of the detector is not impaired.

Determine and record the retention times for benzene and for the internal standard. Calculate the calibration factor, *f*, for benzene using the following Formula (2):

$$f = \frac{m_1 \times A_2}{A_1 \times m_2} \text{Teh STANDARD PREVIEW}$$
(2)

where

- m_1 is the mass of benzene in the calibration sample, in g;
- A_2 is the peak area of the internal standard; simple $\frac{1}{2}$
- 560c6d601be4/sist-en-12177-20
- *A*¹ is the peak area of benzene;
- m_2 is the mass of the internal standard in the calibration sample, in g.

8.3 Determination of density

Determine the density at 15 °C, ρ_s , of the sample in accordance with either EN ISO 3675, EN ISO 3838 or EN ISO 12185 and record the result to the nearest 0,1 kg/m³.

8.4 Preparation of test sample

Cool the sample to between 5 °C and 10 °C. Weigh using a balance (6.3), to the nearest 0,1 mg, the test sample container (6.2) and its rubber septum without sealing it.

Transfer a quantity of the internal standard (5.3) to the test sample container and weigh, to the nearest 0,1 mg, the test sample container with contents and septum, without sealing the sample container. The mass, m_{st} , in g of the internal standard shall amount to between 2 % (m/m) and 5 % (m/m) of the sample, m_{sv} but shall not be less than 0,050 g.

Transfer a quantity, normally between 5 ml and 100 ml, of the cooled sample to the test sample container and seal immediately with the septum. Weigh, to the nearest 0,1 mg, the test sample container and contents. Record the mass, m_s in g of the test sample taken, to the nearest 0,1 mg.

EN 12177:2022 (E)

Record the amount of internal standard in the prepared test sample in % (*m/m*). Homogenize the content of the test sample container by shaking vigorously.

8.5 Introduction of test portion

Inject a suitable quantity of the test sample (8.4) into the gas chromatograph. Ensure that the test portion size does not exceed the capacity of the columns and other components of the gas chromatograph and that the linearity of the detector is not impaired.

8.6 Examination of the chromatogram

Examine the chromatogram and identify the benzene and the internal standard by means of their retention times (see 8.2).

9 Calculation

9.1 Calculation of mass of benzene in the test sample

Calculate the mass, m₃, in g, of benzene in the test sample using Formula (3):

$$m_3 = \frac{f \times A_3 \times m_4}{A_4} \tag{3}$$

where

f is the calibration factor for benzene; *A*₃ is the peak area of benzene; *m*₄ is the mass of the internal standard in the test sample (see 8.4), in g; *A*₄ is the peak area of the internal standard.

9.2 Calculation of volume of benzene in the test sample

Calculate the volume, V₁, in ml, of benzene in the test sample using Formula (4):

$$V_1 = \frac{m_3 \times 1\,000}{\rho_{t,}} \tag{4}$$

where

 m_3 is the mass of benzene in the test sample (see 9.1), in g;

 ρ_{t} is the density of benzene at 15 °C, i.e. 884,3 kg/m³.

9.3 Calculation of volume of test sample

Calculate the volume, V_{2} , in ml, of the test sample using Formula (5):

$$_{2} = \frac{m_{s} \times 1000}{\rho_{s}} \tag{5}$$

where

 $m_{\rm S}$ is the mass of the test sample (see 8.4), in g;