
**Sharps injury protection —
Requirements and test methods —**

**Part 2:
Reusable sharps containers**

*Protection contre les blessures par perforants — Exigences et
méthodes d'essai —*

Partie 2: Conteneurs réutilisables pour perforants

iteh Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO 23907-2:2019](https://standards.iteh.ai/catalog/standards/iso/05d279b2-7a2a-4999-9637-50bca6837ae8/iso-23907-2-2019)

<https://standards.iteh.ai/catalog/standards/iso/05d279b2-7a2a-4999-9637-50bca6837ae8/iso-23907-2-2019>



iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO 23907-2:2019](https://standards.iteh.ai/catalog/standards/iso/05d279b2-7a2a-4999-9637-50bca6837ae8/iso-23907-2-2019)

<https://standards.iteh.ai/catalog/standards/iso/05d279b2-7a2a-4999-9637-50bca6837ae8/iso-23907-2-2019>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2019

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

	Page
Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Requirements and recommendations	4
4.1 General.....	4
4.2 Design and construction.....	4
4.2.1 General.....	4
4.2.2 Container stability.....	4
4.2.3 Strength of handles.....	4
4.2.4 Aperture and closure.....	4
4.2.5 Resistance to penetration.....	5
4.2.6 Resistance to damage or leakage after dropping.....	5
4.2.7 Resistance to damage or leakage after toppling.....	5
4.2.8 Fill line.....	5
4.3 Closure device.....	6
4.4 Monitoring of reuses.....	6
4.5 Cleaning and decontamination.....	6
4.5.1 Cleaning and decontamination process.....	6
4.5.2 Microbiological validation.....	7
5 Lifespan simulation prior to testing	7
5.1 General.....	7
5.2 Conditioning.....	7
5.3 Tumbling with sharps simulation.....	7
5.4 Transport simulations.....	8
5.5 Processing simulations.....	8
6 Test methods	8
6.1 Container stability.....	8
6.2 Strength of handle(s).....	8
6.3 Resistance to penetration.....	9
6.3.1 Apparatus.....	9
6.3.2 Procedure.....	9
6.4 Resistance to damage and leakage after dropping.....	10
6.4.1 Apparatus.....	10
6.4.2 Procedure.....	10
6.5 Resistance to spillage by toppling.....	11
6.5.1 Apparatus.....	11
6.5.2 Procedure.....	11
7 Quality monitoring: Post decontamination quality assurance	11
8 Labelling and marking and instructions for use	12
8.1 Labelling and marking.....	12
8.2 Instructions for use.....	13
Annex A (informative) Microbiological validation	14
Bibliography	17

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 84, *Devices for administration of medicinal products and catheters*.

A list of all parts in the ISO 23907 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

Reusable sharps containers are designed for the containment and disposal of sharps such as scalpel blades, trocars, hypodermic needles and syringes. They are supplied in a wide range of sizes and can be manufactured from a variety of materials. This document does not specify the size range of the containers or the materials selected to manufacture the containers.

Sharps containers can be either single-use or reusable. This document covers reusable sharps containers.

iTeh Standards
(<https://standards.itih.ai>)
Document Preview

[ISO 23907-2:2019](https://standards.itih.ai/catalog/standards/iso/05d279b2-7a2a-4999-9637-50bca6837ae8/iso-23907-2-2019)

<https://standards.itih.ai/catalog/standards/iso/05d279b2-7a2a-4999-9637-50bca6837ae8/iso-23907-2-2019>

Sharps injury protection — Requirements and test methods —

Part 2: Reusable sharps containers

1 Scope

This document specifies requirements for reusable sharps containers intended to hold potentially hazardous sharps medical waste with or without sharps protection features, e.g. scalpel blades, trocars, hypodermic needles and syringes.

This document is applicable to sharps containers that are supplied complete by the manufacturer and to those that are supplied as components intended to be assembled by the user.

It is not applicable to single use sharps containers (refer to ISO 23907-1 for such containers).

This document includes design functionality for user safety, lifespan simulation, cleaning and decontamination, microbiological validation, quality monitoring and performance testing.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 7864, *Sterile hypodermic needles for single use — Requirements and test methods*

<https://standards.iteh.ai/catalog/standards/iso/05d279b2-7a2a-4999-9637-50bca6837ae8/iso-23907-2-2019>

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

3.1

aperture

opening of the sharps container through which *sharps* (3.21) are deposited for disposal

3.2

clean

visually free of soil, debris and organic matter

3.3

closure feature

flap, plug, lid or slide that is intended to close the *aperture* (3.1)

**3.4
decontamination**

use of physical, chemical or thermal means to remove, soiling, and inactivate or destroy pathogens where to the point they are no longer capable of transmitting infectious disease and the container is rendered safe for handling and use

Note 1 to entry: Refer to [4.5](#).

**3.5
fill line**

mark, indicator or feature on the container that represents the *fill volume* ([3.6](#))

**3.6
fill volume**

usable volume determined by the manufacturer and indicated by the *fill line* ([3.5](#)) on the container

**3.7
handle**

appendage, protrusion, flange or recess intended for lifting the container

**3.8
integrally attached**

tethered or joined to the container by a permanent means

**3.9
leak-resistance**

ability of a container to prevent escape of fluid

Note 1 to entry: Refer to [6.4.2.2](#).

**3.10
lifespan**

maximum number of uses and/or processing cycles as validated by the manufacturer

**3.11
manufacturer's allowable gross mass**

maximum mass of the container and contents as recommended by the manufacturer for safe handling and operation

Note 1 to entry: Mass shall be measured in kilograms (kg).

**3.12
needle disconnection feature**

feature allowing single-handed *sharps* ([3.21](#)) disconnection

**3.13
penetration**

movement of a needle through the *test specimen* ([3.25](#)) until the point of the needle exits on the side opposite the point of entry

**3.14
penetration force**

amount of force applied to a hypodermic needle to achieve *penetration* ([3.13](#))

Note 1 to entry: The penetration force is expressed in newtons.

**3.15
permanent closure**

closure feature ([3.3](#)), *integrally attached* ([3.8](#)) to the container, which once activated cannot be re-opened manually

3.16**pocket collectors**

sharps container that has a *fill volume* (3.6) equal to or less than 0,6 l

Note 1 to entry: The primary design considerations for pocket collectors are to prevent penetration of the sharp(s) through the container while providing a compact size that can be carried on the person, such as in the user's pocket. In order to achieve portability and a low profile, these devices have been excluded from certain aspects of the requirements of this document.

3.17**representative sharps**

representative quantities of unused 1 ml to 60 ml syringes-needles in the ratio of 70 % safety devices with safety mechanism activated and 30 % syringe-needles with no safety device

3.18**processing**

validated methods used to render a reusable sharps container, which has been previously used, safe for handling and a subsequent use

3.19**reusable sharps container**

container designed or intended by the manufacturer to be suitable for use, emptying, *processing* (3.18) and subsequent reuse

3.20**secondary stabilizer**

attachment or design feature intended to provide extra stability and prevent the device from toppling over

3.21**sharps**

objects capable of cutting or penetrating skin

EXAMPLE Needles of various types, syringes, scalpels, broken glass, culture slides, culture dishes, broken capillary tubes, broken rigid plastic, exposed ends of dental wires.

3.22**sharps containment area**

surface that directly encloses *sharps* (3.21) for the purposes of container puncture protection while in use and in the final closed configuration

3.23**lifespan simulation**

conditioning of the reusable container by repeated exposure to cycles of representative transport and *decontamination* (3.4) process conditions to represent the *lifespan* (3.10) of the container, as specified by the manufacturer

3.24**temporary closure**

closure feature (3.3) *integrally attached* (3.8) to the container which, once activated for closure, can be re-opened, without being damaged

3.25**test specimen**

portion of the container obtained

Note 1 to entry: Refer to 6.3.2.1.

3.26**total volume of the container**

entire air space in the closed container

4 Requirements and recommendations

4.1 General

The principles of risk assessment, as well as human factors to avoid any inappropriate use such as incorrect assembly and overfilling, shall be considered in the design process of sharps containers.

NOTE ISO 14971 is relevant where containers are classified as a medical device.

The base dominant colour should be yellow unless local regulations state otherwise.

Fill level visibility shall be a design requirement for the containers.

When evaluated in accordance with [Clause 6](#), the container shall show no rupture, leakage, or deterioration that could adversely affect its safe use or functionality.

4.2 Design and construction

4.2.1 General

The materials used in the construction of reusable sharps containers shall be designed and manufactured in a manner that enables full function, safety, and preservation of aesthetics for the expected lifespan of the product.

4.2.2 Container stability

The container shall not topple over when tested in accordance with [6.1](#).

The requirement applies to containers intended for use on a horizontal surface. Sharps containers intended to be used with a secondary stabilizer shall be tested in conjunction with the secondary stabilizer.

Sharps containers equipped with a needle disconnection feature shall have a means whereby the disconnection procedure is achieved with one hand.

Pocket collectors are not required to have needle disconnection feature.

4.2.3 Strength of handles

All sharps containers shall be provided with one or several handles. Where handle(s) is/are included in the design, it shall meet the requirements of [4.2.3](#).

Pocket collectors are not required to have handle(s).

When tested in accordance with [6.2](#), the handle/carrying feature shall not break or detach during testing. The position of the handle(s), finger recesses, protrusions or flanges shall not interfere with the normal use of the container.

Finger recesses, if present, shall be sited above the fill line.

4.2.4 Aperture and closure

4.2.4.1 General

Reusable sharps containers shall be provided with closure features that are integrally attached. The aperture shall be designed to minimize the potential for accidental sharps injuries during placement of sharps into the container. There shall be an indicator or mechanism (preferably visual) to clearly differentiate the permanent and temporary closure engagements.