
**Cigarettes — Determination of water
in total particulate matter from the
mainstream smoke —**

**Part 1:
Gas-chromatographic method**

iTeh STANDARD PREVIEW
(standards.iteh.ai)
*Cigarettes — Dosage de l'eau dans la matière particulaire totale du
courant principal de fumée —
Partie 1: Méthode par chromatographie en phase gazeuse*

[ISO 10362-1:2019](https://standards.iteh.ai/catalog/standards/sist/46a56571-75c6-4f19-9710-8aaf2baa1723/iso-10362-1-2019)

<https://standards.iteh.ai/catalog/standards/sist/46a56571-75c6-4f19-9710-8aaf2baa1723/iso-10362-1-2019>



iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO 10362-1:2019

<https://standards.iteh.ai/catalog/standards/sist/46a56571-75c6-4f19-9710-8aaf2baa1723/iso-10362-1-2019>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2019

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

	Page
Foreword.....	iv
Introduction.....	v
1 Scope.....	1
2 Normative reference.....	1
3 Terms and definitions.....	1
4 Principle.....	1
5 Reagents.....	1
6 Apparatus.....	2
7 Procedure.....	3
7.1 Test portion.....	3
7.2 Setting up the apparatus.....	3
7.3 Calibration of the gas chromatograph.....	3
7.4 Blank test.....	3
7.5 Determination.....	4
8 Alternative gas chromatographic procedures.....	4
8.1 General.....	4
8.2 Packed columns.....	4
8.3 Capillary columns.....	4
8.4 Gas chromatographic conditions.....	4
9 Expression of results.....	5
10 Repeatability and reproducibility.....	6
11 Test report.....	7
Annex A (informative) Use of this method in conjunction with the method of nicotine determination.....	8
Bibliography.....	9

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html. (standards.iteh.ai)

This document was prepared by Technical Committee ISO/TC 126, *Tobacco and tobacco products*.
ISO 10362-1:2019

This third edition cancels and replaces the second edition (ISO 10362-1:1999), which has been technically revised. It also incorporates the Amendment ISO 10362-1:1999/AMD 1:2011. The main changes compared to the previous edition are as follows:

- capillary column procedure as an alternative gas chromatographic procedure has been added;
- the estimates for the repeatability limits (r) and the reproducibility limits (R) has been updated.

A list of all parts in the ISO 10362 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

This document may be considered as part of a set of standards which describes the determination of total and nicotine-free dry particulate matter (NFDPM) in total particulate matter from the mainstream smoke. The set comprises ISO 3308, ISO 3402, ISO 4387, ISO 8243, ISO 10315 and this document.

[Annex A](#) provides information about the use of this method in conjunction or simultaneously with the gas-chromatographic method of nicotine determination described in ISO 10315.

A bibliography is provided.

No machine smoking regime can represent all human smoking behaviour:

- it is recommended that cigarettes also be tested under conditions of a different intensity of machine smoking than those specified in this document;
- machine smoking testing is useful to characterize cigarette emissions for design and regulatory purposes, but communication of machine measurements to smokers can result in misunderstandings about differences in exposure and risk across brands;
- smoke emission data from machine measurements may be used as inputs for product hazard assessment, but they are not intended to be nor are they valid as measures of human exposure or risks. Communicating differences between products in machine measurements as differences in exposure or risk is a misuse of testing using ISO standards.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[ISO 10362-1:2019](#)

<https://standards.iteh.ai/catalog/standards/sist/46a56571-75c6-4f19-9710-8aaf2baa1723/iso-10362-1-2019>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO 10362-1:2019

<https://standards.iteh.ai/catalog/standards/sist/46a56571-75c6-4f19-9710-8aaf2baa1723/iso-10362-1-2019>

Cigarettes — Determination of water in total particulate matter from the mainstream smoke —

Part 1: Gas-chromatographic method

1 Scope

This document specifies a method for the gas-chromatographic determination of water in total particulate matter from the mainstream smoke. The smoking of cigarettes and collection of mainstream smoke are normally carried out in accordance with ISO 4387. However, the method of this document is also applicable to the determination of water in total particulate matter from the mainstream smoke obtained by non-standard smoking.

NOTE In countries not in a position to use the gas-chromatographic method, the determination of water in total particulate matter from the mainstream smoke is performed using the Karl Fischer method (see ISO 10362-2). In such cases, values obtained for water in total particulate matter from the mainstream smoke are used with the addition of a note made in the expression of the results.

iTeh STANDARD PREVIEW

2 Normative reference (standards.iteh.ai)

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3696:1987, *Water for analytical laboratory use — Specification and test methods*

ISO 4387, *Cigarettes — Determination of total and nicotine-free dry particulate matter using a routine analytical smoking machine*

3 Terms and definitions

No terms and definitions are listed in this document.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

4 Principle

The total particulate matter from the mainstream smoke is dissolved in a solvent containing an internal standard. The water content of an aliquot of the solution is determined by gas chromatography, and the water content of the whole of the total particulate matter from the mainstream smoke is calculated.

5 Reagents

Use only reagents of recognized analytical reagent grade.

5.1 Carrier gas, helium or nitrogen (see note in 7.2) or hydrogen (see note in 8.4).

5.2 Propan-2-ol, with maximum water content of 1,0 mg/ml.

5.3 Internal standard, ethanol or methanol (of purity at least 99 %).

5.4 Extraction solution, propan-2-ol (5.2) containing an appropriate concentration of internal standard (5.3), normally in the range of 1 ml/l to 5 ml/l.

Extraction solution not stored in a temperature-controlled laboratory shall be allowed to equilibrate to (22 ± 2) °C before use.

To prevent water being absorbed, the bulk extraction solution container shall be fitted with a water trap and shall be kept sealed. The extraction solution shall be stirred continuously to ensure the homogeneity of the water concentration.

5.5 Reference substance, complying with grade 2 of ISO 3696:1987, or better.

5.6 Calibration solutions

Prepare a series of at least four calibration solutions whose concentrations of added water cover the range expected to be found in the test portion (usually up to 4 mg/ml) by adding weighed amounts of water (5.5) to the extraction solution (5.4). One of these calibration solutions shall be the extraction solution with no added water (extraction solution blank).

To prevent water being absorbed, all solutions shall be kept sealed. The calibration solutions shall be made up using an extraction solution from the same batch used in 7.1.

Calibration solutions are known to be stable for at least one week when properly stored, but the laboratory should verify calibration solution stability.

ISO 10362-1:2019

<https://standards.iteh.ai/catalog/standards/sist/46a56571-75c6-4f19-9710-8aaf2baa1723/iso-10362-1-2019>

6 Apparatus

Usual laboratory apparatus and, in particular, the following items.

6.1 Gas-chromatograph, equipped with a thermal conductivity detector, recorder and integrator or other suitable data-handling equipment.

Glassware and septa for vials should be stored in a desiccator until use.

6.2 Column, of internal diameter between 2 mm and 4 mm and preferably of length 1,5 m to 2 m.

Stationary phase: Porapak Q¹⁾ 150 µm (100 mesh) to 190 µm (80 mesh).

The column is preferably made of deactivated stainless steel but other materials such as glass or nickel may be used. Alternative stationary phases may be used (see [Clause 8](#)).

6.3 Dispensing system, preferably automated, capable of delivering the required volume of extraction solution (5.4).

The dispensing system should be flushed prior to use by dispensing a volume of extraction solution of at least 50 ml which will then be rejected.

1) Porapak Q is a trade names of an example of a suitable products available commercially. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO of these products.

7 Procedure

7.1 Test portion

Prepare the test portion by dissolving the total particulate matter obtained by the machine smoking of a known number of cigarettes in a fixed volume of the extraction solution (5.4) of 20 ml for 44 mm discs or 50 ml for 92 mm discs, ensuring that the disc is fully covered. The volume may be adjusted to give a concentration of water appropriate for the calibration graph (see 7.3) provided that there is an adequate volume for effective extraction of the total particulate matter. Analysis should be performed as soon as possible, but if storage is inevitable then store the sample at between 0 °C and 4 °C. The laboratory should verify test portion stability. For standard smoking, see ISO 4387.

7.2 Setting up the apparatus

Set up the apparatus and operate the gas chromatograph (6.1) in accordance with the manufacturer's instructions. Ensure that the peaks for water, the internal standard and solvent are well resolved, the analysis time being about 4 min. Condition the system just prior to use by injecting a 2 µl aliquot of the extraction solution as a primer.

Suitable operating conditions are as follows:

- column temperature: 170 °C (isothermal);
- injection temperature: 250 °C;
- detector temperature: 250 °C;
- carrier gas: helium at a flow rate of about 30 ml/min;
- injection volume: 2 µl.

NOTE Nitrogen is also used as an alternative carrier gas if the detector sensitivity is sufficiently high.

7.3 Calibration of the gas chromatograph

Inject an aliquot (2 µl) of each of the calibration solutions (5.6) into the gas chromatograph. Record the peak areas of the water and internal standard (5.3). Carry out the determination at least twice.

Calculate the ratio of the water peak to the internal standard peak from the peak area data for each of the calibration solutions including the extraction solution blank. Plot the graph of the water concentrations in accordance with the area ratios, and calculate a linear regression equation of the peak area ratios as a function of the concentration of water from these data.

Perform this full calibration procedure daily. In addition, inject an aliquot of an intermediate concentration standard after every 20 sample determinations. If the calculated concentration for this solution differs by more than 5 % from the original value, repeat the full calibration procedure.

NOTE The regression line does not pass through zero due to water present in the extraction solvent.

If the water content of the solvent exceeds 1,0 mg/ml, the batch should be rejected.

7.4 Blank test

Due to the absorption of water by smoke traps and extraction solution, it is necessary to determine a value for the sample blank. Prepare sample blanks by treating additional smoke traps including filters (at least 2 smoke traps including filters per 100 cigarettes smoked) in the same manner as that used for smoke collection. Place smoke traps, including filters, near the smoking machine during smoking. Extract and analyse the blank test portions together with the test portions.