

**Nadomešča:****SIST EN ISO 10534-2:2002**

---

**Akustika - Ugotavljanje akustičnih lastnosti v Kundtovi cevi - 2. del:  
Dvomikrofonska tehnika za določanje normalnega koeficienta absorpcije zvoka in  
normalne površinske impedance (ISO 10534-2:2023)**

Acoustics - Determination of acoustic properties in impedance tubes - Part 2: Two-  
microphone technique for normal sound absorption coefficient and normal surface  
impedance (ISO 10534-2:2023)

Akustik - Bestimmung der akustischen Eigenschaften in Impedanzrohren - Teil 2: 2-  
Mikrofontechnik für Schallabsorptionsgrad und Oberflächenimpedanz bei senkrechtem  
Einfall (ISO 10534-2:2023)

Acoustique - Détermination des propriétés acoustiques aux tubes d'impédance - Partie  
2: Méthode à deux microphones pour le coefficient d'absorption acoustique normal et  
l'impédance de surface normale (ISO 10534-2:2023)

**Ta slovenski standard je istoveten z: EN ISO 10534-2:2023**

**ICS:**

17.140.01	Akustična merjenja in blaženje hrupa na splošno	Acoustic measurements and noise abatement in general
-----------	--	---

**SIST EN ISO 10534-2:2024****en,fr,de**



EUROPEAN STANDARD  
NORME EUROPÉENNE  
EUROPÄISCHE NORM

**EN ISO 10534-2**

October 2023

ICS 17.140.01

Supersedes EN ISO 10534-2:2001

English Version

**Acoustics - Determination of acoustic properties in  
impedance tubes - Part 2: Two-microphone technique for  
normal sound absorption coefficient and normal surface  
impedance (ISO 10534-2:2023)**

Acoustique - Détermination des propriétés acoustiques  
aux tubes d'impédance - Partie 2: Méthode à deux  
microphones pour le coefficient d'absorption  
acoustique normal et l'impédance de surface normale  
(ISO 10534-2:2023)

Akustik - Bestimmung der akustischen Eigenschaften  
in Impedanzrohren - Teil 2: 2-Mikrofontechnik für  
Standardschallabsorptionsgrad und  
Standardoberflächenimpedanz (ISO 10534-2:2023)

This European Standard was approved by CEN on 2 October 2023.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

<b>Contents</b>	<b>Page</b>
<b>European foreword.....</b>	<b>3</b>

**iTeh Standards**  
**(<https://standards.iteh.ai>)**  
**Document Preview**

[SIST EN ISO 10534-2:2024](https://standards.iteh.ai/catalog/standards/sist/588d4028-5f71-401e-9b0f-f04bf842bf0d/sist-en-iso-10534-2-2024)

<https://standards.iteh.ai/catalog/standards/sist/588d4028-5f71-401e-9b0f-f04bf842bf0d/sist-en-iso-10534-2-2024>

## European foreword

This document (EN ISO 10534-2:2023) has been prepared by Technical Committee ISO/TC 43 "Acoustics" in collaboration with Technical Committee CEN/TC 126 "Acoustic properties of building elements and of buildings" the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by April 2024, and conflicting national standards shall be withdrawn at the latest by April 2024.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 10534-2:2001.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

**iTeh Standards**  
**(<https://standards.iteh.ai>)**  
**Endorsement notice**

The text of ISO 10534-2 has been approved by CEN as EN ISO 10534-2:2023 without any modification.

[SIST EN ISO 10534-2:2024](https://standards.iteh.ai)

<https://standards.iteh.ai/catalog/standards/sist/588d4028-5f71-401e-9b0f-f04bf842bf0d/sist-en-iso-10534-2-2024>



INTERNATIONAL  
STANDARD

ISO  
10534-2

Second edition  
2023-10

---

---

**Acoustics — Determination of acoustic  
properties in impedance tubes —**

Part 2:

**Two-microphone technique for  
normal sound absorption coefficient  
and normal surface impedance**

*Acoustique — Détermination des propriétés acoustiques aux tubes  
d'impédance —*

*Partie 2: Méthode à deux microphones pour le coefficient d'absorption  
sonore normal et l'impédance de surface normale*

Document Preview

[SIST EN ISO 10534-2:2024](https://standards.iteh.ai/catalog/standards/sist/588d4028-5f71-401e-9b0f-f04bf842bf0d/sist-en-iso-10534-2-2024)

<https://standards.iteh.ai/catalog/standards/sist/588d4028-5f71-401e-9b0f-f04bf842bf0d/sist-en-iso-10534-2-2024>



Reference number  
ISO 10534-2:2023(E)

© ISO 2023

ISO 10534-2:2023(E)

# iTeh Standards (<https://standards.iteh.ai>) Document Preview

[SIST EN ISO 10534-2:2024](https://standards.iteh.ai/catalog/standards/sist/588d4028-5f71-401e-9b0f-f04bf842bf0d/sist-en-iso-10534-2-2024)

<https://standards.iteh.ai/catalog/standards/sist/588d4028-5f71-401e-9b0f-f04bf842bf0d/sist-en-iso-10534-2-2024>



## **COPYRIGHT PROTECTED DOCUMENT**

© ISO 2023

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland



# Contents

Page

<b>Foreword</b> .....	<b>iv</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms, definitions and symbols</b> .....	<b>1</b>
<b>4 Principle</b> .....	<b>5</b>
<b>5 Test equipment</b> .....	<b>5</b>
5.1 Construction of the impedance tube.....	5
5.2 Working frequency range.....	6
5.3 Length of the impedance tube.....	7
5.4 Microphones.....	7
5.5 Positions of the microphones.....	7
5.6 Acoustic centre of the microphone.....	8
5.7 Test sample holder.....	8
5.8 Signal processing equipment.....	9
5.9 Loudspeaker.....	9
5.10 Signal generator.....	9
5.11 Thermometer, barometer and relative humidity.....	9
<b>6 Preliminary test and measurements</b> .....	<b>10</b>
<b>7 Test specimen mounting</b> .....	<b>11</b>
<b>8 Test procedure</b> .....	<b>12</b>
8.1 Specification of the reference plane.....	12
8.2 Determination of the sound velocity, wavelength and characteristic impedance.....	12
8.3 Selection of the signal amplitude.....	13
8.4 Selection of the number of averages.....	13
8.5 Correction for microphone mismatch.....	13
8.5.1 General.....	13
8.5.2 Measurement repeated with the channels interchanged.....	13
8.5.3 Predetermined calibration factor.....	14
8.6 Determination of the transfer function between the two locations.....	15
8.6.1 General.....	15
8.6.2 Cross- and autospectra-based estimate.....	15
8.6.3 Frequency-domain deconvolution.....	17
8.6.4 Impulse-response based estimate.....	17
8.7 Determination of the reflection coefficient.....	18
8.8 Determination of the sound absorption coefficient.....	18
8.9 Determination of the specific acoustic impedance ratio.....	18
8.10 Determination of the specific acoustic admittance ratio.....	18
<b>9 Precision</b> .....	<b>19</b>
<b>10 Test report</b> .....	<b>19</b>
<b>Annex A (normative) Preliminary measurements</b> .....	<b>22</b>
<b>Annex B (normative) Procedure for the one-microphone technique</b> .....	<b>24</b>
<b>Annex C (informative) Theoretical background</b> .....	<b>25</b>
<b>Annex D (informative) Error sources</b> .....	<b>27</b>
<b>Annex E (informative) Estimation of diffuse sound absorption coefficient <math>\alpha_{st}</math> of locally reacting absorbers from the results of this document</b> .....	<b>29</b>
<b>Annex F (informative) Estimation of intrinsic properties</b> .....	<b>30</b>
<b>Bibliography</b> .....	<b>32</b>

## ISO 10534-2:2023(E)

### Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at [www.iso.org/patents](http://www.iso.org/patents). ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 43 *Acoustics*, Subcommittee SC 2, *Building acoustics*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 126, *Acoustics properties of building products and of buildings*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO 10534-2:1998), which has been technically revised.

The main changes are as follows:

- the introduction of the measurement procedure to estimate the characteristic properties of porous materials (characteristic impedance, wavenumber, dynamic mass density, dynamic bulk modulus) in an informative annex. The signal processing techniques have been updated since the first version of this document.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

# Acoustics — Determination of acoustic properties in impedance tubes —

## Part 2:

# Two-microphone technique for normal sound absorption coefficient and normal surface impedance

## 1 Scope

This test method covers the use of an impedance tube, two microphone locations and a frequency analysis system for the determination of the sound absorption coefficient of sound absorbing materials for normal incidence sound incidence. It can also be applied for the determination of the acoustical surface impedance or surface admittance of sound absorbing materials. As an extension, it can also be used to assess intrinsic properties of homogeneous acoustical materials such as their characteristic impedance, characteristic wavenumber, dynamic mass density and dynamic bulk modulus.

The test method is similar to the test method specified in ISO 10534-1<sup>[1]</sup> in that it uses an impedance tube with a sound source connected to one end and the test sample mounted in the tube at the other end. However, the measurement technique is different. In this test method, plane waves are generated in a tube by a sound source, and the decomposition of the interference field is achieved by the measurement of acoustic pressures at two fixed locations using wall-mounted microphones or an in-tube traversing microphone, and subsequent calculation of the complex acoustic transfer function and quantities reported in the previous paragraph. The test method is intended to provide an alternative, and generally much faster, measurement technique than that of ISO 10534-1<sup>[1]</sup>.

Normal incidence absorption coefficients from impedance tube measurements are not comparable with random incidence absorption coefficients measured in reverberation rooms according to ISO 354<sup>[2]</sup>. The reverberation room method will (under ideal conditions) determine the sound absorption coefficient for diffuse sound incidence. However, the reverberation room method requires test specimens which are rather large. The impedance tube method is limited to studies at normal and plane incidence and requires samples of the test object which are of the same size as the cross-section of the impedance tube. For materials that are locally reacting only, diffuse incidence sound absorption coefficients can be estimated from measurement results obtained by the impedance tube method (see [Annex E](#)).

Through the whole document, a  $e^{+j\omega t}$  time convention is used.

## 2 Normative references

There are no normative references in this document.

## 3 Terms, definitions and symbols

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

## ISO 10534-2:2023(E)

### 3.1 sound absorption coefficient at normal incidence

$\alpha_n$   
ratio of the sound power dissipated inside the test object to the incident sound power for a plane wave at normal incidence

Note 1 to entry: "Plane wave" here describes a wave whose value, at any moment, is constant over any plane perpendicular to its direction of propagation. "Normal incidence" describes the direction of the longest axis of the impedance tube.

### 3.2 sound pressure reflection coefficient at normal incidence

$r$   
complex ratio of the reflected wave sound pressure amplitude to that of the incident wave in the reference plane for a plane wave at normal incidence

### 3.3 reference plane

cross-section of the impedance tube for which the reflection factor  $r$  or the impedance  $Z$  or the admittance  $G$  are determined and which is usually the surface of the test object, if flat

Note 1 to entry: The reference plane is assumed to be at  $x = 0$ .

### 3.4 normal-incidence surface impedance

$Z$   
ratio of the complex sound pressure  $p(x = 0)$  to the normal component of the complex sound particle velocity  $v(x = 0)$  at an individual frequency in the reference plane defined as  $x = 0$

Note 1 to entry: The particle velocity vector has a positive direction pointing towards the interior of the tested object.

Note 2 to entry:  $Z$  is expressed in newton second per cubic meter ( $\text{Ns/m}^3$ )

### 3.5 normal-incidence surface admittance

$G$   
inverse of the normal-incidence surface impedance  $Z$

Note 1 to entry:  $G$  is expressed in cubic meter per newton per second ( $\text{m}^3/\text{N/s}$ )

### 3.6 wave number in air

$k_0$   
variable, expressed in radian per metre, defined by

$$k_0 = \omega / c_0 = 2\pi f / c_0 = 2\pi / \lambda_0$$

where

$\omega$  is the angular frequency,

$f$  is the frequency,

$c_0$  is the speed of sound in the air,

$\lambda_0$  is the wavelength in air.

Note 1 to entry: In general, the wave number is complex, so that  $k_0 = k'_0 - jk''_0$  where  $k'_0$  is the real component and  $k''_0$  is the imaginary component (which is the attenuation constant).